

QUARTER 3, 2025

SPRAGUE OPERATING RESOURCES LLC

TERMINAL IN SEARSPORT MAINE

Prepared For:

Sprague Operating Resources LLC

185 International Drive,
Portsmouth, NH 03801

Facility:

Sprague Operating Resources LLC

70 Trundy Road,
Searsport, ME 04974

Prepared By:

Montrose Air Quality Services LLC

45 Route 46 East, Suite 601
Pine Brook, NJ 07058

Report Number: 047AA-027966-RT-1051

Date: November 12, 2025

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING - SUMMARY

Flags:	ND	The analyte was not present above the Method Detection Limit									
	J	Estimated Value - The analyte was detected between the Method Detection Limit and Reporting Limit									
	P	Field duplicate(s) exceed 30%RPD									
	Pc	Field duplicate(s) exceed 30%RPD. Concentrations of both samples in duplicate are near the reporting limit									
Sample Code	Tube ID	Benzene		Ethylbenzene		m-/p-Xylene		o-Xylene		Toluene	
		(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag
SPRSEA-1-S-20250626	C57687	0.436		0.272	ND	0.456	J,Pc	0.272	ND	0.721	
SPRSEA-2-S-20250626	C33009	0.514		0.272	ND	0.563	Pc	0.272	ND	1.04	
SPRSEA-3-S-20250626	B19056	0.551		0.3	J	1.24	Pc	0.582		1.15	
SPRSEA-4-S-20250626	C16112	0.515		0.272	ND	0.504	J,Pc	0.272	ND	0.893	
SPRSEA-5-S-20250626	B49576	0.526		0.272	ND	0.466	J,Pc	0.272	ND	1.01	
SPRSEA-6-S-20250626	B38593	0.456		0.272	ND	0.463	J,Pc	0.272	ND	0.829	
SPRSEA-6-D-20250626	B18562	0.458		0.272	ND	0.326	J,Pc	0.272	ND	0.742	
SPRSEA-6-B-20250626	C40533	0.186	ND	0.272	ND	0.272	ND,Pc	0.272	ND	0.24	ND
SPRSEA-7-S-20250626	B50653	0.431		0.272	ND	0.367	J,Pc	0.272	ND	0.671	
SPRSEA-8-S-20250626	C34132	0.421		0.272	ND	0.488	J,Pc	0.272	ND	0.919	
SPRSEA-9-S-20250626	C61442	0.682		0.296	J	1.02	Pc	0.391	J	1.75	
SPRSEA-10-S-20250626	B46360	2.3		1.59		5.05	Pc	2.09		9.7	
SPRSEA-11-S-20250626	C43733	1.18		0.813		2.94	Pc	1.15		4.07	
SPRSEA-12-S-20250626	C00561	0.754		0.407	J	1.33	Pc	0.552	J	2.14	
SPRSEA-12-D-20250626	C57727	0.675		0.402	J	1.41	Pc	0.547	J	2.05	
SPRSEA-12-B-20250626	C20610	0.186	ND	0.272	ND	0.272	ND,Pc	0.272	ND	0.24	ND
SPRSEA-1-S-20250710	C33260	0.437		0.271	ND	0.495	J	0.271	ND,Pc	1.05	P
SPRSEA-2-S-20250710	B17444	0.61		0.475	J	0.801		0.336	J,Pc	1.65	P
SPRSEA-3-S-20250710	C56859	0.516		0.294	J	0.956		0.427	J,Pc	1.49	P
SPRSEA-4-S-20250710	C38901	0.484		0.289	J	0.672		0.271	ND,Pc	1.38	P
SPRSEA-5-S-20250710	C57730	0.383		0.271	ND	0.617		0.271	ND,Pc	1.01	P
SPRSEA-6-S-20250710	C57534	0.362	J	0.271	ND	0.67		0.271	ND,Pc	0.874	P
SPRSEA-6-D-20250710	C01661	0.451		0.309	J	0.543		0.271	ND,Pc	1.23	P
SPRSEA-6-B-20250710	C00592	0.186	ND	0.271	ND	0.271	ND	0.271	ND,Pc	0.24	ND,P
SPRSEA-7-S-20250710	C55400	0.393		0.271	ND	0.341	J	0.271	ND,Pc	0.903	P
SPRSEA-8-S-20250710	C20391	0.577		0.342	J	0.652		0.273	J,Pc	1.58	P
SPRSEA-9-S-20250710	B14580	1.09		0.784		1.81		0.786	Pc	3.65	P

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING - SUMMARY

Flags:	ND	The analyte was not present above the Method Detection Limit									
	J	Estimated Value - The analyte was detected between the Method Detection Limit and Reporting Limit									
	P	Field duplicate(s) exceed 30%RPD									
	Pc	Field duplicate(s) exceed 30%RPD. Concentrations of both samples in duplicate are near the reporting limit									
Sample Code	Tube ID	Benzene		Ethylbenzene		m-/p-Xylene		o-Xylene		Toluene	
		(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag
SPRSEA-10-S-20250710	C43380	3.67		3.13		10.5		4.31	Pc	18.8	P
SPRSEA-11-S-20250710	B12197	1.21		1.28		3.38		1.1	Pc	4.05	P
SPRSEA-12-S-20250710	C36934	0.807		0.745		2.14		0.761	Pc	2.92	P
SPRSEA-12-D-20250710	C43377	0.821		0.871		2.84		1.06	Pc	3.22	P
SPRSEA-12-B-20250710	C43932	0.186	ND	0.271	ND	0.271	ND	0.271	ND,Pc	0.24	ND,P
SPRSEA-1-S-20250724	C60287	0.786		0.405	J	1.56		0.536	J	1.64	
SPRSEA-2-S-20250724	B29752	0.756		0.597		1.46		0.585		1.76	
SPRSEA-3-S-20250724	C60278	0.813		0.354	J	1.29		0.576		1.6	
SPRSEA-4-S-20250724	C38893	0.567		0.278	J	1.04		0.463	J	1.01	
SPRSEA-5-S-20250724	C01908	0.657		0.332	J	0.749		0.352	J	1.03	
SPRSEA-6-S-20250724	C37428	0.632		0.271	ND	0.373	J	0.271	ND	0.774	
SPRSEA-6-D-20250724	C57756	0.511		0.271	ND	0.437	J	0.271	ND	0.647	
SPRSEA-6-B-20250724	C36887	0.186	ND	0.271	ND	0.271	ND	0.271	ND	0.24	ND
SPRSEA-7-S-20250724	B47211	0.679		0.271	ND	0.271	ND	0.271	ND	0.609	
SPRSEA-8-S-20250724	C57437	0.541		0.271	ND	0.449	J	0.271	ND	0.662	
SPRSEA-9-S-20250724	C32832	0.691		0.434	J	1.44		0.598		1.49	
SPRSEA-10-S-20250724	C43550	1.72		1.04		3.82		1.54		5.41	
SPRSEA-11-S-20250724	B50749	1.58		1.02		2.93		1.14		4.6	
SPRSEA-12-S-20250724	C57454	1.31		1.2		4.68		1.59		3.75	
SPRSEA-12-D-20250724	B42495	1.44		1.38		4.81		1.65		4.01	
SPRSEA-12-B-20250724	C39256	0.186	ND	0.271	ND	0.271	ND	0.271	ND	0.24	ND
SPRSEA-1-S-20250807	C43715	0.501		0.289	J,Pc	0.711		0.276	J	1.39	
SPRSEA-2-S-20250807	C00562	0.713		0.48	J,Pc	1.52		0.567		2.13	
SPRSEA-3-S-20250807	C20428	0.721		0.5	J,Pc	1.18		0.428	J	2.11	
SPRSEA-4-S-20250807	C01992	0.352	J	0.271	ND,Pc	0.584		0.271	ND	0.745	
SPRSEA-5-S-20250807	C61499	0.332	J	0.271	ND,Pc	0.683		0.271	ND	0.695	
SPRSEA-6-S-20250807	C01402	0.363	J	0.271	ND,Pc	0.342	J	0.271	ND	0.668	

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING - SUMMARY

Flags:	ND	The analyte was not present above the Method Detection Limit									
	J	Estimated Value - The analyte was detected between the Method Detection Limit and Reporting Limit									
	P	Field duplicate(s) exceed 30%RPD									
	Pc	Field duplicate(s) exceed 30%RPD. Concentrations of both samples in duplicate are near the reporting limit									
Sample Code	Tube ID	Benzene		Ethylbenzene		m-/p-Xylene		o-Xylene		Toluene	
		(ug/m ³)	Flag	(ug/m ³)	Flag	(ug/m ³)	Flag	(ug/m ³)	Flag	(ug/m ³)	Flag
SPRSEA-6-D-20250807	B20636	0.366	J	0.483	J,Pc	0.412	J	0.271	ND	0.782	
SPRSEA-6-B-20250807	B15753	0.186	ND	0.271	ND,Pc	0.271	ND	0.271	ND	0.24	ND
SPRSEA-7-S-20250807	C39271	0.245	J	0.271	ND,Pc	0.271	ND	0.271	ND	0.394	J
SPRSEA-8-S-20250807	B14681	0.518		0.271	ND,Pc	0.271	ND	0.271	ND	0.566	
SPRSEA-9-S-20250807	B46315	0.556		0.533	J,Pc	1.25		0.545	J	1.68	
SPRSEA-10-S-20250807	C37447	1.88		1.66	Pc	5.35		2.11		7.59	
SPRSEA-11-S-20250807	C37014	2.37		1.91	Pc	6.09		2.3		9.93	
SPRSEA-12-S-20250807	C57499	1.31		0.764	Pc	2.06		0.788		4.29	
SPRSEA-12-D-20250807	C36957	1.21		1.09	Pc	2.58		0.927		4.15	
SPRSEA-12-B-20250807	C40632	0.186	ND	0.271	ND,Pc	0.271	ND	0.271	ND	0.312	J
SPRSEA-1-S-20250821	B27217	0.993		1.15		3.54		1.34		4.57	
SPRSEA-2-S-20250821	B46982	0.977		0.842		2.82		1.1		3.75	
SPRSEA-3-S-20250821	B50949	0.687		0.646		2.17		0.93		2.9	
SPRSEA-4-S-20250821	C40149	0.805		1.01		3.05		1.19		3.72	
SPRSEA-5-S-20250821	B49635	0.977		1.06		4.12		1.6		4.95	
SPRSEA-6-S-20250821	B46127	0.87		0.809		3.56		1.43		4.21	
SPRSEA-6-D-20250821	B49528	1		1.04		3.59		1.5		4.47	
SPRSEA-6-B-20250821	C00595	0.188	ND	0.273	ND	0.273	ND	0.273	ND	0.242	ND
SPRSEA-7-S-20250821	B17520	0.978		1.01		3.73		1.47		4.17	
SPRSEA-8-S-20250821	C55400	0.944		0.879		3.23		1.37		4.19	
SPRSEA-9-S-20250821	C56980	1.22		1.25		5.39		2.15		5.84	
SPRSEA-10-S-20250821	C01665	4.68		3.18		12.2		4.91		17.8	
SPRSEA-11-S-20250821	C01392	1.55		1.54		5.62		2.33		7.07	
SPRSEA-12-S-20250821	C70525	1.82		1.58		6.43		2.6		8.04	
SPRSEA-12-D-20250821	C61742	1.79		1.54		6.47		2.51		7.59	
SPRSEA-12-B-20250821	B44953	0.188	ND	0.273	ND	0.273	ND	0.273	ND	0.242	ND
SPRSEA-1-S-20250904	C33505	1.22		0.518	J	1.59	Pc	0.659		2.68	

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING - SUMMARY

Flags:	ND	The analyte was not present above the Method Detection Limit									
	J	Estimated Value - The analyte was detected between the Method Detection Limit and Reporting Limit									
	P	Field duplicate(s) exceed 30%RPD									
	Pc	Field duplicate(s) exceed 30%RPD. Concentrations of both samples in duplicate are near the reporting limit									
Sample Code	Tube ID	Benzene		Ethylbenzene		m-/p-Xylene		o-Xylene		Toluene	
		(ug/m ³)	Flag	(ug/m ³)	Flag	(ug/m ³)	Flag	(ug/m ³)	Flag	(ug/m ³)	Flag
SPRSEA-2-S-20250904	C70563	1.09		0.463	J	1.77	Pc	0.712		2.4	
SPRSEA-3-S-20250904	C01621	1.91		0.384	J	1.21	Pc	0.483	J	2.09	
SPRSEA-4-S-20250904	C38576	1.75		0.274	ND	0.666	Pc	0.274	ND	1.33	
SPRSEA-5-S-20250904	C70808	0.754		0.274	ND	0.934	Pc	0.369	J	1.33	
SPRSEA-6-S-20250904	C65284	0.675		0.274	ND	0.859	Pc	0.354	J	1.12	
SPRSEA-6-D-20250904	C31427	0.834		0.274	ND	0.591	Pc	0.274	ND	0.995	
SPRSEA-6-B-20250904	C55482	0.188	ND	0.274	ND	0.274	ND,Pc	0.274	ND	0.242	ND
SPRSEA-7-S-20250904	B53254	0.968		0.274	ND	0.464	J,Pc	0.274	ND	0.924	
SPRSEA-8-S-20250904	C17220	0.914		0.274	ND	0.605	Pc	0.274	ND	1.18	
SPRSEA-9-S-20250904	C33447	1.12		0.527	J	1.62	Pc	0.695		2.55	
SPRSEA-10-S-20250904	C33770	3.8		2.04		7.27	Pc	2.99		11.1	
SPRSEA-11-S-20250904	C70575	2.8		1.48		5.91	Pc	2.34		8.02	
SPRSEA-12-S-20250904	C69685	2.52		1.37		5.33	Pc	2.16		7.36	
SPRSEA-12-D-20250904	C43700	2.4		1.16		4.16	Pc	1.66		6.71	
SPRSEA-12-B-20250904	C40528	0.188	ND	0.274	ND	0.274	ND,Pc	0.274	ND	0.242	ND
Quarter 3, 2025 Maximum		4.68		3.18		12.2		4.91		18.8	
Quarter 3, 2025 Average		1.05		0.71		2.26		0.92		3.25	
Rolling Annual Maximum		4.68		27.4		119		35.7		18.80	
Rolling Annual Average		0.865		2.12		8.39		2.78		2.78	

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

SAMPLE ID	SAMPLE LOC.	COMPOUND NAME	SAMPLE TYPE	RESULT 3	RESULT UNITS3	MDL3	MDL UNITS3	LAB FLAGS	DETECT FLAG	SAMPLE START DATE	SAMPLE START TIME	SAMPLE END DATE	SAMPLE END TIME
SPRSEA-1-S-20250626	1	Benzene	Sample	0.436	ug/m3	0.186	ug/m3		Y	6/26/2025	11:30	7/10/2025	11:50
SPRSEA-1-S-20250626	1	Ethylbenzene	Sample	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	11:30	7/10/2025	11:50
SPRSEA-1-S-20250626	1	m-/p-Xylenes	Sample	0.456	ug/m3	0.272	ug/m3	J,Pc	Y	6/26/2025	11:30	7/10/2025	11:50
SPRSEA-1-S-20250626	1	o-Xylene	Sample	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	11:30	7/10/2025	11:50
SPRSEA-1-S-20250626	1	Toluene	Sample	0.721	ug/m3	0.24	ug/m3		Y	6/26/2025	11:30	7/10/2025	11:50
SPRSEA-2-S-20250626	2	Benzene	Sample	0.514	ug/m3	0.186	ug/m3		Y	6/26/2025	11:20	7/10/2025	11:40
SPRSEA-2-S-20250626	2	Ethylbenzene	Sample	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	11:20	7/10/2025	11:40
SPRSEA-2-S-20250626	2	m-/p-Xylenes	Sample	0.563	ug/m3	0.272	ug/m3	Pc	Y	6/26/2025	11:20	7/10/2025	11:40
SPRSEA-2-S-20250626	2	o-Xylene	Sample	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	11:20	7/10/2025	11:40
SPRSEA-2-S-20250626	2	Toluene	Sample	1.04	ug/m3	0.24	ug/m3		Y	6/26/2025	11:20	7/10/2025	11:40
SPRSEA-3-S-20250626	3	Benzene	Sample	0.551	ug/m3	0.186	ug/m3		Y	6/26/2025	11:10	7/10/2025	11:30
SPRSEA-3-S-20250626	3	Ethylbenzene	Sample	0.3	ug/m3	0.272	ug/m3	J	Y	6/26/2025	11:10	7/10/2025	11:30
SPRSEA-3-S-20250626	3	m-/p-Xylenes	Sample	1.24	ug/m3	0.272	ug/m3	Pc	Y	6/26/2025	11:10	7/10/2025	11:30
SPRSEA-3-S-20250626	3	o-Xylene	Sample	0.582	ug/m3	0.272	ug/m3		Y	6/26/2025	11:10	7/10/2025	11:30
SPRSEA-3-S-20250626	3	Toluene	Sample	1.15	ug/m3	0.24	ug/m3		Y	6/26/2025	11:10	7/10/2025	11:30
SPRSEA-4-S-20250626	4	Benzene	Sample	0.515	ug/m3	0.186	ug/m3		Y	6/26/2025	11:00	7/10/2025	11:20
SPRSEA-4-S-20250626	4	Ethylbenzene	Sample	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	11:00	7/10/2025	11:20
SPRSEA-4-S-20250626	4	m-/p-Xylenes	Sample	0.504	ug/m3	0.272	ug/m3	J,Pc	Y	6/26/2025	11:00	7/10/2025	11:20
SPRSEA-4-S-20250626	4	o-Xylene	Sample	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	11:00	7/10/2025	11:20
SPRSEA-4-S-20250626	4	Toluene	Sample	0.893	ug/m3	0.24	ug/m3		Y	6/26/2025	11:00	7/10/2025	11:20
SPRSEA-5-S-20250626	5	Benzene	Sample	0.526	ug/m3	0.186	ug/m3		Y	6/26/2025	12:10	7/10/2025	12:20
SPRSEA-5-S-20250626	5	Ethylbenzene	Sample	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	12:10	7/10/2025	12:20
SPRSEA-5-S-20250626	5	m-/p-Xylenes	Sample	0.466	ug/m3	0.272	ug/m3	J,Pc	Y	6/26/2025	12:10	7/10/2025	12:20
SPRSEA-5-S-20250626	5	o-Xylene	Sample	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	12:10	7/10/2025	12:20
SPRSEA-5-S-20250626	5	Toluene	Sample	1.01	ug/m3	0.24	ug/m3		Y	6/26/2025	12:10	7/10/2025	12:20
SPRSEA-6-S-20250626	6	Benzene	Sample	0.456	ug/m3	0.186	ug/m3		Y	6/26/2025	12:20	7/10/2025	12:30
SPRSEA-6-S-20250626	6	Ethylbenzene	Sample	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	12:20	7/10/2025	12:30
SPRSEA-6-S-20250626	6	m-/p-Xylenes	Sample	0.463	ug/m3	0.272	ug/m3	J,Pc	Y	6/26/2025	12:20	7/10/2025	12:30
SPRSEA-6-S-20250626	6	o-Xylene	Sample	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	12:20	7/10/2025	12:30
SPRSEA-6-S-20250626	6	Toluene	Sample	0.829	ug/m3	0.24	ug/m3		Y	6/26/2025	12:20	7/10/2025	12:30
SPRSEA-6-D-20250626	6	Benzene	Duplicate	0.458	ug/m3	0.186	ug/m3		Y	6/26/2025	12:20	7/10/2025	12:30
SPRSEA-6-D-20250626	6	Ethylbenzene	Duplicate	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	12:20	7/10/2025	12:30

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

SAMPLE ID	SAMPLE LOC.	COMPOUND NAME	SAMPLE TYPE	RESULT 3	RESULT UNITS3	MDL3	MDL UNITS3	LAB FLAGS	DETECT FLAG	SAMPLE START DATE	SAMPLE START TIME	SAMPLE END DATE	SAMPLE END TIME
SPRSEA-6-D-20250626	6	m-/p-Xylenes	Duplicate	0.326	ug/m3	0.272	ug/m3	J,Pc	Y	6/26/2025	12:20	7/10/2025	12:30
SPRSEA-6-D-20250626	6	o-Xylene	Duplicate	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	12:20	7/10/2025	12:30
SPRSEA-6-D-20250626	6	Toluene	Duplicate	0.742	ug/m3	0.24	ug/m3		Y	6/26/2025	12:20	7/10/2025	12:30
SPRSEA-6-B-20250626	6	Benzene	Blank	<0.186	ug/m3	0.186	ug/m3	ND	N	6/26/2025	12:20	7/10/2025	12:30
SPRSEA-6-B-20250626	6	Ethylbenzene	Blank	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	12:20	7/10/2025	12:30
SPRSEA-6-B-20250626	6	m-/p-Xylenes	Blank	<0.272	ug/m3	0.272	ug/m3	ND,Pc	N	6/26/2025	12:20	7/10/2025	12:30
SPRSEA-6-B-20250626	6	o-Xylene	Blank	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	12:20	7/10/2025	12:30
SPRSEA-6-B-20250626	6	Toluene	Blank	<0.24	ug/m3	0.24	ug/m3	ND	N	6/26/2025	12:20	7/10/2025	12:30
SPRSEA-7-S-20250626	7	Benzene	Sample	0.431	ug/m3	0.186	ug/m3		Y	6/26/2025	12:30	7/10/2025	12:40
SPRSEA-7-S-20250626	7	Ethylbenzene	Sample	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	12:30	7/10/2025	12:40
SPRSEA-7-S-20250626	7	m-/p-Xylenes	Sample	0.367	ug/m3	0.272	ug/m3	J,Pc	Y	6/26/2025	12:30	7/10/2025	12:40
SPRSEA-7-S-20250626	7	o-Xylene	Sample	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	12:30	7/10/2025	12:40
SPRSEA-7-S-20250626	7	Toluene	Sample	0.671	ug/m3	0.24	ug/m3		Y	6/26/2025	12:30	7/10/2025	12:40
SPRSEA-8-S-20250626	8	Benzene	Sample	0.421	ug/m3	0.186	ug/m3		Y	6/26/2025	12:40	7/10/2025	12:50
SPRSEA-8-S-20250626	8	Ethylbenzene	Sample	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	12:40	7/10/2025	12:50
SPRSEA-8-S-20250626	8	m-/p-Xylenes	Sample	0.488	ug/m3	0.272	ug/m3	J,Pc	Y	6/26/2025	12:40	7/10/2025	12:50
SPRSEA-8-S-20250626	8	o-Xylene	Sample	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	12:40	7/10/2025	12:50
SPRSEA-8-S-20250626	8	Toluene	Sample	0.919	ug/m3	0.24	ug/m3		Y	6/26/2025	12:40	7/10/2025	12:50
SPRSEA-9-S-20250626	9	Benzene	Sample	0.682	ug/m3	0.186	ug/m3		Y	6/26/2025	12:50	7/10/2025	13:00
SPRSEA-9-S-20250626	9	Ethylbenzene	Sample	0.296	ug/m3	0.272	ug/m3	J	Y	6/26/2025	12:50	7/10/2025	13:00
SPRSEA-9-S-20250626	9	m-/p-Xylenes	Sample	1.02	ug/m3	0.272	ug/m3	Pc	Y	6/26/2025	12:50	7/10/2025	13:00
SPRSEA-9-S-20250626	9	o-Xylene	Sample	0.391	ug/m3	0.272	ug/m3	J	Y	6/26/2025	12:50	7/10/2025	13:00
SPRSEA-9-S-20250626	9	Toluene	Sample	1.75	ug/m3	0.24	ug/m3		Y	6/26/2025	12:50	7/10/2025	13:00
SPRSEA-10-S-20250626	10	Benzene	Sample	2.3	ug/m3	0.186	ug/m3		Y	6/26/2025	13:00	7/10/2025	13:10
SPRSEA-10-S-20250626	10	Ethylbenzene	Sample	1.59	ug/m3	0.272	ug/m3		Y	6/26/2025	13:00	7/10/2025	13:10
SPRSEA-10-S-20250626	10	m-/p-Xylenes	Sample	5.05	ug/m3	0.272	ug/m3	Pc	Y	6/26/2025	13:00	7/10/2025	13:10
SPRSEA-10-S-20250626	10	o-Xylene	Sample	2.09	ug/m3	0.272	ug/m3		Y	6/26/2025	13:00	7/10/2025	13:10
SPRSEA-10-S-20250626	10	Toluene	Sample	9.7	ug/m3	0.24	ug/m3		Y	6/26/2025	13:00	7/10/2025	13:10
SPRSEA-11-S-20250626	11	Benzene	Sample	1.18	ug/m3	0.186	ug/m3		Y	6/26/2025	11:55	7/10/2025	12:10
SPRSEA-11-S-20250626	11	Ethylbenzene	Sample	0.813	ug/m3	0.272	ug/m3		Y	6/26/2025	11:55	7/10/2025	12:10
SPRSEA-11-S-20250626	11	m-/p-Xylenes	Sample	2.94	ug/m3	0.272	ug/m3	Pc	Y	6/26/2025	11:55	7/10/2025	12:10

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

SAMPLE ID	SAMPLE LOC.	COMPOUND NAME	SAMPLE TYPE	RESULT 3	RESULT UNITS3	MDL3	MDL UNITS3	LAB FLAGS	DETECT FLAG	SAMPLE START DATE	SAMPLE START TIME	SAMPLE END DATE	SAMPLE END TIME
SPRSEA-11-S-20250626	11	o-Xylene	Sample	1.15	ug/m3	0.272	ug/m3		Y	6/26/2025	11:55	7/10/2025	12:10
SPRSEA-11-S-20250626	11	Toluene	Sample	4.07	ug/m3	0.24	ug/m3		Y	6/26/2025	11:55	7/10/2025	12:10
SPRSEA-12-S-20250626	12	Benzene	Sample	0.754	ug/m3	0.186	ug/m3		Y	6/26/2025	11:45	7/10/2025	12:00
SPRSEA-12-S-20250626	12	Ethylbenzene	Sample	0.407	ug/m3	0.272	ug/m3	J	Y	6/26/2025	11:45	7/10/2025	12:00
SPRSEA-12-S-20250626	12	m-/p-Xylenes	Sample	1.33	ug/m3	0.272	ug/m3	Pc	Y	6/26/2025	11:45	7/10/2025	12:00
SPRSEA-12-S-20250626	12	o-Xylene	Sample	0.552	ug/m3	0.272	ug/m3	J	Y	6/26/2025	11:45	7/10/2025	12:00
SPRSEA-12-S-20250626	12	Toluene	Sample	2.14	ug/m3	0.24	ug/m3		Y	6/26/2025	11:45	7/10/2025	12:00
SPRSEA-12-D-20250626	12	Benzene	Duplicate	0.675	ug/m3	0.186	ug/m3		Y	6/26/2025	11:45	7/10/2025	12:00
SPRSEA-12-D-20250626	12	Ethylbenzene	Duplicate	0.402	ug/m3	0.272	ug/m3	J	Y	6/26/2025	11:45	7/10/2025	12:00
SPRSEA-12-D-20250626	12	m-/p-Xylenes	Duplicate	1.41	ug/m3	0.272	ug/m3	Pc	Y	6/26/2025	11:45	7/10/2025	12:00
SPRSEA-12-D-20250626	12	o-Xylene	Duplicate	0.547	ug/m3	0.272	ug/m3	J	Y	6/26/2025	11:45	7/10/2025	12:00
SPRSEA-12-D-20250626	12	Toluene	Duplicate	2.05	ug/m3	0.24	ug/m3		Y	6/26/2025	11:45	7/10/2025	12:00
SPRSEA-12-B-20250626	12	Benzene	Blank	<0.186	ug/m3	0.186	ug/m3	ND	N	6/26/2025	11:45	7/10/2025	12:00
SPRSEA-12-B-20250626	12	Ethylbenzene	Blank	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	11:45	7/10/2025	12:00
SPRSEA-12-B-20250626	12	m-/p-Xylenes	Blank	<0.272	ug/m3	0.272	ug/m3	ND,Pc	N	6/26/2025	11:45	7/10/2025	12:00
SPRSEA-12-B-20250626	12	o-Xylene	Blank	<0.272	ug/m3	0.272	ug/m3	ND	N	6/26/2025	11:45	7/10/2025	12:00
SPRSEA-12-B-20250626	12	Toluene	Blank	<0.24	ug/m3	0.24	ug/m3	ND	N	6/26/2025	11:45	7/10/2025	12:00
SPRSEA-1-S-20250710	1	Benzene	Sample	0.437	ug/m3	0.186	ug/m3		Y	7/10/2025	11:50	7/24/2025	12:15
SPRSEA-1-S-20250710	1	Ethylbenzene	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	7/10/2025	11:50	7/24/2025	12:15
SPRSEA-1-S-20250710	1	m-/p-Xylenes	Sample	0.495	ug/m3	0.271	ug/m3	J	Y	7/10/2025	11:50	7/24/2025	12:15
SPRSEA-1-S-20250710	1	o-Xylene	Sample	<0.271	ug/m3	0.271	ug/m3	ND,Pc	N	7/10/2025	11:50	7/24/2025	12:15
SPRSEA-1-S-20250710	1	Toluene	Sample	1.05	ug/m3	0.24	ug/m3	P	Y	7/10/2025	11:50	7/24/2025	12:15
SPRSEA-2-S-20250710	2	Benzene	Sample	0.61	ug/m3	0.186	ug/m3		Y	7/10/2025	11:40	7/24/2025	12:05
SPRSEA-2-S-20250710	2	Ethylbenzene	Sample	0.475	ug/m3	0.271	ug/m3	J	Y	7/10/2025	11:40	7/24/2025	12:05
SPRSEA-2-S-20250710	2	m-/p-Xylenes	Sample	0.801	ug/m3	0.271	ug/m3		Y	7/10/2025	11:40	7/24/2025	12:05
SPRSEA-2-S-20250710	2	o-Xylene	Sample	0.336	ug/m3	0.271	ug/m3	J,Pc	Y	7/10/2025	11:40	7/24/2025	12:05
SPRSEA-2-S-20250710	2	Toluene	Sample	1.65	ug/m3	0.24	ug/m3	P	Y	7/10/2025	11:40	7/24/2025	12:05
SPRSEA-3-S-20250710	3	Benzene	Sample	0.516	ug/m3	0.186	ug/m3		Y	7/10/2025	11:30	7/24/2025	11:55
SPRSEA-3-S-20250710	3	Ethylbenzene	Sample	0.294	ug/m3	0.271	ug/m3	J	Y	7/10/2025	11:30	7/24/2025	11:55
SPRSEA-3-S-20250710	3	m-/p-Xylenes	Sample	0.956	ug/m3	0.271	ug/m3		Y	7/10/2025	11:30	7/24/2025	11:55
SPRSEA-3-S-20250710	3	o-Xylene	Sample	0.427	ug/m3	0.271	ug/m3	J,Pc	Y	7/10/2025	11:30	7/24/2025	11:55

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

SAMPLE ID	SAMPLE LOC.	COMPOUND NAME	SAMPLE TYPE	RESULT 3	RESULT UNITS3	MDL3	MDL UNITS3	LAB FLAGS	DETECT FLAG	SAMPLE START DATE	SAMPLE START TIME	SAMPLE END DATE	SAMPLE END TIME
SPRSEA-3-S-20250710	3	Toluene	Sample	1.49	ug/m3	0.24	ug/m3	P	Y	7/10/2025	11:30	7/24/2025	11:55
SPRSEA-4-S-20250710	4	Benzene	Sample	0.484	ug/m3	0.186	ug/m3		Y	7/10/2025	11:20	7/24/2025	11:45
SPRSEA-4-S-20250710	4	Ethylbenzene	Sample	0.289	ug/m3	0.271	ug/m3	J	Y	7/10/2025	11:20	7/24/2025	11:45
SPRSEA-4-S-20250710	4	m-/p-Xylenes	Sample	0.672	ug/m3	0.271	ug/m3		Y	7/10/2025	11:20	7/24/2025	11:45
SPRSEA-4-S-20250710	4	o-Xylene	Sample	<0.271	ug/m3	0.271	ug/m3	ND,Pc	N	7/10/2025	11:20	7/24/2025	11:45
SPRSEA-4-S-20250710	4	Toluene	Sample	1.38	ug/m3	0.24	ug/m3	P	Y	7/10/2025	11:20	7/24/2025	11:45
SPRSEA-5-S-20250710	5	Benzene	Sample	0.383	ug/m3	0.186	ug/m3		Y	7/10/2025	12:20	7/24/2025	12:45
SPRSEA-5-S-20250710	5	Ethylbenzene	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	7/10/2025	12:20	7/24/2025	12:45
SPRSEA-5-S-20250710	5	m-/p-Xylenes	Sample	0.617	ug/m3	0.271	ug/m3		Y	7/10/2025	12:20	7/24/2025	12:45
SPRSEA-5-S-20250710	5	o-Xylene	Sample	<0.271	ug/m3	0.271	ug/m3	ND,Pc	N	7/10/2025	12:20	7/24/2025	12:45
SPRSEA-5-S-20250710	5	Toluene	Sample	1.01	ug/m3	0.24	ug/m3	P	Y	7/10/2025	12:20	7/24/2025	12:45
SPRSEA-6-S-20250710	6	Benzene	Sample	0.362	ug/m3	0.186	ug/m3	J	Y	7/10/2025	12:30	7/24/2025	12:55
SPRSEA-6-S-20250710	6	Ethylbenzene	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	7/10/2025	12:30	7/24/2025	12:55
SPRSEA-6-S-20250710	6	m-/p-Xylenes	Sample	0.67	ug/m3	0.271	ug/m3		Y	7/10/2025	12:30	7/24/2025	12:55
SPRSEA-6-S-20250710	6	o-Xylene	Sample	<0.271	ug/m3	0.271	ug/m3	ND,Pc	N	7/10/2025	12:30	7/24/2025	12:55
SPRSEA-6-S-20250710	6	Toluene	Sample	0.874	ug/m3	0.24	ug/m3	P	Y	7/10/2025	12:30	7/24/2025	12:55
SPRSEA-6-D-20250710	6	Benzene	Duplicate	0.451	ug/m3	0.186	ug/m3		Y	7/10/2025	12:30	7/24/2025	12:55
SPRSEA-6-D-20250710	6	Ethylbenzene	Duplicate	0.309	ug/m3	0.271	ug/m3	J	Y	7/10/2025	12:30	7/24/2025	12:55
SPRSEA-6-D-20250710	6	m-/p-Xylenes	Duplicate	0.543	ug/m3	0.271	ug/m3		Y	7/10/2025	12:30	7/24/2025	12:55
SPRSEA-6-D-20250710	6	o-Xylene	Duplicate	<0.271	ug/m3	0.271	ug/m3	ND,Pc	N	7/10/2025	12:30	7/24/2025	12:55
SPRSEA-6-D-20250710	6	Toluene	Duplicate	1.23	ug/m3	0.24	ug/m3	P	Y	7/10/2025	12:30	7/24/2025	12:55
SPRSEA-6-B-20250710	6	Benzene	Blank	<0.186	ug/m3	0.186	ug/m3	ND	N	7/10/2025	12:30	7/24/2025	12:55
SPRSEA-6-B-20250710	6	Ethylbenzene	Blank	<0.271	ug/m3	0.271	ug/m3	ND	N	7/10/2025	12:30	7/24/2025	12:55
SPRSEA-6-B-20250710	6	m-/p-Xylenes	Blank	<0.271	ug/m3	0.271	ug/m3	ND	N	7/10/2025	12:30	7/24/2025	12:55
SPRSEA-6-B-20250710	6	o-Xylene	Blank	<0.271	ug/m3	0.271	ug/m3	ND,Pc	N	7/10/2025	12:30	7/24/2025	12:55
SPRSEA-6-B-20250710	6	Toluene	Blank	<0.24	ug/m3	0.24	ug/m3	ND,P	N	7/10/2025	12:30	7/24/2025	12:55
SPRSEA-7-S-20250710	7	Benzene	Sample	0.393	ug/m3	0.186	ug/m3		Y	7/10/2025	12:40	7/24/2025	13:05
SPRSEA-7-S-20250710	7	Ethylbenzene	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	7/10/2025	12:40	7/24/2025	13:05
SPRSEA-7-S-20250710	7	m-/p-Xylenes	Sample	0.341	ug/m3	0.271	ug/m3	J	Y	7/10/2025	12:40	7/24/2025	13:05
SPRSEA-7-S-20250710	7	o-Xylene	Sample	<0.271	ug/m3	0.271	ug/m3	ND,Pc	N	7/10/2025	12:40	7/24/2025	13:05
SPRSEA-7-S-20250710	7	Toluene	Sample	0.903	ug/m3	0.24	ug/m3	P	Y	7/10/2025	12:40	7/24/2025	13:05

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

SAMPLE ID	SAMPLE LOC.	COMPOUND NAME	SAMPLE TYPE	RESULT 3	RESULT UNITS3	MDL3	MDL UNITS3	LAB FLAGS	DETECT FLAG	SAMPLE START DATE	SAMPLE START TIME	SAMPLE END DATE	SAMPLE END TIME
SPRSEA-8-S-20250710	8	Benzene	Sample	0.577	ug/m3	0.186	ug/m3		Y	7/10/2025	12:50	7/24/2025	13:15
SPRSEA-8-S-20250710	8	Ethylbenzene	Sample	0.342	ug/m3	0.271	ug/m3	J	Y	7/10/2025	12:50	7/24/2025	13:15
SPRSEA-8-S-20250710	8	m-/p-Xylenes	Sample	0.652	ug/m3	0.271	ug/m3		Y	7/10/2025	12:50	7/24/2025	13:15
SPRSEA-8-S-20250710	8	o-Xylene	Sample	0.273	ug/m3	0.271	ug/m3	J,Pc	Y	7/10/2025	12:50	7/24/2025	13:15
SPRSEA-8-S-20250710	8	Toluene	Sample	1.58	ug/m3	0.24	ug/m3	P	Y	7/10/2025	12:50	7/24/2025	13:15
SPRSEA-9-S-20250710	9	Benzene	Sample	1.09	ug/m3	0.186	ug/m3		Y	7/10/2025	13:00	7/24/2025	13:25
SPRSEA-9-S-20250710	9	Ethylbenzene	Sample	0.784	ug/m3	0.271	ug/m3		Y	7/10/2025	13:00	7/24/2025	13:25
SPRSEA-9-S-20250710	9	m-/p-Xylenes	Sample	1.81	ug/m3	0.271	ug/m3		Y	7/10/2025	13:00	7/24/2025	13:25
SPRSEA-9-S-20250710	9	o-Xylene	Sample	0.786	ug/m3	0.271	ug/m3	Pc	Y	7/10/2025	13:00	7/24/2025	13:25
SPRSEA-9-S-20250710	9	Toluene	Sample	3.65	ug/m3	0.24	ug/m3	P	Y	7/10/2025	13:00	7/24/2025	13:25
SPRSEA-10-S-20250710	10	Benzene	Sample	3.67	ug/m3	0.186	ug/m3		Y	7/10/2025	13:10	7/24/2025	13:35
SPRSEA-10-S-20250710	10	Ethylbenzene	Sample	3.13	ug/m3	0.271	ug/m3		Y	7/10/2025	13:10	7/24/2025	13:35
SPRSEA-10-S-20250710	10	m-/p-Xylenes	Sample	10.5	ug/m3	0.271	ug/m3		Y	7/10/2025	13:10	7/24/2025	13:35
SPRSEA-10-S-20250710	10	o-Xylene	Sample	4.31	ug/m3	0.271	ug/m3	Pc	Y	7/10/2025	13:10	7/24/2025	13:35
SPRSEA-10-S-20250710	10	Toluene	Sample	18.8	ug/m3	0.24	ug/m3	P	Y	7/10/2025	13:10	7/24/2025	13:35
SPRSEA-11-S-20250710	11	Benzene	Sample	1.21	ug/m3	0.186	ug/m3		Y	7/10/2025	12:10	7/24/2025	12:35
SPRSEA-11-S-20250710	11	Ethylbenzene	Sample	1.28	ug/m3	0.271	ug/m3		Y	7/10/2025	12:10	7/24/2025	12:35
SPRSEA-11-S-20250710	11	m-/p-Xylenes	Sample	3.38	ug/m3	0.271	ug/m3		Y	7/10/2025	12:10	7/24/2025	12:35
SPRSEA-11-S-20250710	11	o-Xylene	Sample	1.1	ug/m3	0.271	ug/m3	Pc	Y	7/10/2025	12:10	7/24/2025	12:35
SPRSEA-11-S-20250710	11	Toluene	Sample	4.05	ug/m3	0.24	ug/m3	P	Y	7/10/2025	12:10	7/24/2025	12:35
SPRSEA-12-S-20250710	12	Benzene	Sample	0.807	ug/m3	0.186	ug/m3		Y	7/10/2025	12:00	7/24/2025	12:25
SPRSEA-12-S-20250710	12	Ethylbenzene	Sample	0.745	ug/m3	0.271	ug/m3		Y	7/10/2025	12:00	7/24/2025	12:25
SPRSEA-12-S-20250710	12	m-/p-Xylenes	Sample	2.14	ug/m3	0.271	ug/m3		Y	7/10/2025	12:00	7/24/2025	12:25
SPRSEA-12-S-20250710	12	o-Xylene	Sample	0.761	ug/m3	0.271	ug/m3	Pc	Y	7/10/2025	12:00	7/24/2025	12:25
SPRSEA-12-S-20250710	12	Toluene	Sample	2.92	ug/m3	0.24	ug/m3	P	Y	7/10/2025	12:00	7/24/2025	12:25
SPRSEA-12-D-20250710	12	Benzene	Duplicate	0.821	ug/m3	0.186	ug/m3		Y	7/10/2025	12:00	7/24/2025	12:25
SPRSEA-12-D-20250710	12	Ethylbenzene	Duplicate	0.871	ug/m3	0.271	ug/m3		Y	7/10/2025	12:00	7/24/2025	12:25
SPRSEA-12-D-20250710	12	m-/p-Xylenes	Duplicate	2.84	ug/m3	0.271	ug/m3		Y	7/10/2025	12:00	7/24/2025	12:25
SPRSEA-12-D-20250710	12	o-Xylene	Duplicate	1.06	ug/m3	0.271	ug/m3	Pc	Y	7/10/2025	12:00	7/24/2025	12:25
SPRSEA-12-D-20250710	12	Toluene	Duplicate	3.22	ug/m3	0.24	ug/m3	P	Y	7/10/2025	12:00	7/24/2025	12:25
SPRSEA-12-B-20250710	12	Benzene	Blank	<0.186	ug/m3	0.186	ug/m3	ND	N	7/10/2025	12:00	7/24/2025	12:25

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

SAMPLE ID	SAMPLE LOC.	COMPOUND NAME	SAMPLE TYPE	RESULT 3	RESULT UNITS3	MDL3	MDL UNITS3	LAB FLAGS	DETECT FLAG	SAMPLE START DATE	SAMPLE START TIME	SAMPLE END DATE	SAMPLE END TIME
SPRSEA-12-B-20250710	12	Ethylbenzene	Blank	<0.271	ug/m3	0.271	ug/m3	ND	N	7/10/2025	12:00	7/24/2025	12:25
SPRSEA-12-B-20250710	12	m-/p-Xylenes	Blank	<0.271	ug/m3	0.271	ug/m3	ND	N	7/10/2025	12:00	7/24/2025	12:25
SPRSEA-12-B-20250710	12	o-Xylene	Blank	<0.271	ug/m3	0.271	ug/m3	ND,Pc	N	7/10/2025	12:00	7/24/2025	12:25
SPRSEA-12-B-20250710	12	Toluene	Blank	<0.24	ug/m3	0.24	ug/m3	ND,P	N	7/10/2025	12:00	7/24/2025	12:25
SPRSEA-1-S-20250724	1	Benzene	Sample	0.786	ug/m3	0.186	ug/m3		Y	7/24/2025	12:15	8/7/2025	12:00
SPRSEA-1-S-20250724	1	Ethylbenzene	Sample	0.405	ug/m3	0.271	ug/m3	J	Y	7/24/2025	12:15	8/7/2025	12:00
SPRSEA-1-S-20250724	1	m-/p-Xylenes	Sample	1.56	ug/m3	0.271	ug/m3		Y	7/24/2025	12:15	8/7/2025	12:00
SPRSEA-1-S-20250724	1	o-Xylene	Sample	0.536	ug/m3	0.271	ug/m3	J	Y	7/24/2025	12:15	8/7/2025	12:00
SPRSEA-1-S-20250724	1	Toluene	Sample	1.64	ug/m3	0.24	ug/m3		Y	7/24/2025	12:15	8/7/2025	12:00
SPRSEA-2-S-20250724	2	Benzene	Sample	0.756	ug/m3	0.186	ug/m3		Y	7/24/2025	12:05	8/7/2025	11:50
SPRSEA-2-S-20250724	2	Ethylbenzene	Sample	0.597	ug/m3	0.271	ug/m3		Y	7/24/2025	12:05	8/7/2025	11:50
SPRSEA-2-S-20250724	2	m-/p-Xylenes	Sample	1.46	ug/m3	0.271	ug/m3		Y	7/24/2025	12:05	8/7/2025	11:50
SPRSEA-2-S-20250724	2	o-Xylene	Sample	0.585	ug/m3	0.271	ug/m3		Y	7/24/2025	12:05	8/7/2025	11:50
SPRSEA-2-S-20250724	2	Toluene	Sample	1.76	ug/m3	0.24	ug/m3		Y	7/24/2025	12:05	8/7/2025	11:50
SPRSEA-3-S-20250724	3	Benzene	Sample	0.813	ug/m3	0.186	ug/m3		Y	7/24/2025	11:55	8/7/2025	11:40
SPRSEA-3-S-20250724	3	Ethylbenzene	Sample	0.354	ug/m3	0.271	ug/m3	J	Y	7/24/2025	11:55	8/7/2025	11:40
SPRSEA-3-S-20250724	3	m-/p-Xylenes	Sample	1.29	ug/m3	0.271	ug/m3		Y	7/24/2025	11:55	8/7/2025	11:40
SPRSEA-3-S-20250724	3	o-Xylene	Sample	0.576	ug/m3	0.271	ug/m3		Y	7/24/2025	11:55	8/7/2025	11:40
SPRSEA-3-S-20250724	3	Toluene	Sample	1.6	ug/m3	0.24	ug/m3		Y	7/24/2025	11:55	8/7/2025	11:40
SPRSEA-4-S-20250724	4	Benzene	Sample	0.567	ug/m3	0.186	ug/m3		Y	7/24/2025	11:45	8/7/2025	11:30
SPRSEA-4-S-20250724	4	Ethylbenzene	Sample	0.278	ug/m3	0.271	ug/m3	J	Y	7/24/2025	11:45	8/7/2025	11:30
SPRSEA-4-S-20250724	4	m-/p-Xylenes	Sample	1.04	ug/m3	0.271	ug/m3		Y	7/24/2025	11:45	8/7/2025	11:30
SPRSEA-4-S-20250724	4	o-Xylene	Sample	0.463	ug/m3	0.271	ug/m3	J	Y	7/24/2025	11:45	8/7/2025	11:30
SPRSEA-4-S-20250724	4	Toluene	Sample	1.01	ug/m3	0.24	ug/m3		Y	7/24/2025	11:45	8/7/2025	11:30
SPRSEA-5-S-20250724	5	Benzene	Sample	0.657	ug/m3	0.186	ug/m3		Y	7/24/2025	12:45	8/7/2025	12:30
SPRSEA-5-S-20250724	5	Ethylbenzene	Sample	0.332	ug/m3	0.271	ug/m3	J	Y	7/24/2025	12:45	8/7/2025	12:30
SPRSEA-5-S-20250724	5	m-/p-Xylenes	Sample	0.749	ug/m3	0.271	ug/m3		Y	7/24/2025	12:45	8/7/2025	12:30
SPRSEA-5-S-20250724	5	o-Xylene	Sample	0.352	ug/m3	0.271	ug/m3	J	Y	7/24/2025	12:45	8/7/2025	12:30
SPRSEA-5-S-20250724	5	Toluene	Sample	1.03	ug/m3	0.24	ug/m3		Y	7/24/2025	12:45	8/7/2025	12:30
SPRSEA-6-S-20250724	6	Benzene	Sample	0.632	ug/m3	0.186	ug/m3		Y	7/24/2025	12:55	8/7/2025	12:40
SPRSEA-6-S-20250724	6	Ethylbenzene	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	7/24/2025	12:55	8/7/2025	12:40

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

SAMPLE ID	SAMPLE LOC.	COMPOUND NAME	SAMPLE TYPE	RESULT 3	RESULT UNITS3	MDL3	MDL UNITS3	LAB FLAGS	DETECT FLAG	SAMPLE START DATE	SAMPLE START TIME	SAMPLE END DATE	SAMPLE END TIME
SPRSEA-6-S-20250724	6	m-/p-Xylenes	Sample	0.373	ug/m3	0.271	ug/m3	J	Y	7/24/2025	12:55	8/7/2025	12:40
SPRSEA-6-S-20250724	6	o-Xylene	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	7/24/2025	12:55	8/7/2025	12:40
SPRSEA-6-S-20250724	6	Toluene	Sample	0.774	ug/m3	0.24	ug/m3		Y	7/24/2025	12:55	8/7/2025	12:40
SPRSEA-6-D-20250724	6	Benzene	Duplicate	0.511	ug/m3	0.186	ug/m3		Y	7/24/2025	12:55	8/7/2025	12:40
SPRSEA-6-D-20250724	6	Ethylbenzene	Duplicate	<0.271	ug/m3	0.271	ug/m3	ND	N	7/24/2025	12:55	8/7/2025	12:40
SPRSEA-6-D-20250724	6	m-/p-Xylenes	Duplicate	0.437	ug/m3	0.271	ug/m3	J	Y	7/24/2025	12:55	8/7/2025	12:40
SPRSEA-6-D-20250724	6	o-Xylene	Duplicate	<0.271	ug/m3	0.271	ug/m3	ND	N	7/24/2025	12:55	8/7/2025	12:40
SPRSEA-6-D-20250724	6	Toluene	Duplicate	0.647	ug/m3	0.24	ug/m3		Y	7/24/2025	12:55	8/7/2025	12:40
SPRSEA-6-B-20250724	6	Benzene	Blank	<0.186	ug/m3	0.186	ug/m3	ND	N	7/24/2025	12:55	8/7/2025	12:40
SPRSEA-6-B-20250724	6	Ethylbenzene	Blank	<0.271	ug/m3	0.271	ug/m3	ND	N	7/24/2025	12:55	8/7/2025	12:40
SPRSEA-6-B-20250724	6	m-/p-Xylenes	Blank	<0.271	ug/m3	0.271	ug/m3	ND	N	7/24/2025	12:55	8/7/2025	12:40
SPRSEA-6-B-20250724	6	o-Xylene	Blank	<0.271	ug/m3	0.271	ug/m3	ND	N	7/24/2025	12:55	8/7/2025	12:40
SPRSEA-6-B-20250724	6	Toluene	Blank	<0.24	ug/m3	0.24	ug/m3	ND	N	7/24/2025	12:55	8/7/2025	12:40
SPRSEA-7-S-20250724	7	Benzene	Sample	0.679	ug/m3	0.186	ug/m3		Y	7/24/2025	13:05	8/7/2025	12:50
SPRSEA-7-S-20250724	7	Ethylbenzene	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	7/24/2025	13:05	8/7/2025	12:50
SPRSEA-7-S-20250724	7	m-/p-Xylenes	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	7/24/2025	13:05	8/7/2025	12:50
SPRSEA-7-S-20250724	7	o-Xylene	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	7/24/2025	13:05	8/7/2025	12:50
SPRSEA-7-S-20250724	7	Toluene	Sample	0.609	ug/m3	0.24	ug/m3		Y	7/24/2025	13:05	8/7/2025	12:50
SPRSEA-8-S-20250724	8	Benzene	Sample	0.541	ug/m3	0.186	ug/m3		Y	7/24/2025	13:15	8/7/2025	13:00
SPRSEA-8-S-20250724	8	Ethylbenzene	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	7/24/2025	13:15	8/7/2025	13:00
SPRSEA-8-S-20250724	8	m-/p-Xylenes	Sample	0.449	ug/m3	0.271	ug/m3	J	Y	7/24/2025	13:15	8/7/2025	13:00
SPRSEA-8-S-20250724	8	o-Xylene	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	7/24/2025	13:15	8/7/2025	13:00
SPRSEA-8-S-20250724	8	Toluene	Sample	0.662	ug/m3	0.24	ug/m3		Y	7/24/2025	13:15	8/7/2025	13:00
SPRSEA-9-S-20250724	9	Benzene	Sample	0.691	ug/m3	0.186	ug/m3		Y	7/24/2025	13:25	8/7/2025	13:10
SPRSEA-9-S-20250724	9	Ethylbenzene	Sample	0.434	ug/m3	0.271	ug/m3	J	Y	7/24/2025	13:25	8/7/2025	13:10
SPRSEA-9-S-20250724	9	m-/p-Xylenes	Sample	1.44	ug/m3	0.271	ug/m3		Y	7/24/2025	13:25	8/7/2025	13:10
SPRSEA-9-S-20250724	9	o-Xylene	Sample	0.598	ug/m3	0.271	ug/m3		Y	7/24/2025	13:25	8/7/2025	13:10
SPRSEA-9-S-20250724	9	Toluene	Sample	1.49	ug/m3	0.24	ug/m3		Y	7/24/2025	13:25	8/7/2025	13:10
SPRSEA-10-S-20250724	10	Benzene	Sample	1.72	ug/m3	0.186	ug/m3		Y	7/24/2025	13:35	8/7/2025	13:20
SPRSEA-10-S-20250724	10	Ethylbenzene	Sample	1.04	ug/m3	0.271	ug/m3		Y	7/24/2025	13:35	8/7/2025	13:20
SPRSEA-10-S-20250724	10	m-/p-Xylenes	Sample	3.82	ug/m3	0.271	ug/m3		Y	7/24/2025	13:35	8/7/2025	13:20

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

SAMPLE ID	SAMPLE LOC.	COMPOUND NAME	SAMPLE TYPE	RESULT 3	RESULT UNITS3	MDL3	MDL UNITS3	LAB FLAGS	DETECT FLAG	SAMPLE START DATE	SAMPLE START TIME	SAMPLE END DATE	SAMPLE END TIME
SPRSEA-10-S-20250724	10	o-Xylene	Sample	1.54	ug/m3	0.271	ug/m3		Y	7/24/2025	13:35	8/7/2025	13:20
SPRSEA-10-S-20250724	10	Toluene	Sample	5.41	ug/m3	0.24	ug/m3		Y	7/24/2025	13:35	8/7/2025	13:20
SPRSEA-11-S-20250724	11	Benzene	Sample	1.58	ug/m3	0.186	ug/m3		Y	7/24/2025	12:35	8/7/2025	12:20
SPRSEA-11-S-20250724	11	Ethylbenzene	Sample	1.02	ug/m3	0.271	ug/m3		Y	7/24/2025	12:35	8/7/2025	12:20
SPRSEA-11-S-20250724	11	m-/p-Xylenes	Sample	2.93	ug/m3	0.271	ug/m3		Y	7/24/2025	12:35	8/7/2025	12:20
SPRSEA-11-S-20250724	11	o-Xylene	Sample	1.14	ug/m3	0.271	ug/m3		Y	7/24/2025	12:35	8/7/2025	12:20
SPRSEA-11-S-20250724	11	Toluene	Sample	4.6	ug/m3	0.24	ug/m3		Y	7/24/2025	12:35	8/7/2025	12:20
SPRSEA-12-S-20250724	12	Benzene	Sample	1.31	ug/m3	0.186	ug/m3		Y	7/24/2025	12:25	8/7/2025	12:10
SPRSEA-12-S-20250724	12	Ethylbenzene	Sample	1.2	ug/m3	0.271	ug/m3		Y	7/24/2025	12:25	8/7/2025	12:10
SPRSEA-12-S-20250724	12	m-/p-Xylenes	Sample	4.68	ug/m3	0.271	ug/m3		Y	7/24/2025	12:25	8/7/2025	12:10
SPRSEA-12-S-20250724	12	o-Xylene	Sample	1.59	ug/m3	0.271	ug/m3		Y	7/24/2025	12:25	8/7/2025	12:10
SPRSEA-12-S-20250724	12	Toluene	Sample	3.75	ug/m3	0.24	ug/m3		Y	7/24/2025	12:25	8/7/2025	12:10
SPRSEA-12-D-20250724	12	Benzene	Duplicate	1.44	ug/m3	0.186	ug/m3		Y	7/24/2025	12:25	8/7/2025	12:10
SPRSEA-12-D-20250724	12	Ethylbenzene	Duplicate	1.38	ug/m3	0.271	ug/m3		Y	7/24/2025	12:25	8/7/2025	12:10
SPRSEA-12-D-20250724	12	m-/p-Xylenes	Duplicate	4.81	ug/m3	0.271	ug/m3		Y	7/24/2025	12:25	8/7/2025	12:10
SPRSEA-12-D-20250724	12	o-Xylene	Duplicate	1.65	ug/m3	0.271	ug/m3		Y	7/24/2025	12:25	8/7/2025	12:10
SPRSEA-12-D-20250724	12	Toluene	Duplicate	4.01	ug/m3	0.24	ug/m3		Y	7/24/2025	12:25	8/7/2025	12:10
SPRSEA-12-B-20250724	12	Benzene	Blank	<0.186	ug/m3	0.186	ug/m3	ND	N	7/24/2025	12:25	8/7/2025	12:10
SPRSEA-12-B-20250724	12	Ethylbenzene	Blank	<0.271	ug/m3	0.271	ug/m3	ND	N	7/24/2025	12:25	8/7/2025	12:10
SPRSEA-12-B-20250724	12	m-/p-Xylenes	Blank	<0.271	ug/m3	0.271	ug/m3	ND	N	7/24/2025	12:25	8/7/2025	12:10
SPRSEA-12-B-20250724	12	o-Xylene	Blank	<0.271	ug/m3	0.271	ug/m3	ND	N	7/24/2025	12:25	8/7/2025	12:10
SPRSEA-12-B-20250724	12	Toluene	Blank	<0.24	ug/m3	0.24	ug/m3	ND	N	7/24/2025	12:25	8/7/2025	12:10
SPRSEA-1-S-20250807	1	Benzene	Sample	0.501	ug/m3	0.186	ug/m3		Y	8/7/2025	12:00	8/21/2025	12:10
SPRSEA-1-S-20250807	1	Ethylbenzene	Sample	0.289	ug/m3	0.271	ug/m3	J,Pc	Y	8/7/2025	12:00	8/21/2025	12:10
SPRSEA-1-S-20250807	1	m-/p-Xylenes	Sample	0.711	ug/m3	0.271	ug/m3		Y	8/7/2025	12:00	8/21/2025	12:10
SPRSEA-1-S-20250807	1	o-Xylene	Sample	0.276	ug/m3	0.271	ug/m3	J	Y	8/7/2025	12:00	8/21/2025	12:10
SPRSEA-1-S-20250807	1	Toluene	Sample	1.39	ug/m3	0.24	ug/m3		Y	8/7/2025	12:00	8/21/2025	12:10
SPRSEA-2-S-20250807	2	Benzene	Sample	0.713	ug/m3	0.186	ug/m3		Y	8/7/2025	11:50	8/21/2025	12:00
SPRSEA-2-S-20250807	2	Ethylbenzene	Sample	0.48	ug/m3	0.271	ug/m3	J,Pc	Y	8/7/2025	11:50	8/21/2025	12:00
SPRSEA-2-S-20250807	2	m-/p-Xylenes	Sample	1.52	ug/m3	0.271	ug/m3		Y	8/7/2025	11:50	8/21/2025	12:00
SPRSEA-2-S-20250807	2	o-Xylene	Sample	0.567	ug/m3	0.271	ug/m3		Y	8/7/2025	11:50	8/21/2025	12:00

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

SAMPLE ID	SAMPLE LOC.	COMPOUND NAME	SAMPLE TYPE	RESULT 3	RESULT UNITS3	MDL3	MDL UNITS3	LAB FLAGS	DETECT FLAG	SAMPLE START DATE	SAMPLE START TIME	SAMPLE END DATE	SAMPLE END TIME
SPRSEA-2-S-20250807	2	Toluene	Sample	2.13	ug/m3	0.24	ug/m3		Y	8/7/2025	11:50	8/21/2025	12:00
SPRSEA-3-S-20250807	3	Benzene	Sample	0.721	ug/m3	0.186	ug/m3		Y	8/7/2025	11:40	8/21/2025	11:50
SPRSEA-3-S-20250807	3	Ethylbenzene	Sample	0.5	ug/m3	0.271	ug/m3	J,Pc	Y	8/7/2025	11:40	8/21/2025	11:50
SPRSEA-3-S-20250807	3	m-/p-Xylenes	Sample	1.18	ug/m3	0.271	ug/m3		Y	8/7/2025	11:40	8/21/2025	11:50
SPRSEA-3-S-20250807	3	o-Xylene	Sample	0.428	ug/m3	0.271	ug/m3	J	Y	8/7/2025	11:40	8/21/2025	11:50
SPRSEA-3-S-20250807	3	Toluene	Sample	2.11	ug/m3	0.24	ug/m3		Y	8/7/2025	11:40	8/21/2025	11:50
SPRSEA-4-S-20250807	4	Benzene	Sample	0.352	ug/m3	0.186	ug/m3	J	Y	8/7/2025	11:30	8/21/2025	11:40
SPRSEA-4-S-20250807	4	Ethylbenzene	Sample	<0.271	ug/m3	0.271	ug/m3	ND,Pc	N	8/7/2025	11:30	8/21/2025	11:40
SPRSEA-4-S-20250807	4	m-/p-Xylenes	Sample	0.584	ug/m3	0.271	ug/m3		Y	8/7/2025	11:30	8/21/2025	11:40
SPRSEA-4-S-20250807	4	o-Xylene	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	8/7/2025	11:30	8/21/2025	11:40
SPRSEA-4-S-20250807	4	Toluene	Sample	0.745	ug/m3	0.24	ug/m3		Y	8/7/2025	11:30	8/21/2025	11:40
SPRSEA-5-S-20250807	5	Benzene	Sample	0.332	ug/m3	0.186	ug/m3	J	Y	8/7/2025	12:30	8/21/2025	12:40
SPRSEA-5-S-20250807	5	Ethylbenzene	Sample	<0.271	ug/m3	0.271	ug/m3	ND,Pc	N	8/7/2025	12:30	8/21/2025	12:40
SPRSEA-5-S-20250807	5	m-/p-Xylenes	Sample	0.683	ug/m3	0.271	ug/m3		Y	8/7/2025	12:30	8/21/2025	12:40
SPRSEA-5-S-20250807	5	o-Xylene	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	8/7/2025	12:30	8/21/2025	12:40
SPRSEA-5-S-20250807	5	Toluene	Sample	0.695	ug/m3	0.24	ug/m3		Y	8/7/2025	12:30	8/21/2025	12:40
SPRSEA-6-S-20250807	6	Benzene	Sample	0.363	ug/m3	0.186	ug/m3	J	Y	8/7/2025	12:40	8/21/2025	12:50
SPRSEA-6-S-20250807	6	Ethylbenzene	Sample	<0.271	ug/m3	0.271	ug/m3	ND,Pc	N	8/7/2025	12:40	8/21/2025	12:50
SPRSEA-6-S-20250807	6	m-/p-Xylenes	Sample	0.342	ug/m3	0.271	ug/m3	J	Y	8/7/2025	12:40	8/21/2025	12:50
SPRSEA-6-S-20250807	6	o-Xylene	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	8/7/2025	12:40	8/21/2025	12:50
SPRSEA-6-S-20250807	6	Toluene	Sample	0.668	ug/m3	0.24	ug/m3		Y	8/7/2025	12:40	8/21/2025	12:50
SPRSEA-6-D-20250807	6	Benzene	Duplicate	0.366	ug/m3	0.186	ug/m3	J	Y	8/7/2025	12:40	8/21/2025	12:50
SPRSEA-6-D-20250807	6	Ethylbenzene	Duplicate	0.483	ug/m3	0.271	ug/m3	J,Pc	Y	8/7/2025	12:40	8/21/2025	12:50
SPRSEA-6-D-20250807	6	m-/p-Xylenes	Duplicate	0.412	ug/m3	0.271	ug/m3	J	Y	8/7/2025	12:40	8/21/2025	12:50
SPRSEA-6-D-20250807	6	o-Xylene	Duplicate	<0.271	ug/m3	0.271	ug/m3	ND	N	8/7/2025	12:40	8/21/2025	12:50
SPRSEA-6-D-20250807	6	Toluene	Duplicate	0.782	ug/m3	0.24	ug/m3		Y	8/7/2025	12:40	8/21/2025	12:50
SPRSEA-6-B-20250807	6	Benzene	Blank	<0.186	ug/m3	0.186	ug/m3	ND	N	8/7/2025	12:40	8/21/2025	12:50
SPRSEA-6-B-20250807	6	Ethylbenzene	Blank	<0.271	ug/m3	0.271	ug/m3	ND,Pc	N	8/7/2025	12:40	8/21/2025	12:50
SPRSEA-6-B-20250807	6	m-/p-Xylenes	Blank	<0.271	ug/m3	0.271	ug/m3	ND	N	8/7/2025	12:40	8/21/2025	12:50
SPRSEA-6-B-20250807	6	o-Xylene	Blank	<0.271	ug/m3	0.271	ug/m3	ND	N	8/7/2025	12:40	8/21/2025	12:50
SPRSEA-6-B-20250807	6	Toluene	Blank	<0.24	ug/m3	0.24	ug/m3	ND	N	8/7/2025	12:40	8/21/2025	12:50

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

SAMPLE ID	SAMPLE LOC.	COMPOUND NAME	SAMPLE TYPE	RESULT 3	RESULT UNITS3	MDL3	MDL UNITS3	LAB FLAGS	DETECT FLAG	SAMPLE START DATE	SAMPLE START TIME	SAMPLE END DATE	SAMPLE END TIME
SPRSEA-7-S-20250807	7	Benzene	Sample	0.245	ug/m3	0.186	ug/m3	J	Y	8/7/2025	12:50	8/21/2025	13:00
SPRSEA-7-S-20250807	7	Ethylbenzene	Sample	<0.271	ug/m3	0.271	ug/m3	ND,Pc	N	8/7/2025	12:50	8/21/2025	13:00
SPRSEA-7-S-20250807	7	m-/p-Xylenes	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	8/7/2025	12:50	8/21/2025	13:00
SPRSEA-7-S-20250807	7	o-Xylene	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	8/7/2025	12:50	8/21/2025	13:00
SPRSEA-7-S-20250807	7	Toluene	Sample	0.394	ug/m3	0.24	ug/m3	J	Y	8/7/2025	12:50	8/21/2025	13:00
SPRSEA-8-S-20250807	8	Benzene	Sample	0.518	ug/m3	0.186	ug/m3		Y	8/7/2025	13:00	8/21/2025	13:10
SPRSEA-8-S-20250807	8	Ethylbenzene	Sample	<0.271	ug/m3	0.271	ug/m3	ND,Pc	N	8/7/2025	13:00	8/21/2025	13:10
SPRSEA-8-S-20250807	8	m-/p-Xylenes	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	8/7/2025	13:00	8/21/2025	13:10
SPRSEA-8-S-20250807	8	o-Xylene	Sample	<0.271	ug/m3	0.271	ug/m3	ND	N	8/7/2025	13:00	8/21/2025	13:10
SPRSEA-8-S-20250807	8	Toluene	Sample	0.566	ug/m3	0.24	ug/m3		Y	8/7/2025	13:00	8/21/2025	13:10
SPRSEA-9-S-20250807	9	Benzene	Sample	0.556	ug/m3	0.186	ug/m3		Y	8/7/2025	13:10	8/21/2025	13:20
SPRSEA-9-S-20250807	9	Ethylbenzene	Sample	0.533	ug/m3	0.271	ug/m3	J,Pc	Y	8/7/2025	13:10	8/21/2025	13:20
SPRSEA-9-S-20250807	9	m-/p-Xylenes	Sample	1.25	ug/m3	0.271	ug/m3		Y	8/7/2025	13:10	8/21/2025	13:20
SPRSEA-9-S-20250807	9	o-Xylene	Sample	0.545	ug/m3	0.271	ug/m3	J	Y	8/7/2025	13:10	8/21/2025	13:20
SPRSEA-9-S-20250807	9	Toluene	Sample	1.68	ug/m3	0.24	ug/m3		Y	8/7/2025	13:10	8/21/2025	13:20
SPRSEA-10-S-20250807	10	Benzene	Sample	1.88	ug/m3	0.186	ug/m3		Y	8/7/2025	13:20	8/21/2025	13:30
SPRSEA-10-S-20250807	10	Ethylbenzene	Sample	1.66	ug/m3	0.271	ug/m3	Pc	Y	8/7/2025	13:20	8/21/2025	13:30
SPRSEA-10-S-20250807	10	m-/p-Xylenes	Sample	5.35	ug/m3	0.271	ug/m3		Y	8/7/2025	13:20	8/21/2025	13:30
SPRSEA-10-S-20250807	10	o-Xylene	Sample	2.11	ug/m3	0.271	ug/m3		Y	8/7/2025	13:20	8/21/2025	13:30
SPRSEA-10-S-20250807	10	Toluene	Sample	7.59	ug/m3	0.24	ug/m3		Y	8/7/2025	13:20	8/21/2025	13:30
SPRSEA-11-S-20250807	11	Benzene	Sample	2.37	ug/m3	0.186	ug/m3		Y	8/7/2025	12:20	8/21/2025	12:30
SPRSEA-11-S-20250807	11	Ethylbenzene	Sample	1.91	ug/m3	0.271	ug/m3	Pc	Y	8/7/2025	12:20	8/21/2025	12:30
SPRSEA-11-S-20250807	11	m-/p-Xylenes	Sample	6.09	ug/m3	0.271	ug/m3		Y	8/7/2025	12:20	8/21/2025	12:30
SPRSEA-11-S-20250807	11	o-Xylene	Sample	2.3	ug/m3	0.271	ug/m3		Y	8/7/2025	12:20	8/21/2025	12:30
SPRSEA-11-S-20250807	11	Toluene	Sample	9.93	ug/m3	0.24	ug/m3		Y	8/7/2025	12:20	8/21/2025	12:30
SPRSEA-12-S-20250807	12	Benzene	Sample	1.31	ug/m3	0.186	ug/m3		Y	8/7/2025	12:10	8/21/2025	12:20
SPRSEA-12-S-20250807	12	Ethylbenzene	Sample	0.764	ug/m3	0.271	ug/m3	Pc	Y	8/7/2025	12:10	8/21/2025	12:20
SPRSEA-12-S-20250807	12	m-/p-Xylenes	Sample	2.06	ug/m3	0.271	ug/m3		Y	8/7/2025	12:10	8/21/2025	12:20
SPRSEA-12-S-20250807	12	o-Xylene	Sample	0.788	ug/m3	0.271	ug/m3		Y	8/7/2025	12:10	8/21/2025	12:20
SPRSEA-12-S-20250807	12	Toluene	Sample	4.29	ug/m3	0.24	ug/m3		Y	8/7/2025	12:10	8/21/2025	12:20
SPRSEA-12-D-20250807	12	Benzene	Duplicate	1.21	ug/m3	0.186	ug/m3		Y	8/7/2025	12:10	8/21/2025	12:20

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

SAMPLE ID	SAMPLE LOC.	COMPOUND NAME	SAMPLE TYPE	RESULT 3	RESULT UNITS3	MDL3	MDL UNITS3	LAB FLAGS	DETECT FLAG	SAMPLE START DATE	SAMPLE START TIME	SAMPLE END DATE	SAMPLE END TIME
SPRSEA-12-D-20250807	12	Ethylbenzene	Duplicate	1.09	ug/m3	0.271	ug/m3	Pc	Y	8/7/2025	12:10	8/21/2025	12:20
SPRSEA-12-D-20250807	12	m-/p-Xylenes	Duplicate	2.58	ug/m3	0.271	ug/m3		Y	8/7/2025	12:10	8/21/2025	12:20
SPRSEA-12-D-20250807	12	o-Xylene	Duplicate	0.927	ug/m3	0.271	ug/m3		Y	8/7/2025	12:10	8/21/2025	12:20
SPRSEA-12-D-20250807	12	Toluene	Duplicate	4.15	ug/m3	0.24	ug/m3		Y	8/7/2025	12:10	8/21/2025	12:20
SPRSEA-12-B-20250807	12	Benzene	Blank	<0.186	ug/m3	0.186	ug/m3	ND	N	8/7/2025	12:10	8/21/2025	12:20
SPRSEA-12-B-20250807	12	Ethylbenzene	Blank	<0.271	ug/m3	0.271	ug/m3	ND,Pc	N	8/7/2025	12:10	8/21/2025	12:20
SPRSEA-12-B-20250807	12	m-/p-Xylenes	Blank	<0.271	ug/m3	0.271	ug/m3	ND	N	8/7/2025	12:10	8/21/2025	12:20
SPRSEA-12-B-20250807	12	o-Xylene	Blank	<0.271	ug/m3	0.271	ug/m3	ND	N	8/7/2025	12:10	8/21/2025	12:20
SPRSEA-12-B-20250807	12	Toluene	Blank	0.312	ug/m3	0.24	ug/m3	J	Y	8/7/2025	12:10	8/21/2025	12:20
SPRSEA-1-S-20250821	1	Benzene	Sample	0.993	ug/m3	0.188	ug/m3		Y	8/21/2025	12:10	9/4/2025	11:50
SPRSEA-1-S-20250821	1	Ethylbenzene	Sample	1.15	ug/m3	0.273	ug/m3		Y	8/21/2025	12:10	9/4/2025	11:50
SPRSEA-1-S-20250821	1	m-/p-Xylenes	Sample	3.54	ug/m3	0.273	ug/m3		Y	8/21/2025	12:10	9/4/2025	11:50
SPRSEA-1-S-20250821	1	o-Xylene	Sample	1.34	ug/m3	0.273	ug/m3		Y	8/21/2025	12:10	9/4/2025	11:50
SPRSEA-1-S-20250821	1	Toluene	Sample	4.57	ug/m3	0.242	ug/m3		Y	8/21/2025	12:10	9/4/2025	11:50
SPRSEA-2-S-20250821	2	Benzene	Sample	0.977	ug/m3	0.188	ug/m3		Y	8/21/2025	12:00	9/4/2025	11:40
SPRSEA-2-S-20250821	2	Ethylbenzene	Sample	0.842	ug/m3	0.273	ug/m3		Y	8/21/2025	12:00	9/4/2025	11:40
SPRSEA-2-S-20250821	2	m-/p-Xylenes	Sample	2.82	ug/m3	0.273	ug/m3		Y	8/21/2025	12:00	9/4/2025	11:40
SPRSEA-2-S-20250821	2	o-Xylene	Sample	1.1	ug/m3	0.273	ug/m3		Y	8/21/2025	12:00	9/4/2025	11:40
SPRSEA-2-S-20250821	2	Toluene	Sample	3.75	ug/m3	0.242	ug/m3		Y	8/21/2025	12:00	9/4/2025	11:40
SPRSEA-3-S-20250821	3	Benzene	Sample	0.687	ug/m3	0.188	ug/m3		Y	8/21/2025	11:50	9/4/2025	11:30
SPRSEA-3-S-20250821	3	Ethylbenzene	Sample	0.646	ug/m3	0.273	ug/m3		Y	8/21/2025	11:50	9/4/2025	11:30
SPRSEA-3-S-20250821	3	m-/p-Xylenes	Sample	2.17	ug/m3	0.273	ug/m3		Y	8/21/2025	11:50	9/4/2025	11:30
SPRSEA-3-S-20250821	3	o-Xylene	Sample	0.93	ug/m3	0.273	ug/m3		Y	8/21/2025	11:50	9/4/2025	11:30
SPRSEA-3-S-20250821	3	Toluene	Sample	2.9	ug/m3	0.242	ug/m3		Y	8/21/2025	11:50	9/4/2025	11:30
SPRSEA-4-S-20250821	4	Benzene	Sample	0.805	ug/m3	0.188	ug/m3		Y	8/21/2025	11:40	9/4/2025	11:20
SPRSEA-4-S-20250821	4	Ethylbenzene	Sample	1.01	ug/m3	0.273	ug/m3		Y	8/21/2025	11:40	9/4/2025	11:20
SPRSEA-4-S-20250821	4	m-/p-Xylenes	Sample	3.05	ug/m3	0.273	ug/m3		Y	8/21/2025	11:40	9/4/2025	11:20
SPRSEA-4-S-20250821	4	o-Xylene	Sample	1.19	ug/m3	0.273	ug/m3		Y	8/21/2025	11:40	9/4/2025	11:20
SPRSEA-4-S-20250821	4	Toluene	Sample	3.72	ug/m3	0.242	ug/m3		Y	8/21/2025	11:40	9/4/2025	11:20
SPRSEA-5-S-20250821	5	Benzene	Sample	0.977	ug/m3	0.188	ug/m3		Y	8/21/2025	12:40	9/4/2025	12:20
SPRSEA-5-S-20250821	5	Ethylbenzene	Sample	1.06	ug/m3	0.273	ug/m3		Y	8/21/2025	12:40	9/4/2025	12:20

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

SAMPLE ID	SAMPLE LOC.	COMPOUND NAME	SAMPLE TYPE	RESULT 3	RESULT UNITS3	MDL3	MDL UNITS3	LAB FLAGS	DETECT FLAG	SAMPLE START DATE	SAMPLE START TIME	SAMPLE END DATE	SAMPLE END TIME
SPRSEA-5-S-20250821	5	m-/p-Xylenes	Sample	4.12	ug/m3	0.273	ug/m3		Y	8/21/2025	12:40	9/4/2025	12:20
SPRSEA-5-S-20250821	5	o-Xylene	Sample	1.6	ug/m3	0.273	ug/m3		Y	8/21/2025	12:40	9/4/2025	12:20
SPRSEA-5-S-20250821	5	Toluene	Sample	4.95	ug/m3	0.242	ug/m3		Y	8/21/2025	12:40	9/4/2025	12:20
SPRSEA-6-S-20250821	6	Benzene	Sample	0.87	ug/m3	0.188	ug/m3		Y	8/21/2025	12:50	9/4/2025	12:30
SPRSEA-6-S-20250821	6	Ethylbenzene	Sample	0.809	ug/m3	0.273	ug/m3		Y	8/21/2025	12:50	9/4/2025	12:30
SPRSEA-6-S-20250821	6	m-/p-Xylenes	Sample	3.56	ug/m3	0.273	ug/m3		Y	8/21/2025	12:50	9/4/2025	12:30
SPRSEA-6-S-20250821	6	o-Xylene	Sample	1.43	ug/m3	0.273	ug/m3		Y	8/21/2025	12:50	9/4/2025	12:30
SPRSEA-6-S-20250821	6	Toluene	Sample	4.21	ug/m3	0.242	ug/m3		Y	8/21/2025	12:50	9/4/2025	12:30
SPRSEA-6-D-20250821	6	Benzene	Duplicate	1	ug/m3	0.188	ug/m3		Y	8/21/2025	12:50	9/4/2025	12:30
SPRSEA-6-D-20250821	6	Ethylbenzene	Duplicate	1.04	ug/m3	0.273	ug/m3		Y	8/21/2025	12:50	9/4/2025	12:30
SPRSEA-6-D-20250821	6	m-/p-Xylenes	Duplicate	3.59	ug/m3	0.273	ug/m3		Y	8/21/2025	12:50	9/4/2025	12:30
SPRSEA-6-D-20250821	6	o-Xylene	Duplicate	1.5	ug/m3	0.273	ug/m3		Y	8/21/2025	12:50	9/4/2025	12:30
SPRSEA-6-D-20250821	6	Toluene	Duplicate	4.47	ug/m3	0.242	ug/m3		Y	8/21/2025	12:50	9/4/2025	12:30
SPRSEA-6-B-20250821	6	Benzene	Blank	<0.188	ug/m3	0.188	ug/m3	ND	N	8/21/2025	12:50	9/4/2025	12:30
SPRSEA-6-B-20250821	6	Ethylbenzene	Blank	<0.273	ug/m3	0.273	ug/m3	ND	N	8/21/2025	12:50	9/4/2025	12:30
SPRSEA-6-B-20250821	6	m-/p-Xylenes	Blank	<0.273	ug/m3	0.273	ug/m3	ND	N	8/21/2025	12:50	9/4/2025	12:30
SPRSEA-6-B-20250821	6	o-Xylene	Blank	<0.273	ug/m3	0.273	ug/m3	ND	N	8/21/2025	12:50	9/4/2025	12:30
SPRSEA-6-B-20250821	6	Toluene	Blank	<0.242	ug/m3	0.242	ug/m3	ND	N	8/21/2025	12:50	9/4/2025	12:30
SPRSEA-7-S-20250821	7	Benzene	Sample	0.978	ug/m3	0.188	ug/m3		Y	8/21/2025	13:00	9/4/2025	12:40
SPRSEA-7-S-20250821	7	Ethylbenzene	Sample	1.01	ug/m3	0.273	ug/m3		Y	8/21/2025	13:00	9/4/2025	12:40
SPRSEA-7-S-20250821	7	m-/p-Xylenes	Sample	3.73	ug/m3	0.273	ug/m3		Y	8/21/2025	13:00	9/4/2025	12:40
SPRSEA-7-S-20250821	7	o-Xylene	Sample	1.47	ug/m3	0.273	ug/m3		Y	8/21/2025	13:00	9/4/2025	12:40
SPRSEA-7-S-20250821	7	Toluene	Sample	4.17	ug/m3	0.242	ug/m3		Y	8/21/2025	13:00	9/4/2025	12:40
SPRSEA-8-S-20250821	8	Benzene	Sample	0.944	ug/m3	0.188	ug/m3		Y	8/21/2025	13:10	9/4/2025	12:50
SPRSEA-8-S-20250821	8	Ethylbenzene	Sample	0.879	ug/m3	0.273	ug/m3		Y	8/21/2025	13:10	9/4/2025	12:50
SPRSEA-8-S-20250821	8	m-/p-Xylenes	Sample	3.23	ug/m3	0.273	ug/m3		Y	8/21/2025	13:10	9/4/2025	12:50
SPRSEA-8-S-20250821	8	o-Xylene	Sample	1.37	ug/m3	0.273	ug/m3		Y	8/21/2025	13:10	9/4/2025	12:50
SPRSEA-8-S-20250821	8	Toluene	Sample	4.19	ug/m3	0.242	ug/m3		Y	8/21/2025	13:10	9/4/2025	12:50
SPRSEA-9-S-20250821	9	Benzene	Sample	1.22	ug/m3	0.188	ug/m3		Y	8/21/2025	13:20	9/4/2025	13:00
SPRSEA-9-S-20250821	9	Ethylbenzene	Sample	1.25	ug/m3	0.273	ug/m3		Y	8/21/2025	13:20	9/4/2025	13:00
SPRSEA-9-S-20250821	9	m-/p-Xylenes	Sample	5.39	ug/m3	0.273	ug/m3		Y	8/21/2025	13:20	9/4/2025	13:00

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

SAMPLE ID	SAMPLE LOC.	COMPOUND NAME	SAMPLE TYPE	RESULT 3	RESULT UNITS3	MDL3	MDL UNITS3	LAB FLAGS	DETECT FLAG	SAMPLE START DATE	SAMPLE START TIME	SAMPLE END DATE	SAMPLE END TIME
SPRSEA-9-S-20250821	9	o-Xylene	Sample	2.15	ug/m3	0.273	ug/m3		Y	8/21/2025	13:20	9/4/2025	13:00
SPRSEA-9-S-20250821	9	Toluene	Sample	5.84	ug/m3	0.242	ug/m3		Y	8/21/2025	13:20	9/4/2025	13:00
SPRSEA-10-S-20250821	10	Benzene	Sample	4.68	ug/m3	0.188	ug/m3		Y	8/21/2025	13:30	9/4/2025	13:10
SPRSEA-10-S-20250821	10	Ethylbenzene	Sample	3.18	ug/m3	0.273	ug/m3		Y	8/21/2025	13:30	9/4/2025	13:10
SPRSEA-10-S-20250821	10	m-/p-Xylenes	Sample	12.2	ug/m3	0.273	ug/m3		Y	8/21/2025	13:30	9/4/2025	13:10
SPRSEA-10-S-20250821	10	o-Xylene	Sample	4.91	ug/m3	0.273	ug/m3		Y	8/21/2025	13:30	9/4/2025	13:10
SPRSEA-10-S-20250821	10	Toluene	Sample	17.8	ug/m3	0.242	ug/m3		Y	8/21/2025	13:30	9/4/2025	13:10
SPRSEA-11-S-20250821	11	Benzene	Sample	1.55	ug/m3	0.188	ug/m3		Y	8/21/2025	12:30	9/4/2025	12:10
SPRSEA-11-S-20250821	11	Ethylbenzene	Sample	1.54	ug/m3	0.273	ug/m3		Y	8/21/2025	12:30	9/4/2025	12:10
SPRSEA-11-S-20250821	11	m-/p-Xylenes	Sample	5.62	ug/m3	0.273	ug/m3		Y	8/21/2025	12:30	9/4/2025	12:10
SPRSEA-11-S-20250821	11	o-Xylene	Sample	2.33	ug/m3	0.273	ug/m3		Y	8/21/2025	12:30	9/4/2025	12:10
SPRSEA-11-S-20250821	11	Toluene	Sample	7.07	ug/m3	0.242	ug/m3		Y	8/21/2025	12:30	9/4/2025	12:10
SPRSEA-12-S-20250821	12	Benzene	Sample	1.82	ug/m3	0.188	ug/m3		Y	8/21/2025	12:20	9/4/2025	12:00
SPRSEA-12-S-20250821	12	Ethylbenzene	Sample	1.58	ug/m3	0.273	ug/m3		Y	8/21/2025	12:20	9/4/2025	12:00
SPRSEA-12-S-20250821	12	m-/p-Xylenes	Sample	6.43	ug/m3	0.273	ug/m3		Y	8/21/2025	12:20	9/4/2025	12:00
SPRSEA-12-S-20250821	12	o-Xylene	Sample	2.6	ug/m3	0.273	ug/m3		Y	8/21/2025	12:20	9/4/2025	12:00
SPRSEA-12-S-20250821	12	Toluene	Sample	8.04	ug/m3	0.242	ug/m3		Y	8/21/2025	12:20	9/4/2025	12:00
SPRSEA-12-D-20250821	12	Benzene	Duplicate	1.79	ug/m3	0.188	ug/m3		Y	8/21/2025	12:20	9/4/2025	12:00
SPRSEA-12-D-20250821	12	Ethylbenzene	Duplicate	1.54	ug/m3	0.273	ug/m3		Y	8/21/2025	12:20	9/4/2025	12:00
SPRSEA-12-D-20250821	12	m-/p-Xylenes	Duplicate	6.47	ug/m3	0.273	ug/m3		Y	8/21/2025	12:20	9/4/2025	12:00
SPRSEA-12-D-20250821	12	o-Xylene	Duplicate	2.51	ug/m3	0.273	ug/m3		Y	8/21/2025	12:20	9/4/2025	12:00
SPRSEA-12-D-20250821	12	Toluene	Duplicate	7.59	ug/m3	0.242	ug/m3		Y	8/21/2025	12:20	9/4/2025	12:00
SPRSEA-12-B-20250821	12	Benzene	Blank	<0.188	ug/m3	0.188	ug/m3	ND	N	8/21/2025	12:20	9/4/2025	12:00
SPRSEA-12-B-20250821	12	Ethylbenzene	Blank	<0.273	ug/m3	0.273	ug/m3	ND	N	8/21/2025	12:20	9/4/2025	12:00
SPRSEA-12-B-20250821	12	m-/p-Xylenes	Blank	<0.273	ug/m3	0.273	ug/m3	ND	N	8/21/2025	12:20	9/4/2025	12:00
SPRSEA-12-B-20250821	12	o-Xylene	Blank	<0.273	ug/m3	0.273	ug/m3	ND	N	8/21/2025	12:20	9/4/2025	12:00
SPRSEA-12-B-20250821	12	Toluene	Blank	<0.242	ug/m3	0.242	ug/m3	ND	N	8/21/2025	12:20	9/4/2025	12:00
SPRSEA-1-S-20250904	1	Benzene	Sample	1.22	ug/m3	0.188	ug/m3		Y	9/4/2025	11:50	9/18/2025	11:50
SPRSEA-1-S-20250904	1	Ethylbenzene	Sample	0.518	ug/m3	0.274	ug/m3	J	Y	9/4/2025	11:50	9/18/2025	11:50
SPRSEA-1-S-20250904	1	m-/p-Xylenes	Sample	1.59	ug/m3	0.274	ug/m3	Pc	Y	9/4/2025	11:50	9/18/2025	11:50
SPRSEA-1-S-20250904	1	o-Xylene	Sample	0.659	ug/m3	0.274	ug/m3		Y	9/4/2025	11:50	9/18/2025	11:50

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

SAMPLE ID	SAMPLE LOC.	COMPOUND NAME	SAMPLE TYPE	RESULT 3	RESULT UNITS3	MDL3	MDL UNITS3	LAB FLAGS	DETECT FLAG	SAMPLE START DATE	SAMPLE START TIME	SAMPLE END DATE	SAMPLE END TIME
SPRSEA-1-S-20250904	1	Toluene	Sample	2.68	ug/m3	0.242	ug/m3		Y	9/4/2025	11:50	9/18/2025	11:50
SPRSEA-2-S-20250904	2	Benzene	Sample	1.09	ug/m3	0.188	ug/m3		Y	9/4/2025	11:40	9/18/2025	11:40
SPRSEA-2-S-20250904	2	Ethylbenzene	Sample	0.463	ug/m3	0.274	ug/m3	J	Y	9/4/2025	11:40	9/18/2025	11:40
SPRSEA-2-S-20250904	2	m-/p-Xylenes	Sample	1.77	ug/m3	0.274	ug/m3	Pc	Y	9/4/2025	11:40	9/18/2025	11:40
SPRSEA-2-S-20250904	2	o-Xylene	Sample	0.712	ug/m3	0.274	ug/m3		Y	9/4/2025	11:40	9/18/2025	11:40
SPRSEA-2-S-20250904	2	Toluene	Sample	2.4	ug/m3	0.242	ug/m3		Y	9/4/2025	11:40	9/18/2025	11:40
SPRSEA-3-S-20250904	3	Benzene	Sample	1.91	ug/m3	0.188	ug/m3		Y	9/4/2025	11:30	9/18/2025	11:30
SPRSEA-3-S-20250904	3	Ethylbenzene	Sample	0.384	ug/m3	0.274	ug/m3	J	Y	9/4/2025	11:30	9/18/2025	11:30
SPRSEA-3-S-20250904	3	m-/p-Xylenes	Sample	1.21	ug/m3	0.274	ug/m3	Pc	Y	9/4/2025	11:30	9/18/2025	11:30
SPRSEA-3-S-20250904	3	o-Xylene	Sample	0.483	ug/m3	0.274	ug/m3	J	Y	9/4/2025	11:30	9/18/2025	11:30
SPRSEA-3-S-20250904	3	Toluene	Sample	2.09	ug/m3	0.242	ug/m3		Y	9/4/2025	11:30	9/18/2025	11:30
SPRSEA-4-S-20250904	4	Benzene	Sample	1.75	ug/m3	0.188	ug/m3		Y	9/4/2025	11:20	9/18/2025	11:20
SPRSEA-4-S-20250904	4	Ethylbenzene	Sample	<0.274	ug/m3	0.274	ug/m3	ND	N	9/4/2025	11:20	9/18/2025	11:20
SPRSEA-4-S-20250904	4	m-/p-Xylenes	Sample	0.666	ug/m3	0.274	ug/m3	Pc	Y	9/4/2025	11:20	9/18/2025	11:20
SPRSEA-4-S-20250904	4	o-Xylene	Sample	<0.274	ug/m3	0.274	ug/m3	ND	N	9/4/2025	11:20	9/18/2025	11:20
SPRSEA-4-S-20250904	4	Toluene	Sample	1.33	ug/m3	0.242	ug/m3		Y	9/4/2025	11:20	9/18/2025	11:20
SPRSEA-5-S-20250904	5	Benzene	Sample	0.754	ug/m3	0.188	ug/m3		Y	9/4/2025	12:20	9/18/2025	12:20
SPRSEA-5-S-20250904	5	Ethylbenzene	Sample	<0.274	ug/m3	0.274	ug/m3	ND	N	9/4/2025	12:20	9/18/2025	12:20
SPRSEA-5-S-20250904	5	m-/p-Xylenes	Sample	0.934	ug/m3	0.274	ug/m3	Pc	Y	9/4/2025	12:20	9/18/2025	12:20
SPRSEA-5-S-20250904	5	o-Xylene	Sample	0.369	ug/m3	0.274	ug/m3	J	Y	9/4/2025	12:20	9/18/2025	12:20
SPRSEA-5-S-20250904	5	Toluene	Sample	1.33	ug/m3	0.242	ug/m3		Y	9/4/2025	12:20	9/18/2025	12:20
SPRSEA-6-S-20250904	6	Benzene	Sample	0.675	ug/m3	0.188	ug/m3		Y	9/4/2025	12:30	9/18/2025	12:30
SPRSEA-6-S-20250904	6	Ethylbenzene	Sample	<0.274	ug/m3	0.274	ug/m3	ND	N	9/4/2025	12:30	9/18/2025	12:30
SPRSEA-6-S-20250904	6	m-/p-Xylenes	Sample	0.859	ug/m3	0.274	ug/m3	Pc	Y	9/4/2025	12:30	9/18/2025	12:30
SPRSEA-6-S-20250904	6	o-Xylene	Sample	0.354	ug/m3	0.274	ug/m3	J	Y	9/4/2025	12:30	9/18/2025	12:30
SPRSEA-6-S-20250904	6	Toluene	Sample	1.12	ug/m3	0.242	ug/m3		Y	9/4/2025	12:30	9/18/2025	12:30
SPRSEA-6-D-20250904	6	Benzene	Duplicate	0.834	ug/m3	0.188	ug/m3		Y	9/4/2025	12:30	9/18/2025	12:30
SPRSEA-6-D-20250904	6	Ethylbenzene	Duplicate	<0.274	ug/m3	0.274	ug/m3	ND	N	9/4/2025	12:30	9/18/2025	12:30
SPRSEA-6-D-20250904	6	m-/p-Xylenes	Duplicate	0.591	ug/m3	0.274	ug/m3	Pc	Y	9/4/2025	12:30	9/18/2025	12:30
SPRSEA-6-D-20250904	6	o-Xylene	Duplicate	<0.274	ug/m3	0.274	ug/m3	ND	N	9/4/2025	12:30	9/18/2025	12:30
SPRSEA-6-D-20250904	6	Toluene	Duplicate	0.995	ug/m3	0.242	ug/m3		Y	9/4/2025	12:30	9/18/2025	12:30

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

SAMPLE ID	SAMPLE LOC.	COMPOUND NAME	SAMPLE TYPE	RESULT 3	RESULT UNITS3	MDL3	MDL UNITS3	LAB FLAGS	DETECT FLAG	SAMPLE START DATE	SAMPLE START TIME	SAMPLE END DATE	SAMPLE END TIME
SPRSEA-6-B-20250904	6	Benzene	Blank	<0.188	ug/m3	0.188	ug/m3	ND	N	9/4/2025	12:30	9/18/2025	12:30
SPRSEA-6-B-20250904	6	Ethylbenzene	Blank	<0.274	ug/m3	0.274	ug/m3	ND	N	9/4/2025	12:30	9/18/2025	12:30
SPRSEA-6-B-20250904	6	m-/p-Xylenes	Blank	<0.274	ug/m3	0.274	ug/m3	ND,Pc	N	9/4/2025	12:30	9/18/2025	12:30
SPRSEA-6-B-20250904	6	o-Xylene	Blank	<0.274	ug/m3	0.274	ug/m3	ND	N	9/4/2025	12:30	9/18/2025	12:30
SPRSEA-6-B-20250904	6	Toluene	Blank	<0.242	ug/m3	0.242	ug/m3	ND	N	9/4/2025	12:30	9/18/2025	12:30
SPRSEA-7-S-20250904	7	Benzene	Sample	0.968	ug/m3	0.188	ug/m3		Y	9/4/2025	12:40	9/18/2025	12:40
SPRSEA-7-S-20250904	7	Ethylbenzene	Sample	<0.274	ug/m3	0.274	ug/m3	ND	N	9/4/2025	12:40	9/18/2025	12:40
SPRSEA-7-S-20250904	7	m-/p-Xylenes	Sample	0.464	ug/m3	0.274	ug/m3	J,Pc	Y	9/4/2025	12:40	9/18/2025	12:40
SPRSEA-7-S-20250904	7	o-Xylene	Sample	<0.274	ug/m3	0.274	ug/m3	ND	N	9/4/2025	12:40	9/18/2025	12:40
SPRSEA-7-S-20250904	7	Toluene	Sample	0.924	ug/m3	0.242	ug/m3		Y	9/4/2025	12:40	9/18/2025	12:40
SPRSEA-8-S-20250904	8	Benzene	Sample	0.914	ug/m3	0.188	ug/m3		Y	9/4/2025	12:50	9/18/2025	12:50
SPRSEA-8-S-20250904	8	Ethylbenzene	Sample	<0.274	ug/m3	0.274	ug/m3	ND	N	9/4/2025	12:50	9/18/2025	12:50
SPRSEA-8-S-20250904	8	m-/p-Xylenes	Sample	0.605	ug/m3	0.274	ug/m3	Pc	Y	9/4/2025	12:50	9/18/2025	12:50
SPRSEA-8-S-20250904	8	o-Xylene	Sample	<0.274	ug/m3	0.274	ug/m3	ND	N	9/4/2025	12:50	9/18/2025	12:50
SPRSEA-8-S-20250904	8	Toluene	Sample	1.18	ug/m3	0.242	ug/m3		Y	9/4/2025	12:50	9/18/2025	12:50
SPRSEA-9-S-20250904	9	Benzene	Sample	1.12	ug/m3	0.188	ug/m3		Y	9/4/2025	13:00	9/18/2025	13:00
SPRSEA-9-S-20250904	9	Ethylbenzene	Sample	0.527	ug/m3	0.274	ug/m3	J	Y	9/4/2025	13:00	9/18/2025	13:00
SPRSEA-9-S-20250904	9	m-/p-Xylenes	Sample	1.62	ug/m3	0.274	ug/m3	Pc	Y	9/4/2025	13:00	9/18/2025	13:00
SPRSEA-9-S-20250904	9	o-Xylene	Sample	0.695	ug/m3	0.274	ug/m3		Y	9/4/2025	13:00	9/18/2025	13:00
SPRSEA-9-S-20250904	9	Toluene	Sample	2.55	ug/m3	0.242	ug/m3		Y	9/4/2025	13:00	9/18/2025	13:00
SPRSEA-10-S-20250904	10	Benzene	Sample	3.8	ug/m3	0.188	ug/m3		Y	9/4/2025	13:10	9/18/2025	13:10
SPRSEA-10-S-20250904	10	Ethylbenzene	Sample	2.04	ug/m3	0.274	ug/m3		Y	9/4/2025	13:10	9/18/2025	13:10
SPRSEA-10-S-20250904	10	m-/p-Xylenes	Sample	7.27	ug/m3	0.274	ug/m3	Pc	Y	9/4/2025	13:10	9/18/2025	13:10
SPRSEA-10-S-20250904	10	o-Xylene	Sample	2.99	ug/m3	0.274	ug/m3		Y	9/4/2025	13:10	9/18/2025	13:10
SPRSEA-10-S-20250904	10	Toluene	Sample	11.1	ug/m3	0.242	ug/m3		Y	9/4/2025	13:10	9/18/2025	13:10
SPRSEA-11-S-20250904	11	Benzene	Sample	2.8	ug/m3	0.188	ug/m3		Y	9/4/2025	12:10	9/18/2025	12:10
SPRSEA-11-S-20250904	11	Ethylbenzene	Sample	1.48	ug/m3	0.274	ug/m3		Y	9/4/2025	12:10	9/18/2025	12:10
SPRSEA-11-S-20250904	11	m-/p-Xylenes	Sample	5.91	ug/m3	0.274	ug/m3	Pc	Y	9/4/2025	12:10	9/18/2025	12:10
SPRSEA-11-S-20250904	11	o-Xylene	Sample	2.34	ug/m3	0.274	ug/m3		Y	9/4/2025	12:10	9/18/2025	12:10
SPRSEA-11-S-20250904	11	Toluene	Sample	8.02	ug/m3	0.242	ug/m3		Y	9/4/2025	12:10	9/18/2025	12:10
SPRSEA-12-S-20250904	12	Benzene	Sample	2.52	ug/m3	0.188	ug/m3		Y	9/4/2025	12:00	9/18/2025	12:00

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

SAMPLE ID	SAMPLE LOC.	COMPOUND NAME	SAMPLE TYPE	RESULT 3	RESULT UNITS3	MDL3	MDL UNITS3	LAB FLAGS	DETECT FLAG	SAMPLE START DATE	SAMPLE START TIME	SAMPLE END DATE	SAMPLE END TIME
SPRSEA-12-S-20250904	12	Ethylbenzene	Sample	1.37	ug/m3	0.274	ug/m3		Y	9/4/2025	12:00	9/18/2025	12:00
SPRSEA-12-S-20250904	12	m-/p-Xylenes	Sample	5.33	ug/m3	0.274	ug/m3	Pc	Y	9/4/2025	12:00	9/18/2025	12:00
SPRSEA-12-S-20250904	12	o-Xylene	Sample	2.16	ug/m3	0.274	ug/m3		Y	9/4/2025	12:00	9/18/2025	12:00
SPRSEA-12-S-20250904	12	Toluene	Sample	7.36	ug/m3	0.242	ug/m3		Y	9/4/2025	12:00	9/18/2025	12:00
SPRSEA-12-D-20250904	12	Benzene	Duplicate	2.4	ug/m3	0.188	ug/m3		Y	9/4/2025	12:00	9/18/2025	12:00
SPRSEA-12-D-20250904	12	Ethylbenzene	Duplicate	1.16	ug/m3	0.274	ug/m3		Y	9/4/2025	12:00	9/18/2025	12:00
SPRSEA-12-D-20250904	12	m-/p-Xylenes	Duplicate	4.16	ug/m3	0.274	ug/m3	Pc	Y	9/4/2025	12:00	9/18/2025	12:00
SPRSEA-12-D-20250904	12	o-Xylene	Duplicate	1.66	ug/m3	0.274	ug/m3		Y	9/4/2025	12:00	9/18/2025	12:00
SPRSEA-12-D-20250904	12	Toluene	Duplicate	6.71	ug/m3	0.242	ug/m3		Y	9/4/2025	12:00	9/18/2025	12:00
SPRSEA-12-B-20250904	12	Benzene	Blank	<0.188	ug/m3	0.188	ug/m3	ND	N	9/4/2025	12:00	9/18/2025	12:00
SPRSEA-12-B-20250904	12	Ethylbenzene	Blank	<0.274	ug/m3	0.274	ug/m3	ND	N	9/4/2025	12:00	9/18/2025	12:00
SPRSEA-12-B-20250904	12	m-/p-Xylenes	Blank	<0.274	ug/m3	0.274	ug/m3	ND,Pc	N	9/4/2025	12:00	9/18/2025	12:00
SPRSEA-12-B-20250904	12	o-Xylene	Blank	<0.274	ug/m3	0.274	ug/m3	ND	N	9/4/2025	12:00	9/18/2025	12:00
SPRSEA-12-B-20250904	12	Toluene	Blank	<0.242	ug/m3	0.242	ug/m3	ND	N	9/4/2025	12:00	9/18/2025	12:00

FLM DATA FLAG ABBREVIATIONS - EPA METHOD 325B	
FLAG	EXPLANATION
ND	The analyte was not present above the Method Detection Limit
J	Estimated Value - The analyte was detected between the Method Detection Limit and Reporting Limit
P	Field duplicate(s) exceed 30%RPD
Pc	Field duplicate(s) exceed 30%RPD. Concentrations of both samples in duplicate are near the reporting limit
Note: Meteorological data flagged ND was not available from the airport. Missing data can be due to instrument maintenance, instrument malfunction, data transmission issues, or other factors resulting in missing data.	

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (6/26/25 11:00 to 7/10/25 12:00)**

Date & Time	Wind Speed	Wind Direction	Temperature	Barometric Pressure
	m/s	Deg.	°C	mb
6/26/25 11:00	1.8	185	19.7	1015
6/26/25 12:00	2.2	298	19.2	1015
6/26/25 13:00	1.4	321	19.1	1015
6/26/25 14:00	1.3	343	18.9	1015
6/26/25 15:00	1.0	311	18.9	1014
6/26/25 16:00	2.1	247	19.2	1014
6/26/25 17:00	1.6	285	19.0	1013
6/26/25 18:00	0.8	312	18.8	1013
6/26/25 19:00	1.3	204	17.9	1014
6/26/25 20:00	1.7	248	16.0	1014
6/26/25 21:00	0.8	293	13.8	1015
6/26/25 22:00	1.3	243	12.3	1015
6/26/25 23:00	ND	ND	11.0	1016
6/27/25 0:00	0.8	315	10.4	1016
6/27/25 1:00	ND	ND	9.6	1016
6/27/25 2:00	ND	ND	9.1	1016
6/27/25 3:00	0.3	302	8.5	1016
6/27/25 4:00	ND	ND	8.3	1017
6/27/25 5:00	ND	ND	8.4	1018
6/27/25 6:00	1.5	340	11.5	1018
6/27/25 7:00	ND	ND	14.1	1019
6/27/25 8:00	ND	ND	16.7	1019
6/27/25 9:00	1.8	192	19.1	1019
6/27/25 10:00	1.9	116	21.5	1019
6/27/25 11:00	1.8	272	22.5	1018
6/27/25 12:00	2.1	276	23.0	1018
6/27/25 13:00	2.7	212	23.7	1017
6/27/25 14:00	2.1	258	23.9	1017
6/27/25 15:00	1.4	295	23.6	1017
6/27/25 16:00	3.2	236	23.2	1016
6/27/25 17:00	5.2	187	21.6	1017
6/27/25 18:00	4.5	189	21.2	1017
6/27/25 19:00	3.1	174	19.6	1016
6/27/25 20:00	2.5	186	18.7	1016
6/27/25 21:00	2.0	194	17.9	1017
6/27/25 22:00	1.4	232	17.3	1017
6/27/25 23:00	2.8	182	16.9	1017
6/28/25 0:00	2.3	218	16.1	1017
6/28/25 1:00	0.5	328	14.0	1017
6/28/25 2:00	ND	ND	13.2	1017
6/28/25 3:00	2.9	63	13.0	1016
6/28/25 4:00	2.5	104	13.0	1016
6/28/25 5:00	3.2	102	12.9	1015

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (6/26/25 11:00 to 7/10/25 12:00)**

6/28/25 6:00	3.8	110	12.3	1015
6/28/25 7:00	2.9	129	12.4	1017
6/28/25 8:00	3.9	153	13.4	1015
6/28/25 9:00	5.3	172	13.9	1016
6/28/25 10:00	6.6	172	12.8	1015
6/28/25 11:00	5.2	171	12.0	1014
6/28/25 12:00	3.2	171	12.2	1014
6/28/25 13:00	3.8	144	12.7	1013
6/28/25 14:00	3.8	152	12.9	1013
6/28/25 15:00	2.6	156	12.9	1012
6/28/25 16:00	4.1	146	13.0	1012
6/28/25 17:00	3.8	149	14.0	1010
6/28/25 18:00	2.3	128	13.9	1010
6/28/25 19:00	2.6	130	13.9	1009
6/28/25 20:00	2.2	109	14.0	1009
6/28/25 21:00	1.9	74	13.9	1009
6/28/25 22:00	2.3	79	14.0	1009
6/28/25 23:00	1.5	80	14.0	1009
6/29/25 0:00	1.8	52	13.9	1008
6/29/25 1:00	1.9	50	13.9	1008
6/29/25 2:00	1.7	61	13.9	1007
6/29/25 3:00	1.1	270	13.9	1007
6/29/25 4:00	1.6	133	13.9	1007
6/29/25 5:00	ND	ND	13.9	1007
6/29/25 6:00	1.0	160	13.9	1007
6/29/25 7:00	1.2	307	14.0	1007
6/29/25 8:00	0.8	333	14.5	1007
6/29/25 9:00	1.3	303	15.5	1007
6/29/25 10:00	1.8	302	15.8	1007
6/29/25 11:00	1.8	324	16.2	1007
6/29/25 12:00	2.2	249	17.0	1007
6/29/25 13:00	1.6	299	17.5	1006
6/29/25 14:00	1.4	318	18.0	1006
6/29/25 15:00	0.6	341	18.0	1006
6/29/25 16:00	0.9	353	18.8	1006
6/29/25 17:00	1.0	313	20.7	1006
6/29/25 18:00	ND	ND	20.9	1006
6/29/25 19:00	1.6	127	20.0	1006
6/29/25 20:00	1.3	217	19.0	1006
6/29/25 21:00	0.5	303	18.4	1006
6/29/25 22:00	1.4	245	18.7	1007
6/29/25 23:00	2.0	265	18.4	1007
6/30/25 0:00	2.3	270	17.9	1007
6/30/25 1:00	2.3	259	17.1	1007
6/30/25 2:00	2.4	266	17.0	1007
6/30/25 3:00	0.4	330	16.0	1008

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (6/26/25 11:00 to 7/10/25 12:00)**

6/30/25 4:00	1.0	293	14.2	1008
6/30/25 5:00	ND	ND	14.2	1008
6/30/25 6:00	0.2	342	16.9	1009
6/30/25 7:00	1.0	322	18.9	1009
6/30/25 8:00	1.7	264	21.1	1009
6/30/25 9:00	2.3	269	22.6	1009
6/30/25 10:00	2.1	249	23.8	1008
6/30/25 11:00	2.2	272	25.5	1008
6/30/25 12:00	2.0	257	26.6	1008
6/30/25 13:00	3.0	233	27.1	1007
6/30/25 14:00	2.3	226	27.9	1007
6/30/25 15:00	3.1	221	27.9	1006
6/30/25 16:00	4.7	180	27.6	1006
6/30/25 17:00	5.4	184	26.4	1006
6/30/25 18:00	5.3	175	24.7	1006
6/30/25 19:00	3.3	178	23.3	1006
6/30/25 20:00	3.6	173	21.6	1006
6/30/25 21:00	2.6	190	20.3	1007
6/30/25 22:00	2.8	177	18.3	1006
6/30/25 23:00	2.0	179	17.5	1006
7/1/25 0:00	0.7	291	16.2	1006
7/1/25 1:00	1.8	207	16.1	1005
7/1/25 2:00	1.9	187	16.0	1005
7/1/25 3:00	1.9	178	16.0	1005
7/1/25 4:00	2.0	179	16.0	1004
7/1/25 5:00	2.6	167	15.9	1004
7/1/25 6:00	2.8	178	16.3	1004
7/1/25 7:00	3.5	185	17.2	1003
7/1/25 8:00	4.1	184	17.8	1003
7/1/25 9:00	3.2	179	18.6	1003
7/1/25 10:00	2.8	187	20.8	1002
7/1/25 11:00	4.5	188	21.2	1001
7/1/25 12:00	3.8	192	22.5	1001
7/1/25 13:00	4.0	188	23.7	1000
7/1/25 14:00	4.6	184	23.9	999
7/1/25 15:00	4.3	196	24.1	999
7/1/25 16:00	4.7	177	23.9	998
7/1/25 17:00	4.2	184	23.8	998
7/1/25 18:00	3.8	191	23.1	998
7/1/25 19:00	2.8	184	22.7	998
7/1/25 20:00	2.8	186	21.9	998
7/1/25 21:00	2.3	188	20.9	998
7/1/25 22:00	2.3	193	20.9	999
7/1/25 23:00	2.3	192	20.9	999
7/2/25 0:00	0.8	312	20.8	999
7/2/25 1:00	0.8	296	20.9	998

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (6/26/25 11:00 to 7/10/25 12:00)**

7/2/25 2:00	1.7	218	20.9	998
7/2/25 3:00	1.7	220	20.9	998
7/2/25 4:00	ND	ND	19.6	998
7/2/25 5:00	1.0	276	19.3	999
7/2/25 6:00	1.8	256	21.1	999
7/2/25 7:00	1.7	261	22.7	1000
7/2/25 8:00	2.9	277	25.2	1000
7/2/25 9:00	3.3	286	25.9	1000
7/2/25 10:00	2.7	266	26.3	1000
7/2/25 11:00	3.4	239	27.4	1000
7/2/25 12:00	3.4	240	27.9	1000
7/2/25 13:00	4.1	227	28.0	1000
7/2/25 14:00	4.3	210	28.6	1000
7/2/25 15:00	4.4	208	28.9	999
7/2/25 16:00	3.2	202	28.7	999
7/2/25 17:00	4.1	190	27.9	999
7/2/25 18:00	3.9	191	27.7	999
7/2/25 19:00	4.0	183	26.6	999
7/2/25 20:00	3.0	189	24.8	1000
7/2/25 21:00	2.2	192	23.1	1001
7/2/25 22:00	1.6	204	22.1	1001
7/2/25 23:00	0.5	314	20.6	1001
7/3/25 0:00	1.1	183	18.8	1001
7/3/25 1:00	1.8	165	18.4	1001
7/3/25 2:00	ND	ND	17.8	1001
7/3/25 3:00	ND	ND	17.2	1001
7/3/25 4:00	ND	ND	17.6	1001
7/3/25 5:00	ND	ND	16.6	1001
7/3/25 6:00	1.5	198	18.8	1001
7/3/25 7:00	1.5	185	21.5	1001
7/3/25 8:00	0.8	275	23.6	1001
7/3/25 9:00	1.7	212	26.1	1001
7/3/25 10:00	2.9	159	27.1	1000
7/3/25 11:00	4.5	171	27.9	1000
7/3/25 12:00	5.7	184	28.4	999
7/3/25 13:00	6.4	184	28.2	999
7/3/25 14:00	4.6	187	26.1	999
7/3/25 15:00	4.5	272	20.1	1000
7/3/25 16:00	1.5	170	19.0	1000
7/3/25 17:00	0.2	331	19.8	1000
7/3/25 18:00	1.1	282	19.8	1000
7/3/25 19:00	0.9	306	19.8	1000
7/3/25 20:00	2.9	279	18.4	1001
7/3/25 21:00	2.6	283	17.2	1001
7/3/25 22:00	3.0	289	17.0	1002
7/3/25 23:00	2.1	296	16.4	1003

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (6/26/25 11:00 to 7/10/25 12:00)**

7/4/25 0:00	2.6	276	16.0	1003
7/4/25 1:00	2.6	261	16.0	1003
7/4/25 2:00	2.2	265	15.4	1003
7/4/25 3:00	3.0	262	15.0	1003
7/4/25 4:00	2.7	257	14.9	1003
7/4/25 5:00	2.7	264	15.2	1004
7/4/25 6:00	1.2	291	16.0	1004
7/4/25 7:00	2.4	321	16.7	1005
7/4/25 8:00	4.7	312	18.2	1005
7/4/25 9:00	5.6	305	19.5	1005
7/4/25 10:00	6.9	318	20.9	1005
7/4/25 11:00	5.4	315	21.6	1005
7/4/25 12:00	6.3	295	21.9	1005
7/4/25 13:00	3.9	290	20.9	1005
7/4/25 14:00	5.5	319	19.7	1006
7/4/25 15:00	5.8	329	20.7	1006
7/4/25 16:00	6.7	308	20.8	1007
7/4/25 17:00	6.1	311	20.8	1007
7/4/25 18:00	5.4	303	19.6	1007
7/4/25 19:00	3.4	305	18.2	1008
7/4/25 20:00	2.0	309	17.4	1009
7/4/25 21:00	1.9	264	15.8	1009
7/4/25 22:00	1.9	271	15.6	1009
7/4/25 23:00	1.9	220	15.7	1010
7/5/25 0:00	1.6	226	14.4	1010
7/5/25 1:00	0.9	288	14.7	1010
7/5/25 2:00	2.9	265	15.3	1010
7/5/25 3:00	3.0	269	15.0	1010
7/5/25 4:00	2.5	259	15.0	1010
7/5/25 5:00	3.5	265	15.5	1010
7/5/25 6:00	2.4	268	16.8	1010
7/5/25 7:00	2.8	279	18.4	1011
7/5/25 8:00	3.2	288	20.8	1011
7/5/25 9:00	3.2	288	22.0	1011
7/5/25 10:00	4.0	299	24.3	1011
7/5/25 11:00	3.5	277	25.9	1011
7/5/25 12:00	3.1	313	26.8	1011
7/5/25 13:00	4.2	324	27.0	1010
7/5/25 14:00	3.4	275	26.9	1010
7/5/25 15:00	3.8	271	26.5	1010
7/5/25 16:00	3.7	251	25.9	1010
7/5/25 17:00	3.4	245	24.5	1010
7/5/25 18:00	2.8	231	23.9	1010
7/5/25 19:00	3.2	203	23.3	1009
7/5/25 20:00	2.9	202	22.5	1009
7/5/25 21:00	2.2	205	21.4	1010

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (6/26/25 11:00 to 7/10/25 12:00)**

7/5/25 22:00	2.5	197	20.9	1009
7/5/25 23:00	2.8	194	20.6	1009
7/6/25 0:00	3.1	191	19.1	1008
7/6/25 1:00	3.0	195	18.9	1008
7/6/25 2:00	3.0	184	18.7	1007
7/6/25 3:00	2.5	190	17.9	1006
7/6/25 4:00	2.3	193	17.9	1006
7/6/25 5:00	2.3	186	17.9	1006
7/6/25 6:00	3.3	196	18.7	1006
7/6/25 7:00	3.5	196	20.2	1006
7/6/25 8:00	2.8	199	22.5	1006
7/6/25 9:00	3.4	201	24.4	1006
7/6/25 10:00	3.5	207	28.0	1006
7/6/25 11:00	5.2	241	29.7	1005
7/6/25 12:00	4.4	246	30.3	1005
7/6/25 13:00	3.4	248	30.9	1005
7/6/25 14:00	4.4	249	31.6	1004
7/6/25 15:00	3.4	233	30.9	1004
7/6/25 16:00	4.3	178	30.4	1004
7/6/25 17:00	4.5	172	29.3	1003
7/6/25 18:00	4.2	183	28.3	1003
7/6/25 19:00	4.3	256	26.1	1004
7/6/25 20:00	3.2	125	22.9	1005
7/6/25 21:00	2.2	210	22.9	1005
7/6/25 22:00	2.8	302	23.1	1006
7/6/25 23:00	0.9	257	22.9	1006
7/7/25 0:00	1.7	163	21.9	1006
7/7/25 1:00	1.8	160	21.5	1006
7/7/25 2:00	1.5	192	20.9	1006
7/7/25 3:00	1.7	175	20.3	1006
7/7/25 4:00	1.8	184	19.7	1006
7/7/25 5:00	1.8	186	19.7	1006
7/7/25 6:00	1.1	273	21.3	1006
7/7/25 7:00	1.6	203	23.0	1006
7/7/25 8:00	1.9	150	24.7	1006
7/7/25 9:00	3.0	179	26.6	1006
7/7/25 10:00	3.1	178	27.1	1006
7/7/25 11:00	3.2	172	28.1	1006
7/7/25 12:00	4.5	177	29.0	1005
7/7/25 13:00	5.4	171	29.1	1005
7/7/25 14:00	6.1	176	29.7	1005
7/7/25 15:00	6.3	179	28.7	1004
7/7/25 16:00	5.5	184	27.5	1004
7/7/25 17:00	4.1	178	26.2	1004
7/7/25 18:00	5.3	173	26.9	1004
7/7/25 19:00	4.2	175	25.3	1004

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (6/26/25 11:00 to 7/10/25 12:00)**


7/7/25 20:00	3.6	178	23.6	1005
7/7/25 21:00	2.1	148	22.6	1005
7/7/25 22:00	1.9	134	22.1	1005
7/7/25 23:00	1.3	307	21.7	1006
7/8/25 0:00	1.0	344	21.0	1005
7/8/25 1:00	2.0	354	20.9	1005
7/8/25 2:00	2.9	323	20.2	1005
7/8/25 3:00	3.6	186	19.0	1005
7/8/25 4:00	3.3	59	18.6	1005
7/8/25 5:00	3.0	13	18.1	1006
7/8/25 6:00	4.1	88	18.1	1006
7/8/25 7:00	3.3	25	18.0	1007
7/8/25 8:00	4.2	19	18.2	1007
7/8/25 9:00	4.7	18	20.0	1007
7/8/25 10:00	4.2	31	21.2	1007
7/8/25 11:00	4.5	117	21.9	1008
7/8/25 12:00	5.5	355	21.9	1008
7/8/25 13:00	5.0	81	22.0	1008
7/8/25 14:00	4.0	189	22.3	1008
7/8/25 15:00	5.0	185	23.3	1007
7/8/25 16:00	4.9	155	23.3	1008
7/8/25 17:00	3.8	127	23.1	1008
7/8/25 18:00	3.7	43	23.2	1008
7/8/25 19:00	3.0	46	22.3	1009
7/8/25 20:00	0.9	221	21.1	1009
7/8/25 21:00	ND	ND	20.3	1009
7/8/25 22:00	0.0	360	19.8	1009
7/8/25 23:00	0.0	1	18.4	1009
7/9/25 0:00	0.0	359	17.3	1009
7/9/25 1:00	ND	ND	16.4	1009
7/9/25 2:00	1.5	200	16.2	1009
7/9/25 3:00	1.0	340	15.5	1009
7/9/25 4:00	ND	ND	15.2	1009
7/9/25 5:00	1.5	210	16.5	1009
7/9/25 6:00	ND	ND	17.0	1010
7/9/25 7:00	1.5	200	18.0	1010
7/9/25 8:00	1.2	173	20.5	1010
7/9/25 9:00	1.8	105	22.8	1010
7/9/25 10:00	2.4	65	24.0	1010
7/9/25 11:00	1.2	143	25.2	1010
7/9/25 12:00	1.1	253	26.0	1009
7/9/25 13:00	1.6	210	26.2	1009
7/9/25 14:00	1.3	204	27.1	1008
7/9/25 15:00	1.0	317	27.5	1008
7/9/25 16:00	2.1	202	27.0	1008
7/9/25 17:00	4.4	189	25.2	1008

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (6/26/25 11:00 to 7/10/25 12:00)**

7/9/25 18:00	4.8	174	23.4	1008
7/9/25 19:00	3.7	175	22.4	1009
7/9/25 20:00	3.5	175	21.6	1009
7/9/25 21:00	3.0	194	20.9	1009
7/9/25 22:00	2.5	207	20.2	1009
7/9/25 23:00	1.6	198	19.0	1009
7/10/25 0:00	2.0	221	17.9	1009
7/10/25 1:00	2.2	184	17.9	1009
7/10/25 2:00	1.8	222	17.9	1009
7/10/25 3:00	0.6	258	17.0	1010
7/10/25 4:00	1.7	166	16.9	1010
7/10/25 5:00	1.6	182	16.9	1010
7/10/25 6:00	1.5	177	17.0	1010
7/10/25 7:00	2.0	175	18.0	1010
7/10/25 8:00	1.6	144	19.4	1010
7/10/25 9:00	2.6	132	21.2	1010
7/10/25 10:00	2.9	131	22.1	1010
7/10/25 11:00	3.9	144	22.9	1011
7/10/25 12:00	3.2	187	22.3	1011

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING
Bangor International Airport (BGR) Meteorological Data (6/26/25 11:00 to 7/10/25 12:00)

BGR Wind Rose 6/10/25 11:00 - 7/10/25 12:00



SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (7/10/25 12:00 to 7/24/25 12:00)**

Date & Time	Wind Speed	Wind Direction	Temperature	Barometric Pressure
	m/s	Deg.	°C	mb
7/10/25 12:00	3.2	187	22.3	1011
7/10/25 13:00	2.9	170	22.8	1010
7/10/25 14:00	3.2	140	22.9	1010
7/10/25 15:00	3.6	146	22.9	1010
7/10/25 16:00	3.7	150	22.7	1010
7/10/25 17:00	2.5	163	22.7	1010
7/10/25 18:00	3.0	134	21.6	1010
7/10/25 19:00	3.7	124	20.1	1010
7/10/25 20:00	3.8	126	18.5	1010
7/10/25 21:00	3.8	142	17.9	1010
7/10/25 22:00	2.8	166	17.9	1010
7/10/25 23:00	1.9	179	17.9	1010
7/11/25 0:00	1.5	206	17.9	1010
7/11/25 1:00	1.7	218	17.8	1010
7/11/25 2:00	0.9	291	17.0	1010
7/11/25 3:00	1.5	215	17.4	1010
7/11/25 4:00	1.7	276	17.1	1009
7/11/25 5:00	1.9	238	17.0	1009
7/11/25 6:00	2.1	250	17.0	1010
7/11/25 7:00	1.8	219	17.4	1010
7/11/25 8:00	1.6	176	18.0	1010
7/11/25 9:00	1.4	214	18.7	1010
7/11/25 10:00	0.6	308	19.7	1010
7/11/25 11:00	1.6	174	20.9	1010
7/11/25 12:00	1.5	239	21.3	1010
7/11/25 13:00	2.0	267	22.4	1010
7/11/25 14:00	1.8	276	23.2	1010
7/11/25 15:00	2.5	244	23.7	1009
7/11/25 16:00	0.6	167	24.3	1009
7/11/25 17:00	1.6	226	24.7	1009
7/11/25 18:00	2.7	193	23.9	1009
7/11/25 19:00	2.8	195	23.1	1010
7/11/25 20:00	2.8	196	21.5	1010
7/11/25 21:00	2.8	171	20.3	1011
7/11/25 22:00	2.6	182	19.0	1011
7/11/25 23:00	2.8	190	18.4	1011
7/12/25 0:00	1.9	194	17.5	1012
7/12/25 1:00	2.7	190	16.9	1012
7/12/25 2:00	1.9	164	16.9	1012
7/12/25 3:00	2.0	194	16.7	1012
7/12/25 4:00	2.4	176	16.0	1012
7/12/25 5:00	1.8	186	16.9	1012
7/12/25 6:00	1.9	177	16.9	1013

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (7/10/25 12:00 to 7/24/25 12:00)**

7/12/25 7:00	2.4	186	17.2	1013
7/12/25 8:00	2.5	190	18.1	1013
7/12/25 9:00	2.0	168	19.4	1013
7/12/25 10:00	2.9	182	22.3	1013
7/12/25 11:00	4.0	187	22.3	1013
7/12/25 12:00	4.4	186	21.7	1014
7/12/25 13:00	4.2	173	21.5	1014
7/12/25 14:00	5.1	182	22.4	1013
7/12/25 15:00	5.1	186	21.6	1013
7/12/25 16:00	5.2	179	21.3	1013
7/12/25 17:00	4.6	184	19.3	1013
7/12/25 18:00	4.7	181	18.8	1013
7/12/25 19:00	4.4	176	18.1	1014
7/12/25 20:00	4.0	185	18.0	1014
7/12/25 21:00	3.6	181	17.5	1014
7/12/25 22:00	2.6	181	17.5	1014
7/12/25 23:00	3.5	177	16.9	1014
7/13/25 0:00	3.9	190	16.9	1014
7/13/25 1:00	4.1	187	16.7	1014
7/13/25 2:00	4.2	171	16.0	1013
7/13/25 3:00	4.0	180	16.0	1013
7/13/25 4:00	3.5	188	15.9	1013
7/13/25 5:00	3.6	174	15.9	1013
7/13/25 6:00	3.9	185	16.0	1014
7/13/25 7:00	3.3	181	16.0	1014
7/13/25 8:00	3.9	188	16.8	1014
7/13/25 9:00	4.5	176	18.4	1014
7/13/25 10:00	4.8	184	20.1	1013
7/13/25 11:00	5.1	194	21.5	1013
7/13/25 12:00	6.0	189	22.1	1013
7/13/25 13:00	6.2	185	22.9	1013
7/13/25 14:00	6.1	172	23.4	1012
7/13/25 15:00	6.8	182	22.9	1011
7/13/25 16:00	7.4	185	22.3	1011
7/13/25 17:00	7.0	185	21.1	1011
7/13/25 18:00	6.1	187	18.6	1011
7/13/25 19:00	5.9	182	17.8	1012
7/13/25 20:00	6.7	179	17.0	1012
7/13/25 21:00	5.4	176	16.2	1012
7/13/25 22:00	3.7	183	16.0	1013
7/13/25 23:00	4.6	178	16.0	1013
7/14/25 0:00	4.0	181	16.0	1012
7/14/25 1:00	3.8	176	16.0	1012
7/14/25 2:00	3.7	178	15.9	1011
7/14/25 3:00	3.1	173	15.9	1011
7/14/25 4:00	3.3	173	16.0	1011

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (7/10/25 12:00 to 7/24/25 12:00)**

7/14/25 5:00	3.6	173	16.0	1011
7/14/25 6:00	3.3	176	16.4	1011
7/14/25 7:00	3.5	181	17.1	1011
7/14/25 8:00	3.7	178	18.0	1011
7/14/25 9:00	4.6	177	18.8	1011
7/14/25 10:00	4.4	180	20.2	1011
7/14/25 11:00	5.1	176	21.0	1011
7/14/25 12:00	5.6	184	21.1	1011
7/14/25 13:00	4.7	198	20.8	1011
7/14/25 14:00	5.2	183	21.0	1010
7/14/25 15:00	4.7	189	21.5	1010
7/14/25 16:00	5.4	185	21.9	1010
7/14/25 17:00	3.7	165	21.3	1009
7/14/25 18:00	4.6	176	20.8	1010
7/14/25 19:00	4.3	178	19.9	1010
7/14/25 20:00	3.3	175	19.8	1010
7/14/25 21:00	4.4	183	18.9	1010
7/14/25 22:00	3.3	186	18.9	1009
7/14/25 23:00	2.8	183	18.3	1009
7/15/25 0:00	2.3	184	18.5	1009
7/15/25 1:00	2.1	185	18.0	1009
7/15/25 2:00	0.8	264	18.6	1009
7/15/25 3:00	1.2	249	18.0	1009
7/15/25 4:00	0.8	280	18.0	1009
7/15/25 5:00	1.8	193	18.1	1009
7/15/25 6:00	1.5	173	18.8	1010
7/15/25 7:00	1.7	243	19.1	1010
7/15/25 8:00	1.9	208	20.1	1010
7/15/25 9:00	1.9	225	22.1	1010
7/15/25 10:00	2.4	224	26.8	1010
7/15/25 11:00	2.8	224	27.9	1009
7/15/25 12:00	3.0	254	28.8	1009
7/15/25 13:00	3.2	227	30.0	1009
7/15/25 14:00	3.9	229	30.7	1008
7/15/25 15:00	3.9	233	31.3	1008
7/15/25 16:00	4.1	188	30.8	1007
7/15/25 17:00	4.0	187	29.4	1008
7/15/25 18:00	3.2	183	28.7	1008
7/15/25 19:00	3.0	181	27.2	1008
7/15/25 20:00	2.8	179	25.3	1009
7/15/25 21:00	2.5	189	23.8	1009
7/15/25 22:00	2.2	183	22.7	1009
7/15/25 23:00	2.6	193	21.9	1009
7/16/25 0:00	2.0	198	21.0	1009
7/16/25 1:00	0.5	309	19.7	1009
7/16/25 2:00	0.5	300	18.5	1009

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (7/10/25 12:00 to 7/24/25 12:00)**

7/16/25 3:00	1.2	280	18.1	1009
7/16/25 4:00	1.5	210	18.0	1009
7/16/25 5:00	ND	ND	18.1	1009
7/16/25 6:00	ND	ND	19.9	1010
7/16/25 7:00	1.7	178	22.7	1009
7/16/25 8:00	2.1	187	25.0	1009
7/16/25 9:00	1.6	223	27.4	1009
7/16/25 10:00	2.0	208	29.0	1009
7/16/25 11:00	1.3	277	30.2	1009
7/16/25 12:00	3.3	182	30.9	1008
7/16/25 13:00	4.9	185	31.7	1008
7/16/25 14:00	6.4	186	31.9	1008
7/16/25 15:00	5.9	184	31.9	1007
7/16/25 16:00	6.2	182	31.3	1007
7/16/25 17:00	5.4	176	30.0	1006
7/16/25 18:00	5.1	181	28.6	1006
7/16/25 19:00	5.3	177	27.1	1006
7/16/25 20:00	4.4	179	24.9	1006
7/16/25 21:00	3.4	173	23.0	1006
7/16/25 22:00	4.0	181	21.2	1006
7/16/25 23:00	2.8	197	19.2	1006
7/17/25 0:00	1.6	208	19.4	1006
7/17/25 1:00	2.2	179	19.0	1006
7/17/25 2:00	2.8	182	19.0	1005
7/17/25 3:00	3.5	183	18.9	1005
7/17/25 4:00	3.1	185	18.9	1005
7/17/25 5:00	3.1	184	19.0	1005
7/17/25 6:00	2.3	185	19.0	1005
7/17/25 7:00	2.8	173	20.2	1004
7/17/25 8:00	3.2	206	21.6	1004
7/17/25 9:00	4.0	191	22.9	1004
7/17/25 10:00	4.3	187	23.3	1003
7/17/25 11:00	5.7	175	25.2	1003
7/17/25 12:00	6.1	177	26.6	1001
7/17/25 13:00	7.8	185	27.3	1001
7/17/25 14:00	7.8	174	27.1	1000
7/17/25 15:00	6.8	176	26.7	999
7/17/25 16:00	7.6	175	25.5	999
7/17/25 17:00	5.6	184	23.1	997
7/17/25 18:00	4.8	180	22.5	997
7/17/25 19:00	5.0	172	22.3	997
7/17/25 20:00	3.7	175	21.5	997
7/17/25 21:00	4.2	179	21.0	997
7/17/25 22:00	4.0	182	20.9	996
7/17/25 23:00	4.6	182	20.9	996
7/18/25 0:00	3.8	199	20.9	996

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (7/10/25 12:00 to 7/24/25 12:00)**

7/18/25 1:00	2.3	197	20.9	996
7/18/25 2:00	1.9	201	20.9	996
7/18/25 3:00	2.9	326	21.8	997
7/18/25 4:00	4.5	296	21.6	998
7/18/25 5:00	4.5	302	20.1	1000
7/18/25 6:00	4.0	312	19.2	1002
7/18/25 7:00	7.6	316	20.4	1002
7/18/25 8:00	8.3	328	20.7	1004
7/18/25 9:00	10.0	328	21.0	1004
7/18/25 10:00	8.1	325	21.9	1005
7/18/25 11:00	6.8	334	23.6	1006
7/18/25 12:00	8.0	335	24.5	1006
7/18/25 13:00	6.9	337	25.8	1007
7/18/25 14:00	6.0	314	25.8	1007
7/18/25 15:00	6.3	320	26.0	1008
7/18/25 16:00	6.6	323	25.9	1008
7/18/25 17:00	6.7	331	25.7	1008
7/18/25 18:00	5.7	331	24.4	1008
7/18/25 19:00	4.4	331	22.7	1009
7/18/25 20:00	1.6	299	19.9	1010
7/18/25 21:00	1.9	295	17.9	1010
7/18/25 22:00	2.3	290	17.8	1010
7/18/25 23:00	1.2	298	15.9	1010
7/19/25 0:00	0.9	330	13.8	1011
7/19/25 1:00	1.2	226	13.2	1011
7/19/25 2:00	0.5	292	12.9	1011
7/19/25 3:00	ND	ND	11.8	1011
7/19/25 4:00	ND	ND	11.4	1012
7/19/25 5:00	ND	ND	11.5	1012
7/19/25 6:00	0.0	360	14.3	1012
7/19/25 7:00	2.3	237	17.1	1012
7/19/25 8:00	2.8	250	19.8	1012
7/19/25 9:00	2.3	237	21.9	1011
7/19/25 10:00	2.1	218	23.8	1011
7/19/25 11:00	2.6	247	24.5	1010
7/19/25 12:00	1.8	297	25.8	1010
7/19/25 13:00	4.6	203	26.2	1009
7/19/25 14:00	5.6	215	26.4	1009
7/19/25 15:00	6.5	176	25.8	1009
7/19/25 16:00	6.7	181	24.7	1008
7/19/25 17:00	6.0	179	24.4	1008
7/19/25 18:00	6.0	184	23.7	1008
7/19/25 19:00	4.7	187	22.1	1008
7/19/25 20:00	3.6	187	20.2	1008
7/19/25 21:00	3.5	184	18.8	1008
7/19/25 22:00	2.6	188	17.9	1008

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (7/10/25 12:00 to 7/24/25 12:00)**

7/19/25 23:00	1.6	199	17.8	1007
7/20/25 0:00	0.8	287	17.4	1007
7/20/25 1:00	1.7	204	17.5	1006
7/20/25 2:00	1.1	232	17.0	1006
7/20/25 3:00	1.6	204	17.0	1005
7/20/25 4:00	1.0	242	17.0	1005
7/20/25 5:00	2.5	170	17.2	1004
7/20/25 6:00	2.3	197	18.0	1004
7/20/25 7:00	1.9	263	18.4	1004
7/20/25 8:00	1.6	274	18.9	1004
7/20/25 9:00	0.7	251	19.7	1004
7/20/25 10:00	1.0	268	20.9	1003
7/20/25 11:00	1.9	189	20.5	1002
7/20/25 12:00	2.2	231	19.7	1002
7/20/25 13:00	2.4	155	19.4	1001
7/20/25 14:00	1.7	143	18.9	1001
7/20/25 15:00	2.3	120	19.4	1000
7/20/25 16:00	2.4	150	20.1	1000
7/20/25 17:00	1.2	217	19.9	1000
7/20/25 18:00	1.4	230	19.4	1000
7/20/25 19:00	1.6	201	19.0	999
7/20/25 20:00	1.6	266	18.9	999
7/20/25 21:00	0.4	350	18.0	1000
7/20/25 22:00	1.1	310	17.9	1000
7/20/25 23:00	2.0	280	17.9	999
7/21/25 0:00	2.7	251	17.9	999
7/21/25 1:00	3.2	294	17.7	999
7/21/25 2:00	1.6	307	16.8	998
7/21/25 3:00	1.6	288	16.0	998
7/21/25 4:00	2.1	300	15.8	999
7/21/25 5:00	2.2	293	15.3	999
7/21/25 6:00	3.7	294	15.6	1000
7/21/25 7:00	4.6	310	16.6	1000
7/21/25 8:00	6.1	306	17.0	1001
7/21/25 9:00	7.3	314	17.3	1002
7/21/25 10:00	7.8	325	18.7	1002
7/21/25 11:00	8.2	321	19.8	1002
7/21/25 12:00	5.3	320	20.9	1002
7/21/25 13:00	9.0	322	21.1	1002
7/21/25 14:00	7.9	314	21.9	1002
7/21/25 15:00	6.1	336	18.2	1003
7/21/25 16:00	6.6	327	20.1	1003
7/21/25 17:00	8.3	336	20.9	1004
7/21/25 18:00	6.9	320	19.7	1004
7/21/25 19:00	5.4	312	18.5	1005
7/21/25 20:00	5.0	309	17.9	1006

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (7/10/25 12:00 to 7/24/25 12:00)**


7/21/25 21:00	3.6	295	16.6	1007
7/21/25 22:00	3.6	304	15.9	1008
7/21/25 23:00	2.1	284	14.5	1008
7/22/25 0:00	3.1	274	13.7	1008
7/22/25 1:00	2.9	283	13.0	1009
7/22/25 2:00	3.2	289	12.9	1009
7/22/25 3:00	3.0	264	12.0	1009
7/22/25 4:00	2.3	274	11.8	1009
7/22/25 5:00	2.8	260	11.2	1010
7/22/25 6:00	1.9	286	12.6	1011
7/22/25 7:00	3.0	294	14.8	1012
7/22/25 8:00	3.6	321	17.1	1012
7/22/25 9:00	3.6	337	18.4	1012
7/22/25 10:00	4.5	331	19.5	1012
7/22/25 11:00	4.5	292	20.8	1012
7/22/25 12:00	4.2	306	21.9	1012
7/22/25 13:00	3.5	322	22.7	1012
7/22/25 14:00	2.6	310	23.6	1012
7/22/25 15:00	3.5	333	23.7	1012
7/22/25 16:00	4.6	292	24.3	1012
7/22/25 17:00	3.8	314	24.0	1012
7/22/25 18:00	3.4	313	23.1	1012
7/22/25 19:00	1.7	289	19.8	1012
7/22/25 20:00	0.9	303	17.3	1013
7/22/25 21:00	ND	ND	16.8	1014
7/22/25 22:00	0.4	314	15.4	1014
7/22/25 23:00	0.2	345	15.1	1015
7/23/25 0:00	ND	ND	14.7	1015
7/23/25 1:00	1.0	265	13.9	1015
7/23/25 2:00	0.8	314	12.8	1015
7/23/25 3:00	ND	ND	12.1	1016
7/23/25 4:00	ND	ND	12.1	1016
7/23/25 5:00	ND	ND	11.6	1016
7/23/25 6:00	ND	ND	13.8	1016
7/23/25 7:00	1.5	280	16.7	1017
7/23/25 8:00	1.5	254	19.1	1017
7/23/25 9:00	0.3	306	21.3	1017
7/23/25 10:00	2.1	218	22.7	1017
7/23/25 11:00	3.3	280	23.9	1016
7/23/25 12:00	3.3	257	24.6	1016

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (7/10/25 12:00 to 7/24/25 12:00)**

7/23/25 13:00	3.0	230	25.6	1015
7/23/25 14:00	2.5	270	25.9	1015
7/23/25 15:00	4.5	188	26.0	1015
7/23/25 16:00	5.8	179	25.4	1014
7/23/25 17:00	6.0	190	24.4	1014
7/23/25 18:00	5.9	176	23.2	1014
7/23/25 19:00	5.2	178	21.5	1014
7/23/25 20:00	3.0	183	19.4	1014
7/23/25 21:00	3.4	190	18.5	1015
7/23/25 22:00	3.3	194	17.9	1015
7/23/25 23:00	3.9	192	17.9	1014
7/24/25 0:00	3.1	191	17.2	1014
7/24/25 1:00	3.2	195	16.9	1014
7/24/25 2:00	2.5	189	16.9	1014
7/24/25 3:00	2.5	193	16.9	1013
7/24/25 4:00	2.7	191	16.3	1012
7/24/25 5:00	2.6	185	16.0	1012
7/24/25 6:00	2.7	194	16.8	1012
7/24/25 7:00	3.2	193	17.7	1012
7/24/25 8:00	3.7	178	19.7	1012
7/24/25 9:00	4.5	175	21.6	1011
7/24/25 10:00	3.2	182	23.6	1011
7/24/25 11:00	5.2	182	26.3	1010
7/24/25 12:00	6.7	178	27.9	1009

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING
Bangor International Airport (BGR) Meteorological Data (7/10/25 12:00 to 7/24/25 12:00)

BGR Wind Rose 7/12/25 12:00 - 7/24/25 12:00



SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (7/24/25 12:00 to 8/7/25 12:00)**

Date & Time	Wind Speed	Wind Direction	Temperature	Barometric Pressure
	m/s	Deg.	°C	mb
7/24/25 12:00	6.7	178	27.9	1009
7/24/25 13:00	7.4	178	27.8	1008
7/24/25 14:00	7.3	178	27.3	1007
7/24/25 15:00	6.5	180	25.4	1007
7/24/25 16:00	6.6	179	24.6	1006
7/24/25 17:00	6.8	177	25.6	1006
7/24/25 18:00	6.2	173	25.6	1005
7/24/25 19:00	5.7	177	23.4	1005
7/24/25 20:00	4.1	178	22.4	1005
7/24/25 21:00	4.6	185	21.2	1005
7/24/25 22:00	3.9	194	20.9	1005
7/24/25 23:00	4.0	187	19.8	1005
7/25/25 0:00	3.3	189	19.9	1005
7/25/25 1:00	3.8	190	19.5	1004
7/25/25 2:00	3.5	187	18.9	1004
7/25/25 3:00	3.3	187	18.9	1003
7/25/25 4:00	3.5	185	18.9	1003
7/25/25 5:00	2.8	194	18.9	1003
7/25/25 6:00	3.4	196	19.6	1003
7/25/25 7:00	3.0	189	20.4	1003
7/25/25 8:00	3.2	183	21.4	1002
7/25/25 9:00	3.4	178	23.9	1002
7/25/25 10:00	2.8	197	27.5	1002
7/25/25 11:00	3.1	196	27.2	1001
7/25/25 12:00	2.7	291	26.8	1001
7/25/25 13:00	3.8	267	26.7	1001
7/25/25 14:00	3.2	277	26.4	1001
7/25/25 15:00	3.6	288	27.7	1001
7/25/25 16:00	4.3	302	26.8	1001
7/25/25 17:00	6.2	312	26.6	1002
7/25/25 18:00	5.9	345	26.2	1003
7/25/25 19:00	3.8	340	24.9	1004
7/25/25 20:00	2.7	305	23.4	1005
7/25/25 21:00	4.5	304	22.9	1006
7/25/25 22:00	7.4	335	22.0	1006
7/25/25 23:00	4.6	340	21.0	1007
7/26/25 0:00	2.0	327	18.6	1008
7/26/25 1:00	1.7	326	17.1	1008
7/26/25 2:00	1.1	303	15.9	1009
7/26/25 3:00	2.8	294	15.0	1009
7/26/25 4:00	1.3	330	14.5	1010
7/26/25 5:00	1.9	343	14.3	1010
7/26/25 6:00	0.9	300	16.2	1011

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (7/24/25 12:00 to 8/7/25 12:00)**

7/26/25 7:00	1.6	287	18.0	1011
7/26/25 8:00	3.0	315	21.4	1012
7/26/25 9:00	3.9	314	23.3	1012
7/26/25 10:00	4.5	295	24.7	1012
7/26/25 11:00	4.8	295	25.8	1011
7/26/25 12:00	4.5	307	26.5	1011
7/26/25 13:00	3.7	317	27.1	1011
7/26/25 14:00	6.0	317	27.7	1010
7/26/25 15:00	4.4	326	28.4	1010
7/26/25 16:00	3.5	277	28.4	1010
7/26/25 17:00	3.0	296	27.9	1009
7/26/25 18:00	1.7	307	27.4	1009
7/26/25 19:00	0.8	319	23.9	1009
7/26/25 20:00	1.6	245	21.4	1010
7/26/25 21:00	0.9	272	19.1	1010
7/26/25 22:00	2.4	199	19.0	1010
7/26/25 23:00	2.4	185	17.8	1010
7/27/25 0:00	2.6	182	17.9	1010
7/27/25 1:00	2.4	188	17.1	1010
7/27/25 2:00	2.3	194	16.9	1009
7/27/25 3:00	2.3	186	16.1	1009
7/27/25 4:00	2.6	190	15.9	1009
7/27/25 5:00	2.1	182	15.9	1009
7/27/25 6:00	2.5	180	16.4	1009
7/27/25 7:00	3.7	190	18.1	1009
7/27/25 8:00	3.4	192	20.2	1008
7/27/25 9:00	3.0	198	22.5	1008
7/27/25 10:00	5.2	181	24.3	1008
7/27/25 11:00	5.8	178	26.1	1007
7/27/25 12:00	6.7	182	27.0	1006
7/27/25 13:00	6.8	181	25.5	1006
7/27/25 14:00	6.7	179	23.5	1006
7/27/25 15:00	5.5	175	22.0	1006
7/27/25 16:00	5.5	178	22.2	1005
7/27/25 17:00	4.8	187	22.9	1005
7/27/25 18:00	6.2	195	21.8	1005
7/27/25 19:00	4.7	202	20.4	1005
7/27/25 20:00	4.1	207	18.9	1006
7/27/25 21:00	4.2	210	18.9	1006
7/27/25 22:00	3.5	197	18.9	1006
7/27/25 23:00	2.9	183	18.0	1006
7/28/25 0:00	2.3	176	16.9	1006
7/28/25 1:00	2.4	183	16.9	1006
7/28/25 2:00	1.8	181	16.9	1006
7/28/25 3:00	2.2	197	16.8	1006
7/28/25 4:00	1.7	227	16.9	1006

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (7/24/25 12:00 to 8/7/25 12:00)**

7/28/25 5:00	2.3	189	16.9	1007
7/28/25 6:00	2.3	181	16.9	1006
7/28/25 7:00	2.6	218	17.1	1007
7/28/25 8:00	1.6	220	18.2	1008
7/28/25 9:00	1.3	200	20.1	1008
7/28/25 10:00	1.2	258	23.3	1008
7/28/25 11:00	1.9	204	26.8	1007
7/28/25 12:00	1.4	51	28.0	1007
7/28/25 13:00	1.7	182	29.0	1006
7/28/25 14:00	3.3	273	31.0	1006
7/28/25 15:00	3.9	283	31.9	1006
7/28/25 16:00	3.5	295	31.9	1006
7/28/25 17:00	3.1	326	31.5	1006
7/28/25 18:00	4.0	178	27.9	1006
7/28/25 19:00	4.3	175	25.2	1006
7/28/25 20:00	3.2	173	22.4	1006
7/28/25 21:00	2.3	160	21.6	1007
7/28/25 22:00	2.4	166	20.6	1007
7/28/25 23:00	1.6	180	19.2	1007
7/29/25 0:00	2.2	179	18.9	1006
7/29/25 1:00	2.6	171	18.0	1006
7/29/25 2:00	2.6	184	18.1	1006
7/29/25 3:00	2.6	198	17.9	1005
7/29/25 4:00	3.0	195	17.9	1005
7/29/25 5:00	2.6	188	17.9	1006
7/29/25 6:00	3.4	196	18.5	1005
7/29/25 7:00	2.6	205	20.6	1006
7/29/25 8:00	1.8	235	23.8	1006
7/29/25 9:00	2.3	231	27.3	1006
7/29/25 10:00	4.5	285	29.9	1006
7/29/25 11:00	4.7	272	30.7	1005
7/29/25 12:00	5.2	288	31.5	1005
7/29/25 13:00	4.0	287	32.1	1005
7/29/25 14:00	5.4	301	32.1	1004
7/29/25 15:00	5.4	310	33.0	1004
7/29/25 16:00	6.5	319	32.9	1004
7/29/25 17:00	6.5	328	32.1	1005
7/29/25 18:00	5.0	332	30.9	1005
7/29/25 19:00	3.1	324	28.5	1006
7/29/25 20:00	2.0	328	26.2	1006
7/29/25 21:00	2.7	319	24.7	1007
7/29/25 22:00	2.8	315	23.9	1007
7/29/25 23:00	2.9	263	23.0	1007
7/30/25 0:00	2.0	180	21.6	1008
7/30/25 1:00	2.0	338	20.4	1008
7/30/25 2:00	0.9	169	19.0	1008

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (7/24/25 12:00 to 8/7/25 12:00)**

7/30/25 3:00	0.4	340	17.2	1008
7/30/25 4:00	ND	ND	16.1	1008
7/30/25 5:00	0.3	358	16.0	1008
7/30/25 6:00	0.5	357	17.6	1008
7/30/25 7:00	1.5	309	20.3	1009
7/30/25 8:00	1.5	242	22.8	1008
7/30/25 9:00	2.1	213	24.1	1008
7/30/25 10:00	2.4	217	26.1	1008
7/30/25 11:00	3.2	260	27.8	1008
7/30/25 12:00	3.6	245	29.0	1007
7/30/25 13:00	3.9	235	30.2	1006
7/30/25 14:00	4.7	253	30.6	1006
7/30/25 15:00	4.5	208	30.6	1005
7/30/25 16:00	5.7	189	28.4	1005
7/30/25 17:00	4.3	185	27.4	1005
7/30/25 18:00	2.4	220	26.1	1006
7/30/25 19:00	2.6	242	25.7	1006
7/30/25 20:00	2.5	196	24.0	1006
7/30/25 21:00	2.4	176	23.2	1007
7/30/25 22:00	1.8	266	21.9	1007
7/30/25 23:00	2.2	293	22.8	1007
7/31/25 0:00	2.4	296	22.2	1007
7/31/25 1:00	2.1	283	21.9	1007
7/31/25 2:00	2.0	305	21.7	1008
7/31/25 3:00	1.1	341	20.9	1008
7/31/25 4:00	3.3	217	20.5	1008
7/31/25 5:00	3.1	183	19.1	1009
7/31/25 6:00	3.1	251	19.0	1010
7/31/25 7:00	3.7	50	18.8	1010
7/31/25 8:00	4.1	14	18.9	1011
7/31/25 9:00	4.9	20	19.4	1012
7/31/25 10:00	4.5	39	20.9	1012
7/31/25 11:00	3.4	149	20.9	1012
7/31/25 12:00	2.5	213	21.0	1012
7/31/25 13:00	2.3	172	21.7	1012
7/31/25 14:00	0.9	212	22.3	1012
7/31/25 15:00	0.7	311	22.0	1012
7/31/25 16:00	0.6	325	22.2	1012
7/31/25 17:00	2.0	83	21.9	1013
7/31/25 18:00	2.3	162	20.8	1013
7/31/25 19:00	0.5	312	19.3	1013
7/31/25 20:00	1.3	147	18.9	1014
7/31/25 21:00	1.0	284	18.2	1014
7/31/25 22:00	0.2	336	17.7	1014
7/31/25 23:00	ND	ND	17.1	1014
8/1/25 0:00	1.5	190	16.9	1015

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (7/24/25 12:00 to 8/7/25 12:00)**

8/1/25 1:00	1.2	313	15.9	1015
8/1/25 2:00	1.6	339	15.3	1014
8/1/25 3:00	2.1	51	14.2	1014
8/1/25 4:00	3.2	11	13.9	1015
8/1/25 5:00	3.5	115	13.5	1015
8/1/25 6:00	3.8	16	13.9	1016
8/1/25 7:00	3.9	22	15.7	1016
8/1/25 8:00	3.5	31	17.7	1016
8/1/25 9:00	4.0	69	19.8	1016
8/1/25 10:00	4.1	127	21.5	1016
8/1/25 11:00	4.5	62	22.5	1016
8/1/25 12:00	5.1	121	23.0	1015
8/1/25 13:00	4.8	158	23.7	1015
8/1/25 14:00	4.6	241	23.8	1015
8/1/25 15:00	4.0	151	23.9	1015
8/1/25 16:00	3.7	187	24.0	1014
8/1/25 17:00	3.4	190	23.9	1014
8/1/25 18:00	3.6	72	23.1	1015
8/1/25 19:00	2.7	71	22.1	1015
8/1/25 20:00	1.0	303	18.2	1015
8/1/25 21:00	0.8	348	15.5	1015
8/1/25 22:00	ND	ND	14.5	1015
8/1/25 23:00	2.1	350	13.2	1015
8/2/25 0:00	0.6	349	13.2	1015
8/2/25 1:00	ND	ND	11.4	1015
8/2/25 2:00	0.1	360	10.3	1015
8/2/25 3:00	0.1	360	9.6	1015
8/2/25 4:00	ND	ND	8.9	1015
8/2/25 5:00	ND	ND	8.5	1016
8/2/25 6:00	1.5	240	11.5	1016
8/2/25 7:00	1.9	210	14.2	1017
8/2/25 8:00	1.9	287	17.4	1017
8/2/25 9:00	3.3	323	20.5	1016
8/2/25 10:00	3.3	295	22.5	1016
8/2/25 11:00	4.9	307	23.6	1015
8/2/25 12:00	3.8	338	24.3	1015
8/2/25 13:00	3.2	304	25.3	1014
8/2/25 14:00	3.9	296	26.0	1014
8/2/25 15:00	3.6	302	26.2	1013
8/2/25 16:00	2.8	298	26.0	1013
8/2/25 17:00	2.5	301	25.9	1012
8/2/25 18:00	3.4	211	24.3	1012
8/2/25 19:00	3.4	184	21.5	1013
8/2/25 20:00	2.9	180	19.2	1013
8/2/25 21:00	3.6	194	18.6	1013
8/2/25 22:00	2.9	199	17.7	1013

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (7/24/25 12:00 to 8/7/25 12:00)**

8/2/25 23:00	2.1	186	16.9	1013
8/3/25 0:00	2.3	200	16.6	1013
8/3/25 1:00	1.7	193	15.6	1013
8/3/25 2:00	0.2	342	13.5	1012
8/3/25 3:00	1.1	221	12.6	1012
8/3/25 4:00	0.9	280	12.6	1012
8/3/25 5:00	0.6	291	13.0	1012
8/3/25 6:00	1.1	258	13.7	1013
8/3/25 7:00	2.3	205	16.5	1013
8/3/25 8:00	1.8	223	18.7	1013
8/3/25 9:00	2.4	225	21.8	1013
8/3/25 10:00	2.8	244	23.8	1012
8/3/25 11:00	3.9	189	26.1	1012
8/3/25 12:00	4.5	215	27.4	1011
8/3/25 13:00	4.9	229	27.6	1010
8/3/25 14:00	5.3	219	28.0	1010
8/3/25 15:00	4.6	231	27.6	1010
8/3/25 16:00	5.0	185	26.8	1010
8/3/25 17:00	5.1	207	26.6	1009
8/3/25 18:00	4.3	191	25.4	1010
8/3/25 19:00	3.9	182	23.0	1010
8/3/25 20:00	4.1	188	21.7	1010
8/3/25 21:00	4.0	198	20.7	1011
8/3/25 22:00	2.9	196	19.4	1011
8/3/25 23:00	2.8	195	18.7	1011
8/4/25 0:00	3.3	200	17.9	1011
8/4/25 1:00	1.0	270	16.4	1011
8/4/25 2:00	0.8	296	15.5	1011
8/4/25 3:00	0.9	263	14.7	1011
8/4/25 4:00	0.2	341	13.4	1010
8/4/25 5:00	0.1	360	12.8	1011
8/4/25 6:00	0.2	339	14.5	1012
8/4/25 7:00	0.6	288	17.2	1012
8/4/25 8:00	0.4	300	19.1	1012
8/4/25 9:00	1.4	260	21.8	1012
8/4/25 10:00	2.3	231	24.6	1012
8/4/25 11:00	2.6	231	26.3	1012
8/4/25 12:00	2.9	222	27.7	1012
8/4/25 13:00	3.5	275	28.4	1012
8/4/25 14:00	3.7	303	28.9	1011
8/4/25 15:00	3.8	317	28.7	1011
8/4/25 16:00	4.7	313	28.2	1012
8/4/25 17:00	5.2	307	27.5	1012
8/4/25 18:00	6.4	343	26.2	1013
8/4/25 19:00	4.7	350	24.0	1014
8/4/25 20:00	5.2	344	22.8	1015

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (7/24/25 12:00 to 8/7/25 12:00)**


8/4/25 21:00	3.4	331	21.4	1015
8/4/25 22:00	3.2	338	20.9	1016
8/4/25 23:00	4.6	153	20.8	1017
8/5/25 0:00	4.6	19	19.2	1017
8/5/25 1:00	3.6	19	17.8	1018
8/5/25 2:00	3.7	47	16.5	1018
8/5/25 3:00	3.1	17	15.3	1018
8/5/25 4:00	3.3	286	13.8	1019
8/5/25 5:00	2.0	14	14.0	1020
8/5/25 6:00	2.8	119	14.0	1020
8/5/25 7:00	3.3	188	15.5	1021
8/5/25 8:00	3.0	23	17.3	1021
8/5/25 9:00	3.5	32	19.4	1021
8/5/25 10:00	3.4	104	21.6	1022
8/5/25 11:00	3.8	253	22.9	1022
8/5/25 12:00	4.3	317	23.5	1021
8/5/25 13:00	3.8	151	24.6	1021
8/5/25 14:00	3.2	310	25.1	1021
8/5/25 15:00	3.5	29	25.9	1020
8/5/25 16:00	3.6	116	26.2	1020
8/5/25 17:00	3.4	186	25.9	1020
8/5/25 18:00	3.3	160	25.3	1020
8/5/25 19:00	2.9	24	23.3	1020
8/5/25 20:00	2.3	16	21.6	1021
8/5/25 21:00	1.8	17	18.3	1021
8/5/25 22:00	0.8	355	16.3	1022
8/5/25 23:00	0.8	330	14.3	1022
8/6/25 0:00	0.5	347	13.5	1022
8/6/25 1:00	ND	ND	12.2	1022
8/6/25 2:00	ND	ND	11.6	1022
8/6/25 3:00	0.2	350	11.6	1023
8/6/25 4:00	ND	ND	10.8	1023
8/6/25 5:00	0.5	357	10.8	1024
8/6/25 6:00	0.3	289	12.9	1024
8/6/25 7:00	1.4	87	16.2	1024
8/6/25 8:00	1.0	177	18.9	1024
8/6/25 9:00	1.0	149	21.8	1024
8/6/25 10:00	1.8	270	23.5	1024
8/6/25 11:00	2.4	206	25.1	1024
8/6/25 12:00	3.3	83	25.9	1024
8/6/25 13:00	3.2	318	26.6	1023
8/6/25 14:00	2.3	268	26.8	1023
8/6/25 15:00	1.8	266	27.7	1022
8/6/25 16:00	1.4	259	28.0	1022
8/6/25 17:00	3.1	164	27.1	1022
8/6/25 18:00	4.7	187	24.2	1022

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (7/24/25 12:00 to 8/7/25 12:00)**

8/6/25 19:00	3.8	183	22.3	1022
8/6/25 20:00	2.4	190	20.8	1022
8/6/25 21:00	2.0	195	18.9	1022
8/6/25 22:00	1.7	196	18.3	1022
8/6/25 23:00	0.2	299	16.0	1023
8/7/25 0:00	0.0	316	15.0	1022
8/7/25 1:00	0.9	236	13.9	1022
8/7/25 2:00	ND	ND	13.2	1022
8/7/25 3:00	ND	ND	12.5	1022
8/7/25 4:00	ND	ND	12.2	1022
8/7/25 5:00	ND	ND	12.0	1022
8/7/25 6:00	ND	ND	13.6	1022
8/7/25 7:00	1.0	20	17.0	1022
8/7/25 8:00	1.7	204	19.4	1022
8/7/25 9:00	1.3	201	21.5	1022
8/7/25 10:00	1.0	263	23.3	1021
8/7/25 11:00	1.2	181	25.7	1021
8/7/25 12:00	1.8	195	27.4	1020

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING
Bangor International Airport (BGR) Meteorological Data (7/24/25 12:00 to 8/7/25 12:00)

BGR Wind Rose 7/24/25 12:00 - 8/7/25 12:00



SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (8/7/25 12:00 to 8/21/25 12:00)**

Date & Time	Wind Speed	Wind Direction	Temperature	Barometric Pressure
	m/s	Deg.	°C	mb
8/7/25 12:00	1.8	195	27.4	1020
8/7/25 13:00	2.7	227	28.3	1020
8/7/25 14:00	2.3	198	28.8	1019
8/7/25 15:00	2.0	222	29.1	1018
8/7/25 16:00	2.1	278	29.6	1018
8/7/25 17:00	1.3	225	28.8	1017
8/7/25 18:00	2.9	182	26.6	1018
8/7/25 19:00	3.4	179	23.7	1017
8/7/25 20:00	3.5	192	22.1	1018
8/7/25 21:00	2.8	201	20.7	1018
8/7/25 22:00	1.8	226	19.3	1018
8/7/25 23:00	2.0	186	18.0	1018
8/8/25 0:00	2.1	174	17.3	1018
8/8/25 1:00	1.2	234	15.8	1018
8/8/25 2:00	0.2	343	13.5	1018
8/8/25 3:00	ND	ND	13.6	1017
8/8/25 4:00	ND	ND	13.3	1017
8/8/25 5:00	ND	ND	12.5	1017
8/8/25 6:00	ND	ND	14.5	1017
8/8/25 7:00	1.3	203	17.7	1017
8/8/25 8:00	1.8	208	20.3	1017
8/8/25 9:00	1.4	237	23.0	1017
8/8/25 10:00	0.4	344	25.8	1017
8/8/25 11:00	3.1	174	27.7	1017
8/8/25 12:00	2.9	163	28.9	1017
8/8/25 13:00	3.5	188	29.8	1016
8/8/25 14:00	6.2	175	29.3	1016
8/8/25 15:00	6.5	176	28.5	1016
8/8/25 16:00	7.0	175	27.5	1015
8/8/25 17:00	6.6	178	26.2	1015
8/8/25 18:00	5.8	184	24.8	1015
8/8/25 19:00	4.5	182	22.8	1015
8/8/25 20:00	3.6	190	21.3	1016
8/8/25 21:00	3.0	197	19.7	1017
8/8/25 22:00	2.2	190	18.4	1017
8/8/25 23:00	1.7	227	17.2	1017
8/9/25 0:00	2.8	186	17.0	1017
8/9/25 1:00	2.3	189	16.4	1017
8/9/25 2:00	2.5	180	15.9	1016
8/9/25 3:00	2.4	183	15.9	1016
8/9/25 4:00	2.0	194	15.8	1016
8/9/25 5:00	1.5	206	14.5	1017
8/9/25 6:00	2.8	191	15.9	1017

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (8/7/25 12:00 to 8/21/25 12:00)**

8/9/25 7:00	3.2	187	17.5	1017
8/9/25 8:00	3.5	193	19.9	1017
8/9/25 9:00	3.3	198	22.2	1017
8/9/25 10:00	3.7	200	24.3	1017
8/9/25 11:00	4.0	170	26.3	1016
8/9/25 12:00	5.1	179	26.8	1016
8/9/25 13:00	5.1	181	27.5	1015
8/9/25 14:00	4.9	176	27.8	1015
8/9/25 15:00	6.5	174	26.9	1014
8/9/25 16:00	6.8	173	25.9	1014
8/9/25 17:00	6.6	176	24.5	1014
8/9/25 18:00	5.0	172	22.7	1015
8/9/25 19:00	4.6	174	20.9	1015
8/9/25 20:00	3.2	179	19.0	1015
8/9/25 21:00	2.3	186	18.0	1015
8/9/25 22:00	2.5	211	17.4	1016
8/9/25 23:00	2.7	195	16.9	1016
8/10/25 0:00	2.5	177	16.1	1016
8/10/25 1:00	2.0	186	15.9	1015
8/10/25 2:00	1.9	204	15.7	1015
8/10/25 3:00	1.2	226	15.8	1015
8/10/25 4:00	0.9	262	15.6	1015
8/10/25 5:00	1.1	259	15.5	1015
8/10/25 6:00	1.2	267	15.9	1015
8/10/25 7:00	2.0	197	17.2	1016
8/10/25 8:00	3.0	193	19.3	1016
8/10/25 9:00	2.9	186	21.6	1016
8/10/25 10:00	3.3	190	23.8	1016
8/10/25 11:00	3.2	171	26.4	1015
8/10/25 12:00	4.0	188	28.2	1015
8/10/25 13:00	4.9	190	29.7	1014
8/10/25 14:00	5.6	176	30.8	1014
8/10/25 15:00	6.2	179	30.5	1014
8/10/25 16:00	6.6	183	28.9	1013
8/10/25 17:00	5.4	177	27.6	1013
8/10/25 18:00	4.9	177	26.3	1014
8/10/25 19:00	4.1	185	23.9	1014
8/10/25 20:00	3.9	184	21.9	1014
8/10/25 21:00	3.6	185	20.8	1014
8/10/25 22:00	2.8	200	20.2	1014
8/10/25 23:00	3.4	186	19.2	1014
8/11/25 0:00	3.3	188	18.9	1014
8/11/25 1:00	2.9	182	18.9	1014
8/11/25 2:00	2.0	197	18.9	1014
8/11/25 3:00	1.8	188	18.8	1013
8/11/25 4:00	0.9	269	17.9	1014

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (8/7/25 12:00 to 8/21/25 12:00)**

8/11/25 5:00	1.0	230	18.0	1014
8/11/25 6:00	1.7	199	18.5	1014
8/11/25 7:00	1.6	219	20.7	1015
8/11/25 8:00	1.3	236	23.0	1014
8/11/25 9:00	1.1	230	26.1	1014
8/11/25 10:00	1.7	202	29.1	1014
8/11/25 11:00	2.2	189	32.0	1014
8/11/25 12:00	2.0	228	33.1	1013
8/11/25 13:00	2.7	209	34.2	1012
8/11/25 14:00	3.2	202	34.5	1011
8/11/25 15:00	5.6	193	33.9	1011
8/11/25 16:00	7.2	185	32.8	1011
8/11/25 17:00	6.3	184	31.7	1011
8/11/25 18:00	4.8	187	30.2	1011
8/11/25 19:00	3.9	188	27.8	1011
8/11/25 20:00	2.9	186	25.7	1011
8/11/25 21:00	2.5	192	24.0	1011
8/11/25 22:00	2.8	190	23.0	1011
8/11/25 23:00	1.6	232	22.3	1011
8/12/25 0:00	2.1	192	21.9	1011
8/12/25 1:00	2.4	193	21.6	1011
8/12/25 2:00	2.1	190	20.9	1010
8/12/25 3:00	1.6	250	20.4	1010
8/12/25 4:00	ND	ND	18.8	1010
8/12/25 5:00	ND	ND	18.1	1010
8/12/25 6:00	0.4	185	20.6	1010
8/12/25 7:00	0.6	288	22.7	1010
8/12/25 8:00	0.8	292	25.4	1011
8/12/25 9:00	0.9	259	28.0	1010
8/12/25 10:00	1.7	197	30.5	1010
8/12/25 11:00	1.4	242	32.5	1009
8/12/25 12:00	1.6	220	33.6	1009
8/12/25 13:00	1.3	285	33.4	1008
8/12/25 14:00	2.7	228	35.3	1007
8/12/25 15:00	3.1	211	35.6	1007
8/12/25 16:00	5.6	191	34.5	1006
8/12/25 17:00	5.5	186	32.8	1006
8/12/25 18:00	4.3	173	31.2	1006
8/12/25 19:00	3.5	172	28.7	1006
8/12/25 20:00	2.6	190	26.7	1007
8/12/25 21:00	2.6	183	25.5	1007
8/12/25 22:00	2.4	192	23.9	1007
8/12/25 23:00	2.2	181	22.8	1006
8/13/25 0:00	1.8	215	21.9	1006
8/13/25 1:00	1.9	186	20.8	1006
8/13/25 2:00	1.1	193	19.2	1006

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (8/7/25 12:00 to 8/21/25 12:00)**

8/13/25 3:00	ND	ND	18.8	1005
8/13/25 4:00	1.5	190	18.5	1005
8/13/25 5:00	1.3	204	18.1	1005
8/13/25 6:00	1.5	188	19.0	1005
8/13/25 7:00	1.9	190	21.8	1005
8/13/25 8:00	3.1	183	24.6	1005
8/13/25 9:00	3.2	182	26.9	1005
8/13/25 10:00	3.4	180	29.6	1004
8/13/25 11:00	4.6	187	31.3	1004
8/13/25 12:00	5.4	176	32.6	1003
8/13/25 13:00	6.3	185	32.6	1002
8/13/25 14:00	7.1	183	32.5	1002
8/13/25 15:00	7.6	181	32.2	1001
8/13/25 16:00	7.1	173	30.5	1001
8/13/25 17:00	6.4	178	28.1	1001
8/13/25 18:00	5.9	173	25.9	1002
8/13/25 19:00	4.8	178	23.3	1002
8/13/25 20:00	3.7	194	21.7	1003
8/13/25 21:00	3.5	176	20.9	1003
8/13/25 22:00	3.1	174	20.9	1003
8/13/25 23:00	3.7	173	20.9	1002
8/14/25 0:00	3.5	185	20.9	1002
8/14/25 1:00	2.2	200	20.9	1002
8/14/25 2:00	1.6	186	20.9	1002
8/14/25 3:00	2.0	188	20.9	1002
8/14/25 4:00	1.7	182	20.9	1002
8/14/25 5:00	1.5	220	20.5	1003
8/14/25 6:00	ND	ND	20.5	1003
8/14/25 7:00	0.4	340	21.4	1003
8/14/25 8:00	0.5	340	22.6	1004
8/14/25 9:00	0.4	327	24.0	1005
8/14/25 10:00	2.2	30	26.1	1005
8/14/25 11:00	2.0	106	27.9	1005
8/14/25 12:00	2.4	151	29.0	1004
8/14/25 13:00	2.3	166	29.2	1004
8/14/25 14:00	2.7	243	27.9	1005
8/14/25 15:00	1.5	143	27.6	1005
8/14/25 16:00	1.2	274	28.2	1005
8/14/25 17:00	2.7	201	26.4	1006
8/14/25 18:00	2.1	204	25.2	1007
8/14/25 19:00	2.9	152	24.2	1007
8/14/25 20:00	1.4	160	23.4	1008
8/14/25 21:00	0.3	316	22.4	1009
8/14/25 22:00	1.0	304	21.9	1009
8/14/25 23:00	2.9	245	21.9	1010
8/15/25 0:00	5.6	23	22.1	1010

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (8/7/25 12:00 to 8/21/25 12:00)**

8/15/25 1:00	4.4	59	21.0	1011
8/15/25 2:00	4.0	12	19.0	1011
8/15/25 3:00	3.6	14	17.8	1012
8/15/25 4:00	3.3	10	16.9	1013
8/15/25 5:00	3.6	13	16.3	1014
8/15/25 6:00	3.8	13	16.4	1014
8/15/25 7:00	4.4	20	17.5	1015
8/15/25 8:00	4.7	130	19.1	1016
8/15/25 9:00	4.3	103	21.1	1016
8/15/25 10:00	5.4	129	22.5	1015
8/15/25 11:00	4.4	248	23.8	1015
8/15/25 12:00	3.9	251	24.9	1015
8/15/25 13:00	3.2	220	26.1	1015
8/15/25 14:00	3.1	184	27.1	1014
8/15/25 15:00	2.6	101	27.7	1014
8/15/25 16:00	3.4	87	27.8	1014
8/15/25 17:00	2.2	140	27.9	1014
8/15/25 18:00	2.8	110	27.4	1014
8/15/25 19:00	0.3	290	21.8	1014
8/15/25 20:00	ND	ND	18.6	1014
8/15/25 21:00	1.0	228	17.1	1015
8/15/25 22:00	1.0	281	16.4	1015
8/15/25 23:00	ND	ND	14.9	1015
8/16/25 0:00	ND	ND	13.3	1015
8/16/25 1:00	ND	ND	13.0	1015
8/16/25 2:00	ND	ND	12.3	1014
8/16/25 3:00	ND	ND	11.8	1014
8/16/25 4:00	ND	ND	11.1	1015
8/16/25 5:00	ND	ND	11.0	1015
8/16/25 6:00	ND	ND	12.2	1015
8/16/25 7:00	1.7	162	15.9	1015
8/16/25 8:00	2.8	181	19.8	1014
8/16/25 9:00	3.6	182	23.1	1014
8/16/25 10:00	4.2	198	25.7	1013
8/16/25 11:00	4.6	173	27.1	1013
8/16/25 12:00	5.3	173	28.1	1012
8/16/25 13:00	5.6	183	28.7	1011
8/16/25 14:00	5.9	182	28.3	1010
8/16/25 15:00	7.5	175	27.6	1010
8/16/25 16:00	7.4	180	26.8	1010
8/16/25 17:00	6.7	177	25.6	1010
8/16/25 18:00	5.7	175	23.4	1009
8/16/25 19:00	4.6	174	21.4	1009
8/16/25 20:00	3.9	181	20.4	1009
8/16/25 21:00	3.3	196	18.9	1009
8/16/25 22:00	1.9	230	18.2	1009

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (8/7/25 12:00 to 8/21/25 12:00)**

8/16/25 23:00	2.3	209	17.6	1009
8/17/25 0:00	3.1	180	17.0	1008
8/17/25 1:00	2.2	235	16.7	1008
8/17/25 2:00	1.4	236	16.0	1007
8/17/25 3:00	2.1	176	15.9	1006
8/17/25 4:00	2.8	180	16.3	1006
8/17/25 5:00	2.5	185	16.9	1006
8/17/25 6:00	2.6	197	17.0	1006
8/17/25 7:00	3.4	173	18.3	1005
8/17/25 8:00	3.9	196	20.0	1005
8/17/25 9:00	3.4	197	22.4	1004
8/17/25 10:00	3.5	183	24.0	1004
8/17/25 11:00	3.9	181	26.4	1003
8/17/25 12:00	4.1	170	28.1	1002
8/17/25 13:00	4.7	186	28.9	1002
8/17/25 14:00	4.7	204	28.4	1001
8/17/25 15:00	6.7	128	20.6	1003
8/17/25 16:00	4.6	33	21.7	1004
8/17/25 17:00	4.3	186	20.8	1005
8/17/25 18:00	4.0	220	18.1	1006
8/17/25 19:00	5.6	47	17.4	1006
8/17/25 20:00	7.1	255	16.0	1007
8/17/25 21:00	6.0	47	15.0	1008
8/17/25 22:00	5.1	187	14.1	1008
8/17/25 23:00	3.6	315	13.4	1009
8/18/25 0:00	3.8	304	12.6	1010
8/18/25 1:00	3.5	343	11.9	1010
8/18/25 2:00	4.5	342	11.5	1010
8/18/25 3:00	3.2	325	10.2	1011
8/18/25 4:00	4.4	327	9.4	1011
8/18/25 5:00	3.9	329	9.0	1012
8/18/25 6:00	3.4	333	9.8	1013
8/18/25 7:00	4.9	329	11.5	1013
8/18/25 8:00	4.8	339	13.2	1014
8/18/25 9:00	5.4	344	15.0	1014
8/18/25 10:00	5.7	244	16.5	1014
8/18/25 11:00	5.5	247	17.7	1014
8/18/25 12:00	6.4	341	19.0	1014
8/18/25 13:00	5.5	210	19.8	1014
8/18/25 14:00	6.3	245	20.9	1013
8/18/25 15:00	5.5	212	21.2	1013
8/18/25 16:00	5.7	318	21.8	1014
8/18/25 17:00	5.2	351	21.6	1014
8/18/25 18:00	5.2	337	20.5	1015
8/18/25 19:00	3.3	339	17.8	1015
8/18/25 20:00	2.7	332	15.7	1016

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (8/7/25 12:00 to 8/21/25 12:00)**


8/18/25 21:00	2.3	314	13.4	1016
8/18/25 22:00	2.1	83	13.7	1016
8/18/25 23:00	0.6	317	12.1	1017
8/19/25 0:00	ND	ND	10.3	1017
8/19/25 1:00	1.4	308	9.5	1017
8/19/25 2:00	0.4	337	8.7	1017
8/19/25 3:00	0.6	351	8.3	1018
8/19/25 4:00	0.5	345	7.7	1018
8/19/25 5:00	0.5	349	7.6	1018
8/19/25 6:00	1.3	316	9.0	1019
8/19/25 7:00	1.0	316	11.9	1019
8/19/25 8:00	1.8	79	13.6	1019
8/19/25 9:00	1.7	189	16.2	1019
8/19/25 10:00	2.5	232	18.1	1019
8/19/25 11:00	3.0	224	19.1	1018
8/19/25 12:00	2.9	100	20.0	1017
8/19/25 13:00	2.5	216	21.0	1017
8/19/25 14:00	2.0	110	21.7	1016
8/19/25 15:00	1.4	284	22.0	1015
8/19/25 16:00	2.3	249	22.2	1015
8/19/25 17:00	1.3	174	22.0	1015
8/19/25 18:00	2.5	173	21.0	1015
8/19/25 19:00	3.4	165	18.0	1015
8/19/25 20:00	3.1	176	16.4	1015
8/19/25 21:00	2.6	183	15.9	1015
8/19/25 22:00	2.9	192	15.7	1016
8/19/25 23:00	2.0	185	14.9	1016
8/20/25 0:00	1.4	202	13.9	1016
8/20/25 1:00	1.9	181	13.6	1015
8/20/25 2:00	1.5	213	13.3	1015
8/20/25 3:00	1.5	180	12.9	1015
8/20/25 4:00	1.5	215	12.7	1015
8/20/25 5:00	ND	ND	12.0	1015
8/20/25 6:00	ND	ND	12.5	1015
8/20/25 7:00	1.5	201	14.1	1015
8/20/25 8:00	2.6	201	16.0	1015
8/20/25 9:00	1.5	237	17.5	1016
8/20/25 10:00	1.4	221	18.9	1016
8/20/25 11:00	2.6	173	20.6	1015
8/20/25 12:00	1.9	168	21.1	1015
8/20/25 13:00	2.4	188	21.6	1015
8/20/25 14:00	2.9	188	21.3	1015
8/20/25 15:00	4.0	187	21.0	1014
8/20/25 16:00	3.9	188	20.9	1014
8/20/25 17:00	4.4	183	20.7	1014
8/20/25 18:00	3.6	188	19.3	1014

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (8/7/25 12:00 to 8/21/25 12:00)**

8/20/25 19:00	2.4	190	17.6	1015
8/20/25 20:00	2.0	182	16.0	1015
8/20/25 21:00	1.4	201	13.8	1015
8/20/25 22:00	1.8	203	13.2	1015
8/20/25 23:00	0.6	190	11.8	1015
8/21/25 0:00	ND	ND	10.2	1015
8/21/25 1:00	ND	ND	9.3	1014
8/21/25 2:00	ND	ND	8.9	1014
8/21/25 3:00	1.1	225	8.9	1014
8/21/25 4:00	0.9	270	8.6	1014
8/21/25 5:00	0.8	273	8.3	1014
8/21/25 6:00	ND	ND	8.9	1014
8/21/25 7:00	1.3	253	11.8	1014
8/21/25 8:00	1.4	220	14.5	1014
8/21/25 9:00	1.7	228	17.4	1014
8/21/25 10:00	1.4	107	19.7	1014
8/21/25 11:00	1.8	119	21.4	1013
8/21/25 12:00	2.5	138	22.4	1012

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING
Bangor International Airport (BGR) Meteorological Data (8/7/25 12:00 to 8/21/25 12:00)

BGR Wind Rose 8/7/25 12:00 - 8/21/25 12:00



SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (8/21/25 12:00 to 9/4/25 12:00)**

Date & Time	Wind Speed	Wind Direction	Temperature	Barometric Pressure
	m/s	Deg.	°C	mb
8/21/25 12:00	2.5	138	22.4	1012
8/21/25 13:00	2.7	164	23.5	1012
8/21/25 14:00	2.4	86	24.4	1011
8/21/25 15:00	0.7	262	24.8	1010
8/21/25 16:00	3.6	151	24.7	1010
8/21/25 17:00	4.8	181	22.6	1010
8/21/25 18:00	3.5	193	21.4	1010
8/21/25 19:00	2.4	195	19.3	1010
8/21/25 20:00	1.2	209	16.4	1010
8/21/25 21:00	0.8	275	13.8	1010
8/21/25 22:00	ND	ND	12.6	1009
8/21/25 23:00	ND	ND	11.9	1009
8/22/25 0:00	0.5	353	11.5	1009
8/22/25 1:00	0.7	352	10.3	1008
8/22/25 2:00	0.2	334	10.1	1008
8/22/25 3:00	ND	ND	9.4	1007
8/22/25 4:00	1.1	345	9.2	1007
8/22/25 5:00	0.3	325	9.3	1007
8/22/25 6:00	1.0	256	10.6	1008
8/22/25 7:00	2.6	252	13.7	1007
8/22/25 8:00	5.0	222	17.5	1007
8/22/25 9:00	4.3	87	20.6	1007
8/22/25 10:00	4.4	37	23.0	1007
8/22/25 11:00	5.2	186	24.3	1007
8/22/25 12:00	4.7	250	25.4	1007
8/22/25 13:00	5.3	261	26.1	1007
8/22/25 14:00	6.0	316	26.7	1006
8/22/25 15:00	5.3	221	27.0	1006
8/22/25 16:00	4.4	154	26.9	1006
8/22/25 17:00	2.7	134	26.7	1006
8/22/25 18:00	2.2	51	25.9	1006
8/22/25 19:00	0.2	329	21.1	1007
8/22/25 20:00	ND	ND	18.5	1008
8/22/25 21:00	2.8	218	17.9	1008
8/22/25 22:00	3.1	181	17.1	1009
8/22/25 23:00	2.9	179	16.2	1009
8/23/25 0:00	2.8	196	15.7	1009
8/23/25 1:00	2.8	178	14.7	1009
8/23/25 2:00	2.6	178	14.8	1009
8/23/25 3:00	1.1	267	13.9	1009
8/23/25 4:00	1.2	251	13.4	1009
8/23/25 5:00	1.3	201	12.8	1010
8/23/25 6:00	1.8	188	13.6	1010

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (8/21/25 12:00 to 9/4/25 12:00)**

8/23/25 7:00	2.0	214	15.9	1010
8/23/25 8:00	3.7	186	18.3	1010
8/23/25 9:00	4.5	184	21.4	1010
8/23/25 10:00	4.7	187	23.5	1010
8/23/25 11:00	6.0	186	24.9	1009
8/23/25 12:00	6.5	183	26.6	1009
8/23/25 13:00	6.7	173	27.3	1008
8/23/25 14:00	6.7	174	27.7	1008
8/23/25 15:00	7.9	181	26.4	1008
8/23/25 16:00	7.4	181	25.4	1008
8/23/25 17:00	7.0	175	23.5	1008
8/23/25 18:00	7.2	173	21.2	1008
8/23/25 19:00	6.5	176	19.1	1008
8/23/25 20:00	5.9	171	18.1	1009
8/23/25 21:00	6.9	173	18.0	1008
8/23/25 22:00	6.6	177	18.0	1008
8/23/25 23:00	5.7	173	17.1	1008
8/24/25 0:00	5.8	177	17.0	1009
8/24/25 1:00	5.5	169	17.0	1009
8/24/25 2:00	4.5	174	16.9	1008
8/24/25 3:00	4.8	177	17.0	1008
8/24/25 4:00	4.2	183	16.9	1008
8/24/25 5:00	4.6	182	16.9	1008
8/24/25 6:00	4.6	186	17.0	1008
8/24/25 7:00	5.2	187	18.1	1008
8/24/25 8:00	5.1	180	19.7	1008
8/24/25 9:00	5.4	181	21.4	1008
8/24/25 10:00	7.7	174	22.3	1008
8/24/25 11:00	8.4	174	23.7	1008
8/24/25 12:00	9.6	179	23.9	1007
8/24/25 13:00	9.8	180	23.9	1007
8/24/25 14:00	9.7	171	23.9	1007
8/24/25 15:00	7.6	182	23.6	1007
8/24/25 16:00	7.0	182	23.2	1007
8/24/25 17:00	6.4	172	21.9	1007
8/24/25 18:00	6.0	172	20.7	1007
8/24/25 19:00	5.7	173	19.1	1007
8/24/25 20:00	5.8	170	18.1	1007
8/24/25 21:00	5.9	165	18.0	1007
8/24/25 22:00	4.4	175	17.9	1007
8/24/25 23:00	4.2	169	17.9	1007
8/25/25 0:00	4.0	167	18.0	1006
8/25/25 1:00	4.9	177	17.9	1006
8/25/25 2:00	4.3	171	17.8	1006
8/25/25 3:00	3.2	166	17.9	1005
8/25/25 4:00	3.6	164	18.0	1005

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (8/21/25 12:00 to 9/4/25 12:00)**

8/25/25 5:00	3.8	175	18.0	1005
8/25/25 6:00	3.7	165	18.0	1004
8/25/25 7:00	3.3	151	18.7	1004
8/25/25 8:00	4.4	139	18.9	1004
8/25/25 9:00	3.9	159	18.9	1003
8/25/25 10:00	4.4	176	18.9	1003
8/25/25 11:00	5.4	181	19.1	1003
8/25/25 12:00	5.2	182	19.1	1002
8/25/25 13:00	5.3	176	19.0	1002
8/25/25 14:00	6.1	179	19.0	1001
8/25/25 15:00	4.5	188	19.0	1001
8/25/25 16:00	4.0	183	19.0	1001
8/25/25 17:00	4.4	158	19.0	1001
8/25/25 18:00	3.4	168	18.4	1001
8/25/25 19:00	3.5	181	18.0	1001
8/25/25 20:00	2.6	199	17.9	1002
8/25/25 21:00	2.3	192	17.9	1002
8/25/25 22:00	2.7	188	17.9	1002
8/25/25 23:00	2.0	213	17.8	1002
8/26/25 0:00	1.5	207	17.0	1002
8/26/25 1:00	1.3	225	17.6	1001
8/26/25 2:00	1.3	276	17.0	1001
8/26/25 3:00	1.8	244	16.9	1001
8/26/25 4:00	1.4	286	16.3	1001
8/26/25 5:00	2.5	247	16.0	1002
8/26/25 6:00	1.8	265	16.0	1002
8/26/25 7:00	1.1	282	16.5	1003
8/26/25 8:00	1.8	247	17.7	1003
8/26/25 9:00	2.9	287	19.9	1003
8/26/25 10:00	3.4	281	21.5	1003
8/26/25 11:00	3.5	264	22.0	1003
8/26/25 12:00	3.9	255	23.1	1003
8/26/25 13:00	4.3	255	23.2	1003
8/26/25 14:00	3.8	278	23.3	1003
8/26/25 15:00	4.6	269	23.7	1003
8/26/25 16:00	4.1	268	23.1	1004
8/26/25 17:00	4.1	271	22.5	1004
8/26/25 18:00	4.1	267	21.6	1005
8/26/25 19:00	2.3	259	19.8	1005
8/26/25 20:00	2.3	281	18.4	1006
8/26/25 21:00	1.8	268	16.1	1007
8/26/25 22:00	1.3	264	14.6	1007
8/26/25 23:00	1.1	258	13.0	1007
8/27/25 0:00	0.9	306	12.2	1007
8/27/25 1:00	ND	ND	10.7	1007
8/27/25 2:00	1.5	267	11.8	1007

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (8/21/25 12:00 to 9/4/25 12:00)**

8/27/25 3:00	1.9	256	11.9	1007
8/27/25 4:00	1.6	256	11.1	1007
8/27/25 5:00	1.2	289	10.4	1007
8/27/25 6:00	1.7	289	11.9	1008
8/27/25 7:00	1.7	288	14.1	1008
8/27/25 8:00	0.5	322	16.8	1008
8/27/25 9:00	2.8	288	18.8	1008
8/27/25 10:00	2.0	299	20.6	1007
8/27/25 11:00	3.2	294	21.6	1007
8/27/25 12:00	3.1	296	22.1	1006
8/27/25 13:00	3.5	277	21.7	1006
8/27/25 14:00	4.0	295	22.2	1006
8/27/25 15:00	3.0	317	22.0	1006
8/27/25 16:00	3.3	334	22.1	1006
8/27/25 17:00	2.5	329	21.1	1006
8/27/25 18:00	1.7	290	19.1	1006
8/27/25 19:00	2.5	198	16.7	1007
8/27/25 20:00	1.9	241	15.9	1008
8/27/25 21:00	1.7	280	15.5	1009
8/27/25 22:00	1.5	290	15.1	1009
8/27/25 23:00	2.6	257	13.9	1009
8/28/25 0:00	2.2	261	12.4	1009
8/28/25 1:00	2.7	267	12.9	1009
8/28/25 2:00	2.6	260	12.8	1009
8/28/25 3:00	1.6	227	11.8	1010
8/28/25 4:00	1.2	250	10.0	1010
8/28/25 5:00	ND	ND	8.9	1010
8/28/25 6:00	ND	ND	10.0	1011
8/28/25 7:00	0.3	322	12.6	1011
8/28/25 8:00	1.7	272	15.7	1012
8/28/25 9:00	1.8	252	17.9	1012
8/28/25 10:00	2.2	262	20.4	1012
8/28/25 11:00	2.5	225	21.7	1012
8/28/25 12:00	4.1	226	22.7	1011
8/28/25 13:00	4.2	219	23.5	1011
8/28/25 14:00	5.4	178	23.3	1010
8/28/25 15:00	6.3	179	23.1	1010
8/28/25 16:00	6.9	176	22.3	1010
8/28/25 17:00	6.6	178	21.4	1010
8/28/25 18:00	5.6	177	19.1	1010
8/28/25 19:00	3.8	169	17.3	1011
8/28/25 20:00	3.2	165	15.8	1011
8/28/25 21:00	2.4	161	14.5	1011
8/28/25 22:00	1.9	173	14.0	1011
8/28/25 23:00	1.7	148	13.1	1011
8/29/25 0:00	0.3		12.6	1011

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (8/21/25 12:00 to 9/4/25 12:00)**

8/29/25 1:00	1.3	160	12.8	1010
8/29/25 2:00	0.8	320	12.4	1010
8/29/25 3:00	ND	ND	11.2	1010
8/29/25 4:00	1.5	150	12.1	1009
8/29/25 5:00	2.0	149	14.6	1010
8/29/25 6:00	3.1	170	15.9	1010
8/29/25 7:00	3.1	187	16.5	1010
8/29/25 8:00	4.1	178	17.7	1010
8/29/25 9:00	4.9	177	19.2	1009
8/29/25 10:00	4.9	175	21.1	1009
8/29/25 11:00	7.1	185	20.0	1009
8/29/25 12:00	4.5	190	16.8	1009
8/29/25 13:00	5.1	147	16.5	1009
8/29/25 14:00	3.1	153	16.4	1009
8/29/25 15:00	4.5	134	17.3	1008
8/29/25 16:00	3.4	154	17.9	1008
8/29/25 17:00	3.9	186	17.9	1008
8/29/25 18:00	3.5	170	17.8	1008
8/29/25 19:00	2.1	162	16.9	1007
8/29/25 20:00	2.4	142	16.9	1007
8/29/25 21:00	2.5	137	16.6	1006
8/29/25 22:00	2.3	149	15.9	1006
8/29/25 23:00	2.1	162	15.9	1006
8/30/25 0:00	1.6	148	15.9	1006
8/30/25 1:00	1.6	160	15.9	1006
8/30/25 2:00	2.1	130	15.3	1005
8/30/25 3:00	2.2	140	15.6	1005
8/30/25 4:00	1.6	137	15.2	1004
8/30/25 5:00	1.9	144	15.8	1004
8/30/25 6:00	2.2	160	15.5	1005
8/30/25 7:00	2.3	194	15.8	1005
8/30/25 8:00	3.1	212	16.0	1005
8/30/25 9:00	2.8	254	16.6	1005
8/30/25 10:00	4.2	309	17.3	1005
8/30/25 11:00	4.0	311	18.1	1005
8/30/25 12:00	4.0	290	18.3	1006
8/30/25 13:00	3.5	282	19.0	1006
8/30/25 14:00	3.4	283	19.5	1006
8/30/25 15:00	3.5	280	20.1	1006
8/30/25 16:00	2.9	282	20.6	1006
8/30/25 17:00	4.1	253	19.9	1006
8/30/25 18:00	2.5	281	17.7	1007
8/30/25 19:00	2.4	265	16.3	1008
8/30/25 20:00	3.1	246	15.9	1009
8/30/25 21:00	2.6	249	14.9	1009
8/30/25 22:00	2.5	258	13.3	1010

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (8/21/25 12:00 to 9/4/25 12:00)**

8/30/25 23:00	2.0	247	12.2	1010
8/31/25 0:00	2.5	263	11.7	1010
8/31/25 1:00	0.5	328	10.6	1010
8/31/25 2:00	1.5	300	9.8	1010
8/31/25 3:00	1.6	272	9.4	1011
8/31/25 4:00	ND	ND	8.3	1011
8/31/25 5:00	1.8	230	8.5	1012
8/31/25 6:00	2.8	265	10.7	1012
8/31/25 7:00	2.2	264	12.3	1013
8/31/25 8:00	2.7	289	14.8	1013
8/31/25 9:00	3.3	318	17.2	1014
8/31/25 10:00	2.4	342	18.8	1014
8/31/25 11:00	2.5	269	20.8	1014
8/31/25 12:00	2.2	299	21.4	1014
8/31/25 13:00	3.0	272	22.9	1013
8/31/25 14:00	3.2	267	23.9	1013
8/31/25 15:00	4.1	299	24.7	1013
8/31/25 16:00	3.8	211	24.3	1013
8/31/25 17:00	3.0	329	24.2	1014
8/31/25 18:00	2.3	325	22.6	1014
8/31/25 19:00	1.9	298	19.3	1015
8/31/25 20:00	0.8	330	18.0	1015
8/31/25 21:00	0.1	360	15.7	1016
8/31/25 22:00	ND	ND	15.1	1016
8/31/25 23:00	ND	ND	13.3	1016
9/1/25 0:00	ND	ND	12.8	1017
9/1/25 1:00	ND	ND	12.2	1017
9/1/25 2:00	1.3	ND	11.5	1017
9/1/25 3:00	ND	ND	10.9	1017
9/1/25 4:00	0.6	350	10.7	1018
9/1/25 5:00	ND	ND	10.1	1018
9/1/25 6:00	0.0	359	10.5	1018
9/1/25 7:00	0.5	360	13.6	1019
9/1/25 8:00	1.3	228	16.8	1019
9/1/25 9:00	1.7	163	19.2	1019
9/1/25 10:00	1.7	292	22.0	1019
9/1/25 11:00	1.7	171	23.6	1019
9/1/25 12:00	2.1	228	24.8	1019
9/1/25 13:00	3.0	265	25.9	1018
9/1/25 14:00	2.6	187	26.1	1018
9/1/25 15:00	1.5	244	26.4	1017
9/1/25 16:00	0.9	273	25.9	1017
9/1/25 17:00	2.6	200	25.9	1017
9/1/25 18:00	2.5	265	24.5	1017
9/1/25 19:00	2.6	186	20.9	1017
9/1/25 20:00	2.2	191	18.9	1017

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (8/21/25 12:00 to 9/4/25 12:00)**


9/1/25 21:00	2.0	185	18.0	1017
9/1/25 22:00	1.8	237	16.9	1017
9/1/25 23:00	0.3	319	14.3	1017
9/2/25 0:00	1.5	210	14.1	1017
9/2/25 1:00	ND	ND	13.3	1016
9/2/25 2:00	ND	ND	12.9	1016
9/2/25 3:00	ND	ND	12.8	1016
9/2/25 4:00	ND	ND	12.0	1016
9/2/25 5:00	0.6	337	12.2	1016
9/2/25 6:00	ND	ND	12.5	1016
9/2/25 7:00	ND	ND	14.1	1016
9/2/25 8:00	ND	ND	16.6	1016
9/2/25 9:00	0.0	ND	18.6	1015
9/2/25 10:00	0.9	110	20.9	1015
9/2/25 11:00	0.7	225	22.7	1014
9/2/25 12:00	0.6	340	24.2	1014
9/2/25 13:00	1.5	260	25.5	1013
9/2/25 14:00	1.3	273	25.4	1012
9/2/25 15:00	1.7	219	25.7	1012
9/2/25 16:00	3.1	195	25.8	1011
9/2/25 17:00	3.3	187	25.1	1011
9/2/25 18:00	3.8	189	22.6	1011
9/2/25 19:00	3.6	179	20.3	1011
9/2/25 20:00	2.9	174	18.4	1011
9/2/25 21:00	2.7	176	17.3	1011
9/2/25 22:00	2.0	189	15.9	1010
9/2/25 23:00	0.8	251	14.2	1010
9/3/25 0:00	1.6	182	13.7	1010
9/3/25 1:00	1.1	236	13.2	1010
9/3/25 2:00	0.4	318	12.5	1009
9/3/25 3:00	ND	ND	12.0	1009
9/3/25 4:00	2.4	186	12.7	1009
9/3/25 5:00	2.8	186	13.0	1008
9/3/25 6:00	1.8	186	13.4	1009
9/3/25 7:00	2.2	188	13.9	1008
9/3/25 8:00	3.5	197	15.3	1008
9/3/25 9:00	4.1	203	16.9	1008
9/3/25 10:00	4.4	196	19.6	1008
9/3/25 11:00	3.3	192	21.4	1008
9/3/25 12:00	4.6	182	22.5	1007
9/3/25 13:00	4.7	183	23.2	1006
9/3/25 14:00	4.9	181	23.9	1005
9/3/25 15:00	5.4	194	23.3	1005
9/3/25 16:00	5.6	189	22.7	1005
9/3/25 17:00	6.1	187	21.4	1005
9/3/25 18:00	4.9	175	19.4	1005

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (8/21/25 12:00 to 9/4/25 12:00)**

9/3/25 19:00	4.4	180	17.8	1005
9/3/25 20:00	4.7	186	16.9	1005
9/3/25 21:00	2.4	182	16.1	1006
9/3/25 22:00	1.9	188	15.5	1006
9/3/25 23:00	0.4	324	13.9	1006
9/4/25 0:00	1.1	248	13.8	1006
9/4/25 1:00	2.2	201	13.6	1006
9/4/25 2:00	2.4	195	14.0	1006
9/4/25 3:00	2.0	228	13.9	1006
9/4/25 4:00	1.7	220	14.0	1006
9/4/25 5:00	0.3	323	13.9	1006
9/4/25 6:00	0.6	325	14.8	1006
9/4/25 7:00	1.6	202	15.8	1007
9/4/25 8:00	2.4	203	16.6	1007
9/4/25 9:00	3.0	202	18.9	1007
9/4/25 10:00	3.5	189	21.2	1007
9/4/25 11:00	4.5	184	22.6	1006
9/4/25 12:00	5.8	180	23.7	1006

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING
Bangor International Airport (BGR) Meteorological Data (8/21/25 12:00 to 9/4/25 12:00)

BGR Wind Rose 8/21/25 12:00 - 9/4/25 12:00



SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (9/4/25 12:00 to 9/18/25 12:00)**

Date & Time	Wind Speed	Wind Direction	Temperature	Barometric Pressure
	m/s	Deg.	°C	mb
9/4/25 12:00	5.8	180	23.7	1006
9/4/25 13:00	6.5	180	23.9	1006
9/4/25 14:00	6.3	197	23.9	1005
9/4/25 15:00	6.5	190	23.2	1005
9/4/25 16:00	6.0	181	22.5	1005
9/4/25 17:00	5.8	188	21.3	1005
9/4/25 18:00	5.1	189	19.3	1005
9/4/25 19:00	5.2	176	17.3	1005
9/4/25 20:00	3.1	169	16.6	1006
9/4/25 21:00	2.9	177	15.9	1006
9/4/25 22:00	2.7	177	16.0	1006
9/4/25 23:00	2.9	162	15.9	1006
9/5/25 0:00	2.3	157	16.2	1006
9/5/25 1:00	3.2	175	17.0	1006
9/5/25 2:00	3.8	169	17.0	1006
9/5/25 3:00	3.4	164	17.0	1006
9/5/25 4:00	4.6	165	17.2	1006
9/5/25 5:00	4.6	177	17.9	1007
9/5/25 6:00	4.8	177	17.9	1007
9/5/25 7:00	5.2	179	18.0	1007
9/5/25 8:00	4.7	169	18.0	1007
9/5/25 9:00	5.7	176	18.2	1008
9/5/25 10:00	5.4	177	19.2	1008
9/5/25 11:00	6.1	173	21.3	1007
9/5/25 12:00	6.1	179	22.0	1007
9/5/25 13:00	7.1	181	21.0	1007
9/5/25 14:00	7.3	178	20.3	1007
9/5/25 15:00	6.7	178	20.2	1006
9/5/25 16:00	6.0	179	19.3	1006
9/5/25 17:00	5.4	179	19.0	1006
9/5/25 18:00	4.7	175	19.0	1006
9/5/25 19:00	6.0	172	18.4	1006
9/5/25 20:00	6.0	171	18.0	1006
9/5/25 21:00	5.2	174	18.0	1006
9/5/25 22:00	4.8	170	18.4	1006
9/5/25 23:00	3.7	175	18.8	1006
9/6/25 0:00	3.3	168	18.1	1006
9/6/25 1:00	3.4	183	18.0	1005
9/6/25 2:00	4.1	180	18.0	1005
9/6/25 3:00	4.2	175	18.0	1006
9/6/25 4:00	4.2	184	18.0	1006
9/6/25 5:00	4.0	181	18.0	1006
9/6/25 6:00	2.5	213	18.4	1007

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (9/4/25 12:00 to 9/18/25 12:00)**

9/6/25 7:00	2.4	180	18.9	1008
9/6/25 8:00	2.7	193	19.4	1008
9/6/25 9:00	2.4	167	20.7	1008
9/6/25 10:00	2.9	186	21.5	1008
9/6/25 11:00	4.0	168	22.5	1008
9/6/25 12:00	4.6	174	23.7	1007
9/6/25 13:00	5.2	170	24.0	1006
9/6/25 14:00	5.4	192	24.3	1006
9/6/25 15:00	5.5	171	23.9	1005
9/6/25 16:00	6.1	174	22.9	1004
9/6/25 17:00	6.2	201	20.7	1005
9/6/25 18:00	5.3	341	16.5	1007
9/6/25 19:00	3.7	179	15.8	1007
9/6/25 20:00	3.4	32	14.8	1008
9/6/25 21:00	3.5	156	14.1	1008
9/6/25 22:00	3.1	194	14.0	1008
9/6/25 23:00	1.9	72	14.0	1008
9/7/25 0:00	2.0	310	14.0	1009
9/7/25 1:00	2.2	329	14.0	1009
9/7/25 2:00	2.2	336	14.0	1008
9/7/25 3:00	1.1	334	14.0	1008
9/7/25 4:00	1.5	353	14.0	1009
9/7/25 5:00	1.0	ND	14.0	1009
9/7/25 6:00	ND	ND	14.0	1009
9/7/25 7:00	ND	ND	14.0	1010
9/7/25 8:00	0.4	315	14.1	1010
9/7/25 9:00	1.9	164	14.9	1010
9/7/25 10:00	2.1	179	14.8	1010
9/7/25 11:00	2.3	156	14.9	1010
9/7/25 12:00	1.9	176	14.8	1010
9/7/25 13:00	3.1	191	14.9	1009
9/7/25 14:00	2.1	188	15.0	1008
9/7/25 15:00	1.2	210	15.4	1008
9/7/25 16:00	1.7	248	15.9	1008
9/7/25 17:00	1.7	260	16.0	1008
9/7/25 18:00	ND	ND	15.3	1008
9/7/25 19:00	1.0	267	14.4	1009
9/7/25 20:00	2.0	187	14.0	1009
9/7/25 21:00	1.9	246	14.0	1010
9/7/25 22:00	1.5	217	14.0	1011
9/7/25 23:00	2.3	218	13.9	1011
9/8/25 0:00	1.0	273	13.9	1011
9/8/25 1:00	1.0	293	13.9	1011
9/8/25 2:00	1.9	255	13.4	1012
9/8/25 3:00	2.4	260	12.9	1012
9/8/25 4:00	2.1	282	11.9	1013

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (9/4/25 12:00 to 9/18/25 12:00)**

9/8/25 5:00	2.0	274	11.0	1013
9/8/25 6:00	2.5	262	10.9	1014
9/8/25 7:00	2.3	257	12.1	1015
9/8/25 8:00	2.6	280	14.2	1016
9/8/25 9:00	3.2	295	16.2	1016
9/8/25 10:00	2.9	286	17.4	1016
9/8/25 11:00	2.7	281	18.3	1016
9/8/25 12:00	3.5	294	18.7	1016
9/8/25 13:00	3.4	287	19.1	1016
9/8/25 14:00	3.8	293	19.0	1016
9/8/25 15:00	2.9	281	19.7	1016
9/8/25 16:00	2.8	285	19.0	1016
9/8/25 17:00	1.6	289	18.8	1017
9/8/25 18:00	0.3	326	16.3	1017
9/8/25 19:00	0.4	349	13.4	1017
9/8/25 20:00	0.6	284	12.4	1018
9/8/25 21:00	ND	ND	11.5	1019
9/8/25 22:00	ND	ND	11.2	1019
9/8/25 23:00	0.5	320	10.8	1020
9/9/25 0:00	ND	ND	9.6	1020
9/9/25 1:00	ND	ND	8.3	1021
9/9/25 2:00	ND	ND	7.4	1021
9/9/25 3:00	ND	ND	6.8	1021
9/9/25 4:00	ND	ND	6.4	1021
9/9/25 5:00	ND	ND	6.2	1021
9/9/25 6:00	0.5	300	6.2	1022
9/9/25 7:00	ND	ND	7.9	1022
9/9/25 8:00	2.1	280	10.2	1023
9/9/25 9:00	2.1	280	12.5	1023
9/9/25 10:00	0.3	ND	17.4	1023
9/9/25 11:00	0.7	223	18.8	1022
9/9/25 12:00	1.9	225	19.6	1022
9/9/25 13:00	1.8	274	20.6	1022
9/9/25 14:00	2.6	229	21.1	1021
9/9/25 15:00	3.0	174	21.1	1021
9/9/25 16:00	3.4	185	20.7	1021
9/9/25 17:00	4.3	190	19.3	1021
9/9/25 18:00	3.4	194	17.4	1021
9/9/25 19:00	2.1	185	15.5	1021
9/9/25 20:00	2.0	182	13.9	1021
9/9/25 21:00	1.5	183	12.8	1021
9/9/25 22:00	0.3	314	11.0	1021
9/9/25 23:00	ND	ND	9.0	1021
9/10/25 0:00	1.5	170	9.0	1021
9/10/25 1:00	2.0	177	9.8	1021
9/10/25 2:00	ND	ND	7.9	1021

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (9/4/25 12:00 to 9/18/25 12:00)**

9/10/25 3:00	0.4	309	7.4	1021
9/10/25 4:00	0.2	340	6.9	1020
9/10/25 5:00	ND	ND	7.0	1020
9/10/25 6:00	ND	ND	6.7	1020
9/10/25 7:00	ND	ND	7.6	1021
9/10/25 8:00	ND	ND	9.4	1020
9/10/25 9:00	1.0	ND	13.5	1020
9/10/25 10:00	1.5	146	17.4	1020
9/10/25 11:00	1.6	200	18.8	1019
9/10/25 12:00	2.1	167	19.6	1018
9/10/25 13:00	2.2	210	20.6	1017
9/10/25 14:00	2.0	197	21.0	1016
9/10/25 15:00	3.0	206	21.2	1016
9/10/25 16:00	2.5	171	21.0	1015
9/10/25 17:00	2.3	183	20.8	1014
9/10/25 18:00	2.5	186	19.1	1014
9/10/25 19:00	2.4	172	16.9	1014
9/10/25 20:00	2.1	193	14.9	1014
9/10/25 21:00	1.0	250	12.4	1013
9/10/25 22:00	0.9	310	11.3	1013
9/10/25 23:00	0.3	330	10.2	1013
9/11/25 0:00	ND	ND	8.9	1012
9/11/25 1:00	ND	ND	8.4	1012
9/11/25 2:00	ND	ND	7.4	1011
9/11/25 3:00	ND	ND	6.9	1010
9/11/25 4:00	0.1	360	6.7	1010
9/11/25 5:00	0.2	355	6.2	1010
9/11/25 6:00	ND	ND	6.6	1010
9/11/25 7:00	1.5	340	9.4	1010
9/11/25 8:00	0.2	355	12.5	1010
9/11/25 9:00	1.1	339	16.3	1010
9/11/25 10:00	1.7	196	18.8	1010
9/11/25 11:00	3.2	311	21.8	1010
9/11/25 12:00	3.6	242	23.0	1010
9/11/25 13:00	4.9	346	23.3	1009
9/11/25 14:00	6.3	316	23.3	1009
9/11/25 15:00	7.3	304	23.8	1009
9/11/25 16:00	6.2	345	23.4	1010
9/11/25 17:00	4.5	344	22.9	1010
9/11/25 18:00	4.0	339	20.9	1010
9/11/25 19:00	4.0	242	18.5	1011
9/11/25 20:00	2.7	185	17.4	1012
9/11/25 21:00	3.6	220	16.8	1012
9/11/25 22:00	3.8	80	15.8	1013
9/11/25 23:00	1.4	193	13.2	1013
9/12/25 0:00	0.3	355	11.2	1014

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (9/4/25 12:00 to 9/18/25 12:00)**

9/12/25 1:00	0.6	330	9.9	1014
9/12/25 2:00	0.0	360	8.7	1015
9/12/25 3:00	0.6	289	8.5	1015
9/12/25 4:00	ND	ND	6.9	1015
9/12/25 5:00	0.2	352	6.5	1016
9/12/25 6:00	0.3	347	6.8	1016
9/12/25 7:00	0.9	326	9.5	1017
9/12/25 8:00	2.7	308	12.4	1017
9/12/25 9:00	3.5	89	15.1	1017
9/12/25 10:00	2.4	148	16.4	1017
9/12/25 11:00	3.2	279	17.7	1017
9/12/25 12:00	4.2	329	19.1	1016
9/12/25 13:00	3.8	312	20.4	1016
9/12/25 14:00	4.1	315	20.9	1015
9/12/25 15:00	2.8	260	21.3	1015
9/12/25 16:00	3.1	310	21.6	1015
9/12/25 17:00	1.9	330	20.8	1015
9/12/25 18:00	1.2	308	16.3	1015
9/12/25 19:00	ND	ND	13.5	1015
9/12/25 20:00	ND	ND	12.1	1015
9/12/25 21:00	ND	343	11.1	1015
9/12/25 22:00	0.5	305	9.5	1015
9/12/25 23:00	1.5	230	8.1	1015
9/13/25 0:00	0.8	290	8.3	1015
9/13/25 1:00	ND	ND	7.4	1015
9/13/25 2:00	ND	ND	6.8	1015
9/13/25 3:00	ND	ND	6.6	1015
9/13/25 4:00	ND	ND	6.6	1015
9/13/25 5:00	ND	ND	6.4	1015
9/13/25 6:00	0.8	317	6.9	1015
9/13/25 7:00	2.4	190	9.2	1014
9/13/25 8:00	2.0	167	12.7	1014
9/13/25 9:00	3.6	175	16.5	1014
9/13/25 10:00	4.8	186	18.2	1014
9/13/25 11:00	5.0	181	19.0	1013
9/13/25 12:00	4.0	178	20.1	1013
9/13/25 13:00	5.1	190	20.6	1012
9/13/25 14:00	5.0	177	20.6	1011
9/13/25 15:00	5.2	179	20.2	1011
9/13/25 16:00	5.9	174	18.8	1010
9/13/25 17:00	5.4	175	17.7	1010
9/13/25 18:00	4.2	181	16.6	1010
9/13/25 19:00	3.5	177	15.9	1010
9/13/25 20:00	2.6	190	15.1	1010
9/13/25 21:00	2.2	191	14.0	1011
9/13/25 22:00	2.5	198	13.8	1011

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (9/4/25 12:00 to 9/18/25 12:00)**

9/13/25 23:00	1.8	226	12.9	1010
9/14/25 0:00	0.7	261	12.1	1010
9/14/25 1:00	1.5	181	11.5	1010
9/14/25 2:00	1.1	240	10.1	1009
9/14/25 3:00	1.6	198	10.3	1009
9/14/25 4:00	1.4	210	11.4	1009
9/14/25 5:00	0.4	318	11.9	1009
9/14/25 6:00	1.6	186	12.1	1009
9/14/25 7:00	1.5	188	13.0	1009
9/14/25 8:00	2.1	196	14.0	1009
9/14/25 9:00	1.4	271	16.3	1009
9/14/25 10:00	1.0	250	18.8	1009
9/14/25 11:00	2.1	186	21.2	1009
9/14/25 12:00	2.2	205	22.8	1009
9/14/25 13:00	3.2	274	23.4	1009
9/14/25 14:00	2.1	270	23.8	1009
9/14/25 15:00	4.9	253	20.2	1009
9/14/25 16:00	0.3	352	21.2	1010
9/14/25 17:00	ND	ND	21.0	1010
9/14/25 18:00	0.2	356	18.7	1010
9/14/25 19:00	1.4	228	18.4	1011
9/14/25 20:00	ND	ND	17.9	1012
9/14/25 21:00	2.3	70	17.8	1013
9/14/25 22:00	3.1	185	17.9	1013
9/14/25 23:00	3.6	81	16.7	1014
9/15/25 0:00	3.3	243	15.3	1014
9/15/25 1:00	3.4	152	13.8	1015
9/15/25 2:00	1.4	355	12.0	1015
9/15/25 3:00	0.9	242	10.4	1016
9/15/25 4:00	0.7	344	10.4	1016
9/15/25 5:00	1.3	345	10.0	1017
9/15/25 6:00	0.6	314	9.2	1018
9/15/25 7:00	1.7	228	11.6	1018
9/15/25 8:00	2.7	56	14.5	1019
9/15/25 9:00	3.2	25	17.1	1019
9/15/25 10:00	2.0	61	18.6	1019
9/15/25 11:00	2.4	120	20.2	1019
9/15/25 12:00	2.7	155	21.2	1018
9/15/25 13:00	2.2	184	22.0	1018
9/15/25 14:00	1.2	247	22.7	1017
9/15/25 15:00	0.5	287	23.0	1017
9/15/25 16:00	1.0	278	22.9	1017
9/15/25 17:00	1.0	327	22.2	1017
9/15/25 18:00	0.3	346	18.0	1017
9/15/25 19:00	0.2	342	16.2	1018
9/15/25 20:00	1.7	236	16.1	1018

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (9/4/25 12:00 to 9/18/25 12:00)**


9/15/25 21:00	0.8	277	14.3	1018
9/15/25 22:00	0.7	303	13.2	1018
9/15/25 23:00	ND	ND	11.6	1018
9/16/25 0:00	ND	ND	10.8	1018
9/16/25 1:00	1.5	ND	11.0	1018
9/16/25 2:00	1.5	190	10.4	1018
9/16/25 3:00	1.5	345	10.3	1018
9/16/25 4:00	ND	ND	9.0	1018
9/16/25 5:00	ND	ND	8.5	1018
9/16/25 6:00	1.9	333	8.4	1019
9/16/25 7:00	1.6	154	10.6	1019
9/16/25 8:00	2.2	113	12.1	1020
9/16/25 9:00	1.3	252	14.9	1020
9/16/25 10:00	1.5	340	18.9	1020
9/16/25 11:00	2.0	104	21.6	1019
9/16/25 12:00	2.2	274	22.7	1019
9/16/25 13:00	2.6	257	23.4	1018
9/16/25 14:00	2.4	274	24.0	1018
9/16/25 15:00	1.0	200	24.3	1017
9/16/25 16:00	1.4	322	24.0	1017
9/16/25 17:00	1.7	263	23.3	1017
9/16/25 18:00	0.7	339	19.0	1017
9/16/25 19:00	1.7	219	17.4	1017
9/16/25 20:00	1.2	183	16.8	1017
9/16/25 21:00	1.1	243	14.6	1017
9/16/25 22:00	ND	ND	13.1	1017
9/16/25 23:00	ND	ND	12.4	1017
9/17/25 0:00	ND	ND	12.0	1017
9/17/25 1:00	ND	ND	11.4	1017
9/17/25 2:00	1.5	190	11.3	1017
9/17/25 3:00	ND	ND	10.8	1016
9/17/25 4:00	ND	ND	10.1	1016
9/17/25 5:00	ND	ND	9.5	1016
9/17/25 6:00	1.5	235	9.1	1016
9/17/25 7:00	ND	ND	11.2	1016
9/17/25 8:00	1.4	197	13.9	1016
9/17/25 9:00	1.9	190	16.7	1016
9/17/25 10:00	2.2	178	20.0	1016
9/17/25 11:00	2.3	181	21.9	1015
9/17/25 12:00	2.3	190	22.2	1015
9/17/25 13:00	2.9	183	22.8	1014
9/17/25 14:00	4.1	180	22.0	1013
9/17/25 15:00	4.3	176	22.1	1013
9/17/25 16:00	4.3	176	21.9	1013
9/17/25 17:00	4.6	181	20.6	1012
9/17/25 18:00	4.0	171	18.1	1012

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING**Bangor International Airport (BGR) Meteorological Data (9/4/25 12:00 to 9/18/25 12:00)**

9/17/25 19:00	3.0	171	16.8	1012
9/17/25 20:00	3.9	175	16.0	1011
9/17/25 21:00	2.5	184	14.6	1011
9/17/25 22:00	1.9	195	13.9	1011
9/17/25 23:00	0.5	311	12.3	1011
9/18/25 0:00	2.0	192	13.0	1011
9/18/25 1:00	0.4	321	12.6	1010
9/18/25 2:00	1.0	249	12.7	1009
9/18/25 3:00	2.0	183	13.9	1009
9/18/25 4:00	1.8	217	13.5	1009
9/18/25 5:00	1.6	198	13.6	1008
9/18/25 6:00	2.1	203	13.9	1008
9/18/25 7:00	2.9	197	14.0	1008
9/18/25 8:00	3.1	189	15.7	1008
9/18/25 9:00	3.0	179	16.9	1008
9/18/25 10:00	2.6	190	18.5	1007
9/18/25 11:00	2.5	172	20.6	1007
9/18/25 12:00	3.1	189	22.4	1006

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING
Bangor International Airport (BGR) Meteorological Data (9/4/25 12:00 to 9/18/25 12:00)

BGR Wind Rose 9/4/25 12:00 - 9/18/25 12:00



SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

Terminal Fenceline Perimeter Length = 3,158 m

Monitor Location Method: EPA Method 325A Option 2

Terminal Fenceline Area = 107 Acres

Spacing Between Monitors: 263.2 m ($\pm$ 26.3 m)

Sampling Station	Target Compounds	Latitude	Longitude	Distances From Adjacent Sites (m)	Notes
Site 1	BTEX	44.452543°	-68.898865°	to Site 2 = 278.8 to Site 12 = 260.0	Site 1 is situated adjacent to the dock area and located 8.5 m inside fenceline for safe access.
Site 2	BTEX	44.452323°	-68.902000°	to Site 1 = 278.8 to Site 3 = 239.5	Site 2 is 2 m inside the fenceline to avoid brush and vegetation and allow safe access.
Site 3	BTEX	44.452339°	-68.905029°	to Site 2 = 239.5 to Site 4 = 279.1	Site 3 is located between two warehouses that abut the fenceline and is 2.5 m inside the fenceline to avoid brush and vegetation and allow safe access.
Site 4	BTEX	44.453422°	-68.906708°	to Site 3 = 279.1 to Site 5 = 285.8	Site 4 is located 17.3 m inside the fenceline to avoid swampy terrain, dense brush, and allow safe access.
Site 5	BTEX	44.455975°	-68.906269°	to Site 4 = 285.8 to Site 6 = 239.5	Site 5 is on the fenceline.
Site 6	BTEX	44.458018°	-68.906154°	to Site 5 = 239.5 to Site 7 = 262.7	Site 6 is on the fenceline.
Site 7	BTEX	44.460344°	-68.905925°	to Site 6 = 262.7 to Site 7 = 267.3	Site 7 is on the fenceline.
Site 8	BTEX	44.460944°	-68.903657°	to Site 7 = 267.3 to Site 9 = 264.9	Site 8 is on the fenceline.

SPRAGUE SEARSPORT MAINE TERMINAL FENCELINE MONITORING

Terminal Fenceline Perimeter Length = 3,158 m

Monitor Location Method: EPA Method 325A Option 2

Terminal Fenceline Area = 107 Acres

Spacing Between Monitors: 263.2 m ($\pm$ 26.3 m)

Sampling Station	Target Compounds	Latitude	Longitude	Distances From Adjacent Sites (m)	Notes
Site 9	BTEX	44.459783°	-68.901525°	to Site 8 = 264.9 to Site 10 = 262.1	Site 9 is on the fenceline.
Site 10	BTEX	44.457551°	-68.901666°	to Site 9 = 262.1 to Site 11 = 264.1	Site 10 is situated 14 m inside the fenceline to provide safe access away from railroad tracks located along this entire section of the fenceline.
Site 11	BTEX	44.455160°	-68.901370°	to Site 10 = 264.1 to Site 12 = 254.2	Site 11 is situated 21.6 m inside the fenceline because terrain and a long run of large pipes prevent closer access to the fenceline in this area.
Site 12	BTEX	44.454803°	-68.899471°	to Site 11 = 254.2 to Site 1 = 260.0	Site 12 is situated 5.8 m inside the fenceline to avoid potentially dangerous terrain adjacent to the asphalt detention pond.

Aerial View of the Sprague Searsport Terminal Fenceline Sampling Locations



APPENDIX A – LAB RESULTS

Sprague - Searsport

70 Trundy Road
Searsport, ME 04974

Sampling Event 25 Sprague - Searsport

Client Project# PROJ-027966

Samples Received: 7/14/2025

Analytical Report 2025GC301

EPA Method 325B Analysis

Report Issue Date: 7/24/2025

I certify that to the best of my knowledge all analytical data presented in this report have been checked for completeness, accuracy, errors and legibility in addition to having been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s). This report shall not be reproduced except in full without approval of the laboratory. This will provide assurance that parts of the report are not taken out of context.

Amendment(s):

Signature:



QA Review by Isabel Obando Marrero, Data Reviewer



Matt Cavanaugh
mcavanaugh@montrose-env.com / www.enthalpy.com
O: (919) 850-4392
Enthalpy Analytical
800 Capitola Drive Suite 1 Durham, NC 27713

Table of Contents

Case Narrative	3
Results	6
Summary of Results	7
Detailed Results	8
QC Data	11
Chromatograms	14
Initial Calibration	66
Sample Custody	69
Chain of Custody	70

Narrative Summary



Enthalpy Analytical Narrative Summary

Company	Montrose Air Quality Services, LLC - New Jersey
Job No.	2025GC301-1
Client ID.	PROJ-027966 Site: Sprague - Searsport

1. Custody

The samples were received at Enthalpy Analytical on July 14, 2025 at 25.2 °C, which is above the method recommended 23.0 °C,. The samples were received in good condition. Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC

Table 1 - Sample Inventory

Sample ID	Tube ID	Sample Type
SPRSEA-1-S-20250626	C57687	Sample
SPRSEA-2-S-20250626	C33009	Sample
SPRSEA-3-S-20250626	B19056	Sample
SPRSEA-4-S-20250626	C16112	Sample
SPRSEA-5-S-20250626	B49576	Sample
SPRSEA-6-S-20250626	B38593	Sample
SPRSEA-6-D-20250626	B18562	Duplicate
SPRSEA-6-B-20250626	C40533	Blank
SPRSEA-7-S-20250626	B50653	Sample
SPRSEA-8-S-20250626	C34132	Sample
SPRSEA-9-S-20250626	C61442	Sample
SPRSEA-10-S-20250626	B46360	Sample
SPRSEA-11-S-20250626	C43733	Sample
SPRSEA-12-S-20250626	C00561	Sample
SPRSEA-12-D-20250626	C57727	Duplicate
SPRSEA-12-B-20250626	C20610	Blank

2. Analysis

The samples were analyzed for Benzene, Toluene, Ethylbenzene, m-/p-Xylenes, and o-Xylene using EPA Method 325B – Volatile Organic Compounds from Fugitive and Area Sources by Thermal Desorption and GC/MS. A copy of the acquisition method M325B-TD35 is not included in this report but may be available upon request.

The sample tube media used for this sampling period was CarbopackX. All calibration standards and laboratory QC were prepared using the same media.

3. Calibration

All BFB tune criteria have been met for this analysis.

The initial calibration (P041125A_CC252679) met all 30% RSD criteria. The initial calibration verification met 30% difference criteria. The continuing calibration verifications met 30% difference criteria. The initial and continuing calibration raw data are not included in this report but are available upon request.

Enthalpy Analytical Narrative Summary

Company	Montrose Air Quality Services, LLC - New Jersey
Job No.	2025GC301-1
Client ID.	PROJ-027966 Site: Sprague - Searsport

5. QC Notes

All quality control criteria required by the method and/or the laboratory SOP have been met unless noted otherwise below.

The primary sample SPRSEA-6-S-20250626 (tube ID B38593) and its corresponding duplicate SPRSEA-6-D-20250626 (tube ID B18562) failed to meet the 30% difference criterion for m-/p-Xylenes as specified by the method. However, the concentrations of the analyte in both the sample and the duplicate were less than two times the reporting limit of the instrument's calibration curve. Therefore, the percent difference observed may not suggest the data set has been negatively affected. All samples in the data set have been flagged "Pc" for m-/p-Xylenes to denote this failure.

6. Reporting Notes

All tubes used for this sampling period met the method criteria for number of uses; no tube exceeded 50 field uses.

As specified in EPA Method 325B, the response factor of the daily continuing calibration standard was used to quantitate all field samples and blanks.

All samples were reported as amount in ng catch, and concentration in ug/m³ and ppbv.

The results presented in this report are representative of the samples as provided to the laboratory. These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, located at 800 Capitola Drive, Suite 1, Durham NC, 27713 is accredited by the Louisiana Department of Environmental Quality (LDEQ) for EPA Method 325B for all analytes included in this report under **Certificate Number 04010**.

Results

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC301-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Summary

Sample Code	Tube ID	Benzene		Toluene		Ethylbenzene		m-/p-Xylenes		o-Xylene	
		(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag
SPRSEA-1-S-20250626	C57687	0.436		0.721		0.272	ND	0.456	J,Pc	0.272	ND
SPRSEA-2-S-20250626	C33009	0.514		1.04		0.272	ND	0.563	Pc	0.272	ND
SPRSEA-3-S-20250626	B19056	0.551		1.15		0.300	J	1.24	Pc	0.582	
SPRSEA-4-S-20250626	C16112	0.515		0.893		0.272	ND	0.504	J,Pc	0.272	ND
SPRSEA-5-S-20250626	B49576	0.526		1.01		0.272	ND	0.466	J,Pc	0.272	ND
SPRSEA-6-S-20250626	B38593	0.456		0.829		0.272	ND	0.463	J,Pc	0.272	ND
SPRSEA-6-D-20250626	B18562	0.458		0.742		0.272	ND	0.326	J,Pc	0.272	ND
SPRSEA-6-B-20250626	C40533	0.186	ND	0.240	ND	0.272	ND	0.272	ND,Pc	0.272	ND
SPRSEA-7-S-20250626	B50653	0.431		0.671		0.272	ND	0.367	J,Pc	0.272	ND
SPRSEA-8-S-20250626	C34132	0.421		0.919		0.272	ND	0.488	J,Pc	0.272	ND
SPRSEA-9-S-20250626	C61442	0.682		1.75		0.296	J	1.02	Pc	0.391	J
SPRSEA-10-S-20250626	B46360	2.30		9.70		1.59		5.05	Pc	2.09	
SPRSEA-11-S-20250626	C43733	1.18		4.07		0.813		2.94	Pc	1.15	
SPRSEA-12-S-20250626	C00561	0.754		2.14		0.407	J	1.33	Pc	0.552	J
SPRSEA-12-D-20250626	C57727	0.675		2.05		0.402	J	1.41	Pc	0.547	J
SPRSEA-12-B-20250626	C20610	0.186	ND	0.240	ND	0.272	ND	0.272	ND,Pc	0.272	ND

J: Estimated Value - The analyte was detected between the Method Detection Limit and Reporting Limit

ND: The analyte was not present above the Method Detection Limit

Pc: Field duplicate(s) exceed 30%RPD. Concentrations of both samples in duplicate are near the reporting limit

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC301-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Benzene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp (°F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250626	C57687	0.436	0.137	5.85	68.4	0.665	20180	0.186	0.370	0.0584	0.116		P2504481.D	2025-07-16 17:16	0.868	8.402	46215	821922	90.3	8.349	-5.2%
SPRSEA-2-S-20250626	C33009	0.514	0.161	6.89	68.4	0.665	20180	0.186	0.370	0.0584	0.116		P2504482.D	2025-07-16 17:54	0.868	8.402	54939	828560	90.3	8.349	-4.4%
SPRSEA-3-S-20250626	B19056	0.551	0.173	7.39	68.4	0.665	20180	0.186	0.370	0.0584	0.116		P2504483.D	2025-07-16 18:31	0.868	8.402	58322	820506	90.3	8.349	-5.4%
SPRSEA-4-S-20250626	C16112	0.515	0.161	6.91	68.4	0.665	20180	0.186	0.370	0.0584	0.116		P2504484.D	2025-07-16 19:08	0.868	8.408	55140	829911	90.3	8.349	-4.3%
SPRSEA-5-S-20250626	B49576	0.526	0.165	7.05	68.4	0.665	20170	0.186	0.370	0.0584	0.116		P2504485.D	2025-07-16 19:46	0.868	8.414	55753	822537	90.3	8.349	-5.1%
SPRSEA-6-S-20250626	B38593	0.456	0.143	6.11	68.4	0.665	20170	0.186	0.370	0.0584	0.116		P2504486.D	2025-07-16 20:24	0.868	8.397	48274	820759	90.3	8.349	-5.3%
SPRSEA-6-D-20250626	B18562	0.458	0.143	6.14	68.4	0.665	20170	0.186	0.370	0.0584	0.116		P2504487.D	2025-07-16 21:01	0.868	8.397	47824	810156	90.3	8.349	-6.6%
SPRSEA-6-B-20250626	C40533	0.186	0.0584		68.4	0.665	20170	0.186	0.370	0.0584	0.116	ND	P2504480.D	2025-07-16 16:26	0.868	8.414	12089	855593	90.3	8.355	-1.3%
SPRSEA-7-S-20250626	B50653	0.431	0.135	5.78	68.4	0.665	20170	0.186	0.370	0.0584	0.116		P2504488.D	2025-07-16 21:38	0.868	8.402	45542	819034	90.3	8.343	-5.5%
SPRSEA-8-S-20250626	C34132	0.421	0.132	5.64	68.4	0.665	20170	0.186	0.370	0.0584	0.116		P2504489.D	2025-07-16 22:15	0.868	8.402	43579	803114	90.3	8.343	-7.4%
SPRSEA-9-S-20250626	C61442	0.682	0.213	9.14	68.4	0.665	20170	0.186	0.370	0.0584	0.116		P2504490.D	2025-07-16 22:53	0.868	8.402	72841	828777	90.3	8.349	-4.4%
SPRSEA-10-S-20250626	B46360	2.30	0.720	30.8	68.4	0.665	20170	0.186	0.370	0.0584	0.116		P2504492.D	2025-07-17 00:30	0.868	8.396	249067	840058	90.3	8.349	-3.1%
SPRSEA-11-S-20250626	C43733	1.18	0.370	15.8	68.4	0.665	20175	0.186	0.370	0.0584	0.116		P2504493.D	2025-07-17 01:07	0.868	8.396	126456	829840	90.3	8.343	-4.3%
SPRSEA-12-S-20250626	C00561	0.754	0.236	10.1	68.4	0.665	20175	0.186	0.370	0.0584	0.116		P2504494.D	2025-07-17 01:44	0.868	8.402	81712	840620	90.3	8.349	-3.0%
SPRSEA-12-D-20250626	C57727	0.675	0.212	9.06	68.4	0.665	20175	0.186	0.370	0.0584	0.116		P2504495.D	2025-07-17 02:22	0.868	8.408	71519	820996	90.3	8.349	-5.3%
SPRSEA-12-B-20250626	C20610	0.186	0.0584		68.4	0.665	20175	0.186	0.370	0.0584	0.116	ND	P2504496.D	2025-07-17 02:59	0.868	8.408	16409	851049	90.3	8.349	-1.8%

Toluene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp (°F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250626	C57687	0.721	0.191	7.51	68.4	0.516	20180	0.240	0.515	0.0638	0.137		P2504481.D	2025-07-16 17:16	0.874	11.002	60013	963154	105.3	10.913	-9.4%
SPRSEA-2-S-20250626	C33009	1.04	0.276	10.8	68.4	0.516	20180	0.240	0.515	0.0638	0.137		P2504482.D	2025-07-16 17:54	0.874	11.002	87483	975003	105.3	10.913	-8.3%
SPRSEA-3-S-20250626	B19056	1.15	0.306	12.0	68.4	0.516	20180	0.240	0.515	0.0638	0.137		P2504483.D	2025-07-16 18:31	0.874	11.002	96840	972496	105.3	10.913	-8.6%
SPRSEA-4-S-20250626	C16112	0.893	0.237	9.30	68.4	0.516	20180	0.240	0.515	0.0638	0.137		P2504484.D	2025-07-16 19:08	0.874	11.008	75301	975607	105.3	10.913	-8.3%
SPRSEA-5-S-20250626	B49576	1.01	0.267	10.5	68.4	0.516	20170	0.240	0.516	0.0638	0.137		P2504485.D	2025-07-16 19:46	0.874	11.002	84414	971059	105.3	10.913	-8.7%
SPRSEA-6-S-20250626	B38593	0.829	0.220	8.62	68.4	0.516	20170	0.240	0.516	0.0638	0.137		P2504486.D	2025-07-16 20:24	0.874	11.008	70410	983419	105.3	10.913	-7.5%
SPRSEA-6-D-20250626	B18562	0.742	0.197	7.72	68.4	0.516	20170	0.240	0.516	0.0638	0.137		P2504487.D	2025-07-16 21:01	0.874	11.008	63050	984060	105.3	10.913	-7.5%
SPRSEA-6-B-20250626	C40533	0.240	0.0638		68.4	0.516	20170	0.240	0.516	0.0638	0.137	ND	P2504480.D	2025-07-16 16:26	0.874	11.002	10065	1053898	105.3	10.913	-0.9%
SPRSEA-7-S-20250626	B50653	0.671	0.178	6.98	68.4	0.516	20170	0.240	0.516	0.0638	0.137		P2504488.D	2025-07-16 21:38	0.874	11.002	57150	986454	105.3	10.913	-7.2%
SPRSEA-8-S-20250626	C34132	0.919	0.244	9.56	68.4	0.516	20170	0.240	0.516	0.0638	0.137		P2504489.D	2025-07-16 22:15	0.874	11.002	78385	987527	105.3	10.907	-7.1%
SPRSEA-9-S-20250626	C61442	1.75	0.465	18.2	68.4	0.516	20170	0.240	0.516	0.0638	0.137		P2504490.D	2025-07-16 22:53	0.874	11.002	152327	1006254	105.3	10.907	-5.4%
SPRSEA-10-S-20250626	B46360	9.70	2.58	101	68.4	0.516	20170	0.240	0.516	0.0638	0.137		P2504492.D	2025-07-17 00:30	0.874	11.002	844609	1007928	105.3	10.907	-5.2%
SPRSEA-11-S-20250626	C43733	4.07	1.08	42.4	68.4	0.516	20175	0.240	0.516	0.0638	0.137		P2504493.D	2025-07-17 01:07	0.874	11.002	357099	1014980	105.3	10.907	-4.6%
SPRSEA-12-S-20250626	C00561	2.14	0.568	22.3	68.4	0.516	20175	0.240	0.516	0.0638	0.137		P2504494.D	2025-07-17 01:44	0.874	11.002	181145	979944	105.3	10.907	-7.9%
SPRSEA-12-D-20250626	C57727	2.05	0.543	21.3	68.4	0.516	20175	0.240	0.516	0.0638	0.137		P2504495.D	2025-07-17 02:22	0.874	11.008	174027	984681	105.3	10.907	-7.4%
SPRSEA-12-B-20250626	C20610	0.240	0.0638		68.4	0.516	20175	0.240	0.516	0.0638	0.137	ND	P2504496.D	2025-07-17 02:59	0.874	11.002	8864	971733	105.3	10.907	-8.6%

Ethylbenzene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp (°F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250626	C57687	0.272	0.0626		68.4	0.456	20180	0.272	0.560	0.0626	0.129	ND	P2504481.D	2025-07-16 17:16	1.016	13.133	13159	963154	105.3	10.913	-9.4%

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC301-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Ethylbenzene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-2-S-20250626	C33009	0.272	0.0626		68.4	0.456	20180	0.272	0.560	0.0626	0.129	ND	P2504482.D	2025-07-16 17:54	1.016	13.133	21927	975003	105.3	10.913	-8.3%
SPRSEA-3-S-20250626	B19056	0.300	0.0691	2.76	68.4	0.456	20180	0.272	0.560	0.0626	0.129	J	P2504483.D	2025-07-16 18:31	1.016	13.133	25910	972496	105.3	10.913	-8.6%
SPRSEA-4-S-20250626	C16112	0.272	0.0626		68.4	0.456	20180	0.272	0.560	0.0626	0.129	ND	P2504484.D	2025-07-16 19:08	1.016	13.127	14745	975607	105.3	10.913	-8.3%
SPRSEA-5-S-20250626	B49576	0.272	0.0626		68.4	0.456	20170	0.272	0.560	0.0626	0.129	ND	P2504485.D	2025-07-16 19:46	1.016	13.133	16675	971059	105.3	10.913	-8.7%
SPRSEA-6-S-20250626	B38593	0.272	0.0626		68.4	0.456	20170	0.272	0.560	0.0626	0.129	ND	P2504486.D	2025-07-16 20:24	1.016	13.133	13855	983419	105.3	10.913	-7.5%
SPRSEA-6-D-20250626	B18562	0.272	0.0626		68.4	0.456	20170	0.272	0.560	0.0626	0.129	ND	P2504487.D	2025-07-16 21:01	1.016	13.127	10615	984060	105.3	10.913	-7.5%
SPRSEA-6-B-20250626	C40533	0.272	0.0626		68.4	0.456	20170	0.272	0.560	0.0626	0.129	ND	P2504480.D	2025-07-16 16:26	1.016	13.145	4530	1053898	105.3	10.913	-0.9%
SPRSEA-7-S-20250626	B50653	0.272	0.0626		68.4	0.456	20170	0.272	0.560	0.0626	0.129	ND	P2504488.D	2025-07-16 21:38	1.016	13.127	9023	986454	105.3	10.913	-7.2%
SPRSEA-8-S-20250626	C34132	0.272	0.0626		68.4	0.456	20170	0.272	0.560	0.0626	0.129	ND	P2504489.D	2025-07-16 22:15	1.016	13.133	16105	987527	105.3	10.907	-7.1%
SPRSEA-9-S-20250626	C61442	0.296	0.0682	2.72	68.4	0.456	20170	0.272	0.560	0.0626	0.129	J	P2504490.D	2025-07-16 22:53	1.016	13.133	26448	1006254	105.3	10.907	-5.4%
SPRSEA-10-S-20250626	B46360	1.59	0.366	14.6	68.4	0.456	20170	0.272	0.560	0.0626	0.129		P2504492.D	2025-07-17 00:30	1.016	13.127	142332	1007928	105.3	10.907	-5.2%
SPRSEA-11-S-20250626	C43733	0.813	0.187	7.48	68.4	0.456	20175	0.272	0.560	0.0626	0.129		P2504493.D	2025-07-17 01:07	1.016	13.133	73312	1014980	105.3	10.907	-4.6%
SPRSEA-12-S-20250626	C00561	0.407	0.0938	3.75	68.4	0.456	20175	0.272	0.560	0.0626	0.129	J	P2504494.D	2025-07-17 01:44	1.016	13.127	35461	979944	105.3	10.907	-7.9%
SPRSEA-12-D-20250626	C57727	0.402	0.0926	3.70	68.4	0.456	20175	0.272	0.560	0.0626	0.129	J	P2504495.D	2025-07-17 02:22	1.016	13.133	35141	984681	105.3	10.907	-7.4%
SPRSEA-12-B-20250626	C20610	0.272	0.0626		68.4	0.456	20175	0.272	0.560	0.0626	0.129	ND	P2504496.D	2025-07-17 02:59	1.016	13.139	1130	971733	105.3	10.907	-8.6%

m-/p-Xylenes

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250626	C57687	0.456	0.105	4.20	68.4	0.456	20180	0.272	0.529	0.0626	0.122	J,Pc	P2504481.D	2025-07-16 17:16	0.730	13.305	28039	963154	105.3	10.913	-9.4%
SPRSEA-2-S-20250626	C33009	0.563	0.130	5.19	68.4	0.456	20180	0.272	0.529	0.0626	0.122	Pc	P2504482.D	2025-07-16 17:54	0.730	13.305	35069	975003	105.3	10.913	-8.3%
SPRSEA-3-S-20250626	B19056	1.24	0.286	11.4	68.4	0.456	20180	0.272	0.529	0.0626	0.122	Pc	P2504483.D	2025-07-16 18:31	0.730	13.311	77034	972496	105.3	10.913	-8.6%
SPRSEA-4-S-20250626	C16112	0.504	0.116	4.64	68.4	0.456	20180	0.272	0.529	0.0626	0.122	J,Pc	P2504484.D	2025-07-16 19:08	0.730	13.305	31380	975607	105.3	10.913	-8.3%
SPRSEA-5-S-20250626	B49576	0.466	0.107	4.29	68.4	0.456	20170	0.272	0.529	0.0626	0.122	J,Pc	P2504485.D	2025-07-16 19:46	0.730	13.317	28866	971059	105.3	10.913	-8.7%
SPRSEA-6-S-20250626	B38593	0.463	0.107	4.26	68.4	0.456	20170	0.272	0.529	0.0626	0.122	J,Pc	P2504486.D	2025-07-16 20:24	0.730	13.311	29061	983419	105.3	10.913	-7.5%
SPRSEA-6-D-20250626	B18562	0.326	0.0752	3.00	68.4	0.456	20170	0.272	0.529	0.0626	0.122	J,Pc	P2504487.D	2025-07-16 21:01	0.730	13.311	20478	984060	105.3	10.913	-7.5%
SPRSEA-6-B-20250626	C40533	0.272	0.0626		68.4	0.456	20170	0.272	0.529	0.0626	0.122	ND,Pc	P2504480.D	2025-07-16 16:26	0.730	13.323	2776	1053898	105.3	10.913	-0.9%
SPRSEA-7-S-20250626	B50653	0.367	0.0845	3.37	68.4	0.456	20170	0.272	0.529	0.0626	0.122	J,Pc	P2504488.D	2025-07-16 21:38	0.730	13.305	23073	986454	105.3	10.913	-7.2%
SPRSEA-8-S-20250626	C34132	0.488	0.112	4.49	68.4	0.456	20170	0.272	0.529	0.0626	0.122	J,Pc	P2504489.D	2025-07-16 22:15	0.730	13.299	30740	987527	105.3	10.907	-7.1%
SPRSEA-9-S-20250626	C61442	1.02	0.234	9.35	68.4	0.456	20170	0.272	0.529	0.0626	0.122	Pc	P2504490.D	2025-07-16 22:53	0.730	13.305	65236	1006254	105.3	10.907	-5.4%
SPRSEA-10-S-20250626	B46360	5.05	1.16	46.5	68.4	0.456	20170	0.272	0.529	0.0626	0.122	Pc	P2504492.D	2025-07-17 00:30	0.730	13.305	324898	1007928	105.3	10.907	-5.2%
SPRSEA-11-S-20250626	C43733	2.94	0.679	27.1	68.4	0.456	20175	0.272	0.529	0.0626	0.122	Pc	P2504493.D	2025-07-17 01:07	0.730	13.305	190772	1014980	105.3	10.907	-4.6%
SPRSEA-12-S-20250626	C00561	1.33	0.307	12.3	68.4	0.456	20175	0.272	0.529	0.0626	0.122	Pc	P2504494.D	2025-07-17 01:44	0.730	13.311	83326	979944	105.3	10.907	-7.9%
SPRSEA-12-D-20250626	C57727	1.41	0.325	13.0	68.4	0.456	20175	0.272	0.529	0.0626	0.122	Pc	P2504495.D	2025-07-17 02:22	0.730	13.311	88687	984681	105.3	10.907	-7.4%
SPRSEA-12-B-20250626	C20610	0.272	0.0626		68.4	0.456	20175	0.272	0.529	0.0626	0.122	ND,Pc	P2504496.D	2025-07-17 02:59	0.730	13.329	363	971733	105.3	10.907	-8.6%

o-Xylene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250626	C57687	0.272	0.0626		68.4	0.456	20180	0.272	0.552	0.0626	0.127	ND	P2504481.D	2025-07-16 17:16	0.760	13.798	9531	963154	105.3	10.913	-9.4%
SPRSEA-2-S-20250626	C33009	0.272	0.0626		68.4	0.456	20180	0.272	0.552	0.0626	0.127	ND	P2504482.D	2025-07-16 17:54	0.760	13.798	15736	975003	105.3	10.913	-8.3%

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC301-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

o-Xylene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-3-S-20250626	B19056	0.582	0.134	5.36	68.4	0.456	20180	0.272	0.552	0.0626	0.127		P2504483.D	2025-07-16 18:31	0.760	13.804	37638	972496	105.3	10.913	-8.6%
SPRSEA-4-S-20250626	C16112	0.272	0.0626		68.4	0.456	20180	0.272	0.552	0.0626	0.127	ND	P2504484.D	2025-07-16 19:08	0.760	13.798	14049	975607	105.3	10.913	-8.3%
SPRSEA-5-S-20250626	B49576	0.272	0.0626		68.4	0.456	20170	0.272	0.552	0.0626	0.127	ND	P2504485.D	2025-07-16 19:46	0.760	13.797	13993	971059	105.3	10.913	-8.7%
SPRSEA-6-S-20250626	B38593	0.272	0.0626		68.4	0.456	20170	0.272	0.552	0.0626	0.127	ND	P2504486.D	2025-07-16 20:24	0.760	13.792	10310	983419	105.3	10.913	-7.5%
SPRSEA-6-D-20250626	B18562	0.272	0.0626		68.4	0.456	20170	0.272	0.552	0.0626	0.127	ND	P2504487.D	2025-07-16 21:01	0.760	13.804	7722	984060	105.3	10.913	-7.5%
SPRSEA-6-B-20250626	C40533	0.272	0.0626		68.4	0.456	20170	0.272	0.552	0.0626	0.127	ND	P2504480.D	2025-07-16 16:26	0.760	13.810	1912	1053898	105.3	10.913	-0.9%
SPRSEA-7-S-20250626	B50653	0.272	0.0626		68.4	0.456	20170	0.272	0.552	0.0626	0.127	ND	P2504488.D	2025-07-16 21:38	0.760	13.792	8480	986454	105.3	10.913	-7.2%
SPRSEA-8-S-20250626	C34132	0.272	0.0626		68.4	0.456	20170	0.272	0.552	0.0626	0.127	ND	P2504489.D	2025-07-16 22:15	0.760	13.804	15517	987527	105.3	10.907	-7.1%
SPRSEA-9-S-20250626	C61442	0.391	0.0901	3.60	68.4	0.456	20170	0.272	0.552	0.0626	0.127	J	P2504490.D	2025-07-16 22:53	0.760	13.792	26137	1006254	105.3	10.907	-5.4%
SPRSEA-10-S-20250626	B46360	2.09	0.481	19.2	68.4	0.456	20170	0.272	0.552	0.0626	0.127		P2504492.D	2025-07-17 00:30	0.760	13.798	139652	1007928	105.3	10.907	-5.2%
SPRSEA-11-S-20250626	C43733	1.15	0.266	10.6	68.4	0.456	20175	0.272	0.552	0.0626	0.127		P2504493.D	2025-07-17 01:07	0.760	13.792	77868	1014980	105.3	10.907	-4.6%
SPRSEA-12-S-20250626	C00561	0.552	0.127	5.08	68.4	0.456	20175	0.272	0.552	0.0626	0.127	J	P2504494.D	2025-07-17 01:44	0.760	13.792	35930	979944	105.3	10.907	-7.9%
SPRSEA-12-D-20250626	C57727	0.547	0.126	5.04	68.4	0.456	20175	0.272	0.552	0.0626	0.127	J	P2504495.D	2025-07-17 02:22	0.760	13.804	35789	984681	105.3	10.907	-7.4%
SPRSEA-12-B-20250626	C20610	0.272	0.0626		68.4	0.456	20175	0.272	0.552	0.0626	0.127	ND	P2504496.D	2025-07-17 02:59	0.760	14.243	0	971733	105.3	10.907	-8.6%

J: Estimated Value - The analyte was detected between the Method Detection Limit and Reporting Limit

ND: The analyte was not present above the Method Detection Limit

Pc: Field duplicate(s) exceed 30%RPD. Concentrations of both samples in duplicate are near the reporting limit

QC Data



Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey
Job No.: 2025GC301-1 EPA Method 325B Analysis
Client No.: PROJ-027966 Site: Sprague - Searsport

QC Samples

Field Sample Type	Sample Code	Benzene		Toluene		Ethylbenzene		m-/p-Xylenes		o-Xylene	
Blanks (ug/m³)	SPRSEA-6-B-20250626	ND	Pass	ND	Pass	ND	Pass	ND	Pass	ND	Pass
	SPRSEA-12-B-20250626	ND	Pass	ND	Pass	ND	Pass	ND	Pass	ND	Pass
Duplicates (difference)	SPRSEA-6-D-20250626	0.36%	Pass	11%	Pass	ND	Pass	35%	Fail	ND	Pass
	SPRSEA-12-D-20250626	11%	Pass	4.5%	Pass	1.4%	Pass	5.8%	Pass	0.88%	Pass

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC301-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Benzene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5 REC	P2504478.D	C40664	Cal	0.868	0.896	0.868	-3.1%	13%		Pass	
2025GC301 Method Blank-1	P2504479.D	C43688	Blank		0.896	0.868			0.11%	Pass	ND
M325B CCV 5 REC	P2504491.D	C65345	Check	0.856	0.896	0.868	-4.4%		-0.90%	Pass	
M325B CCV 5	P2504497.D	C57478	Check	0.862	0.896	0.868	-3.8%		-2.1%	Pass	

Toluene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5 REC	P2504478.D	C40664	Cal	0.874	0.887	0.874	-1.4%	11%		Pass	
2025GC301 Method Blank-1	P2504479.D	C43688	Blank		0.887	0.874			-0.13%	Pass	ND
M325B CCV 5 REC	P2504491.D	C65345	Check	0.891	0.887	0.874	0.44%		-3.2%	Pass	
M325B CCV 5	P2504497.D	C57478	Check	0.864	0.887	0.874	-2.5%		-4.8%	Pass	

Ethylbenzene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5 REC	P2504478.D	C40664	Cal	1.016	1.034	1.016	-1.7%	11%		Pass	
2025GC301 Method Blank-1	P2504479.D	C43688	Blank		1.034	1.016			-0.13%	Pass	ND
M325B CCV 5 REC	P2504491.D	C65345	Check	1.034	1.034	1.016	0.027%		-3.2%	Pass	
M325B CCV 5	P2504497.D	C57478	Check	1.062	1.034	1.016	2.7%		-4.8%	Pass	

m-/p-Xylenes Calibration and Blanks


Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5 REC	P2504478.D	C40664	Cal	0.730	0.792	0.730	-7.8%	11%		Pass	
2025GC301 Method Blank-1	P2504479.D	C43688	Blank		0.792	0.730			-0.13%	Pass	ND
M325B CCV 5 REC	P2504491.D	C65345	Check	0.848	0.792	0.730	7.1%		-3.2%	Pass	
M325B CCV 5	P2504497.D	C57478	Check	0.846	0.792	0.730	6.8%		-4.8%	Pass	

o-Xylene Calibration and Blanks





Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5 REC	P2504478.D	C40664	Cal	0.760	0.833	0.760	-8.8%	11%		Pass	
2025GC301 Method Blank-1	P2504479.D	C43688	Blank		0.833	0.760			-0.13%	Pass	ND
M325B CCV 5 REC	P2504491.D	C65345	Check	0.827	0.833	0.760	-0.75%		-3.2%	Pass	
M325B CCV 5	P2504497.D	C57478	Check	0.846	0.833	0.760	1.5%		-4.8%	Pass	

Chromatograms

Name 2025GC301 Method Blank-1
Comment C43688; Recollect
Data File P2504479.D
Acq. Date-Time 7/16/2025 3:51:44 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxen X
Analyze Quant Version 12.1
Report Quant Version 12.1


Sample Chromatogram

Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.349	8.367	867,925	
Benzene	benzene-d6 (IS)	8.402	8.426	13,773	
Toluene-d8 (IS)		10.913	10.931	1,062,018	
Toluene	Toluene-d8 (IS)	11.014	11.020	17,788	
Ethylbenzene	Toluene-d8 (IS)	13.133	13.151	11,138	
m-/p-Xylenes	Toluene-d8 (IS)	13.305	13.341	10,266	
o-Xylene	Toluene-d8 (IS)	13.804	13.809	5,702	


benzene-d6 (IS)**Benzene**

Toluene-d8 (IS)


+ EIC (98.1) Scan P2504479.D




+ Scan (10.866-11.062 min, 33 scans) P2504479.D

**Toluene**


+ EIC (91.1) Scan P2504479.D




+ Scan (10.962-11.056 min, 16 scans) P2504479.D

**Ethylbenzene**


+ EIC (91.1) Scan P2504479.D




+ Scan (13.111-13.186 min, 13 scans) P2504479.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2504479.D




+ Scan (13.277-13.382 min, 17 scans) P2504479.D




o-Xylene

+ EIC (91.1) Scan P2504479.D




+ Scan (13.773-13.832 min, 10 scans) P2504479.D





Name SPRSEA-6-B-20250626
Comment C40533; Recollect
Data File P2504480.D
Acq. Date-Time 7/16/2025 4:26:52 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.355	8.367	855,593	
Benzene	benzene-d6 (IS)	8.414	8.426	12,089	
Toluene-d8 (IS)		10.913	10.931	1,053,898	
Toluene	Toluene-d8 (IS)	11.002	11.020	10,065	
Ethylbenzene	Toluene-d8 (IS)	13.145	13.151	4,530	
m-/p-Xylenes	Toluene-d8 (IS)	13.323	13.341	2,776	
o-Xylene	Toluene-d8 (IS)	13.810	13.809	1,912	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2504480.D




+ Scan (10.872-11.085 min, 37 scans) P2504480.D

**Toluene**


+ EIC (91.1) Scan P2504480.D




+ Scan (10.975-11.079 min, 18 scans) P2504480.D

**Ethylbenzene**


+ EIC (91.1) Scan P2504480.D




+ Scan (13.115-13.174 min, 10 scans) P2504480.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2504480.D




+ Scan (13.293-13.351 min, 10 scans) P2504480.D




o-Xylene

+ EIC (91.1) Scan P2504480.D




+ Scan (13.786-13.833 min, 9 scans) P2504480.D





Name SPRSEA-1-S-20250626
Comment C57687
Data File P2504481.D
Acq. Date-Time 7/16/2025 5:16:49 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxen X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.349	8.367	821,922	
Benzene	benzene-d6 (IS)	8.402	8.426	46,215	
Toluene-d8 (IS)		10.913	10.931	963,154	
Toluene	Toluene-d8 (IS)	11.002	11.020	60,013	
Ethylbenzene	Toluene-d8 (IS)	13.133	13.151	13,159	
m-/p-Xylenes	Toluene-d8 (IS)	13.305	13.341	28,039	
o-Xylene	Toluene-d8 (IS)	13.798	13.809	9,531	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2504481.D




+ Scan (10.868-11.026 min, 27 scans) P2504481.D

**Toluene**


+ EIC (91.1) Scan P2504481.D




+ Scan (10.956-11.091 min, 22 scans) P2504481.D

**Ethylbenzene**


+ EIC (91.1) Scan P2504481.D




+ Scan (13.107-13.174 min, 11 scans) P2504481.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2504481.D




+ Scan (13.281-13.412 min, 23 scans) P2504481.D




o-Xylene

+ EIC (91.1) Scan P2504481.D




+ Scan (13.769-13.838 min, 11 scans) P2504481.D





Name SPRSEA-2-S-20250626
Comment C33009
Data File P2504482.D
Acq. Date-Time 7/16/2025 5:54:09 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.349	8.367	828,560	
Benzene	benzene-d6 (IS)	8.402	8.426	54,939	
Toluene-d8 (IS)		10.913	10.931	975,003	
Toluene	Toluene-d8 (IS)	11.002	11.020	87,483	
Ethylbenzene	Toluene-d8 (IS)	13.133	13.151	21,927	
m-/p-Xylenes	Toluene-d8 (IS)	13.305	13.341	35,069	
o-Xylene	Toluene-d8 (IS)	13.798	13.809	15,736	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2504482.D




+ Scan (10.860-11.014 min, 26 scans) P2504482.D

**Toluene**


+ EIC (91.1) Scan P2504482.D




+ Scan (10.962-11.120 min, 26 scans) P2504482.D

**Ethylbenzene**


+ EIC (91.1) Scan P2504482.D




+ Scan (13.104-13.200 min, 16 scans) P2504482.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2504482.D




+ Scan (13.277-13.436 min, 27 scans) P2504482.D




o-Xylene

+ EIC (91.1) Scan P2504482.D




+ Scan (13.769-13.845 min, 13 scans) P2504482.D





Name SPRSEA-3-S-20250626
Comment B19056
Data File P2504483.D
Acq. Date-Time 7/16/2025 6:31:29 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.349	8.367	820,506	
Benzene	benzene-d6 (IS)	8.402	8.426	58,322	
Toluene-d8 (IS)		10.913	10.931	972,496	
Toluene	Toluene-d8 (IS)	11.002	11.020	96,840	
Ethylbenzene	Toluene-d8 (IS)	13.133	13.151	25,910	
m-/p-Xylenes	Toluene-d8 (IS)	13.311	13.341	77,034	
o-Xylene	Toluene-d8 (IS)	13.804	13.809	37,638	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2504483.D




+ Scan (10.863-11.050 min, 32 scans) P2504483.D

**Toluene**


+ EIC (91.1) Scan P2504483.D




+ Scan (10.966-11.121 min, 27 scans) P2504483.D

**Ethylbenzene**


+ EIC (91.1) Scan P2504483.D




+ Scan (13.110-13.204 min, 16 scans) P2504483.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2504483.D




+ Scan (13.271-13.430 min, 27 scans) P2504483.D




o-Xylene

+ EIC (91.1) Scan P2504483.D




+ Scan (13.768-13.878 min, 18 scans) P2504483.D





Name SPRSEA-4-S-20250626
Comment C16112
Data File P2504484.D
Acq. Date-Time 7/16/2025 7:08:48 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.349	8.367	829,911	
Benzene	benzene-d6 (IS)	8.408	8.426	55,140	
Toluene-d8 (IS)		10.913	10.931	975,607	
Toluene	Toluene-d8 (IS)	11.008	11.020	75,301	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	14,745	
m-/p-Xylenes	Toluene-d8 (IS)	13.305	13.341	31,380	
o-Xylene	Toluene-d8 (IS)	13.798	13.809	14,049	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2504484.D




+ Scan (10.866-11.014 min, 26 scans) P2504484.D

**Toluene**


+ EIC (91.1) Scan P2504484.D




+ Scan (10.973-11.111 min, 23 scans) P2504484.D

**Ethylbenzene**


+ EIC (91.1) Scan P2504484.D




+ Scan (13.100-13.184 min, 14 scans) P2504484.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2504484.D




+ Scan (13.274-13.400 min, 22 scans) P2504484.D




o-Xylene

+ EIC (91.1) Scan P2504484.D




+ Scan (13.768-13.859 min, 16 scans) P2504484.D





Name SPRSEA-5-S-20250626
Comment B49576
Data File P2504485.D
Acq. Date-Time 7/16/2025 7:46:41 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.349	8.367	822,537	
Benzene	benzene-d6 (IS)	8.414	8.426	55,753	
Toluene-d8 (IS)		10.913	10.931	971,059	
Toluene	Toluene-d8 (IS)	11.002	11.020	84,414	
Ethylbenzene	Toluene-d8 (IS)	13.133	13.151	16,675	
m-/p-Xylenes	Toluene-d8 (IS)	13.317	13.341	28,866	
o-Xylene	Toluene-d8 (IS)	13.797	13.809	13,993	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2504485.D




+ Scan (10.872-11.061 min, 32 scans) P2504485.D

**Toluene**


+ EIC (91.1) Scan P2504485.D




+ Scan (10.963-11.114 min, 25 scans) P2504485.D

**Ethylbenzene**


+ EIC (91.1) Scan P2504485.D




+ Scan (13.112-13.174 min, 11 scans) P2504485.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2504485.D




+ Scan (13.275-13.406 min, 23 scans) P2504485.D




o-Xylene

+ EIC (91.1) Scan P2504485.D




+ Scan (13.765-13.863 min, 17 scans) P2504485.D





Name SPRSEA-6-S-20250626
Comment B38593
Data File P2504486.D
Acq. Date-Time 7/16/2025 8:24:00 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.349	8.367	820,759	
Benzene	benzene-d6 (IS)	8.397	8.426	48,274	
Toluene-d8 (IS)		10.913	10.931	983,419	
Toluene	Toluene-d8 (IS)	11.008	11.020	70,410	
Ethylbenzene	Toluene-d8 (IS)	13.133	13.151	13,855	
m-/p-Xylenes	Toluene-d8 (IS)	13.311	13.341	29,061	
o-Xylene	Toluene-d8 (IS)	13.792	13.809	10,310	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2504486.D




+ Scan (10.860-11.056 min, 34 scans) P2504486.D

**Toluene**


+ EIC (91.1) Scan P2504486.D




+ Scan (10.973-11.067 min, 16 scans) P2504486.D

**Ethylbenzene**


+ EIC (91.1) Scan P2504486.D




+ Scan (13.106-13.163 min, 10 scans) P2504486.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2504486.D




+ Scan (13.264-13.399 min, 22 scans) P2504486.D




o-Xylene

+ EIC (91.1) Scan P2504486.D




+ Scan (13.765-13.827 min, 10 scans) P2504486.D





Name SPRSEA-6-D-20250626
Comment B18562
Data File P2504487.D
Acq. Date-Time 7/16/2025 9:01:18 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.349	8.367	810,156	
Benzene	benzene-d6 (IS)	8.397	8.426	47,824	
Toluene-d8 (IS)		10.913	10.931	984,060	
Toluene	Toluene-d8 (IS)	11.008	11.020	63,050	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	10,615	
m-/p-Xylenes	Toluene-d8 (IS)	13.311	13.341	20,478	
o-Xylene	Toluene-d8 (IS)	13.804	13.809	7,722	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2504487.D




+ Scan (10.869-11.050 min, 31 scans) P2504487.D

**Toluene**


+ EIC (91.1) Scan P2504487.D




+ Scan (10.961-11.109 min, 25 scans) P2504487.D

**Ethylbenzene**


+ EIC (91.1) Scan P2504487.D




+ Scan (13.104-13.196 min, 15 scans) P2504487.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2504487.D




+ Scan (13.273-13.382 min, 19 scans) P2504487.D




o-Xylene

+ EIC (91.1) Scan P2504487.D




+ Scan (13.766-13.827 min, 11 scans) P2504487.D





Name SPRSEA-7-S-20250626
Comment B50653
Data File P2504488.D
Acq. Date-Time 7/16/2025 9:38:36 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.343	8.367	819,034	
Benzene	benzene-d6 (IS)	8.402	8.426	45,542	
Toluene-d8 (IS)		10.913	10.931	986,454	
Toluene	Toluene-d8 (IS)	11.002	11.020	57,150	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	9,023	
m-/p-Xylenes	Toluene-d8 (IS)	13.305	13.341	23,073	
o-Xylene	Toluene-d8 (IS)	13.792	13.809	8,480	m

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2504488.D




+ Scan (10.867-11.020 min, 26 scans) P2504488.D

**Toluene**


+ EIC (91.1) Scan P2504488.D




+ Scan (10.971-11.079 min, 18 scans) P2504488.D

**Ethylbenzene**


+ EIC (91.1) Scan P2504488.D




+ Scan (13.105-13.163 min, 10 scans) P2504488.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2504488.D




+ Scan (13.276-13.400 min, 21 scans) P2504488.D




o-Xylene

+ EIC (91.1) Scan P2504488.D




+ Scan (13.770-13.833 min, 11 scans) P2504488.D





Name SPRSEA-8-S-20250626
Comment C34132
Data File P2504489.D
Acq. Date-Time 7/16/2025 10:15:55 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.343	8.367	803,114	
Benzene	benzene-d6 (IS)	8.402	8.426	43,579	
Toluene-d8 (IS)		10.907	10.931	987,527	
Toluene	Toluene-d8 (IS)	11.002	11.020	78,385	
Ethylbenzene	Toluene-d8 (IS)	13.133	13.151	16,105	
m-/p-Xylenes	Toluene-d8 (IS)	13.299	13.341	30,740	
o-Xylene	Toluene-d8 (IS)	13.804	13.809	15,517	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2504489.D




+ Scan (10.866-11.044 min, 30 scans) P2504489.D

**Toluene**


+ EIC (91.1) Scan P2504489.D




+ Scan (10.961-11.115 min, 26 scans) P2504489.D

**Ethylbenzene**


+ EIC (91.1) Scan P2504489.D




+ Scan (13.103-13.174 min, 13 scans) P2504489.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2504489.D




+ Scan (13.271-13.392 min, 20 scans) P2504489.D




o-Xylene

+ EIC (91.1) Scan P2504489.D




+ Scan (13.768-13.868 min, 16 scans) P2504489.D





Name SPRSEA-9-S-20250626
Comment C61442
Data File P2504490.D
Acq. Date-Time 7/16/2025 10:53:12 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.349	8.367	828,777	
Benzene	benzene-d6 (IS)	8.402	8.426	72,841	
Toluene-d8 (IS)		10.907	10.931	1,006,254	
Toluene	Toluene-d8 (IS)	11.002	11.020	152,327	
Ethylbenzene	Toluene-d8 (IS)	13.133	13.151	26,448	
m-/p-Xylenes	Toluene-d8 (IS)	13.305	13.341	65,236	
o-Xylene	Toluene-d8 (IS)	13.792	13.809	26,137	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2504490.D




+ Scan (10.860-11.050 min, 33 scans) P2504490.D

**Toluene**


+ EIC (91.1) Scan P2504490.D




+ Scan (10.957-11.139 min, 31 scans) P2504490.D

**Ethylbenzene**


+ EIC (91.1) Scan P2504490.D




+ Scan (13.087-13.186 min, 16 scans) P2504490.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2504490.D




+ Scan (13.268-13.376 min, 19 scans) P2504490.D




o-Xylene

+ EIC (91.1) Scan P2504490.D




+ Scan (13.763-13.833 min, 12 scans) P2504490.D



Name SPRSEA-10-S-20250626
Comment B46360
Data File P2504492.D
Acq. Date-Time 7/17/2025 12:30:22 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1


Sample Chromatogram




Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.349	8.367	840,058	
Benzene	benzene-d6 (IS)	8.396	8.426	249,067	
Toluene-d8 (IS)		10.907	10.931	1,007,928	
Toluene	Toluene-d8 (IS)	11.002	11.020	844,609	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	142,332	
m-/p-Xylenes	Toluene-d8 (IS)	13.305	13.341	324,898	
o-Xylene	Toluene-d8 (IS)	13.798	13.809	139,652	

benzene-d6 (IS)

+ EIC (84.1) Scan P2504492.D




+ Scan (8.349-8.349 min, 1 scans) P2504492.D




Benzene

+ EIC (78.1) Scan P2504492.D




+ Scan (8.396-8.396 min, 1 scans) P2504492.D




Toluene-d8 (IS)


+ EIC (98.1) Scan P2504492.D




+ Scan (10.862-11.050 min, 32 scans) P2504492.D

**Toluene**


+ EIC (91.1) Scan P2504492.D




+ Scan (10.956-11.139 min, 31 scans) P2504492.D

**Ethylbenzene**


+ EIC (91.1) Scan P2504492.D




+ Scan (13.091-13.237 min, 25 scans) P2504492.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2504492.D




+ Scan (13.264-13.477 min, 36 scans) P2504492.D




o-Xylene

+ EIC (91.1) Scan P2504492.D




+ Scan (13.761-13.899 min, 24 scans) P2504492.D





Name SPRSEA-11-S-20250626
Comment C43733
Data File P2504493.D
Acq. Date-Time 7/17/2025 1:07:39 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.343	8.367	829,840	
Benzene	benzene-d6 (IS)	8.396	8.426	126,456	
Toluene-d8 (IS)		10.907	10.931	1,014,980	
Toluene	Toluene-d8 (IS)	11.002	11.020	357,099	
Ethylbenzene	Toluene-d8 (IS)	13.133	13.151	73,312	
m-/p-Xylenes	Toluene-d8 (IS)	13.305	13.341	190,772	
o-Xylene	Toluene-d8 (IS)	13.792	13.809	77,868	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2504493.D




+ Scan (10.854-11.056 min, 35 scans) P2504493.D

**Toluene**


+ EIC (91.1) Scan P2504493.D




+ Scan (10.961-11.139 min, 30 scans) P2504493.D

**Ethylbenzene**


+ EIC (91.1) Scan P2504493.D




+ Scan (13.094-13.204 min, 19 scans) P2504493.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2504493.D




+ Scan (13.260-13.483 min, 38 scans) P2504493.D




o-Xylene

+ EIC (91.1) Scan P2504493.D




+ Scan (13.768-13.893 min, 22 scans) P2504493.D





Name SPRSEA-12-S-20250626
Comment C00561
Data File P2504494.D
Acq. Date-Time 7/17/2025 1:44:56 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.349	8.367	840,620	
Benzene	benzene-d6 (IS)	8.402	8.426	81,712	
Toluene-d8 (IS)		10.907	10.931	979,944	
Toluene	Toluene-d8 (IS)	11.002	11.020	181,145	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	35,461	
m-/p-Xylenes	Toluene-d8 (IS)	13.311	13.341	83,326	
o-Xylene	Toluene-d8 (IS)	13.792	13.809	35,930	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2504494.D




+ Scan (10.861-11.038 min, 30 scans) P2504494.D

**Toluene**


+ EIC (91.1) Scan P2504494.D




+ Scan (10.961-11.073 min, 19 scans) P2504494.D

**Ethylbenzene**


+ EIC (91.1) Scan P2504494.D




+ Scan (13.103-13.198 min, 17 scans) P2504494.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2504494.D




+ Scan (13.265-13.448 min, 31 scans) P2504494.D




o-Xylene

+ EIC (91.1) Scan P2504494.D




+ Scan (13.760-13.849 min, 15 scans) P2504494.D





Name SPRSEA-12-D-20250626
Comment C57727
Data File P2504495.D
Acq. Date-Time 7/17/2025 2:22:12 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.349	8.367	820,996	
Benzene	benzene-d6 (IS)	8.408	8.426	71,519	
Toluene-d8 (IS)		10.907	10.931	984,681	
Toluene	Toluene-d8 (IS)	11.008	11.020	174,027	
Ethylbenzene	Toluene-d8 (IS)	13.133	13.151	35,141	
m-/p-Xylenes	Toluene-d8 (IS)	13.311	13.341	88,687	
o-Xylene	Toluene-d8 (IS)	13.804	13.809	35,789	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2504495.D




+ Scan (10.866-11.050 min, 31 scans) P2504495.D

**Toluene**


+ EIC (91.1) Scan P2504495.D




+ Scan (10.959-11.120 min, 27 scans) P2504495.D

**Ethylbenzene**


+ EIC (91.1) Scan P2504495.D




+ Scan (13.092-13.196 min, 17 scans) P2504495.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2504495.D




+ Scan (13.275-13.409 min, 23 scans) P2504495.D




o-Xylene

+ EIC (91.1) Scan P2504495.D




+ Scan (13.763-13.851 min, 15 scans) P2504495.D




Name SPRSEA-12-B-20250626
Comment C20610
Data File P2504496.D
Acq. Date-Time 7/17/2025 2:59:31 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram




Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.349	8.367	851,049	
Benzene	benzene-d6 (IS)	8.408	8.426	16,409	
Toluene-d8 (IS)		10.907	10.931	971,733	
Toluene	Toluene-d8 (IS)	11.002	11.020	8,864	
Ethylbenzene	Toluene-d8 (IS)	13.139	13.151	1,130	m
m-/p-Xylenes	Toluene-d8 (IS)	13.329	13.341	363	
o-Xylene	Toluene-d8 (IS)	14.243	13.809	ND	m

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2504496.D




+ Scan (10.867-10.978 min, 19 scans) P2504496.D

**Toluene**


+ EIC (91.1) Scan P2504496.D




+ Scan (10.964-11.058 min, 16 scans) P2504496.D

**Ethylbenzene**


+ EIC (91.1) Scan P2504496.D




+ Scan (13.114-13.145 min, 5 scans) P2504496.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2504496.D




+ Scan (13.323-13.335 min, 3 scans) P2504496.D




o-Xylene

+ EIC (91.1) Scan P2504496.D



+ Scan (14.243-14.243 min, 1 scans) P2504496.D



Initial Calibration

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC301-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Calibration Curves

Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
P041125A_CC252679	Benzene	1	P2502258.D	4.96	47440	90.3	798559	1.081	0.21
P041125A_CC252679	Benzene	2	P2502259.D	9.92	81580	90.3	784474	0.946	0.056
P041125A_CC252679	Benzene	3	P2502260.D	19.85	154990	90.3	766291	0.920	0.027
P041125A_CC252679	Benzene	4	P2502261.D	39.70	307006	90.3	776135	0.900	0.0043
P041125A_CC252679	Benzene	5	P2502262.D	99.25	682063	90.3	768529	0.807	-0.099
P041125A_CC252679	Benzene	6	P2502263.D	198.49	1383026	90.3	766391	0.821	-0.084
P041125A_CC252679	Benzene	7	P2502264.D	595.47	3847989	90.3	733331	0.796	-0.11
						Avg:	770530	0.896	
						%RSD:	2.6%	11.2%	
P041125A_CC252679	Toluene	1	P2502258.D	5.37	53462	105.3	981002	1.069	0.21
P041125A_CC252679	Toluene	2	P2502259.D	10.73	93718	105.3	977723	0.940	0.061
P041125A_CC252679	Toluene	3	P2502260.D	21.46	177517	105.3	952641	0.914	0.031
P041125A_CC252679	Toluene	4	P2502261.D	42.92	366700	105.3	958584	0.938	0.058
P041125A_CC252679	Toluene	5	P2502262.D	107.31	751020	105.3	937774	0.786	-0.11
P041125A_CC252679	Toluene	6	P2502263.D	214.62	1544431	105.3	950212	0.797	-0.1
P041125A_CC252679	Toluene	7	P2502264.D	643.86	4297415	105.3	923104	0.761	-0.14
						Avg:	954434	0.887	
						%RSD:	2.2%	12.5%	
P041125A_CC252679	Ethylbenzene	1	P2502258.D	5.16	56966	105.3	981002	1.185	0.15
P041125A_CC252679	Ethylbenzene	2	P2502259.D	10.32	108870	105.3	977723	1.137	0.1
P041125A_CC252679	Ethylbenzene	3	P2502260.D	20.63	213920	105.3	952641	1.146	0.11
P041125A_CC252679	Ethylbenzene	4	P2502261.D	41.26	422263	105.3	958584	1.124	0.088
P041125A_CC252679	Ethylbenzene	5	P2502262.D	103.15	801782	105.3	937774	0.873	-0.16
P041125A_CC252679	Ethylbenzene	6	P2502263.D	206.30	1736097	105.3	950212	0.932	-0.098
P041125A_CC252679	Ethylbenzene	7	P2502264.D	618.90	4543663	105.3	923104	0.837	-0.19
						Avg:	954434	1.034	
						%RSD:	2.2%	14.2%	
P041125A_CC252679	m-/p-Xylenes	1	P2502258.D	4.87	38899	105.3	981002	0.857	0.083
P041125A_CC252679	m-/p-Xylenes	2	P2502259.D	9.74	73807	105.3	977723	0.816	0.03
P041125A_CC252679	m-/p-Xylenes	3	P2502260.D	19.48	153415	105.3	952641	0.870	0.099
P041125A_CC252679	m-/p-Xylenes	4	P2502261.D	38.97	310930	105.3	958584	0.876	0.11
P041125A_CC252679	m-/p-Xylenes	5	P2502262.D	97.42	591523	105.3	937774	0.682	-0.14
P041125A_CC252679	m-/p-Xylenes	6	P2502263.D	194.84	1339681	105.3	950212	0.762	-0.038
P041125A_CC252679	m-/p-Xylenes	7	P2502264.D	584.52	3480142	105.3	923104	0.679	-0.14
						Avg:	954434	0.792	
						%RSD:	2.2%	10.8%	
P041125A_CC252679	o-Xylene	1	P2502258.D	5.08	44039	105.3	981002	0.930	0.12
P041125A_CC252679	o-Xylene	2	P2502259.D	10.16	84765	105.3	977723	0.898	0.078
P041125A_CC252679	o-Xylene	3	P2502260.D	20.32	174712	105.3	952641	0.950	0.14

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey
Job No.: 2025GC301-1 EPA Method 325B Analysis
Client No.: PROJ-027966 Site: Sprague - Searsport

Calibration Curves

Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
P041125A_CC252679	o-Xylene	4	P2502261.D	40.64	330869	105.3	958584	0.894	0.073
P041125A_CC252679	o-Xylene	5	P2502262.D	101.61	648711	105.3	937774	0.717	-0.14
P041125A_CC252679	o-Xylene	6	P2502263.D	203.21	1413541	105.3	950212	0.771	-0.075
P041125A_CC252679	o-Xylene	7	P2502264.D	609.64	3599376	105.3	923104	0.673	-0.19
						Avg:	954434	0.833	
						%RSD:	2.2%	13.3%	

Calibration Curves

Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
P041125A_CC252679	Benzene	ICV	P2502265.D	64.22	465786	90.3	753743	0.869	-3.0%
P041125A_CC252679	Toluene	ICV	P2502265.D	76.59	610878	105.3	957560	0.877	-1.1%
P041125A_CC252679	Ethylbenzene	ICV	P2502265.D	86.22	753297	105.3	957560	0.961	-7.1%
P041125A_CC252679	m-/p-Xylenes	ICV	P2502265.D	89.75	613496	105.3	957560	0.752	-5.1%
P041125A_CC252679	o-Xylene	ICV	P2502265.D	88.34	625434	105.3	957560	0.778	-6.6%

Sample Custody



EPA Method 325 A/B
Field Test Data Sheet and
Chain of Custody Record
Page #1 of #1

- ☐ Standard Turn Around Time (10 business days)
☐ Rush Turn Around Time
 • All TATs Subject to Approval by Enthalpy Analytical, Inc.
 • Unless otherwise specified, sample tubes will be conditioned for re-use 3 business days after submission of results

Site Name: Sprague Searsport	Client Name: Montrose Ave	PO#:
Site Address: 20 Trundy Road	Project Number: # 027966	Sample Event #
City: Searsport	Project Manager: Haig Brochu	Sorbent:
State: Maine	Email Address: haigbrochu@Montrose-env.com	
Zip: 04974	Telephone #: 207.441.0025	

Location	Sample ID (Tube ID)	Sample, Blank or Duplicate	Start Date	Start Time	Stop Date	Stop Time	Deployed/Collected by	Ave. Pressure (inHg)	Avg. Ambient Temp. (°F)
1	C 57687	S	6/26/25	1130	7/10/25	1150	HPB		
2	C 33009	S		1120		1140			
3	B 19056	S		1110		1130			
4	C 16112	S		1100		1120			
5	B 49576	S		1210		1220			
6	B 38593	S		1220		1230			
6	B 18562	D		1220		1230			
6	C 40533	B		1220		1230			
7	B 50653	S		1230		1240			
8	C 34132	S		1240		1250			
9	C 61442	S		1250		1300			
10	B 46360	S		1300		1310			
11	C 43733	S		11:55		1210			
12	C 00561	S		1145		1200			
12	C 57727	D		1145		1200			
12	C 20610	B	6/26/25	1145	7/10/25	1200	HPB		

Relinquished By (printed): Haig Brochu	Relinquished By (signature):	Relinquished Date: 7/10/2025	Relinquished Time: 16:30
Received By (printed): Ryn Flood	Received By (signature):	Receipt Date: 7/14/25	Receipt Time: 3:26pm
Sample Condition Upon Receipt: good	Compound List:	Custody Seal Intact? Y/N: Y	Delivery tracking #
Ice Temp: Blank Temp: 25.2	Fluke 4	Add Custody Seal # below: 24H11352	

Comments:

**This Is The Last Page
Of This Report.**

Sprague - Searsport

70 Trundy Road
Searsport, ME 04974

Sampling Event 26 Sprague - Searsport

Client Project# PROJ-027966

Samples Received: 7/29/2025

Analytical Report 2025GC302

EPA Method 325B Analysis

Report Issue Date: 8/7/2025

I certify that to the best of my knowledge all analytical data presented in this report have been checked for completeness, accuracy, errors and legibility in addition to having been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s). This report shall not be reproduced except in full without approval of the laboratory. This will provide assurance that parts of the report are not taken out of context.

Amendment(s):

Signature:



QA REVIEW PERFORMED BY

Brianna Berry
QA Associate I



Matt Cavanaugh
Matthew.Cavanaugh@enthalpy.com / www.enthalpy.com
O: (919) 850-4392
Enthalpy Analytical
800 Capitola Drive Suite 1 Durham, NC 27713

Table of Contents

Case Narrative	3
Results	6
Summary of Results	7
Detailed Results	8
QC Data	11
Chromatograms	14
Initial Calibration	66
Sample Custody	69
Chain of Custody	70

Narrative Summary



Enthalpy Analytical Narrative Summary

Company	Montrose Air Quality Services, LLC - New Jersey
Job No.	2025GC302-1
Client ID.	PROJ-027966 Site: Sprague - Searsport

1. Custody

The samples were received at Enthalpy Analytical on July 29, 2025 at 25.8 °C, which is above the method recommended 23.0 °C. The samples were received in good condition. Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC

Table 1 - Sample Inventory

Sample ID	Tube ID	Sample Type
SPRSEA-1-S-20250710	C33260	Sample
SPRSEA-2-S-20250710	B17444	Sample
SPRSEA-3-S-20250710	C56859	Sample
SPRSEA-4-S-20250710	C38901	Sample
SPRSEA-5-S-20250710	C57730	Sample
SPRSEA-6-S-20250710	C57534	Sample
SPRSEA-6-D-20250710	C01661	Duplicate
SPRSEA-6-B-20250710	C00592	Blank
SPRSEA-7-S-20250710	C55400	Sample
SPRSEA-8-S-20250710	C20391	Sample
SPRSEA-9-S-20250710	B14580	Sample
SPRSEA-10-S-20250710	C43380	Sample
SPRSEA-11-S-20250710	B12197	Sample
SPRSEA-12-S-20250710	C36934	Sample
SPRSEA-12-D-20250710	C43377	Duplicate
SPRSEA-12-B-20250710	C43932	Blank

2. Analysis

The samples were analyzed for Benzene, Toluene, Ethylbenzene, m-/p-Xylenes, and o-Xylene using EPA Method 325B – Volatile Organic Compounds from Fugitive and Area Sources by Thermal Desorption and GC/MS. A copy of the acquisition method M325B-MTD is not included in this report but may be available upon request.

The sample tube media used for this sampling period was CarbopackX. All calibration standards and laboratory QC were prepared using the same media.

3. Calibration

All BFB tune criteria have been met for this analysis.

The initial calibration (E050525A_CC252679) met all 30% RSD criteria. The initial calibration verification met 30% difference criteria. The continuing calibration verifications met 30% difference criteria. The initial and continuing calibration raw data are not included in this report but are available upon request.

Enthalpy Analytical Narrative Summary

Company	Montrose Air Quality Services, LLC - New Jersey
Job No.	2025GC302-1
Client ID.	PROJ-027966 Site: Sprague - Searsport

5. QC Notes

All quality control criteria required by the method and/or the laboratory SOP have been met unless noted otherwise below.

The primary sample SPRSEA-12-S-20250710 (tube ID C36934) and its corresponding duplicate SPRSEA-12-D-20250710 (tube ID C43377) failed to meet the 30% difference criterion for o-Xylene as specified by the method. However, the concentrations of the analyte in both the sample and the duplicate were less than two times the reporting limit of the instrument's calibration curve. Therefore, the percent difference observed may not suggest the data set has been negatively affected. All samples in the data set have been flagged "Pc" for o-Xylene to denote this failure.

The primary sample SPRSEA-6-S-20250710 (tube ID C57534) and its corresponding duplicate SPRSEA-6-D-20250710 (tube ID C01661) failed to meet the 30% difference criterion for Toluene as specified by the method. All samples in the data set have been flagged "P" for Toluene to denote this failure.

6. Reporting Notes

All tubes used for this sampling period met the method criteria for number of uses; no tube exceeded 50 field uses.

As specified in EPA Method 325B, the response factor of the daily continuing calibration standard was used to quantitate all field samples and blanks.

All samples were reported as amount in ng catch, and concentration in ug/m³ and ppbv.

The results presented in this report are representative of the samples as provided to the laboratory. These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, located at 800 Capitola Drive, Suite 1, Durham NC, 27713 is accredited by the Louisiana Department of Environmental Quality (LDEQ) for EPA Method 325B for all analytes included in this report under **Certificate Number 04010**.

Results

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC302-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Summary

Sample Code	Tube ID	Benzene		Toluene		Ethylbenzene		m-/p-Xylenes		o-Xylene	
		(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag
SPRSEA-1-S-20250710	C33260	0.437		1.05	P	0.271	ND	0.495	J	0.271	ND,Pc
SPRSEA-2-S-20250710	B17444	0.610		1.65	P	0.475	J	0.801		0.336	J,Pc
SPRSEA-3-S-20250710	C56859	0.516		1.49	P	0.294	J	0.956		0.427	J,Pc
SPRSEA-4-S-20250710	C38901	0.484		1.38	P	0.289	J	0.672		0.271	ND,Pc
SPRSEA-5-S-20250710	C57730	0.383		1.01	P	0.271	ND	0.617		0.271	ND,Pc
SPRSEA-6-S-20250710	C57534	0.362	J	0.874	P	0.271	ND	0.670		0.271	ND,Pc
SPRSEA-6-D-20250710	C01661	0.451		1.23	P	0.309	J	0.543		0.271	ND,Pc
SPRSEA-6-B-20250710	C00592	0.186	ND	0.240	ND,P	0.271	ND	0.271	ND	0.271	ND,Pc
SPRSEA-7-S-20250710	C55400	0.393		0.903	P	0.271	ND	0.341	J	0.271	ND,Pc
SPRSEA-8-S-20250710	C20391	0.577		1.58	P	0.342	J	0.652		0.273	J,Pc
SPRSEA-9-S-20250710	B14580	1.09		3.65	P	0.784		1.81		0.786	Pc
SPRSEA-10-S-20250710	C43380	3.67		18.8	P	3.13		10.5		4.31	Pc
SPRSEA-11-S-20250710	B12197	1.21		4.05	P	1.28		3.38		1.10	Pc
SPRSEA-12-S-20250710	C36934	0.807		2.92	P	0.745		2.14		0.761	Pc
SPRSEA-12-D-20250710	C43377	0.821		3.22	P	0.871		2.84		1.06	Pc
SPRSEA-12-B-20250710	C43932	0.186	ND	0.240	ND,P	0.271	ND	0.271	ND	0.271	ND,Pc

J: Estimated Value - The analyte was detected between the Method Detection Limit and Reporting Limit

ND: The analyte was not present above the Method Detection Limit

P: Field duplicate(s) exceed 30%RPD

Pc: Field duplicate(s) exceed 30%RPD. Concentrations of both samples in duplicate are near the reporting limit

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC302-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Benzene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250710	C33260	0.437	0.137	5.86	68.4	0.665	20185	0.186	0.368	0.0584	0.115		E2502140.d	2025-07-31 01:16	0.784	8.109	42108	496490	54.2	8.052	0.1%
SPRSEA-2-S-20250710	B17444	0.610	0.191	8.18	68.4	0.665	20185	0.186	0.368	0.0584	0.115		E2502141.d	2025-07-31 01:41	0.784	8.103	58428	493749	54.2	8.045	-0.5%
SPRSEA-3-S-20250710	C56859	0.516	0.162	6.92	68.4	0.665	20185	0.186	0.368	0.0584	0.115		E2502142.d	2025-07-31 02:08	0.784	8.110	49681	496202	54.2	8.045	0.0%
SPRSEA-4-S-20250710	C38901	0.484	0.152	6.50	68.4	0.665	20185	0.186	0.368	0.0584	0.115		E2502143.d	2025-07-31 02:34	0.784	8.110	46518	494734	54.2	8.052	-0.3%
SPRSEA-5-S-20250710	C57730	0.383	0.120	5.13	68.4	0.665	20185	0.186	0.368	0.0584	0.115		E2502144.d	2025-07-31 02:59	0.784	8.110	37330	502560	54.2	8.052	1.3%
SPRSEA-6-S-20250710	C57534	0.362	0.113	4.86	68.4	0.665	20185	0.186	0.368	0.0584	0.115	J	E2502145.d	2025-07-31 03:25	0.784	8.110	35070	499176	54.2	8.045	0.6%
SPRSEA-6-D-20250710	C01661	0.451	0.141	6.05	68.4	0.665	20185	0.186	0.368	0.0584	0.115		E2502146.d	2025-07-31 03:51	0.784	8.110	43769	499746	54.2	8.052	0.7%
SPRSEA-6-B-20250710	C00592	0.186	0.0584		68.4	0.665	20185	0.186	0.368	0.0584	0.115	ND	E2502139.d	2025-07-31 00:50	0.784	8.102	7198	497429	54.2	8.045	0.3%
SPRSEA-7-S-20250710	C55400	0.393	0.123	5.27	68.4	0.665	20185	0.186	0.368	0.0584	0.115		E2502147.d	2025-07-31 04:17	0.784	8.109	37961	497648	54.2	8.052	0.3%
SPRSEA-8-S-20250710	C20391	0.577	0.181	7.74	68.4	0.665	20185	0.186	0.368	0.0584	0.115		E2502148.d	2025-07-31 04:42	0.784	8.110	55780	498117	54.2	8.045	0.4%
SPRSEA-9-S-20250710	B14580	1.09	0.343	14.7	68.4	0.665	20185	0.186	0.368	0.0584	0.115		E2502149.d	2025-07-31 05:08	0.784	8.110	105303	496100	54.2	8.052	0.0%
SPRSEA-10-S-20250710	C43380	3.67	1.15	49.2	68.4	0.665	20185	0.186	0.368	0.0584	0.115		E2502151.d	2025-07-31 05:59	0.784	8.110	351364	493424	54.2	8.052	-0.5%
SPRSEA-11-S-20250710	B12197	1.21	0.379	16.2	68.4	0.665	20185	0.186	0.368	0.0584	0.115		E2502152.d	2025-07-31 06:25	0.784	8.110	117131	498565	54.2	8.052	0.5%
SPRSEA-12-S-20250710	C36934	0.807	0.253	10.8	68.4	0.665	20185	0.186	0.368	0.0584	0.115		E2502153.d	2025-07-31 06:51	0.784	8.110	78548	501332	54.2	8.052	1.1%
SPRSEA-12-D-20250710	C43377	0.821	0.257	11.0	68.4	0.665	20185	0.186	0.368	0.0584	0.115		E2502154.d	2025-07-31 07:17	0.784	8.110	79826	501003	54.2	8.052	1.0%
SPRSEA-12-B-20250710	C43932	0.186	0.0584		68.4	0.665	20185	0.186	0.368	0.0584	0.115	ND	E2502155.d	2025-07-31 07:43	0.784	8.110	14246	500845	54.2	8.045	1.0%

Toluene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250710	C33260	1.05	0.280	11.0	68.4	0.516	20185	0.240	0.513	0.0638	0.136	P	E2502140.d	2025-07-31 01:16	0.838	10.846	81560	560798	63.2	10.745	-0.4%
SPRSEA-2-S-20250710	B17444	1.65	0.439	17.2	68.4	0.516	20185	0.240	0.513	0.0638	0.136	P	E2502141.d	2025-07-31 01:41	0.838	10.846	126564	553685	63.2	10.746	-1.7%
SPRSEA-3-S-20250710	C56859	1.49	0.397	15.6	68.4	0.516	20185	0.240	0.513	0.0638	0.136	P	E2502142.d	2025-07-31 02:08	0.838	10.846	116994	567031	63.2	10.746	0.7%
SPRSEA-4-S-20250710	C38901	1.38	0.368	14.4	68.4	0.516	20185	0.240	0.513	0.0638	0.136	P	E2502143.d	2025-07-31 02:34	0.838	10.846	107285	560833	63.2	10.746	-0.4%
SPRSEA-5-S-20250710	C57730	1.01	0.268	10.5	68.4	0.516	20185	0.240	0.513	0.0638	0.136	P	E2502144.d	2025-07-31 02:59	0.838	10.846	78508	563313	63.2	10.746	0.1%
SPRSEA-6-S-20250710	C57534	0.874	0.232	9.10	68.4	0.516	20185	0.240	0.513	0.0638	0.136	P	E2502145.d	2025-07-31 03:25	0.838	10.846	67890	562560	63.2	10.746	-0.1%
SPRSEA-6-D-20250710	C01661	1.23	0.327	12.8	68.4	0.516	20185	0.240	0.513	0.0638	0.136	P	E2502146.d	2025-07-31 03:51	0.838	10.846	95112	558483	63.2	10.746	-0.8%
SPRSEA-6-B-20250710	C00592	0.240	0.0638		68.4	0.516	20185	0.240	0.513	0.0638	0.136	ND,P	E2502139.d	2025-07-31 00:50	0.838	10.846	7202	565495	63.2	10.746	0.4%
SPRSEA-7-S-20250710	C55400	0.903	0.240	9.40	68.4	0.516	20185	0.240	0.513	0.0638	0.136	P	E2502147.d	2025-07-31 04:17	0.838	10.846	70205	562706	63.2	10.753	-0.1%
SPRSEA-8-S-20250710	C20391	1.58	0.420	16.5	68.4	0.516	20185	0.240	0.513	0.0638	0.136	P	E2502148.d	2025-07-31 04:42	0.838	10.846	124501	570289	63.2	10.746	1.3%
SPRSEA-9-S-20250710	B14580	3.65	0.970	38.0	68.4	0.516	20185	0.240	0.513	0.0638	0.136	P	E2502149.d	2025-07-31 05:08	0.838	10.846	285141	565042	63.2	10.753	0.4%
SPRSEA-10-S-20250710	C43380	18.8	5.00	196	68.4	0.516	20185	0.240	0.513	0.0638	0.136	P	E2502151.d	2025-07-31 05:59	0.838	10.846	1452224	558657	63.2	10.746	-0.8%
SPRSEA-11-S-20250710	B12197	4.05	1.08	42.2	68.4	0.516	20185	0.240	0.513	0.0638	0.136	P	E2502152.d	2025-07-31 06:25	0.838	10.846	315539	563851	63.2	10.746	0.1%
SPRSEA-12-S-20250710	C36934	2.92	0.775	30.4	68.4	0.516	20185	0.240	0.513	0.0638	0.136	P	E2502153.d	2025-07-31 06:51	0.838	10.846	229833	569892	63.2	10.746	1.2%
SPRSEA-12-D-20250710	C43377	3.22	0.854	33.5	68.4	0.516	20185	0.240	0.513	0.0638	0.136	P	E2502154.d	2025-07-31 07:17	0.838	10.846	251735	566494	63.2	10.746	0.6%
SPRSEA-12-B-20250710	C43932	0.240	0.0638		68.4	0.516	20185	0.240	0.513	0.0638	0.136	ND,P	E2502155.d	2025-07-31 07:43	0.838	10.846	11564	571448	63.2	10.753	1.5%

Ethylbenzene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250710	C33260	0.271	0.0626		68.4	0.456	20185	0.271	0.557	0.0626	0.128	ND	E2502140.d	2025-07-31 01:16	1.009	13.038	17068	560798	63.2	10.745	-0.4%

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey
Job No.: 2025GC302-1 EPA Method 325B Analysis
Client No.: PROJ-027966 Site: Sprague - Searsport

Ethylbenzene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-2-S-20250710	B17444	0.475	0.110	4.38	68.4	0.456	20185	0.271	0.557	0.0626	0.128	J	E2502141.d	2025-07-31 01:41	1.009	13.038	38708	553685	63.2	10.746	-1.7%
SPRSEA-3-S-20250710	C56859	0.294	0.0678	2.71	68.4	0.456	20185	0.271	0.557	0.0626	0.128	J	E2502142.d	2025-07-31 02:08	1.009	13.038	24530	567031	63.2	10.746	0.7%
SPRSEA-4-S-20250710	C38901	0.289	0.0666	2.66	68.4	0.456	20185	0.271	0.557	0.0626	0.128	J	E2502143.d	2025-07-31 02:34	1.009	13.038	23835	560833	63.2	10.746	-0.4%
SPRSEA-5-S-20250710	C57730	0.271	0.0626		68.4	0.456	20185	0.271	0.557	0.0626	0.128	ND	E2502144.d	2025-07-31 02:59	1.009	13.038	16149	563313	63.2	10.746	0.1%
SPRSEA-6-S-20250710	C57534	0.271	0.0626		68.4	0.456	20185	0.271	0.557	0.0626	0.128	ND	E2502145.d	2025-07-31 03:25	1.009	13.038	17550	562560	63.2	10.746	-0.1%
SPRSEA-6-D-20250710	C01661	0.309	0.0712	2.85	68.4	0.456	20185	0.271	0.557	0.0626	0.128	J	E2502146.d	2025-07-31 03:51	1.009	13.038	25382	558483	63.2	10.746	-0.8%
SPRSEA-6-B-20250710	C00592	0.271	0.0626		68.4	0.456	20185	0.271	0.557	0.0626	0.128	ND	E2502139.d	2025-07-31 00:50	1.009	13.038	926	565495	63.2	10.746	0.4%
SPRSEA-7-S-20250710	C55400	0.271	0.0626		68.4	0.456	20185	0.271	0.557	0.0626	0.128	ND	E2502147.d	2025-07-31 04:17	1.009	13.038	11269	562706	63.2	10.753	-0.1%
SPRSEA-8-S-20250710	C20391	0.342	0.0789	3.15	68.4	0.456	20185	0.271	0.557	0.0625	0.128	J	E2502148.d	2025-07-31 04:42	1.009	13.038	28715	570289	63.2	10.746	1.3%
SPRSEA-9-S-20250710	B14580	0.784	0.181	7.22	68.4	0.456	20185	0.271	0.557	0.0625	0.128		E2502149.d	2025-07-31 05:08	1.009	13.038	65167	565042	63.2	10.753	0.4%
SPRSEA-10-S-20250710	C43380	3.13	0.720	28.8	68.4	0.456	20185	0.271	0.557	0.0625	0.128		E2502151.d	2025-07-31 05:59	1.009	13.038	256833	558657	63.2	10.746	-0.8%
SPRSEA-11-S-20250710	B12197	1.28	0.296	11.8	68.4	0.456	20185	0.271	0.557	0.0626	0.128		E2502152.d	2025-07-31 06:25	1.009	13.038	106544	563851	63.2	10.746	0.1%
SPRSEA-12-S-20250710	C36934	0.745	0.172	6.86	68.4	0.456	20185	0.271	0.557	0.0626	0.128		E2502153.d	2025-07-31 06:51	1.009	13.038	62460	569892	63.2	10.746	1.2%
SPRSEA-12-D-20250710	C43377	0.871	0.201	8.02	68.4	0.456	20185	0.271	0.557	0.0626	0.128		E2502154.d	2025-07-31 07:17	1.009	13.038	72541	566494	63.2	10.746	0.6%
SPRSEA-12-B-20250710	C43932	0.271	0.0626		68.4	0.456	20185	0.271	0.557	0.0626	0.128	ND	E2502155.d	2025-07-31 07:43	1.009	13.031	2384	571448	63.2	10.753	1.5%

m-/p-Xylenes

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250710	C33260	0.495	0.114	4.56	68.4	0.456	20185	0.271	0.526	0.0626	0.121	J	E2502140.d	2025-07-31 01:16	0.857	13.209	34682	560798	63.2	10.745	-0.4%
SPRSEA-2-S-20250710	B17444	0.801	0.185	7.38	68.4	0.456	20185	0.271	0.526	0.0626	0.121		E2502141.d	2025-07-31 01:41	0.857	13.210	55459	553685	63.2	10.746	-1.7%
SPRSEA-3-S-20250710	C56859	0.956	0.220	8.81	68.4	0.456	20185	0.271	0.526	0.0626	0.121		E2502142.d	2025-07-31 02:08	0.857	13.217	67786	567031	63.2	10.746	0.7%
SPRSEA-4-S-20250710	C38901	0.672	0.155	6.19	68.4	0.456	20185	0.271	0.526	0.0626	0.121		E2502143.d	2025-07-31 02:34	0.857	13.210	47107	560833	63.2	10.746	-0.4%
SPRSEA-5-S-20250710	C57730	0.617	0.142	5.68	68.4	0.456	20185	0.271	0.526	0.0626	0.121		E2502144.d	2025-07-31 02:59	0.857	13.210	43439	563313	63.2	10.746	0.1%
SPRSEA-6-S-20250710	C57534	0.670	0.154	6.17	68.4	0.456	20185	0.271	0.526	0.0626	0.121		E2502145.d	2025-07-31 03:25	0.857	13.210	47097	562560	63.2	10.746	-0.1%
SPRSEA-6-D-20250710	C01661	0.543	0.125	5.01	68.4	0.456	20185	0.271	0.526	0.0626	0.121		E2502146.d	2025-07-31 03:51	0.857	13.210	37946	558483	63.2	10.746	-0.8%
SPRSEA-6-B-20250710	C00592	0.271	0.0626		68.4	0.456	20185	0.271	0.526	0.0626	0.121	ND	E2502139.d	2025-07-31 00:50	0.857	13.231	1058	565495	63.2	10.746	0.4%
SPRSEA-7-S-20250710	C55400	0.341	0.0785	3.14	68.4	0.456	20185	0.271	0.526	0.0626	0.121	J	E2502147.d	2025-07-31 04:17	0.857	13.210	23979	562706	63.2	10.753	-0.1%
SPRSEA-8-S-20250710	C20391	0.652	0.150	6.00	68.4	0.456	20185	0.271	0.526	0.0625	0.121		E2502148.d	2025-07-31 04:42	0.857	13.210	46467	570289	63.2	10.746	1.3%
SPRSEA-9-S-20250710	B14580	1.81	0.416	16.6	68.4	0.456	20185	0.271	0.526	0.0625	0.121		E2502149.d	2025-07-31 05:08	0.857	13.210	127680	565042	63.2	10.753	0.4%
SPRSEA-10-S-20250710	C43380	10.5	2.41	96.5	68.4	0.456	20185	0.271	0.526	0.0625	0.121		E2502151.d	2025-07-31 05:59	0.857	13.210	731948	558657	63.2	10.746	-0.8%
SPRSEA-11-S-20250710	B12197	3.38	0.778	31.1	68.4	0.456	20185	0.271	0.526	0.0626	0.121		E2502152.d	2025-07-31 06:25	0.857	13.210	237921	563851	63.2	10.746	0.1%
SPRSEA-12-S-20250710	C36934	2.14	0.494	19.7	68.4	0.456	20185	0.271	0.526	0.0626	0.121		E2502153.d	2025-07-31 06:51	0.857	13.210	152693	569892	63.2	10.746	1.2%
SPRSEA-12-D-20250710	C43377	2.84	0.655	26.2	68.4	0.456	20185	0.271	0.526	0.0626	0.121		E2502154.d	2025-07-31 07:17	0.857	13.210	201268	566494	63.2	10.746	0.6%
SPRSEA-12-B-20250710	C43932	0.271	0.0626		68.4	0.456	20185	0.271	0.526	0.0626	0.121	ND	E2502155.d	2025-07-31 07:43	0.857	13.203	1276	571448	63.2	10.753	1.5%

o-Xylene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250710	C33260	0.271	0.0626		68.4	0.456	20185	0.271	0.549	0.0626	0.126	ND,Pc	E2502140.d	2025-07-31 01:16	0.842	13.718	13630	560798	63.2	10.745	-0.4%
SPRSEA-2-S-20250710	B17444	0.336	0.0774	3.09	68.4	0.456	20185	0.271	0.549	0.0626	0.126	J,Pc	E2502141.d	2025-07-31 01:41	0.842	13.711	22826	553685	63.2	10.746	-1.7%

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC302-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

o-Xylene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-3-S-20250710	C56859	0.427	0.0984	3.93	68.4	0.456	20185	0.271	0.549	0.0626	0.126	J,Pc	E2502142.d	2025-07-31 02:08	0.842	13.718	29718	567031	63.2	10.746	0.7%
SPRSEA-4-S-20250710	C38901	0.271	0.0626		68.4	0.456	20185	0.271	0.549	0.0626	0.126	ND,Pc	E2502143.d	2025-07-31 02:34	0.842	13.718	17984	560833	63.2	10.746	-0.4%
SPRSEA-5-S-20250710	C57730	0.271	0.0626		68.4	0.456	20185	0.271	0.549	0.0626	0.126	ND,Pc	E2502144.d	2025-07-31 02:59	0.842	13.711	16755	563313	63.2	10.746	0.1%
SPRSEA-6-S-20250710	C57534	0.271	0.0626		68.4	0.456	20185	0.271	0.549	0.0626	0.126	ND,Pc	E2502145.d	2025-07-31 03:25	0.842	13.711	16075	562560	63.2	10.746	-0.1%
SPRSEA-6-D-20250710	C01661	0.271	0.0626		68.4	0.456	20185	0.271	0.549	0.0626	0.126	ND,Pc	E2502146.d	2025-07-31 03:51	0.842	13.718	13584	558483	63.2	10.746	-0.8%
SPRSEA-6-B-20250710	C00592	0.271	0.0626		68.4	0.456	20185	0.271	0.549	0.0626	0.126	ND,Pc	E2502139.d	2025-07-31 00:50	0.842	13.718	610	565495	63.2	10.746	0.4%
SPRSEA-7-S-20250710	C55400	0.271	0.0626		68.4	0.456	20185	0.271	0.549	0.0626	0.126	ND,Pc	E2502147.d	2025-07-31 04:17	0.842	13.711	8687	562706	63.2	10.753	-0.1%
SPRSEA-8-S-20250710	C20391	0.273	0.0629	2.51	68.4	0.456	20185	0.271	0.549	0.0625	0.126	J,Pc	E2502148.d	2025-07-31 04:42	0.842	13.718	19105	570289	63.2	10.746	1.3%
SPRSEA-9-S-20250710	B14580	0.786	0.181	7.24	68.4	0.456	20185	0.271	0.549	0.0625	0.126	Pc	E2502149.d	2025-07-31 05:08	0.842	13.711	54523	565042	63.2	10.753	0.4%
SPRSEA-10-S-20250710	C43380	4.31	0.993	39.7	68.4	0.456	20185	0.271	0.549	0.0625	0.126	Pc	E2502151.d	2025-07-31 05:59	0.842	13.711	295436	558657	63.2	10.746	-0.8%
SPRSEA-11-S-20250710	B12197	1.10	0.253	10.1	68.4	0.456	20185	0.271	0.549	0.0626	0.126	Pc	E2502152.d	2025-07-31 06:25	0.842	13.718	76097	563851	63.2	10.746	0.1%
SPRSEA-12-S-20250710	C36934	0.761	0.175	7.01	68.4	0.456	20185	0.271	0.549	0.0626	0.126	Pc	E2502153.d	2025-07-31 06:51	0.842	13.711	53260	569892	63.2	10.746	1.2%
SPRSEA-12-D-20250710	C43377	1.06	0.243	9.73	68.4	0.456	20185	0.271	0.549	0.0626	0.126	Pc	E2502154.d	2025-07-31 07:17	0.842	13.718	73467	566494	63.2	10.746	0.6%
SPRSEA-12-B-20250710	C43932	0.271	0.0626		68.4	0.456	20185	0.271	0.549	0.0626	0.126	ND,Pc	E2502155.d	2025-07-31 07:43	0.842	13.718	449	571448	63.2	10.753	1.5%

J: Estimated Value - The analyte was detected between the Method Detection Limit and Reporting Limit

ND: The analyte was not present above the Method Detection Limit

P: Field duplicate(s) exceed 30%RPD

Pc: Field duplicate(s) exceed 30%RPD. Concentrations of both samples in duplicate are near the reporting limit

QC Data



Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey
Job No.: 2025GC302-1 EPA Method 325B Analysis
Client No.: PROJ-027966 Site: Sprague - Searsport

QC Samples

Field Sample Type	Sample Code	Benzene		Toluene		Ethylbenzene		m-/p-Xylenes		o-Xylene	
Blanks (ug/m³)	SPRSEA-6-B-20250710	ND	Pass	ND	Pass	ND	Pass	ND	Pass	ND	Pass
	SPRSEA-12-B-20250710	ND	Pass	ND	Pass	ND	Pass	ND	Pass	ND	Pass
Duplicates (difference)	SPRSEA-6-D-20250710	22%	Pass	34%	Fail	ND	Pass	21%	Pass	ND	Pass
	SPRSEA-12-D-20250710	1.7%	Pass	9.7%	Pass	16%	Pass	28%	Pass	32%	Fail

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC302-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Benzene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	E2502137.d	C43295	Cal	0.784	0.877	0.784	-11%	-4.1%		Pass	
2025GC302 Method Blank-1	E2502138.d	C39279	Blank		0.877	0.784			0.60%	Pass	ND
M325B CCV 5 REC	E2502150.d	B15702	Check	0.835	0.877	0.784	-4.8%		-0.95%	Pass	
M325B CCV 5 REC	E2502156.d	C43842	Check	0.790	0.877	0.784	-9.9%		0.17%	Pass	

Toluene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	E2502137.d	C43295	Cal	0.838	0.989	0.838	-15%	3.5%		Pass	
2025GC302 Method Blank-1	E2502138.d	C39279	Blank		0.989	0.838			-1.2%	Pass	ND
M325B CCV 5 REC	E2502150.d	B15702	Check	0.976	0.989	0.838	-1.3%		-1.6%	Pass	
M325B CCV 5 REC	E2502156.d	C43842	Check	0.853	0.989	0.838	-14%		-0.13%	Pass	

Ethylbenzene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	E2502137.d	C43295	Cal	1.009	1.195	1.009	-16%	3.5%		Pass	
2025GC302 Method Blank-1	E2502138.d	C39279	Blank		1.195	1.009			-1.2%	Pass	ND
M325B CCV 5 REC	E2502150.d	B15702	Check	1.077	1.195	1.009	-9.8%		-1.6%	Pass	
M325B CCV 5 REC	E2502156.d	C43842	Check	1.009	1.195	1.009	-16%		-0.13%	Pass	

m-/p-Xylenes Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	E2502137.d	C43295	Cal	0.857	1.037	0.857	-17%	3.5%		Pass	
2025GC302 Method Blank-1	E2502138.d	C39279	Blank		1.037	0.857			-1.2%	Pass	ND
M325B CCV 5 REC	E2502150.d	B15702	Check	0.905	1.037	0.857	-13%		-1.6%	Pass	
M325B CCV 5 REC	E2502156.d	C43842	Check	0.856	1.037	0.857	-18%		-0.13%	Pass	


o-Xylene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	E2502137.d	C43295	Cal	0.842	0.950	0.842	-11%	3.5%		Pass	
2025GC302 Method Blank-1	E2502138.d	C39279	Blank		0.950	0.842			-1.2%	Pass	ND
M325B CCV 5 REC	E2502150.d	B15702	Check	0.896	0.950	0.842	-5.7%		-1.6%	Pass	
M325B CCV 5 REC	E2502156.d	C43842	Check	0.828	0.950	0.842	-13%		-0.13%	Pass	

Chromatograms



Name 2025GC302 Method Blank-1
Comment C39279
Data File E2502138.d
Acq. Date-Time 7/31/2025 12:24:22 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	499,027	
Benzene	benzene-d6 (IS)	8.102	8.110	4,661	
Toluene-d8 (IS)		10.746	10.753	556,200	
Toluene	Toluene-d8 (IS)	10.846	10.846	6,983	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	1,727	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	1,736	
o-Xylene	Toluene-d8 (IS)	13.718	13.718	1,124	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502138.d




+ Scan (10.703-10.882 min, 26 scans) E2502138.d

**Toluene**


+ EIC (91.1) Scan E2502138.d




+ Scan (10.817-10.889 min, 10 scans) E2502138.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502138.d




+ Scan (12.999-13.070 min, 10 scans) E2502138.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502138.d




+ Scan (13.196-13.272 min, 10 scans) E2502138.d




o-Xylene

+ EIC (91.1) Scan E2502138.d




+ Scan (13.690-13.738 min, 7 scans) E2502138.d





Name SPRSEA-6-B-20250710
Comment C00592
Data File E2502139.d
Acq. Date-Time 7/31/2025 12:50:11 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	497,429	
Benzene	benzene-d6 (IS)	8.102	8.110	7,198	
Toluene-d8 (IS)		10.746	10.753	565,495	
Toluene	Toluene-d8 (IS)	10.846	10.846	7,202	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	926	
m-/p-Xylenes	Toluene-d8 (IS)	13.231	13.217	1,058	
o-Xylene	Toluene-d8 (IS)	13.718	13.718	610	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502139.d




+ Scan (10.703-10.875 min, 25 scans) E2502139.d

**Toluene**


+ EIC (91.1) Scan E2502139.d




+ Scan (10.811-10.875 min, 9 scans) E2502139.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502139.d




+ Scan (13.018-13.066 min, 6 scans) E2502139.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502139.d




+ Scan (13.195-13.260 min, 10 scans) E2502139.d




o-Xylene

+ EIC (91.1) Scan E2502139.d




+ Scan (13.697-13.746 min, 7 scans) E2502139.d





Name SPRSEA-1-S-20250710
Comment C33260
Data File E2502140.d
Acq. Date-Time 7/31/2025 1:16:00 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.052	8.052	496,490	
Benzene	benzene-d6 (IS)	8.109	8.110	42,108	
Toluene-d8 (IS)		10.745	10.753	560,798	
Toluene	Toluene-d8 (IS)	10.846	10.846	81,560	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	17,068	
m-/p-Xylenes	Toluene-d8 (IS)	13.209	13.217	34,682	
o-Xylene	Toluene-d8 (IS)	13.718	13.718	13,630	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502140.d




+ Scan (10.702-10.881 min, 26 scans) E2502140.d

**Toluene**


+ EIC (91.1) Scan E2502140.d




+ Scan (10.803-10.917 min, 16 scans) E2502140.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502140.d




+ Scan (13.004-13.070 min, 9 scans) E2502140.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502140.d




+ Scan (13.177-13.279 min, 14 scans) E2502140.d




o-Xylene

+ EIC (91.1) Scan E2502140.d




+ Scan (13.679-13.754 min, 11 scans) E2502140.d





Name SPRSEA-2-S-20250710
Comment B17444
Data File E2502141.d
Acq. Date-Time 7/31/2025 1:41:49 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	493,749	
Benzene	benzene-d6 (IS)	8.103	8.110	58,428	
Toluene-d8 (IS)		10.746	10.753	553,685	
Toluene	Toluene-d8 (IS)	10.846	10.846	126,564	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	38,708	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	55,459	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	22,826	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502141.d




+ Scan (10.703-10.875 min, 25 scans) E2502141.d

**Toluene**


+ EIC (91.1) Scan E2502141.d




+ Scan (10.796-10.932 min, 20 scans) E2502141.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502141.d




+ Scan (13.001-13.073 min, 10 scans) E2502141.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502141.d




+ Scan (13.179-13.332 min, 22 scans) E2502141.d




o-Xylene

+ EIC (91.1) Scan E2502141.d




+ Scan (13.684-13.776 min, 13 scans) E2502141.d





Name SPRSEA-3-S-20250710
Comment C56859
Data File E2502142.d
Acq. Date-Time 7/31/2025 2:08:25 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	496,202	
Benzene	benzene-d6 (IS)	8.110	8.110	49,681	
Toluene-d8 (IS)		10.746	10.753	567,031	
Toluene	Toluene-d8 (IS)	10.846	10.846	116,994	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	24,530	
m-/p-Xylenes	Toluene-d8 (IS)	13.217	13.217	67,786	
o-Xylene	Toluene-d8 (IS)	13.718	13.718	29,718	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502142.d




+ Scan (10.703-10.867 min, 24 scans) E2502142.d

**Toluene**


+ EIC (91.1) Scan E2502142.d




+ Scan (10.796-10.910 min, 17 scans) E2502142.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502142.d




+ Scan (12.995-13.095 min, 15 scans) E2502142.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502142.d




+ Scan (13.173-13.281 min, 15 scans) E2502142.d




o-Xylene

+ EIC (91.1) Scan E2502142.d




+ Scan (13.675-13.768 min, 14 scans) E2502142.d





Name SPRSEA-4-S-20250710
Comment C38901
Data File E2502143.d
Acq. Date-Time 7/31/2025 2:34:13 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.052	8.052	494,734	
Benzene	benzene-d6 (IS)	8.110	8.110	46,518	
Toluene-d8 (IS)		10.746	10.753	560,833	
Toluene	Toluene-d8 (IS)	10.846	10.846	107,285	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	23,835	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	47,107	
o-Xylene	Toluene-d8 (IS)	13.718	13.718	17,984	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502143.d




+ Scan (10.703-10.867 min, 24 scans) E2502143.d

**Toluene**


+ EIC (91.1) Scan E2502143.d




+ Scan (10.798-10.896 min, 14 scans) E2502143.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502143.d




+ Scan (13.003-13.095 min, 13 scans) E2502143.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502143.d




+ Scan (13.175-13.293 min, 16 scans) E2502143.d




o-Xylene

+ EIC (91.1) Scan E2502143.d




+ Scan (13.684-13.765 min, 11 scans) E2502143.d





Name SPRSEA-5-S-20250710
Comment C57730
Data File E2502144.d
Acq. Date-Time 7/31/2025 2:59:53 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.052	8.052	502,560	
Benzene	benzene-d6 (IS)	8.110	8.110	37,330	
Toluene-d8 (IS)		10.746	10.753	563,313	
Toluene	Toluene-d8 (IS)	10.846	10.846	78,508	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	16,149	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	43,439	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	16,755	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502144.d




+ Scan (10.703-10.846 min, 21 scans) E2502144.d

**Toluene**


+ EIC (91.1) Scan E2502144.d




+ Scan (10.796-10.938 min, 19 scans) E2502144.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502144.d




+ Scan (12.993-13.107 min, 16 scans) E2502144.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502144.d




+ Scan (13.167-13.274 min, 15 scans) E2502144.d




o-Xylene

+ EIC (91.1) Scan E2502144.d




+ Scan (13.684-13.746 min, 8 scans) E2502144.d





Name SPRSEA-6-S-20250710
Comment C57534
Data File E2502145.d
Acq. Date-Time 7/31/2025 3:25:41 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	499,176	
Benzene	benzene-d6 (IS)	8.110	8.110	35,070	
Toluene-d8 (IS)		10.746	10.753	562,560	
Toluene	Toluene-d8 (IS)	10.846	10.846	67,890	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	17,550	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	47,097	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	16,075	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502145.d




+ Scan (10.703-10.867 min, 24 scans) E2502145.d

**Toluene**


+ EIC (91.1) Scan E2502145.d




+ Scan (10.810-10.918 min, 16 scans) E2502145.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502145.d




+ Scan (13.004-13.102 min, 14 scans) E2502145.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502145.d




+ Scan (13.174-13.267 min, 13 scans) E2502145.d




o-Xylene

+ EIC (91.1) Scan E2502145.d




+ Scan (13.683-13.745 min, 9 scans) E2502145.d





Name SPRSEA-6-D-20250710
Comment C01661
Data File E2502146.d
Acq. Date-Time 7/31/2025 3:51:27 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.052	8.052	499,746	
Benzene	benzene-d6 (IS)	8.110	8.110	43,769	
Toluene-d8 (IS)		10.746	10.753	558,483	
Toluene	Toluene-d8 (IS)	10.846	10.846	95,112	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	25,382	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	37,946	
o-Xylene	Toluene-d8 (IS)	13.718	13.718	13,584	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502146.d




+ Scan (10.703-10.879 min, 25 scans) E2502146.d

**Toluene**


+ EIC (91.1) Scan E2502146.d




+ Scan (10.803-10.939 min, 20 scans) E2502146.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502146.d




+ Scan (12.995-13.079 min, 12 scans) E2502146.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502146.d




+ Scan (13.177-13.274 min, 14 scans) E2502146.d




o-Xylene

+ EIC (91.1) Scan E2502146.d




+ Scan (13.685-13.747 min, 9 scans) E2502146.d





Name SPRSEA-7-S-20250710
Comment C55400
Data File E2502147.d
Acq. Date-Time 7/31/2025 4:17:11 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.052	8.052	497,648	
Benzene	benzene-d6 (IS)	8.109	8.110	37,961	
Toluene-d8 (IS)		10.753	10.753	562,706	
Toluene	Toluene-d8 (IS)	10.846	10.846	70,205	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	11,269	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	23,979	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	8,687	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502147.d




+ Scan (10.695-10.846 min, 22 scans) E2502147.d

**Toluene**


+ EIC (91.1) Scan E2502147.d




+ Scan (10.796-10.932 min, 20 scans) E2502147.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502147.d




+ Scan (12.995-13.070 min, 11 scans) E2502147.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502147.d




+ Scan (13.177-13.266 min, 12 scans) E2502147.d




o-Xylene

+ EIC (91.1) Scan E2502147.d




+ Scan (13.678-13.766 min, 12 scans) E2502147.d





Name SPRSEA-8-S-20250710
Comment C20391
Data File E2502148.d
Acq. Date-Time 7/31/2025 4:42:53 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	498,117	m
Benzene	benzene-d6 (IS)	8.110	8.110	55,780	
Toluene-d8 (IS)		10.746	10.753	570,289	
Toluene	Toluene-d8 (IS)	10.846	10.846	124,501	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	28,715	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	46,467	
o-Xylene	Toluene-d8 (IS)	13.718	13.718	19,105	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502148.d




+ Scan (10.703-10.882 min, 26 scans) E2502148.d

**Toluene**


+ EIC (91.1) Scan E2502148.d




+ Scan (10.803-10.932 min, 18 scans) E2502148.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502148.d




+ Scan (13.003-13.079 min, 10 scans) E2502148.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502148.d




+ Scan (13.181-13.281 min, 13 scans) E2502148.d




o-Xylene

+ EIC (91.1) Scan E2502148.d




+ Scan (13.680-13.761 min, 11 scans) E2502148.d





Name SPRSEA-9-S-20250710
Comment B14580
Data File E2502149.d
Acq. Date-Time 7/31/2025 5:08:47 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.052	8.052	496,100	
Benzene	benzene-d6 (IS)	8.110	8.110	105,303	
Toluene-d8 (IS)		10.753	10.753	565,042	
Toluene	Toluene-d8 (IS)	10.846	10.846	285,141	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	65,167	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	127,680	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	54,523	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502149.d




+ Scan (10.696-10.839 min, 21 scans) E2502149.d

**Toluene**


+ EIC (91.1) Scan E2502149.d




+ Scan (10.803-10.925 min, 18 scans) E2502149.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502149.d




+ Scan (12.995-13.108 min, 16 scans) E2502149.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502149.d




+ Scan (13.174-13.339 min, 24 scans) E2502149.d




o-Xylene

+ EIC (91.1) Scan E2502149.d




+ Scan (13.683-13.754 min, 10 scans) E2502149.d





Name SPRSEA-10-S-20250710
Comment C43380
Data File E2502151.d
Acq. Date-Time 7/31/2025 5:59:20 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.052	8.052	493,424	
Benzene	benzene-d6 (IS)	8.110	8.110	351,364	
Toluene-d8 (IS)		10.746	10.753	558,657	
Toluene	Toluene-d8 (IS)	10.846	10.846	1,452,224	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	256,833	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	731,948	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	295,436	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502151.d




+ Scan (10.703-10.832 min, 19 scans) E2502151.d

**Toluene**


+ EIC (91.1) Scan E2502151.d




+ Scan (10.796-10.932 min, 20 scans) E2502151.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502151.d




+ Scan (12.996-13.109 min, 16 scans) E2502151.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502151.d




+ Scan (13.167-13.339 min, 25 scans) E2502151.d




o-Xylene

+ EIC (91.1) Scan E2502151.d




+ Scan (13.677-13.810 min, 18 scans) E2502151.d





Name SPRSEA-11-S-20250710
Comment B12197
Data File E2502152.d
Acq. Date-Time 7/31/2025 6:25:57 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.052	8.052	498,565	
Benzene	benzene-d6 (IS)	8.110	8.110	117,131	
Toluene-d8 (IS)		10.746	10.753	563,851	
Toluene	Toluene-d8 (IS)	10.846	10.846	315,539	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	106,544	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	237,921	
o-Xylene	Toluene-d8 (IS)	13.718	13.718	76,097	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502152.d




+ Scan (10.703-10.867 min, 24 scans) E2502152.d

**Toluene**


+ EIC (91.1) Scan E2502152.d




+ Scan (10.796-10.932 min, 20 scans) E2502152.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502152.d




+ Scan (13.000-13.095 min, 14 scans) E2502152.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502152.d




+ Scan (13.174-13.360 min, 26 scans) E2502152.d




o-Xylene

+ EIC (91.1) Scan E2502152.d




+ Scan (13.663-13.802 min, 19 scans) E2502152.d





Name SPRSEA-12-S-20250710
Comment C36934
Data File E2502153.d
Acq. Date-Time 7/31/2025 6:51:36 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.052	8.052	501,332	
Benzene	benzene-d6 (IS)	8.110	8.110	78,548	
Toluene-d8 (IS)		10.746	10.753	569,892	
Toluene	Toluene-d8 (IS)	10.846	10.846	229,833	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	62,460	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	152,693	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	53,260	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502153.d




+ Scan (10.703-10.882 min, 26 scans) E2502153.d

**Toluene**


+ EIC (91.1) Scan E2502153.d




+ Scan (10.796-10.910 min, 17 scans) E2502153.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502153.d




+ Scan (13.002-13.085 min, 11 scans) E2502153.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502153.d




+ Scan (13.174-13.316 min, 20 scans) E2502153.d




o-Xylene

+ EIC (91.1) Scan E2502153.d




+ Scan (13.675-13.776 min, 15 scans) E2502153.d





Name SPRSEA-12-D-20250710
Comment C43377
Data File E2502154.d
Acq. Date-Time 7/31/2025 7:17:19 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.052	8.052	501,003	
Benzene	benzene-d6 (IS)	8.110	8.110	79,826	
Toluene-d8 (IS)		10.746	10.753	566,494	
Toluene	Toluene-d8 (IS)	10.846	10.846	251,735	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	72,541	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	201,268	
o-Xylene	Toluene-d8 (IS)	13.718	13.718	73,467	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502154.d




+ Scan (10.703-10.867 min, 24 scans) E2502154.d

**Toluene**


+ EIC (91.1) Scan E2502154.d




+ Scan (10.796-10.925 min, 19 scans) E2502154.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502154.d




+ Scan (13.002-13.090 min, 12 scans) E2502154.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502154.d




+ Scan (13.174-13.314 min, 20 scans) E2502154.d




o-Xylene

+ EIC (91.1) Scan E2502154.d




+ Scan (13.678-13.789 min, 15 scans) E2502154.d





Name SPRSEA-12-B-20250710
Comment C43932
Data File E2502155.d
Acq. Date-Time 7/31/2025 7:43:06 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	500,845	
Benzene	benzene-d6 (IS)	8.110	8.110	14,246	
Toluene-d8 (IS)		10.753	10.753	571,448	
Toluene	Toluene-d8 (IS)	10.846	10.846	11,564	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	2,384	
m-/p-Xylenes	Toluene-d8 (IS)	13.203	13.217	1,276	
o-Xylene	Toluene-d8 (IS)	13.718	13.718	449	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502155.d




+ Scan (10.696-10.839 min, 21 scans) E2502155.d

**Toluene**


+ EIC (91.1) Scan E2502155.d




+ Scan (10.810-10.882 min, 10 scans) E2502155.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502155.d




+ Scan (13.005-13.065 min, 8 scans) E2502155.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502155.d




+ Scan (13.189-13.258 min, 9 scans) E2502155.d




o-Xylene

+ EIC (91.1) Scan E2502155.d



+ Scan (13.701-13.737 min, 5 scans) E2502155.d



Initial Calibration



Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC302-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Calibration Curves

Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
E050525A_CC252679	Benzene	1	E2500043.d	4.94	48593	54.2	531168	1.004	0.14
E050525A_CC252679	Benzene	2	E2500044.d	10.06	88289	54.2	525272	0.905	0.032
E050525A_CC252679	Benzene	3	E2500045.d	20.12	167722	54.2	519365	0.869	-0.0086
E050525A_CC252679	Benzene	4	E2500046.d	40.25	314317	54.2	495891	0.853	-0.027
E050525A_CC252679	Benzene	5	E2500047.d	100.62	771973	54.2	516636	0.804	-0.083
E050525A_CC252679	Benzene	6	E2500048.d	201.24	1566630	54.2	517234	0.815	-0.07
E050525A_CC252679	Benzene	7	E2500049.d	603.73	5084416	54.2	514456	0.887	0.011
						Avg:	517146	0.877	
						%RSD:	2.1%	7.6%	
E050525A_CC252679	Toluene	1	E2500043.d	5.34	57611	63.2	569179	1.198	0.21
E050525A_CC252679	Toluene	2	E2500044.d	10.88	96506	63.2	555358	1.009	0.02
E050525A_CC252679	Toluene	3	E2500045.d	21.76	183454	63.2	550972	0.967	-0.023
E050525A_CC252679	Toluene	4	E2500046.d	43.52	328667	63.2	514484	0.927	-0.063
E050525A_CC252679	Toluene	5	E2500047.d	108.80	817722	63.2	541924	0.876	-0.11
E050525A_CC252679	Toluene	6	E2500048.d	217.60	1753197	63.2	536168	0.949	-0.04
E050525A_CC252679	Toluene	7	E2500049.d	652.80	5564266	63.2	538797	0.999	0.01
						Avg:	543840	0.989	
						%RSD:	3.2%	10.3%	
E050525A_CC252679	Ethylbenzene	1	E2500043.d	5.13	60169	63.2	569179	1.301	0.089
E050525A_CC252679	Ethylbenzene	2	E2500044.d	10.46	105835	63.2	555358	1.151	-0.037
E050525A_CC252679	Ethylbenzene	3	E2500045.d	20.92	208215	63.2	550972	1.141	-0.045
E050525A_CC252679	Ethylbenzene	4	E2500046.d	41.83	392499	63.2	514484	1.152	-0.036
E050525A_CC252679	Ethylbenzene	5	E2500047.d	104.58	995735	63.2	541924	1.110	-0.071
E050525A_CC252679	Ethylbenzene	6	E2500048.d	209.16	2327735	63.2	536168	1.311	0.097
E050525A_CC252679	Ethylbenzene	7	E2500049.d	627.49	6406466	63.2	538797	1.197	0.0018
						Avg:	543840	1.195	
						%RSD:	3.2%	6.7%	
E050525A_CC252679	m-/p-Xylenes	1	E2500043.d	4.85	58394	63.2	569179	1.337	0.29
E050525A_CC252679	m-/p-Xylenes	2	E2500044.d	9.88	88399	63.2	555358	1.018	-0.019
E050525A_CC252679	m-/p-Xylenes	3	E2500045.d	19.75	172744	63.2	550972	1.003	-0.034
E050525A_CC252679	m-/p-Xylenes	4	E2500046.d	39.51	305031	63.2	514484	0.948	-0.086
E050525A_CC252679	m-/p-Xylenes	5	E2500047.d	98.77	730099	63.2	541924	0.862	-0.17
E050525A_CC252679	m-/p-Xylenes	6	E2500048.d	197.54	1739458	63.2	536168	1.037	-4.1E-05
E050525A_CC252679	m-/p-Xylenes	7	E2500049.d	592.63	5345328	63.2	538797	1.057	0.019
						Avg:	543840	1.037	
						%RSD:	3.2%	14.2%	
E050525A_CC252679	o-Xylene	1	E2500043.d	5.05	50277	63.2	569179	1.104	0.16
E050525A_CC252679	o-Xylene	2	E2500044.d	10.30	81187	63.2	555358	0.896	-0.056
E050525A_CC252679	o-Xylene	3	E2500045.d	20.60	161444	63.2	550972	0.898	-0.054

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC302-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Calibration Curves

Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
E050525A_CC252679	o-Xylene	4	E2500046.d	41.21	299918	63.2	514484	0.894	-0.059
E050525A_CC252679	o-Xylene	5	E2500047.d	103.02	797704	63.2	541924	0.903	-0.05
E050525A_CC252679	o-Xylene	6	E2500048.d	206.03	1692581	63.2	536168	0.968	0.019
E050525A_CC252679	o-Xylene	7	E2500049.d	618.10	5206672	63.2	538797	0.988	0.04
						Avg:	543840	0.950	
						%RSD:	3.2%	8.2%	

Calibration Curves

Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
E050525A_CC252679	Benzene	ICV	E2500050.d	64.90	531194	54.2	505187	0.878	0.1%
E050525A_CC252679	Toluene	ICV	E2500050.d	77.40	626119	63.2	533535	0.958	-3.2%
E050525A_CC252679	Ethylbenzene	ICV	E2500050.d	87.14	736386	63.2	533535	1.001	-16.0%
E050525A_CC252679	m-/p-Xylenes	ICV	E2500050.d	90.70	588321	63.2	533535	0.768	-26.0%
E050525A_CC252679	o-Xylene	ICV	E2500050.d	89.27	625693	63.2	533535	0.830	-13.0%

M325B PDF Report ver.20250630

Sample Custody






EPA Method 325 A/B
Field Test Data Sheet and
Chain of Custody Record

Page # 1 of # 1

- ☐ Standard Turn Around Time (10 business days)
- ☐ Rush Turn Around Time
- All TATs Subject to Approval by Enthalpy Analytical, Inc.
- Unless otherwise specified, sample tubes will be conditioned for re-use 3 business days after submission of results

Site Name: Sprague Seaport		Client Name: Montrose Air		PO#:					
Site Address: 70 Trundy Road		Project Number: # 027966		Sample Event #					
City: Seaport		Project Manager: Harg Brock		Sorbent:					
State: Maine		Email Address: hargbrock@montrose-env.com							
Zip: 04974		Telephone #: 207-441-0025							
Location	Sample ID (Tube ID)	Sample, Blank or Duplicate	Start Date	Start Time	Stop Date	Stop Time	Deployed/ Collected by	Ave. Pressure (inHg)	Avg. Ambient Temp. (°F)
1	C 33260	S	7/10/25	1150	7/24/25	1215	HFB		
2	B 17444	S		1140		1205			
3	C 56859	S		1130		1155			
4	C 38901	S		1120		1145			
5	C 57730	S		1220		1245			
6	C 57534	S		1230		1245			
6	C 01661	D		1230		1255			
6	C 00592	B		1230		1255			
7	C 55400	S		1240		1305			
8	C 20391	S		1250		1315			
9	B 14580	S		1300		1325			
10	C 43380	S		1310		1335			
11	B 12197	S		1210		1235			
12	C 36934	S		1200		1225			
12	C 43377	D		1200		1225			
12	C 43932	B	7/10/25	1200	7/24/25	1225	HFB		
Relinquished By (printed): Harg Brock			Relinquished By (signature): 			Relinquished Date: 7/24/2025		Relinquished Time: 1630	
Received By (printed): Daniel Simpson			Received By (signature): 			Receipt Date: 7/29/25		Receipt Time: 10:38 AM	
Sample Condition Upon Receipt: Good			Compound List:			Custody Seal intact? Y/N:		Delivery tracking #	
e Temp: 25.8			Blank Temp: Fuke 4			Add Custody Seal # below:			
Comments:									
800-1 Capitola Drive • Durham, NC 27713 • (919) 850-4382 • FAX (919) 850-9012 • www.enthalpy.com									

**This Is The Last Page
Of This Report.**

Sprague - Searsport

70 Trundy Road
Searsport, ME 04974

Sampling Event 27 Sprague - Searsport

Client Project# PROJ-027966

Samples Received: 8/11/2025

Analytical Report 2025GC303

EPA Method 325B Analysis

Report Issue Date: 8/19/2025

I certify that to the best of my knowledge all analytical data presented in this report have been checked for completeness, accuracy, errors and legibility in addition to having been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s). This report shall not be reproduced except in full without approval of the laboratory. This will provide assurance that parts of the report are not taken out of context.

Amendment(s):

Signature:


QA Review by Isabel Obando Marrero, Data Reviewer



Matt Cavanaugh
Matthew.Cavanaugh@enthalpy.com / www.enthalpy.com
O: (919) 850-4392
Enthalpy Analytical
800 Capitola Drive Suite 1 Durham, NC 27713

Table of Contents

Case Narrative	3
Results	6
Summary of Results	7
Detailed Results	8
QC Data	11
Chromatograms	15
Initial Calibration	67
Sample Custody	70
Chain of Custody	71

Narrative Summary



Enthalpy Analytical Narrative Summary

Company	Montrose Air Quality Services, LLC - New Jersey
Job No.	2025GC303-1
Client ID.	PROJ-027966 Site: Sprague - Searsport

1. Custody

The samples were received at Enthalpy Analytical on August 11, 2025 at 23.3 °C, which is above the method recommended 23.0 °C. The samples were received in good condition. Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC

Table 1 - Sample Inventory

Sample ID	Tube ID	Sample Type
SPRSEA-1-S-20250724	C60287	Sample
SPRSEA-2-S-20250724	B29752	Sample
SPRSEA-3-S-20250724	C60278	Sample
SPRSEA-4-S-20250724	C38893	Sample
SPRSEA-5-S-20250724	C01908	Sample
SPRSEA-6-S-20250724	C37428	Sample
SPRSEA-6-D-20250724	C57756	Duplicate
SPRSEA-6-B-20250724	C36887	Blank
SPRSEA-7-S-20250724	B47211	Sample
SPRSEA-8-S-20250724	C57437	Sample
SPRSEA-9-S-20250724	C32832	Sample
SPRSEA-10-S-20250724	C43550	Sample
SPRSEA-11-S-20250724	B50749	Sample
SPRSEA-12-S-20250724	C57454	Sample
SPRSEA-12-D-20250724	B42495	Duplicate
SPRSEA-12-B-20250724	C39256	Blank

2. Analysis

The samples were analyzed for Benzene, Toluene, Ethylbenzene, m-/p-Xylenes, and o-Xylene using EPA Method 325B – Volatile Organic Compounds from Fugitive and Area Sources by Thermal Desorption and GC/MS. A copy of the acquisition method M325B-MTD is not included in this report but may be available upon request.

The sample tube media used for this sampling period was CarbopackX. All calibration standards and laboratory QC were prepared using the same media.

3. Calibration

All BFB tune criteria have been met for this analysis.

The initial calibration (E050525A_CC252679_QT_CC185154) met all 30% RSD criteria. The initial calibration verification met 30% difference criteria. The continuing calibration verifications met 30% difference criteria. The initial and continuing calibration raw data are not included in this report but are available upon request.

Enthalpy Analytical Narrative Summary

Company	Montrose Air Quality Services, LLC - New Jersey
Job No.	2025GC303-1
Client ID.	PROJ-027966 Site: Sprague - Searsport

5. QC Notes

All quality control criteria required by the method and/or the laboratory SOP have been met unless noted otherwise below.

6. Reporting Notes

All tubes used for this sampling period met the method criteria for number of uses; no tube exceeded 50 field uses.

As specified in EPA Method 325B, the response factor of the daily continuing calibration standard was used to quantitate all field samples and blanks.

All samples were reported as amount in ng catch, and concentration in ug/m³ and ppbv.

The results presented in this report are representative of the samples as provided to the laboratory. These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, located at 800 Capitola Drive, Suite 1, Durham NC, 27713 is accredited by the Louisiana Department of Environmental Quality (LDEQ) for EPA Method 325B for all analytes included in this report under **Certificate Number 04010**.

Results

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC303-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Summary

Sample Code	Tube ID	Benzene		Toluene		Ethylbenzene		m-/p-Xylenes		o-Xylene	
		(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag
SPRSEA-1-S-20250724	C60287	0.786		1.64		0.405	J	1.56		0.536	J
SPRSEA-2-S-20250724	B29752	0.756		1.76		0.597		1.46		0.585	
SPRSEA-3-S-20250724	C60278	0.813		1.60		0.354	J	1.29		0.576	
SPRSEA-4-S-20250724	C38893	0.567		1.01		0.278	J	1.04		0.463	J
SPRSEA-5-S-20250724	C01908	0.657		1.03		0.332	J	0.749		0.352	J
SPRSEA-6-S-20250724	C37428	0.632		0.774		0.271	ND	0.373	J	0.271	ND
SPRSEA-6-D-20250724	C57756	0.511		0.647		0.271	ND	0.437	J	0.271	ND
SPRSEA-6-B-20250724	C36887	0.186	ND	0.240	ND	0.271	ND	0.271	ND	0.271	ND
SPRSEA-7-S-20250724	B47211	0.679		0.609		0.271	ND	0.271	ND	0.271	ND
SPRSEA-8-S-20250724	C57437	0.541		0.662		0.271	ND	0.449	J	0.271	ND
SPRSEA-9-S-20250724	C32832	0.691		1.49		0.434	J	1.44		0.598	
SPRSEA-10-S-20250724	C43550	1.72		5.41		1.04		3.82		1.54	
SPRSEA-11-S-20250724	B50749	1.58		4.60		1.02		2.93		1.14	
SPRSEA-12-S-20250724	C57454	1.31		3.75		1.20		4.68		1.59	
SPRSEA-12-D-20250724	B42495	1.44		4.01		1.38		4.81		1.65	
SPRSEA-12-B-20250724	C39256	0.186	ND	0.240	ND	0.271	ND	0.271	ND	0.271	ND

J: Estimated Value - The analyte was detected between the Method Detection Limit and Reporting Limit

ND: The analyte was not present above the Method Detection Limit

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey
Job No.: 2025GC303-1 EPA Method 325B Analysis
Client No.: PROJ-027966 Site: Sprague - Searsport

Benzene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp (°F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250724	C60287	0.786	0.246	10.5	70.3	0.666	20145	0.186	0.368	0.0584	0.115		E2502495.d	2025-08-12 18:48	0.989	8.102	89852	466762	54.2	8.045	-5.7%
SPRSEA-2-S-20250724	B29752	0.756	0.237	10.1	70.3	0.666	20145	0.186	0.368	0.0584	0.115		E2502496.d	2025-08-12 19:13	0.989	8.103	86289	466439	54.2	8.045	-5.7%
SPRSEA-3-S-20250724	C60278	0.813	0.255	10.9	70.3	0.666	20145	0.186	0.368	0.0584	0.115		E2502497.d	2025-08-12 19:39	0.989	8.103	93107	467553	54.2	8.045	-5.5%
SPRSEA-4-S-20250724	C38893	0.567	0.177	7.60	70.3	0.666	20145	0.186	0.368	0.0584	0.115		E2502498.d	2025-08-12 20:06	0.989	8.102	64396	464250	54.2	8.045	-6.2%
SPRSEA-5-S-20250724	C01908	0.657	0.206	8.81	70.3	0.666	20145	0.186	0.368	0.0584	0.115		E2502499.d	2025-08-12 20:31	0.989	8.103	75515	469652	54.2	8.045	-5.1%
SPRSEA-6-S-20250724	C37428	0.632	0.198	8.48	70.3	0.666	20145	0.186	0.368	0.0584	0.115		E2502500.d	2025-08-12 20:57	0.989	8.103	72551	468780	54.2	8.045	-5.3%
SPRSEA-6-D-20250724	C57756	0.511	0.160	6.85	70.3	0.666	20145	0.186	0.368	0.0584	0.115		E2502501.d	2025-08-12 21:22	0.989	8.103	58982	471846	54.2	8.045	-4.6%
SPRSEA-6-B-20250724	C36887	0.186	0.0584		70.3	0.666	20145	0.186	0.368	0.0584	0.115	ND	E2502494.d	2025-08-12 18:22	0.989	8.103	9849	468184	54.2	8.045	-5.4%
SPRSEA-7-S-20250724	B47211	0.679	0.213	9.11	70.3	0.666	20145	0.186	0.368	0.0584	0.115		E2502502.d	2025-08-12 21:48	0.989	8.103	77615	466560	54.2	8.045	-5.7%
SPRSEA-8-S-20250724	C57437	0.541	0.169	7.26	70.3	0.666	20145	0.186	0.368	0.0584	0.115		E2502503.d	2025-08-12 22:13	0.989	8.103	62138	469088	54.2	8.045	-5.2%
SPRSEA-9-S-20250724	C32832	0.691	0.217	9.27	70.3	0.666	20145	0.186	0.368	0.0584	0.115		E2502504.d	2025-08-12 22:39	0.989	8.102	78589	464342	54.2	8.045	-6.2%
SPRSEA-10-S-20250724	C43550	1.72	0.537	23.0	70.3	0.666	20145	0.186	0.368	0.0584	0.115		E2502506.d	2025-08-12 23:30	0.989	8.102	195817	466194	54.2	8.045	-5.8%
SPRSEA-11-S-20250724	B50749	1.58	0.495	21.2	70.3	0.666	20145	0.186	0.368	0.0584	0.115		E2502507.d	2025-08-12 23:56	0.989	8.103	171140	442184	54.2	8.045	-10.6%
SPRSEA-12-S-20250724	C57454	1.31	0.410	17.6	70.3	0.666	20145	0.186	0.368	0.0584	0.115		E2502508.d	2025-08-13 00:22	0.989	8.103	143196	446611	54.2	8.045	-9.7%
SPRSEA-12-D-20250724	B42495	1.44	0.450	19.3	70.3	0.666	20145	0.186	0.368	0.0584	0.115		E2502509.d	2025-08-13 00:48	0.989	8.103	155291	441386	54.2	8.045	-10.8%
SPRSEA-12-B-20250724	C39256	0.186	0.0584		70.3	0.666	20145	0.186	0.368	0.0584	0.115	ND	E2502510.d	2025-08-13 01:14	0.989	8.102	8913	453022	54.2	8.045	-8.4%

Toluene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp (°F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250724	C60287	1.64	0.436	17.1	70.3	0.517	20145	0.240	0.513	0.0638	0.136		E2502495.d	2025-08-12 18:48	1.152	10.839	167137	535948	63.2	10.746	-3.7%
SPRSEA-2-S-20250724	B29752	1.76	0.468	18.3	70.3	0.517	20145	0.240	0.513	0.0638	0.136		E2502496.d	2025-08-12 19:13	1.152	10.839	179110	535962	63.2	10.746	-3.7%
SPRSEA-3-S-20250724	C60278	1.60	0.426	16.7	70.3	0.517	20145	0.240	0.513	0.0638	0.136		E2502497.d	2025-08-12 19:39	1.152	10.846	163508	537387	63.2	10.746	-3.4%
SPRSEA-4-S-20250724	C38893	1.01	0.268	10.5	70.3	0.517	20145	0.240	0.513	0.0638	0.136		E2502498.d	2025-08-12 20:06	1.152	10.839	103515	539815	63.2	10.746	-3.0%
SPRSEA-5-S-20250724	C01908	1.03	0.275	10.8	70.3	0.517	20145	0.240	0.513	0.0638	0.136		E2502499.d	2025-08-12 20:31	1.152	10.839	104902	534156	63.2	10.746	-4.0%
SPRSEA-6-S-20250724	C37428	0.774	0.206	8.06	70.3	0.517	20145	0.240	0.513	0.0638	0.136		E2502500.d	2025-08-12 20:57	1.152	10.846	79514	541089	63.2	10.746	-2.8%
SPRSEA-6-D-20250724	C57756	0.647	0.172	6.73	70.3	0.517	20145	0.240	0.513	0.0638	0.136		E2502501.d	2025-08-12 21:22	1.152	10.839	65466	533175	63.2	10.746	-4.2%
SPRSEA-6-B-20250724	C36887	0.240	0.0638		70.3	0.517	20145	0.240	0.513	0.0638	0.136	ND	E2502494.d	2025-08-12 18:22	1.152	10.839	17169	541980	63.2	10.746	-2.6%
SPRSEA-7-S-20250724	B47211	0.609	0.162	6.34	70.3	0.517	20145	0.240	0.513	0.0638	0.136		E2502502.d	2025-08-12 21:48	1.152	10.846	62285	539076	63.2	10.746	-3.1%
SPRSEA-8-S-20250724	C57437	0.662	0.176	6.89	70.3	0.517	20145	0.240	0.513	0.0638	0.136		E2502503.d	2025-08-12 22:13	1.152	10.839	66860	532190	63.2	10.746	-4.4%
SPRSEA-9-S-20250724	C32832	1.49	0.396	15.5	70.3	0.517	20145	0.240	0.513	0.0638	0.136		E2502504.d	2025-08-12 22:39	1.152	10.839	150508	532248	63.2	10.746	-4.4%
SPRSEA-10-S-20250724	C43550	5.41	1.44	56.3	70.3	0.517	20145	0.240	0.513	0.0638	0.136		E2502506.d	2025-08-12 23:30	1.152	10.846	513767	500671	63.2	10.746	-10.0%
SPRSEA-11-S-20250724	B50749	4.60	1.22	47.9	70.3	0.517	20145	0.240	0.513	0.0638	0.136		E2502507.d	2025-08-12 23:56	1.152	10.846	435947	499157	63.2	10.746	-10.3%
SPRSEA-12-S-20250724	C57454	3.75	0.997	39.1	70.3	0.517	20145	0.240	0.513	0.0638	0.136		E2502508.d	2025-08-13 00:22	1.152	10.846	360569	505973	63.2	10.746	-9.1%
SPRSEA-12-D-20250724	B42495	4.01	1.06	41.7	70.3	0.517	20145	0.240	0.513	0.0638	0.136		E2502509.d	2025-08-13 00:48	1.152	10.839	382728	503386	63.2	10.746	-9.5%
SPRSEA-12-B-20250724	C39256	0.240	0.0638		70.3	0.517	20145	0.240	0.513	0.0638	0.136	ND	E2502510.d	2025-08-13 01:14	1.152	10.846	12717	508298	63.2	10.745	-8.7%

Ethylbenzene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp (°F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250724	C60287	0.405	0.0932	3.73	70.3	0.457	20145	0.271	0.557	0.0626	0.128	J	E2502495.d	2025-08-12 18:48	1.207	13.031	38145	535948	63.2	10.746	-3.7%

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey
Job No.: 2025GC303-1 EPA Method 325B Analysis
Client No.: PROJ-027966 Site: Sprague - Searsport

Ethylbenzene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-2-S-20250724	B29752	0.597	0.138	5.50	70.3	0.457	20145	0.271	0.557	0.0626	0.128		E2502496.d	2025-08-12 19:13	1.207	13.031	56307	535962	63.2	10.746	-3.7%
SPRSEA-3-S-20250724	C60278	0.354	0.0817	3.26	70.3	0.457	20145	0.271	0.557	0.0626	0.128	J	E2502497.d	2025-08-12 19:39	1.207	13.031	33508	537387	63.2	10.746	-3.4%
SPRSEA-4-S-20250724	C38893	0.278	0.0640	2.56	70.3	0.457	20145	0.271	0.557	0.0626	0.128	J	E2502498.d	2025-08-12 20:06	1.207	13.031	26358	539815	63.2	10.746	-3.0%
SPRSEA-5-S-20250724	C01908	0.332	0.0764	3.05	70.3	0.457	20145	0.271	0.557	0.0626	0.128	J	E2502499.d	2025-08-12 20:31	1.207	13.031	31175	534156	63.2	10.746	-4.0%
SPRSEA-6-S-20250724	C37428	0.271	0.0626		70.3	0.457	20145	0.271	0.557	0.0626	0.128	ND	E2502500.d	2025-08-12 20:57	1.207	13.031	19171	541089	63.2	10.746	-2.8%
SPRSEA-6-D-20250724	C57756	0.271	0.0626		70.3	0.457	20145	0.271	0.557	0.0626	0.128	ND	E2502501.d	2025-08-12 21:22	1.207	13.031	14068	533175	63.2	10.746	-4.2%
SPRSEA-6-B-20250724	C36887	0.271	0.0626		70.3	0.457	20145	0.271	0.557	0.0626	0.128	ND	E2502494.d	2025-08-12 18:22	1.207	13.038	4979	541980	63.2	10.746	-2.6%
SPRSEA-7-S-20250724	B47211	0.271	0.0626		70.3	0.457	20145	0.271	0.557	0.0626	0.128	ND	E2502502.d	2025-08-12 21:48	1.207	13.038	10480	539076	63.2	10.746	-3.1%
SPRSEA-8-S-20250724	C57437	0.271	0.0626		70.3	0.457	20145	0.271	0.557	0.0626	0.128	ND	E2502503.d	2025-08-12 22:13	1.207	13.038	14293	532190	63.2	10.746	-4.4%
SPRSEA-9-S-20250724	C32832	0.434	0.100	4.00	70.3	0.457	20145	0.271	0.557	0.0626	0.128	J	E2502504.d	2025-08-12 22:39	1.207	13.031	40670	532248	63.2	10.746	-4.4%
SPRSEA-10-S-20250724	C43550	1.04	0.239	9.55	70.3	0.457	20145	0.271	0.557	0.0626	0.128		E2502506.d	2025-08-12 23:30	1.207	13.038	91320	500671	63.2	10.746	-10.0%
SPRSEA-11-S-20250724	B50749	1.02	0.235	9.41	70.3	0.457	20145	0.271	0.557	0.0626	0.128		E2502507.d	2025-08-12 23:56	1.207	13.031	89698	499157	63.2	10.746	-10.3%
SPRSEA-12-S-20250724	C57454	1.20	0.277	11.1	70.3	0.457	20145	0.271	0.557	0.0626	0.128		E2502508.d	2025-08-13 00:22	1.207	13.031	107035	505973	63.2	10.746	-9.1%
SPRSEA-12-D-20250724	B42495	1.38	0.317	12.7	70.3	0.457	20145	0.271	0.557	0.0626	0.128		E2502509.d	2025-08-13 00:48	1.207	13.031	121865	503386	63.2	10.746	-9.5%
SPRSEA-12-B-20250724	C39256	0.271	0.0626		70.3	0.457	20145	0.271	0.557	0.0626	0.128	ND	E2502510.d	2025-08-13 01:14	1.207	13.038	2854	508298	63.2	10.745	-8.7%

m-/p-Xylenes

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250724	C60287	1.56	0.359	14.4	70.3	0.457	20145	0.271	0.526	0.0626	0.121		E2502495.d	2025-08-12 18:48	0.913	13.210	111234	535948	63.2	10.746	-3.7%
SPRSEA-2-S-20250724	B29752	1.46	0.337	13.5	70.3	0.457	20145	0.271	0.526	0.0626	0.121		E2502496.d	2025-08-12 19:13	0.913	13.210	104398	535962	63.2	10.746	-3.7%
SPRSEA-3-S-20250724	C60278	1.29	0.298	11.9	70.3	0.457	20145	0.271	0.526	0.0626	0.121		E2502497.d	2025-08-12 19:39	0.913	13.210	92483	537387	63.2	10.746	-3.4%
SPRSEA-4-S-20250724	C38893	1.04	0.239	9.55	70.3	0.457	20145	0.271	0.526	0.0626	0.121		E2502498.d	2025-08-12 20:06	0.913	13.210	74488	539815	63.2	10.746	-3.0%
SPRSEA-5-S-20250724	C01908	0.749	0.172	6.89	70.3	0.457	20145	0.271	0.526	0.0626	0.121		E2502499.d	2025-08-12 20:31	0.913	13.210	53214	534156	63.2	10.746	-4.0%
SPRSEA-6-S-20250724	C37428	0.373	0.0860	3.44	70.3	0.457	20145	0.271	0.526	0.0626	0.121	J	E2502500.d	2025-08-12 20:57	0.913	13.210	26866	541089	63.2	10.746	-2.8%
SPRSEA-6-D-20250724	C57756	0.437	0.101	4.02	70.3	0.457	20145	0.271	0.526	0.0626	0.121	J	E2502501.d	2025-08-12 21:22	0.913	13.210	30980	533175	63.2	10.746	-4.2%
SPRSEA-6-B-20250724	C36887	0.271	0.0626		70.3	0.457	20145	0.271	0.526	0.0626	0.121	ND	E2502494.d	2025-08-12 18:22	0.913	13.217	902	541980	63.2	10.746	-2.6%
SPRSEA-7-S-20250724	B47211	0.271	0.0626		70.3	0.457	20145	0.271	0.526	0.0626	0.121	ND	E2502502.d	2025-08-12 21:48	0.913	13.210	17352	539076	63.2	10.746	-3.1%
SPRSEA-8-S-20250724	C57437	0.449	0.104	4.14	70.3	0.457	20145	0.271	0.526	0.0626	0.121	J	E2502503.d	2025-08-12 22:13	0.913	13.210	31821	532190	63.2	10.746	-4.4%
SPRSEA-9-S-20250724	C32832	1.44	0.331	13.2	70.3	0.457	20145	0.271	0.526	0.0626	0.121		E2502504.d	2025-08-12 22:39	0.913	13.210	101716	532248	63.2	10.746	-4.4%
SPRSEA-10-S-20250724	C43550	3.82	0.880	35.2	70.3	0.457	20145	0.271	0.526	0.0626	0.121		E2502506.d	2025-08-12 23:30	0.913	13.210	254566	500671	63.2	10.746	-10.0%
SPRSEA-11-S-20250724	B50749	2.93	0.675	27.0	70.3	0.457	20145	0.271	0.526	0.0626	0.121		E2502507.d	2025-08-12 23:56	0.913	13.210	194454	499157	63.2	10.746	-10.3%
SPRSEA-12-S-20250724	C57454	4.68	1.08	43.1	70.3	0.457	20145	0.271	0.526	0.0626	0.121		E2502508.d	2025-08-13 00:22	0.913	13.210	315196	505973	63.2	10.746	-9.1%
SPRSEA-12-D-20250724	B42495	4.81	1.11	44.3	70.3	0.457	20145	0.271	0.526	0.0626	0.121		E2502509.d	2025-08-13 00:48	0.913	13.210	322353	503386	63.2	10.746	-9.5%
SPRSEA-12-B-20250724	C39256	0.271	0.0626		70.3	0.457	20145	0.271	0.526	0.0626	0.121	ND	E2502510.d	2025-08-13 01:14	0.913	13.217	632	508298	63.2	10.745	-8.7%

o-Xylene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250724	C60287	0.536	0.123	4.93	70.3	0.457	20145	0.271	0.549	0.0626	0.126	J	E2502495.d	2025-08-12 18:48	0.902	13.711	37762	535948	63.2	10.746	-3.7%
SPRSEA-2-S-20250724	B29752	0.585	0.135	5.39	70.3	0.457	20145	0.271	0.549	0.0626	0.126		E2502496.d	2025-08-12 19:13	0.902	13.711	41227	535962	63.2	10.746	-3.7%

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC303-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

o-Xylene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-3-S-20250724	C60278	0.576	0.133	5.31	70.3	0.457	20145	0.271	0.549	0.0626	0.126		E2502497.d	2025-08-12 19:39	0.902	13.711	40736	537387	63.2	10.746	-3.4%
SPRSEA-4-S-20250724	C38893	0.463	0.107	4.27	70.3	0.457	20145	0.271	0.549	0.0626	0.126	J	E2502498.d	2025-08-12 20:06	0.902	13.711	32879	539815	63.2	10.746	-3.0%
SPRSEA-5-S-20250724	C01908	0.352	0.0811	3.24	70.3	0.457	20145	0.271	0.549	0.0626	0.126	J	E2502499.d	2025-08-12 20:31	0.902	13.711	24735	534156	63.2	10.746	-4.0%
SPRSEA-6-S-20250724	C37428	0.271	0.0626		70.3	0.457	20145	0.271	0.549	0.0626	0.126	ND	E2502500.d	2025-08-12 20:57	0.902	13.711	10998	541089	63.2	10.746	-2.8%
SPRSEA-6-D-20250724	C57756	0.271	0.0626		70.3	0.457	20145	0.271	0.549	0.0626	0.126	ND	E2502501.d	2025-08-12 21:22	0.902	13.711	13019	533175	63.2	10.746	-4.2%
SPRSEA-6-B-20250724	C36887	0.271	0.0626		70.3	0.457	20145	0.271	0.549	0.0626	0.126	ND	E2502494.d	2025-08-12 18:22	0.902	13.718	577	541980	63.2	10.746	-2.6%
SPRSEA-7-S-20250724	B47211	0.271	0.0626		70.3	0.457	20145	0.271	0.549	0.0626	0.126	ND	E2502502.d	2025-08-12 21:48	0.902	13.711	6497	539076	63.2	10.746	-3.1%
SPRSEA-8-S-20250724	C57437	0.271	0.0626		70.3	0.457	20145	0.271	0.549	0.0626	0.126	ND	E2502503.d	2025-08-12 22:13	0.902	13.711	12372	532248	63.2	10.746	-4.4%
SPRSEA-9-S-20250724	C32832	0.598	0.138	5.51	70.3	0.457	20145	0.271	0.549	0.0626	0.126		E2502504.d	2025-08-12 22:39	0.902	13.711	41868	532248	63.2	10.746	-4.4%
SPRSEA-10-S-20250724	C43550	1.54	0.354	14.1	70.3	0.457	20145	0.271	0.549	0.0626	0.126		E2502506.d	2025-08-12 23:30	0.902	13.711	101070	500671	63.2	10.746	-10.0%
SPRSEA-11-S-20250724	B50749	1.14	0.263	10.5	70.3	0.457	20145	0.271	0.549	0.0626	0.126		E2502507.d	2025-08-12 23:56	0.902	13.711	74901	499157	63.2	10.746	-10.3%
SPRSEA-12-S-20250724	C57454	1.59	0.366	14.6	70.3	0.457	20145	0.271	0.549	0.0626	0.126		E2502508.d	2025-08-13 00:22	0.902	13.711	105627	505973	63.2	10.746	-9.1%
SPRSEA-12-D-20250724	B42495	1.65	0.380	15.2	70.3	0.457	20145	0.271	0.549	0.0626	0.126		E2502509.d	2025-08-13 00:48	0.902	13.711	109123	503386	63.2	10.746	-9.5%
SPRSEA-12-B-20250724	C39256	0.271	0.0626		70.3	0.457	20145	0.271	0.549	0.0626	0.126	ND	E2502510.d	2025-08-13 01:14	0.902	13.732	401	508298	63.2	10.745	-8.7%

J: Estimated Value - The analyte was detected between the Method Detection Limit and Reporting Limit

ND: The analyte was not present above the Method Detection Limit

QC Data



Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey
Job No.: 2025GC303-1 EPA Method 325B Analysis
Client No.: PROJ-027966 Site: Sprague - Searsport

QC Samples

Field Sample Type	Sample Code	Benzene		Toluene		Ethylbenzene		m-/p-Xylenes		o-Xylene	
Blanks (ug/m³)	SPRSEA-6-B-20250724	ND	Pass	ND	Pass	ND	Pass	ND	Pass	ND	Pass
	SPRSEA-12-B-20250724	ND	Pass	ND	Pass	ND	Pass	ND	Pass	ND	Pass
Duplicates (difference)	SPRSEA-6-D-20250724	21%	Pass	18%	Pass	ND	Pass	16%	Pass	ND	Pass
	SPRSEA-12-D-20250724	9.3%	Pass	6.5%	Pass	13%	Pass	2.8%	Pass	3.8%	Pass

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC303-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Benzene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	E2502469.d	C71464	Cal	0.989	0.877	0.989	13%	-4.3%		Pass	
M325B CCV 5	E2502492.d	C71469	Check	0.981	0.877	0.989	12%		-5.3%	Pass	
2025GC303 Method Blank-1	E2502493.d	B45080	Blank		0.877	0.989			-4.6%	Pass	ND
M325B CCV 5	E2502505.d	C00655	Check	0.990	0.877	0.989	13%		-9.1%	Pass	
M325B CCV 5	E2502511.d	C34191	Check	0.978	0.877	0.989	12%		-8.7%	Pass	

Toluene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	E2502469.d	C71464	Cal	1.152	0.989	1.152	16%	2.3%		Pass	
M325B CCV 5	E2502492.d	C71469	Check	1.125	0.989	1.152	14%		-2.8%	Pass	
2025GC303 Method Blank-1	E2502493.d	B45080	Blank		0.989	1.152			-2.4%	Pass	ND
M325B CCV 5	E2502505.d	C00655	Check	1.157	0.989	1.152	17%		-9.8%	Pass	
M325B CCV 5	E2502511.d	C34191	Check	1.122	0.989	1.152	13%		-8.0%	Pass	

Ethylbenzene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	E2502469.d	C71464	Cal	1.207	1.195	1.207	1.0%	2.3%		Pass	
M325B CCV 5	E2502492.d	C71469	Check	1.225	1.195	1.207	2.5%		-2.8%	Pass	
2025GC303 Method Blank-1	E2502493.d	B45080	Blank		1.195	1.207			-2.4%	Pass	ND
M325B CCV 5	E2502505.d	C00655	Check	1.343	1.195	1.207	12%		-9.8%	Pass	
M325B CCV 5	E2502511.d	C34191	Check	1.157	1.195	1.207	-3.1%		-8.0%	Pass	

m-/p-Xylenes Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	E2502469.d	C71464	Cal	0.913	1.037	0.913	-12%	2.3%		Pass	
M325B CCV 5	E2502492.d	C71469	Check	0.924	1.037	0.913	-11%		-2.8%	Pass	
2025GC303 Method Blank-1	E2502493.d	B45080	Blank		1.037	0.913			-2.4%	Pass	ND
M325B CCV 5	E2502505.d	C00655	Check	1.024	1.037	0.913	-1.3%		-9.8%	Pass	
M325B CCV 5	E2502511.d	C34191	Check	0.850	1.037	0.913	-18%		-8.0%	Pass	

o-Xylene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	E2502469.d	C71464	Cal	0.902	0.950	0.902	-5.1%	2.3%		Pass	
M325B CCV 5	E2502492.d	C71469	Check	0.918	0.950	0.902	-3.3%		-2.8%	Pass	
2025GC303 Method Blank-1	E2502493.d	B45080	Blank		0.950	0.902			-2.4%	Pass	ND

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey
Job No.: 2025GC303-1 EPA Method 325B Analysis
Client No.: PROJ-027966 Site: Sprague - Searsport


o-Xylene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	E2502505.d	C00655	Check	0.968	0.950	0.902	1.9%		-9.8%	Pass	
M325B CCV 5	E2502511.d	C34191	Check	0.860	0.950	0.902	-9.4%		-8.0%	Pass	

Chromatograms



Name 2025GC303 Method Blank-1
Comment B45080
Data File E2502493.d
Acq. Date-Time 8/12/2025 5:57:07 PM
Acq. Method File M325B-MTD
Tube Sorbent Carboxen
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	472,098	
Benzene	benzene-d6 (IS)	8.102	8.110	4,812	
Toluene-d8 (IS)		10.746	10.753	543,198	
Toluene	Toluene-d8 (IS)	10.846	10.846	7,031	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	1,910	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	1,642	
o-Xylene	Toluene-d8 (IS)	13.718	13.718	1,201	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502493.d




+ Scan (10.695-10.839 min, 21 scans) E2502493.d

**Toluene**


+ EIC (91.1) Scan E2502493.d




+ Scan (10.803-10.882 min, 12 scans) E2502493.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502493.d




+ Scan (13.007-13.063 min, 8 scans) E2502493.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502493.d




+ Scan (13.179-13.252 min, 10 scans) E2502493.d




o-Xylene

+ EIC (91.1) Scan E2502493.d




+ Scan (13.690-13.741 min, 7 scans) E2502493.d





Name SPRSEA-6-B-20250724
Comment C36887
Data File E2502494.d
Acq. Date-Time 8/12/2025 6:22:44 PM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	468,184	
Benzene	benzene-d6 (IS)	8.103	8.110	9,849	
Toluene-d8 (IS)		10.746	10.753	541,980	
Toluene	Toluene-d8 (IS)	10.839	10.846	17,169	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	4,979	
m-/p-Xylenes	Toluene-d8 (IS)	13.217	13.217	902	
o-Xylene	Toluene-d8 (IS)	13.718	13.718	577	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502494.d




+ Scan (10.696-10.839 min, 21 scans) E2502494.d

**Toluene**


+ EIC (91.1) Scan E2502494.d




+ Scan (10.810-10.889 min, 11 scans) E2502494.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502494.d




+ Scan (13.004-13.065 min, 8 scans) E2502494.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502494.d




+ Scan (13.196-13.260 min, 10 scans) E2502494.d




o-Xylene

+ EIC (91.1) Scan E2502494.d




+ Scan (13.686-13.738 min, 7 scans) E2502494.d





Name SPRSEA-1-S-20250724
Comment C60287
Data File E2502495.d
Acq. Date-Time 8/12/2025 6:48:17 PM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	466,762	
Benzene	benzene-d6 (IS)	8.102	8.110	89,852	
Toluene-d8 (IS)		10.746	10.753	535,948	
Toluene	Toluene-d8 (IS)	10.839	10.846	167,137	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	38,145	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	111,234	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	37,762	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502495.d




+ Scan (10.703-10.832 min, 19 scans) E2502495.d

**Toluene**


+ EIC (91.1) Scan E2502495.d




+ Scan (10.803-10.889 min, 13 scans) E2502495.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502495.d




+ Scan (13.002-13.074 min, 10 scans) E2502495.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502495.d




+ Scan (13.174-13.277 min, 14 scans) E2502495.d




o-Xylene

+ EIC (91.1) Scan E2502495.d




+ Scan (13.677-13.783 min, 15 scans) E2502495.d





Name SPRSEA-2-S-20250724
Comment B29752
Data File E2502496.d
Acq. Date-Time 8/12/2025 7:13:53 PM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	466,439	
Benzene	benzene-d6 (IS)	8.103	8.110	86,289	
Toluene-d8 (IS)		10.746	10.753	535,962	
Toluene	Toluene-d8 (IS)	10.839	10.846	179,110	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	56,307	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	104,398	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	41,227	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502496.d




+ Scan (10.703-10.832 min, 19 scans) E2502496.d

**Toluene**


+ EIC (91.1) Scan E2502496.d




+ Scan (10.796-10.910 min, 17 scans) E2502496.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502496.d




+ Scan (12.990-13.087 min, 13 scans) E2502496.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502496.d




+ Scan (13.169-13.289 min, 17 scans) E2502496.d




o-Xylene

+ EIC (91.1) Scan E2502496.d




+ Scan (13.679-13.754 min, 10 scans) E2502496.d





Name SPRSEA-3-S-20250724
Comment C60278
Data File E2502497.d
Acq. Date-Time 8/12/2025 7:39:29 PM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	467,553	
Benzene	benzene-d6 (IS)	8.103	8.110	93,107	
Toluene-d8 (IS)		10.746	10.753	537,387	
Toluene	Toluene-d8 (IS)	10.846	10.846	163,508	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	33,508	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	92,483	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	40,736	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502497.d




+ Scan (10.703-10.825 min, 18 scans) E2502497.d

**Toluene**


+ EIC (91.1) Scan E2502497.d




+ Scan (10.803-10.937 min, 18 scans) E2502497.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502497.d




+ Scan (13.003-13.068 min, 9 scans) E2502497.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502497.d




+ Scan (13.160-13.276 min, 16 scans) E2502497.d




o-Xylene

+ EIC (91.1) Scan E2502497.d




+ Scan (13.675-13.761 min, 13 scans) E2502497.d





Name SPRSEA-4-S-20250724
Comment C38893
Data File E2502498.d
Acq. Date-Time 8/12/2025 8:06:01 PM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	464,250	
Benzene	benzene-d6 (IS)	8.102	8.110	64,396	
Toluene-d8 (IS)		10.746	10.753	539,815	
Toluene	Toluene-d8 (IS)	10.839	10.846	103,515	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	26,358	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	74,488	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	32,879	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502498.d




+ Scan (10.703-10.839 min, 20 scans) E2502498.d

**Toluene**


+ EIC (91.1) Scan E2502498.d




+ Scan (10.798-10.897 min, 14 scans) E2502498.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502498.d




+ Scan (12.995-13.074 min, 12 scans) E2502498.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502498.d




+ Scan (13.175-13.274 min, 14 scans) E2502498.d




o-Xylene

+ EIC (91.1) Scan E2502498.d




+ Scan (13.679-13.774 min, 13 scans) E2502498.d





Name SPRSEA-5-S-20250724
Comment C01908
Data File E2502499.d
Acq. Date-Time 8/12/2025 8:31:41 PM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	469,652	m
Benzene	benzene-d6 (IS)	8.103	8.110	75,515	
Toluene-d8 (IS)		10.746	10.753	534,156	
Toluene	Toluene-d8 (IS)	10.839	10.846	104,902	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	31,175	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	53,214	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	24,735	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502499.d




+ Scan (10.703-10.832 min, 19 scans) E2502499.d

**Toluene**


+ EIC (91.1) Scan E2502499.d




+ Scan (10.796-10.909 min, 16 scans) E2502499.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502499.d




+ Scan (12.997-13.083 min, 12 scans) E2502499.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502499.d




+ Scan (13.170-13.289 min, 17 scans) E2502499.d




o-Xylene

+ EIC (91.1) Scan E2502499.d




+ Scan (13.673-13.760 min, 12 scans) E2502499.d





Name SPRSEA-6-S-20250724
Comment C37428
Data File E2502500.d
Acq. Date-Time 8/12/2025 8:57:11 PM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	468,780	
Benzene	benzene-d6 (IS)	8.103	8.110	72,551	
Toluene-d8 (IS)		10.746	10.753	541,089	
Toluene	Toluene-d8 (IS)	10.846	10.846	79,514	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	19,171	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	26,866	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	10,998	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502500.d




+ Scan (10.703-10.825 min, 18 scans) E2502500.d

**Toluene**


+ EIC (91.1) Scan E2502500.d




+ Scan (10.789-10.907 min, 17 scans) E2502500.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502500.d




+ Scan (13.002-13.088 min, 12 scans) E2502500.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502500.d




+ Scan (13.169-13.267 min, 14 scans) E2502500.d




o-Xylene

+ EIC (91.1) Scan E2502500.d




+ Scan (13.668-13.754 min, 12 scans) E2502500.d





Name SPRSEA-6-D-20250724
Comment C57756
Data File E2502501.d
Acq. Date-Time 8/12/2025 9:22:44 PM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	471,846	
Benzene	benzene-d6 (IS)	8.103	8.110	58,982	
Toluene-d8 (IS)		10.746	10.753	533,175	
Toluene	Toluene-d8 (IS)	10.839	10.846	65,466	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	14,068	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	30,980	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	13,019	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502501.d




+ Scan (10.696-10.832 min, 20 scans) E2502501.d

**Toluene**


+ EIC (91.1) Scan E2502501.d




+ Scan (10.803-10.881 min, 11 scans) E2502501.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502501.d




+ Scan (13.002-13.083 min, 12 scans) E2502501.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502501.d




+ Scan (13.176-13.271 min, 13 scans) E2502501.d




o-Xylene

+ EIC (91.1) Scan E2502501.d




+ Scan (13.680-13.740 min, 9 scans) E2502501.d





Name SPRSEA-7-S-20250724
Comment B47211
Data File E2502502.d
Acq. Date-Time 8/12/2025 9:48:21 PM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	466,560	
Benzene	benzene-d6 (IS)	8.103	8.110	77,615	
Toluene-d8 (IS)		10.746	10.753	539,076	
Toluene	Toluene-d8 (IS)	10.846	10.846	62,285	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	10,480	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	17,352	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	6,497	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502502.d




+ Scan (10.703-10.825 min, 18 scans) E2502502.d

**Toluene**


+ EIC (91.1) Scan E2502502.d




+ Scan (10.791-10.908 min, 16 scans) E2502502.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502502.d




+ Scan (12.996-13.090 min, 13 scans) E2502502.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502502.d




+ Scan (13.167-13.266 min, 14 scans) E2502502.d




o-Xylene

+ EIC (91.1) Scan E2502502.d




+ Scan (13.683-13.747 min, 8 scans) E2502502.d





Name SPRSEA-8-S-20250724
Comment C57437
Data File E2502503.d
Acq. Date-Time 8/12/2025 10:13:58 PM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	469,088	
Benzene	benzene-d6 (IS)	8.103	8.110	62,138	
Toluene-d8 (IS)		10.746	10.753	532,190	
Toluene	Toluene-d8 (IS)	10.839	10.846	66,860	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	14,293	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	31,821	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	12,372	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502503.d




+ Scan (10.703-10.832 min, 19 scans) E2502503.d

**Toluene**


+ EIC (91.1) Scan E2502503.d




+ Scan (10.799-10.903 min, 15 scans) E2502503.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502503.d




+ Scan (13.004-13.072 min, 9 scans) E2502503.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502503.d




+ Scan (13.174-13.272 min, 14 scans) E2502503.d




o-Xylene

+ EIC (91.1) Scan E2502503.d




+ Scan (13.681-13.746 min, 9 scans) E2502503.d





Name SPRSEA-9-S-20250724
Comment C32832
Data File E2502504.d
Acq. Date-Time 8/12/2025 10:39:38 PM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	464,342	
Benzene	benzene-d6 (IS)	8.102	8.110	78,589	
Toluene-d8 (IS)		10.746	10.753	532,248	
Toluene	Toluene-d8 (IS)	10.839	10.846	150,508	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	40,670	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	101,716	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	41,868	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502504.d




+ Scan (10.695-10.824 min, 19 scans) E2502504.d

**Toluene**


+ EIC (91.1) Scan E2502504.d




+ Scan (10.796-10.918 min, 18 scans) E2502504.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502504.d




+ Scan (12.990-13.085 min, 13 scans) E2502504.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502504.d




+ Scan (13.176-13.303 min, 18 scans) E2502504.d




o-Xylene

+ EIC (91.1) Scan E2502504.d




+ Scan (13.678-13.749 min, 10 scans) E2502504.d





Name SPRSEA-10-S-20250724
Comment C43550
Data File E2502506.d
Acq. Date-Time 8/12/2025 11:30:46 PM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	466,194	
Benzene	benzene-d6 (IS)	8.102	8.110	195,817	
Toluene-d8 (IS)		10.746	10.753	500,671	
Toluene	Toluene-d8 (IS)	10.846	10.846	513,767	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	91,320	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	254,566	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	101,070	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502506.d




+ Scan (10.703-10.824 min, 18 scans) E2502506.d

**Toluene**


+ EIC (91.1) Scan E2502506.d




+ Scan (10.789-10.932 min, 21 scans) E2502506.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502506.d




+ Scan (12.991-13.073 min, 11 scans) E2502506.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502506.d




+ Scan (13.171-13.308 min, 19 scans) E2502506.d




o-Xylene

+ EIC (91.1) Scan E2502506.d




+ Scan (13.674-13.783 min, 16 scans) E2502506.d





Name SPRSEA-11-S-20250724
Comment B50749
Data File E2502507.d
Acq. Date-Time 8/12/2025 11:56:20 PM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	442,184	
Benzene	benzene-d6 (IS)	8.103	8.110	171,140	
Toluene-d8 (IS)		10.746	10.753	499,157	
Toluene	Toluene-d8 (IS)	10.846	10.846	435,947	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	89,698	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	194,454	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	74,901	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502507.d




+ Scan (10.703-10.825 min, 18 scans) E2502507.d

**Toluene**


+ EIC (91.1) Scan E2502507.d




+ Scan (10.789-10.925 min, 20 scans) E2502507.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502507.d




+ Scan (12.989-13.099 min, 15 scans) E2502507.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502507.d




+ Scan (13.167-13.339 min, 24 scans) E2502507.d




o-Xylene

+ EIC (91.1) Scan E2502507.d




+ Scan (13.675-13.766 min, 13 scans) E2502507.d





Name SPRSEA-12-S-20250724
Comment C57454
Data File E2502508.d
Acq. Date-Time 8/13/2025 12:22:48 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	446,611	
Benzene	benzene-d6 (IS)	8.103	8.110	143,196	
Toluene-d8 (IS)		10.746	10.753	505,973	
Toluene	Toluene-d8 (IS)	10.846	10.846	360,569	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	107,035	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	315,196	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	105,627	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502508.d




+ Scan (10.703-10.817 min, 17 scans) E2502508.d

**Toluene**


+ EIC (91.1) Scan E2502508.d




+ Scan (10.791-10.925 min, 19 scans) E2502508.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502508.d




+ Scan (12.988-13.095 min, 16 scans) E2502508.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502508.d




+ Scan (13.175-13.308 min, 18 scans) E2502508.d




o-Xylene

+ EIC (91.1) Scan E2502508.d




+ Scan (13.675-13.790 min, 16 scans) E2502508.d





Name SPRSEA-12-D-20250724
Comment B42495
Data File E2502509.d
Acq. Date-Time 8/13/2025 12:48:26 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	441,386	
Benzene	benzene-d6 (IS)	8.103	8.110	155,291	
Toluene-d8 (IS)		10.746	10.753	503,386	
Toluene	Toluene-d8 (IS)	10.839	10.846	382,728	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	121,865	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	322,353	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	109,123	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502509.d




+ Scan (10.696-10.839 min, 20 scans) E2502509.d

**Toluene**


+ EIC (91.1) Scan E2502509.d




+ Scan (10.796-10.968 min, 25 scans) E2502509.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502509.d




+ Scan (12.998-13.080 min, 11 scans) E2502509.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502509.d




+ Scan (13.170-13.339 min, 23 scans) E2502509.d




o-Xylene

+ EIC (91.1) Scan E2502509.d




+ Scan (13.668-13.754 min, 13 scans) E2502509.d





Name SPRSEA-12-B-20250724
Comment C39256
Data File E2502510.d
Acq. Date-Time 8/13/2025 1:14:03 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	453,022	
Benzene	benzene-d6 (IS)	8.102	8.110	8,913	
Toluene-d8 (IS)		10.745	10.753	508,298	
Toluene	Toluene-d8 (IS)	10.846	10.846	12,717	
Ethylbenzene	Toluene-d8 (IS)	13.038	13.038	2,854	
m-/p-Xylenes	Toluene-d8 (IS)	13.217	13.217	632	
o-Xylene	Toluene-d8 (IS)	13.732	13.718	401	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502510.d




+ Scan (10.702-10.831 min, 19 scans) E2502510.d

**Toluene**


+ EIC (91.1) Scan E2502510.d




+ Scan (10.811-10.891 min, 11 scans) E2502510.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502510.d




+ Scan (13.002-13.072 min, 10 scans) E2502510.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502510.d




+ Scan (13.190-13.231 min, 6 scans) E2502510.d




o-Xylene

+ EIC (91.1) Scan E2502510.d



+ Scan (13.697-13.739 min, 5 scans) E2502510.d



Initial Calibration



Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC303-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Calibration Curves

Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
E050525A_CC252679_QT_CC185154	Benzene	1	E2500043.d	4.94	48593	54.2	531168	1.004	0.14
E050525A_CC252679_QT_CC185154	Benzene	2	E2500044.d	10.06	88289	54.2	525272	0.905	0.032
E050525A_CC252679_QT_CC185154	Benzene	3	E2500045.d	20.12	167722	54.2	519365	0.869	-0.0086
E050525A_CC252679_QT_CC185154	Benzene	4	E2500046.d	40.25	314317	54.2	495891	0.853	-0.027
E050525A_CC252679_QT_CC185154	Benzene	5	E2500047.d	100.62	771973	54.2	516636	0.804	-0.083
E050525A_CC252679_QT_CC185154	Benzene	6	E2500048.d	201.24	1566630	54.2	517234	0.815	-0.07
E050525A_CC252679_QT_CC185154	Benzene	7	E2500049.d	603.73	5084416	54.2	514456	0.887	0.011
E050525A_CC252679_QT_CC185154	Benzene	QT		120.51		54.2			
						Avg:	517146	0.877	
						%RSD:	2.1%	7.6%	
E050525A_CC252679_QT_CC185154	Toluene	1	E2500043.d	5.34	57611	63.2	569179	1.198	0.21
E050525A_CC252679_QT_CC185154	Toluene	2	E2500044.d	10.88	96506	63.2	555358	1.009	0.02
E050525A_CC252679_QT_CC185154	Toluene	3	E2500045.d	21.76	183454	63.2	550972	0.967	-0.023
E050525A_CC252679_QT_CC185154	Toluene	4	E2500046.d	43.52	328667	63.2	514484	0.927	-0.063
E050525A_CC252679_QT_CC185154	Toluene	5	E2500047.d	108.80	817722	63.2	541924	0.876	-0.11
E050525A_CC252679_QT_CC185154	Toluene	6	E2500048.d	217.60	1753197	63.2	536168	0.949	-0.04
E050525A_CC252679_QT_CC185154	Toluene	7	E2500049.d	652.80	5564266	63.2	538797	0.999	0.01
E050525A_CC252679_QT_CC185154	Toluene	QT		105.84		63.2			
						Avg:	543840	0.989	
						%RSD:	3.2%	10.3%	
E050525A_CC252679_QT_CC185154	Ethylbenzene	1	E2500043.d	5.13	60169	63.2	569179	1.301	0.089
E050525A_CC252679_QT_CC185154	Ethylbenzene	2	E2500044.d	10.46	105835	63.2	555358	1.151	-0.037
E050525A_CC252679_QT_CC185154	Ethylbenzene	3	E2500045.d	20.92	208215	63.2	550972	1.141	-0.045
E050525A_CC252679_QT_CC185154	Ethylbenzene	4	E2500046.d	41.83	392499	63.2	514484	1.152	-0.036
E050525A_CC252679_QT_CC185154	Ethylbenzene	5	E2500047.d	104.58	995735	63.2	541924	1.110	-0.071
E050525A_CC252679_QT_CC185154	Ethylbenzene	6	E2500048.d	209.16	2327735	63.2	536168	1.311	0.097
E050525A_CC252679_QT_CC185154	Ethylbenzene	7	E2500049.d	627.49	6406466	63.2	538797	1.197	0.0018
E050525A_CC252679_QT_CC185154	Ethylbenzene	QT		110.01		63.2			
						Avg:	543840	1.195	
						%RSD:	3.2%	6.7%	

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC303-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Calibration Curves

Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	1	E2500043.d	4.85	58394	63.2	569179	1.337	0.29
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	2	E2500044.d	9.88	88399	63.2	555358	1.018	-0.019
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	3	E2500045.d	19.75	172744	63.2	550972	1.003	-0.034
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	4	E2500046.d	39.51	305031	63.2	514484	0.948	-0.086
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	5	E2500047.d	98.77	730099	63.2	541924	0.862	-0.17
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	6	E2500048.d	197.54	1739458	63.2	536168	1.037	-4.1E-05
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	7	E2500049.d	592.63	5345328	63.2	538797	1.057	0.019
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	QT		123.29		63.2			
						Avg:	543840	1.037	
						%RSD:	3.2%	14.2%	
E050525A_CC252679_QT_CC185154	o-Xylene	1	E2500043.d	5.05	50277	63.2	569179	1.104	0.16
E050525A_CC252679_QT_CC185154	o-Xylene	2	E2500044.d	10.30	81187	63.2	555358	0.896	-0.056
E050525A_CC252679_QT_CC185154	o-Xylene	3	E2500045.d	20.60	161444	63.2	550972	0.898	-0.054
E050525A_CC252679_QT_CC185154	o-Xylene	4	E2500046.d	41.21	299918	63.2	514484	0.894	-0.059
E050525A_CC252679_QT_CC185154	o-Xylene	5	E2500047.d	103.02	797704	63.2	541924	0.903	-0.05
E050525A_CC252679_QT_CC185154	o-Xylene	6	E2500048.d	206.03	1692581	63.2	536168	0.968	0.019
E050525A_CC252679_QT_CC185154	o-Xylene	7	E2500049.d	618.10	5206672	63.2	538797	0.988	0.04
E050525A_CC252679_QT_CC185154	o-Xylene	QT		114.65		63.2			
						Avg:	543840	0.950	
						%RSD:	3.2%	8.2%	

Calibration Curves

Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
E050525A_CC252679_QT_CC185154	Benzene	ICV	E2500050.d	64.90	531194	54.2	505187	0.878	0.1%
E050525A_CC252679_QT_CC185154	Toluene	ICV	E2500050.d	77.40	626119	63.2	533535	0.958	-3.2%
E050525A_CC252679_QT_CC185154	Ethylbenzene	ICV	E2500050.d	87.14	736386	63.2	533535	1.001	-16.0%
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	ICV	E2500050.d	90.70	588321	63.2	533535	0.768	-26.0%
E050525A_CC252679_QT_CC185154	o-Xylene	ICV	E2500050.d	89.27	625693	63.2	533535	0.830	-13.0%

M325B PDF Report ver.20250630

Sample Custody



EPA Method 325 A/B
Field Test Data Sheet and
Chain of Custody Record

Page # 1 of # 1

- ☐ Standard Turn Around Time (10 business days)
- ☐ Rush Turn Around Time
- All TATs Subject to Approval by Enthalpy Analytical, Inc.
- Unless otherwise specified, sample tubes will be conditioned for re-use 3 business days after submission of results

[illegible]

**This Is The Last Page
Of This Report.**

Sprague - Searsport

70 Trundy Road
Searsport, ME 04974

Sampling Event 28 Sprague - Searsport

Client Project# PROJ-027966

Samples Received: 8/25/2025

Analytical Report 2025GC304

EPA Method 325B Analysis

Report Issue Date: 9/4/2025

I certify that to the best of my knowledge all analytical data presented in this report have been checked for completeness, accuracy, errors and legibility in addition to having been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s). This report shall not be reproduced except in full without approval of the laboratory. This will provide assurance that parts of the report are not taken out of context.

Amendment(s):

Signature:



QA Review by Isabel Obando Marrero, Data Reviewer



Matt Cavanaugh
Matthew.Cavanaugh@enthalpy.com / www.enthalpy.com
O: (919) 850-4392
Enthalpy Analytical
800 Capitola Drive Suite 1 Durham, NC 27713

Table of Contents

Case Narrative	3
Results	6
Summary of Results	7
Detailed Results	8
QC Data	11
Chromatograms	15
Initial Calibration	67
Sample Custody	70
Chain of Custody	71

Narrative Summary



Enthalpy Analytical Narrative Summary

Company	Montrose Air Quality Services, LLC - New Jersey
Job No.	2025GC304-1
Client ID.	PROJ-027966 Site: Sprague - Searsport

1. Custody

The samples were received at Enthalpy Analytical on August 25, 2025 at 23.3 °C, which is above the method recommended 23.0 °C. The samples were received in good condition. Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC

Table 1 - Sample Inventory

Sample ID	Tube ID	Sample Type
SPRSEA-1-S-20250807	C43715	Sample
SPRSEA-2-S-20250807	C00562	Sample
SPRSEA-3-S-20250807	C20428	Sample
SPRSEA-4-S-20250807	C01992	Sample
SPRSEA-5-S-20250807	C61499	Sample
SPRSEA-6-S-20250807	C01402	Sample
SPRSEA-6-D-20250807	B20636	Duplicate
SPRSEA-6-B-20250807	B15753	Blank
SPRSEA-7-S-20250807	C39271	Sample
SPRSEA-8-S-20250807	B14681	Sample
SPRSEA-9-S-20250807	B46315	Sample
SPRSEA-10-S-20250807	C37447	Sample
SPRSEA-11-S-20250807	C37014	Sample
SPRSEA-12-S-20250807	C57499	Sample
SPRSEA-12-D-20250807	C36957	Duplicate
SPRSEA-12-B-20250807	C40632	Blank

2. Analysis

The samples were analyzed for Benzene, Toluene, Ethylbenzene, m-/p-Xylenes, and o-Xylene using EPA Method 325B – Volatile Organic Compounds from Fugitive and Area Sources by Thermal Desorption and GC/MS. A copy of the acquisition method M325B-MTD is not included in this report but may be available upon request.

The sample tube media used for this sampling period was CarbopackX. All calibration standards and laboratory QC were prepared using the same media.

3. Calibration

All BFB tune criteria have been met for this analysis.

The initial calibration (E050525A_CC252679_QT_CC185154) met all 30% RSD criteria. The initial calibration verification met 30% difference criteria. The continuing calibration verifications met 30% difference criteria. The initial and continuing calibration raw data are not included in this report but are available upon request.

Enthalpy Analytical Narrative Summary

Company	Montrose Air Quality Services, LLC - New Jersey
Job No.	2025GC304-1
Client ID.	PROJ-027966 Site: Sprague - Searsport

5. QC Notes

All quality control criteria required by the method and/or the laboratory SOP have been met unless noted otherwise below.

The primary sample SPRSEA-12-S-20250807 (tube ID C57499) and its corresponding duplicate SPRSEA-12-D-20250807 (tube ID C36957) failed to meet the 30% difference criterion for Ethylbenzene as specified by the method. However, the concentrations of the analyte in both the sample and the duplicate were less than two times the reporting limit of the instrument's calibration curve. Therefore, the percent difference observed may not suggest the data set has been negatively affected. All samples in the data set have been flagged "Pc" for Ethylbenzene to denote this failure.

The primary sample SPRSEA-6-S-20250807 (tube ID C01402) and its corresponding duplicate SPRSEA-6-D-20250807 (tube ID B20636) failed to meet the 30% difference criterion for Ethylbenzene as specified by the method. However, the concentrations of the analyte in both the sample and the duplicate were less than two times the reporting limit of the instrument's calibration curve. Therefore, the percent difference observed may not suggest the data set has been negatively affected. All samples in the data set have been flagged "Pc" for Ethylbenzene to denote this failure.

6. Reporting Notes

All tubes used for this sampling period met the method criteria for number of uses; no tube exceeded 50 field uses.

As specified in EPA Method 325B, the response factor of the daily continuing calibration standard was used to quantitate all field samples and blanks.

All samples were reported as amount in ng catch, and concentration in ug/m³ and ppbv.

The results presented in this report are representative of the samples as provided to the laboratory. These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, located at 800 Capitola Drive, Suite 1, Durham NC, 27713 is accredited by the Louisiana Department of Environmental Quality (LDEQ) for EPA Method 325B for all analytes included in this report under **Certificate Number 04010**.

Results

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC304-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Summary

Sample Code	Tube ID	Benzene		Toluene		Ethylbenzene		m-/p-Xylenes		o-Xylene	
		(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag
SPRSEA-1-S-20250807	C43715	0.501		1.39		0.289	J,Pc	0.711		0.276	J
SPRSEA-2-S-20250807	C00562	0.713		2.13		0.480	J,Pc	1.52		0.567	
SPRSEA-3-S-20250807	C20428	0.721		2.11		0.500	J,Pc	1.18		0.428	J
SPRSEA-4-S-20250807	C01992	0.352	J	0.745		0.271	ND,Pc	0.584		0.271	ND
SPRSEA-5-S-20250807	C61499	0.332	J	0.695		0.271	ND,Pc	0.683		0.271	ND
SPRSEA-6-S-20250807	C01402	0.363	J	0.668		0.271	ND,Pc	0.342	J	0.271	ND
SPRSEA-6-D-20250807	B20636	0.366	J	0.782		0.483	J,Pc	0.412	J	0.271	ND
SPRSEA-6-B-20250807	B15753	0.186	ND	0.240	ND	0.271	ND,Pc	0.271	ND	0.271	ND
SPRSEA-7-S-20250807	C39271	0.245	J	0.394	J	0.271	ND,Pc	0.271	ND	0.271	ND
SPRSEA-8-S-20250807	B14681	0.518		0.566		0.271	ND,Pc	0.271	ND	0.271	ND
SPRSEA-9-S-20250807	B46315	0.556		1.68		0.533	J,Pc	1.25		0.545	J
SPRSEA-10-S-20250807	C37447	1.88		7.59		1.66	Pc	5.35		2.11	
SPRSEA-11-S-20250807	C37014	2.37		9.93		1.91	Pc	6.09		2.30	
SPRSEA-12-S-20250807	C57499	1.31		4.29		0.764	Pc	2.06		0.788	
SPRSEA-12-D-20250807	C36957	1.21		4.15		1.09	Pc	2.58		0.927	
SPRSEA-12-B-20250807	C40632	0.186	ND	0.312	J	0.271	ND,Pc	0.271	ND	0.271	ND

J: Estimated Value - The analyte was detected between the Method Detection Limit and Reporting Limit

ND: The analyte was not present above the Method Detection Limit

Pc: Field duplicate(s) exceed 30%RPD. Concentrations of both samples in duplicate are near the reporting limit

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC304-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Benzene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp (°F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250807	C43715	0.501	0.157	6.73	69.6	0.665	20170	0.186	0.368	0.0583	0.115		E2502898.d	2025-08-26 01:47	1.030	8.102	65850	514952	54.2	8.045	6.8%
SPRSEA-2-S-20250807	C00562	0.713	0.223	9.57	69.6	0.665	20170	0.186	0.368	0.0583	0.115		E2502899.d	2025-08-26 02:13	1.030	8.102	92722	509514	54.2	8.045	5.6%
SPRSEA-3-S-20250807	C20428	0.721	0.226	9.67	69.6	0.665	20170	0.186	0.368	0.0583	0.115		E2502900.d	2025-08-26 02:38	1.030	8.102	94856	516046	54.2	8.045	7.0%
SPRSEA-4-S-20250807	C01992	0.352	0.110	4.72	69.7	0.665	20170	0.186	0.368	0.0583	0.115	J	E2502901.d	2025-08-26 03:05	1.030	8.103	46138	514023	54.2	8.045	6.6%
SPRSEA-5-S-20250807	C61499	0.332	0.104	4.46	69.6	0.665	20170	0.186	0.368	0.0583	0.115	J	E2502902.d	2025-08-26 03:31	1.030	8.103	44064	520171	54.2	8.045	7.8%
SPRSEA-6-S-20250807	C01402	0.363	0.114	4.87	69.6	0.665	20170	0.186	0.368	0.0583	0.115	J	E2502903.d	2025-08-26 03:56	1.030	8.102	48341	521834	54.2	8.045	8.2%
SPRSEA-6-D-20250807	B20636	0.366	0.115	4.91	69.6	0.665	20170	0.186	0.368	0.0583	0.115	J	E2502904.d	2025-08-26 04:22	1.030	8.103	49038	525789	54.2	8.045	9.0%
SPRSEA-6-B-20250807	B15753	0.186	0.0583		69.6	0.665	20170	0.186	0.368	0.0583	0.115	ND	E2502897.d	2025-08-26 01:22	1.030	8.102	13116	511459	54.2	8.045	6.0%
SPRSEA-7-S-20250807	C39271	0.245	0.0767	3.29	69.6	0.665	20170	0.186	0.368	0.0583	0.115	J	E2502905.d	2025-08-26 04:47	1.030	8.103	32771	524471	54.2	8.045	8.7%
SPRSEA-8-S-20250807	B14681	0.518	0.162	6.95	69.6	0.665	20170	0.186	0.368	0.0583	0.115		E2502906.d	2025-08-26 05:13	1.030	8.103	69513	526214	54.2	8.045	9.1%
SPRSEA-9-S-20250807	B46315	0.556	0.174	7.47	69.6	0.665	20170	0.186	0.368	0.0583	0.115		E2502907.d	2025-08-26 05:38	1.030	8.102	74571	525513	54.2	8.045	8.9%
SPRSEA-10-S-20250807	C37447	1.88	0.589	25.2	69.6	0.665	20170	0.186	0.368	0.0583	0.115		E2502916.d	2025-08-26 10:38	1.030	8.103	248622	518598	54.2	8.045	7.5%
SPRSEA-11-S-20250807	C37014	2.37	0.744	31.9	69.6	0.665	20170	0.186	0.368	0.0583	0.115		E2502917.d	2025-08-26 11:03	1.030	8.110	314323	519043	54.2	8.045	7.6%
SPRSEA-12-S-20250807	C57499	1.31	0.410	17.6	69.6	0.665	20170	0.186	0.368	0.0583	0.115		E2502918.d	2025-08-26 11:28	1.030	8.102	173197	518898	54.2	8.045	7.6%
SPRSEA-12-D-20250807	C36957	1.21	0.380	16.3	69.6	0.665	20170	0.186	0.368	0.0583	0.115		E2502919.d	2025-08-26 11:52	1.030	8.102	159734	515759	54.2	8.045	6.9%
SPRSEA-12-B-20250807	C40632	0.186	0.0583		69.6	0.665	20170	0.186	0.368	0.0583	0.115	ND	E2502920.d	2025-08-26 12:17	1.030	8.102	21343	522322	54.2	8.045	8.3%

Toluene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp (°F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250807	C43715	1.39	0.368	14.4	69.6	0.516	20170	0.240	0.513	0.0637	0.136		E2502898.d	2025-08-26 01:47	1.112	10.839	142953	562393	63.2	10.746	4.9%
SPRSEA-2-S-20250807	C00562	2.13	0.566	22.2	69.6	0.516	20170	0.240	0.513	0.0637	0.136		E2502899.d	2025-08-26 02:13	1.112	10.839	222381	568451	63.2	10.746	6.1%
SPRSEA-3-S-20250807	C20428	2.11	0.561	22.0	69.6	0.516	20170	0.240	0.513	0.0637	0.136		E2502900.d	2025-08-26 02:38	1.112	10.839	220381	568359	63.2	10.746	6.0%
SPRSEA-4-S-20250807	C01992	0.745	0.198	7.76	69.7	0.516	20170	0.240	0.513	0.0637	0.136		E2502901.d	2025-08-26 03:05	1.112	10.839	78079	571519	63.2	10.746	6.6%
SPRSEA-5-S-20250807	C61499	0.695	0.184	7.24	69.6	0.516	20170	0.240	0.513	0.0637	0.136		E2502902.d	2025-08-26 03:31	1.112	10.839	73305	575281	63.2	10.746	7.3%
SPRSEA-6-S-20250807	C01402	0.668	0.177	6.96	69.6	0.516	20170	0.240	0.513	0.0637	0.136		E2502903.d	2025-08-26 03:56	1.112	10.839	69902	570190	63.2	10.746	6.4%
SPRSEA-6-D-20250807	B20636	0.782	0.208	8.15	69.6	0.516	20170	0.240	0.513	0.0637	0.136		E2502904.d	2025-08-26 04:22	1.112	10.839	82232	573177	63.2	10.746	6.9%
SPRSEA-6-B-20250807	B15753	0.240	0.0637		69.6	0.516	20170	0.240	0.513	0.0637	0.136	ND	E2502897.d	2025-08-26 01:22	1.112	10.838	13542	561514	63.2	10.745	4.8%
SPRSEA-7-S-20250807	C39271	0.394	0.105	4.11	69.6	0.516	20170	0.240	0.513	0.0637	0.136	J	E2502905.d	2025-08-26 04:47	1.112	10.839	41287	571071	63.2	10.746	6.6%
SPRSEA-8-S-20250807	B14681	0.566	0.150	5.90	69.6	0.516	20170	0.240	0.513	0.0637	0.136		E2502906.d	2025-08-26 05:13	1.112	10.839	58952	567879	63.2	10.746	6.0%
SPRSEA-9-S-20250807	B46315	1.68	0.446	17.5	69.6	0.516	20170	0.240	0.513	0.0637	0.136		E2502907.d	2025-08-26 05:38	1.112	10.839	179652	582874	63.2	10.745	8.8%
SPRSEA-10-S-20250807	C37447	7.59	2.02	79.0	69.6	0.516	20170	0.240	0.513	0.0637	0.136		E2502916.d	2025-08-26 10:38	1.112	10.839	792536	569372	63.2	10.746	6.2%
SPRSEA-11-S-20250807	C37014	9.93	2.64	103	69.6	0.516	20170	0.240	0.513	0.0637	0.136		E2502917.d	2025-08-26 11:03	1.112	10.839	1036316	568737	63.2	10.746	6.1%
SPRSEA-12-S-20250807	C57499	4.29	1.14	44.7	69.6	0.516	20170	0.240	0.513	0.0637	0.136		E2502918.d	2025-08-26 11:28	1.112	10.846	445835	566495	63.2	10.746	5.7%
SPRSEA-12-D-20250807	C36957	4.15	1.10	43.2	69.6	0.516	20170	0.240	0.513	0.0637	0.136		E2502919.d	2025-08-26 11:52	1.112	10.846	438996	577249	63.2	10.746	7.7%
SPRSEA-12-B-20250807	C40632	0.312	0.0829	3.25	69.6	0.516	20170	0.240	0.513	0.0637	0.136	J	E2502920.d	2025-08-26 12:17	1.112	10.839	33067	577148	63.2	10.745	7.7%

Ethylbenzene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp (°F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250807	C43715	0.289	0.0665	2.66	69.6	0.457	20170	0.271	0.557	0.0625	0.128	J,Pc	E2502898.d	2025-08-26 01:47	1.134	13.031	26849	562393	63.2	10.746	4.9%

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey
Job No.: 2025GC304-1 EPA Method 325B Analysis
Client No.: PROJ-027966 Site: Sprague - Searsport

Ethylbenzene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-2-S-20250807	C00562	0.480	0.111	4.43	69.6	0.457	20170	0.271	0.557	0.0625	0.128	J,Pc	E2502899.d	2025-08-26 02:13	1.134	13.031	45178	568451	63.2	10.746	6.1%
SPRSEA-3-S-20250807	C20428	0.500	0.115	4.61	69.6	0.457	20170	0.271	0.557	0.0625	0.128	J,Pc	E2502900.d	2025-08-26 02:38	1.134	13.031	47018	568359	63.2	10.746	6.0%
SPRSEA-4-S-20250807	C01992	0.271	0.0625		69.7	0.457	20170	0.271	0.557	0.0625	0.128	ND,Pc	E2502901.d	2025-08-26 03:05	1.134	13.031	23362	571519	63.2	10.746	6.6%
SPRSEA-5-S-20250807	C61499	0.271	0.0625		69.6	0.457	20170	0.271	0.557	0.0625	0.128	ND,Pc	E2502902.d	2025-08-26 03:31	1.134	13.031	15565	575281	63.2	10.746	7.3%
SPRSEA-6-S-20250807	C01402	0.271	0.0625		69.6	0.457	20170	0.271	0.557	0.0625	0.128	ND,Pc	E2502903.d	2025-08-26 03:56	1.134	13.031	15311	570190	63.2	10.746	6.4%
SPRSEA-6-D-20250807	B20636	0.483	0.111	4.45	69.6	0.457	20170	0.271	0.557	0.0625	0.128	J,Pc	E2502904.d	2025-08-26 04:22	1.134	13.031	45838	573177	63.2	10.746	6.9%
SPRSEA-6-B-20250807	B15753	0.271	0.0625		69.6	0.457	20170	0.271	0.557	0.0625	0.128	ND,Pc	E2502897.d	2025-08-26 01:22	1.134	13.030	1918	561514	63.2	10.745	4.8%
SPRSEA-7-S-20250807	C39271	0.271	0.0625		69.6	0.457	20170	0.271	0.557	0.0625	0.128	ND,Pc	E2502905.d	2025-08-26 04:47	1.134	13.031	9670	571071	63.2	10.746	6.6%
SPRSEA-8-S-20250807	B14681	0.271	0.0625		69.6	0.457	20170	0.271	0.557	0.0625	0.128	ND,Pc	E2502906.d	2025-08-26 05:13	1.134	13.031	15370	567879	63.2	10.746	6.0%
SPRSEA-9-S-20250807	B46315	0.533	0.123	4.91	69.6	0.457	20170	0.271	0.557	0.0625	0.128	J,Pc	E2502907.d	2025-08-26 05:38	1.134	13.030	51396	582874	63.2	10.745	8.8%
SPRSEA-10-S-20250807	C37447	1.66	0.383	15.3	69.6	0.457	20170	0.271	0.557	0.0625	0.128	Pc	E2502916.d	2025-08-26 10:38	1.134	13.031	156544	569372	63.2	10.746	6.2%
SPRSEA-11-S-20250807	C37014	1.91	0.439	17.6	69.6	0.457	20170	0.271	0.557	0.0625	0.128	Pc	E2502917.d	2025-08-26 11:03	1.134	13.031	179281	568737	63.2	10.746	6.1%
SPRSEA-12-S-20250807	C57499	0.764	0.176	7.04	69.6	0.457	20170	0.271	0.557	0.0625	0.128	Pc	E2502918.d	2025-08-26 11:28	1.134	13.031	71629	566495	63.2	10.746	5.7%
SPRSEA-12-D-20250807	C36957	1.09	0.252	10.1	69.6	0.457	20170	0.271	0.557	0.0625	0.128	Pc	E2502919.d	2025-08-26 11:52	1.134	13.031	104237	577249	63.2	10.746	7.7%
SPRSEA-12-B-20250807	C40632	0.271	0.0625		69.6	0.457	20170	0.271	0.557	0.0625	0.128	ND,Pc	E2502920.d	2025-08-26 12:17	1.134	13.031	7878	577148	63.2	10.745	7.7%

m-/p-Xylenes

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250807	C43715	0.711	0.164	6.56	69.6	0.457	20170	0.271	0.526	0.0625	0.121		E2502898.d	2025-08-26 01:47	0.812	13.210	47363	562393	63.2	10.746	4.9%
SPRSEA-2-S-20250807	C00562	1.52	0.350	14.0	69.6	0.457	20170	0.271	0.526	0.0625	0.121		E2502899.d	2025-08-26 02:13	0.812	13.210	102086	568451	63.2	10.746	6.1%
SPRSEA-3-S-20250807	C20428	1.18	0.271	10.8	69.6	0.457	20170	0.271	0.526	0.0625	0.121		E2502900.d	2025-08-26 02:38	0.812	13.210	79101	568359	63.2	10.746	6.0%
SPRSEA-4-S-20250807	C01992	0.584	0.135	5.38	69.7	0.457	20170	0.271	0.526	0.0625	0.121		E2502901.d	2025-08-26 03:05	0.812	13.210	39509	571519	63.2	10.746	6.6%
SPRSEA-5-S-20250807	C61499	0.683	0.157	6.29	69.6	0.457	20170	0.271	0.526	0.0625	0.121		E2502902.d	2025-08-26 03:31	0.812	13.210	46511	575281	63.2	10.746	7.3%
SPRSEA-6-S-20250807	C01402	0.342	0.0789	3.16	69.6	0.457	20170	0.271	0.526	0.0625	0.121	J	E2502903.d	2025-08-26 03:56	0.812	13.210	23115	570190	63.2	10.746	6.4%
SPRSEA-6-D-20250807	B20636	0.412	0.0949	3.80	69.6	0.457	20170	0.271	0.526	0.0625	0.121	J	E2502904.d	2025-08-26 04:22	0.812	13.210	27955	573177	63.2	10.746	6.9%
SPRSEA-6-B-20250807	B15753	0.271	0.0625		69.6	0.457	20170	0.271	0.526	0.0625	0.121	ND	E2502897.d	2025-08-26 01:22	0.812	13.202	1491	561514	63.2	10.745	4.8%
SPRSEA-7-S-20250807	C39271	0.271	0.0625		69.6	0.457	20170	0.271	0.526	0.0625	0.121	ND	E2502905.d	2025-08-26 04:47	0.812	13.210	15001	571071	63.2	10.746	6.6%
SPRSEA-8-S-20250807	B14681	0.271	0.0625		69.6	0.457	20170	0.271	0.526	0.0625	0.121	ND	E2502906.d	2025-08-26 05:13	0.812	13.210	17118	567879	63.2	10.746	6.0%
SPRSEA-9-S-20250807	B46315	1.25	0.288	11.5	69.6	0.457	20170	0.271	0.526	0.0625	0.121		E2502907.d	2025-08-26 05:38	0.812	13.210	86142	582874	63.2	10.745	8.8%
SPRSEA-10-S-20250807	C37447	5.35	1.23	49.3	69.6	0.457	20170	0.271	0.526	0.0625	0.121		E2502916.d	2025-08-26 10:38	0.812	13.210	360398	569372	63.2	10.746	6.2%
SPRSEA-11-S-20250807	C37014	6.09	1.40	56.1	69.6	0.457	20170	0.271	0.526	0.0625	0.121		E2502917.d	2025-08-26 11:03	0.812	13.210	410052	568737	63.2	10.746	6.1%
SPRSEA-12-S-20250807	C57499	2.06	0.474	18.9	69.6	0.457	20170	0.271	0.526	0.0625	0.121		E2502918.d	2025-08-26 11:28	0.812	13.210	137864	566495	63.2	10.746	5.7%
SPRSEA-12-D-20250807	C36957	2.58	0.594	23.8	69.6	0.457	20170	0.271	0.526	0.0625	0.121		E2502919.d	2025-08-26 11:52	0.812	13.210	176245	577249	63.2	10.746	7.7%
SPRSEA-12-B-20250807	C40632	0.271	0.0625		69.6	0.457	20170	0.271	0.526	0.0625	0.121	ND	E2502920.d	2025-08-26 12:17	0.812	13.210	3689	577148	63.2	10.745	7.7%

o-Xylene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250807	C43715	0.276	0.0635	2.54	69.6	0.457	20170	0.271	0.549	0.0625	0.126	J	E2502898.d	2025-08-26 01:47	0.845	13.711	19111	562393	63.2	10.746	4.9%
SPRSEA-2-S-20250807	C00562	0.567	0.131	5.22	69.6	0.457	20170	0.271	0.549	0.0625	0.126		E2502899.d	2025-08-26 02:13	0.845	13.711	39696	568451	63.2	10.746	6.1%

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC304-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

o-Xylene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-3-S-20250807	C20428	0.428	0.0987	3.95	69.6	0.457	20170	0.271	0.549	0.0625	0.126	J	E2502900.d	2025-08-26 02:38	0.845	13.711	30015	568359	63.2	10.746	6.0%
SPRSEA-4-S-20250807	C01992	0.271	0.0625		69.7	0.457	20170	0.271	0.549	0.0625	0.126	ND	E2502901.d	2025-08-26 03:05	0.845	13.711	16923	571519	63.2	10.746	6.6%
SPRSEA-5-S-20250807	C61499	0.271	0.0625		69.6	0.457	20170	0.271	0.549	0.0625	0.126	ND	E2502902.d	2025-08-26 03:31	0.845	13.711	18625	575281	63.2	10.746	7.3%
SPRSEA-6-S-20250807	C01402	0.271	0.0625		69.6	0.457	20170	0.271	0.549	0.0625	0.126	ND	E2502903.d	2025-08-26 03:56	0.845	13.711	9642	570190	63.2	10.746	6.4%
SPRSEA-6-D-20250807	B20636	0.271	0.0625		69.6	0.457	20170	0.271	0.549	0.0625	0.126	ND	E2502904.d	2025-08-26 04:22	0.845	13.711	14018	573177	63.2	10.746	6.9%
SPRSEA-6-B-20250807	B15753	0.271	0.0625		69.6	0.457	20170	0.271	0.549	0.0625	0.126	ND	E2502897.d	2025-08-26 01:22	0.845	13.704	662	561514	63.2	10.745	4.8%
SPRSEA-7-S-20250807	C39271	0.271	0.0625		69.6	0.457	20170	0.271	0.549	0.0625	0.126	ND	E2502905.d	2025-08-26 04:47	0.845	13.711	6478	571071	63.2	10.746	6.6%
SPRSEA-8-S-20250807	B14681	0.271	0.0625		69.6	0.457	20170	0.271	0.549	0.0625	0.126	ND	E2502906.d	2025-08-26 05:13	0.845	13.711	8401	567879	63.2	10.746	6.0%
SPRSEA-9-S-20250807	B46315	0.545	0.126	5.02	69.6	0.457	20170	0.271	0.549	0.0625	0.126	J	E2502907.d	2025-08-26 05:38	0.845	13.711	39164	582874	63.2	10.745	8.8%
SPRSEA-10-S-20250807	C37447	2.11	0.486	19.4	69.6	0.457	20170	0.271	0.549	0.0625	0.126		E2502916.d	2025-08-26 10:38	0.845	13.711	147876	569372	63.2	10.746	6.2%
SPRSEA-11-S-20250807	C37014	2.30	0.529	21.2	69.6	0.457	20170	0.271	0.549	0.0625	0.126		E2502917.d	2025-08-26 11:03	0.845	13.711	161045	568737	63.2	10.746	6.1%
SPRSEA-12-S-20250807	C57499	0.788	0.182	7.26	69.6	0.457	20170	0.271	0.549	0.0625	0.126		E2502918.d	2025-08-26 11:28	0.845	13.711	54994	566495	63.2	10.746	5.7%
SPRSEA-12-D-20250807	C36957	0.927	0.214	8.54	69.6	0.457	20170	0.271	0.549	0.0625	0.126		E2502919.d	2025-08-26 11:52	0.845	13.711	65921	577249	63.2	10.746	7.7%
SPRSEA-12-B-20250807	C40632	0.271	0.0625		69.6	0.457	20170	0.271	0.549	0.0625	0.126	ND	E2502920.d	2025-08-26 12:17	0.845	13.711	1503	577148	63.2	10.745	7.7%

J: Estimated Value - The analyte was detected between the Method Detection Limit and Reporting Limit

ND: The analyte was not present above the Method Detection Limit

Pc: Field duplicate(s) exceed 30%RPD. Concentrations of both samples in duplicate are near the reporting limit

QC Data

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey
Job No.: 2025GC304-1 EPA Method 325B Analysis
Client No.: PROJ-027966 Site: Sprague - Searsport

QC Samples

Field Sample Type	Sample Code	Benzene		Toluene		Ethylbenzene		m-/p-Xylenes		o-Xylene	
Blanks (ug/m³)	SPRSEA-6-B-20250807	ND	Pass	ND	Pass	ND	Pass	ND	Pass	ND	Pass
	SPRSEA-12-B-20250807	ND	Pass	0.312	Pass	ND	Pass	ND	Pass	ND	Pass
Duplicates (difference)	SPRSEA-6-D-20250807	0.68%	Pass	16%	Pass	ND	Fail	18%	Pass	ND	Pass
	SPRSEA-12-D-20250807	7.5%	Pass	3.4%	Pass	35%	Fail	23%	Pass	16%	Pass

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC304-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Benzene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5 REC	E2502872.d	C68611	Cal	1.030		1.030	17%	-6.7%		Pass	
M325B CCV 5	E2502895.d	B27830	Check	0.973		1.030	11%		5.0%	Pass	
2025GC304 Method Blank-1	E2502896.d	C33495	Blank			1.030			6.5%	Pass	ND
M325B CCV 5	E2502908.d	C69795	Check	0.969		1.030	10%		9.2%	Pass	
M325B CCV 5 REC	E2502915.d	C70783	Check	0.987		1.030	13%		8.1%	Pass	
M325B CCV 5 REC	E2502921.d	B27830	Check	0.983		1.030	12%		6.9%	Pass	

Toluene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5 REC	E2502872.d	C68611	Cal	1.112		1.112	12%	-1.5%		Pass	
M325B CCV 5	E2502895.d	B27830	Check	1.129		1.112	14%		3.9%	Pass	
2025GC304 Method Blank-1	E2502896.d	C33495	Blank			1.112			5.3%	Pass	ND
M325B CCV 5	E2502908.d	C69795	Check	1.105		1.112	12%		9.4%	Pass	
M325B CCV 5 REC	E2502915.d	C70783	Check	1.133		1.112	15%		7.9%	Pass	
M325B CCV 5 REC	E2502921.d	B27830	Check	1.149		1.112	16%		5.4%	Pass	

Ethylbenzene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5 REC	E2502872.d	C68611	Cal	1.134		1.134	-5.1%	-1.5%		Pass	
M325B CCV 5	E2502895.d	B27830	Check	1.180		1.134	-1.2%		3.9%	Pass	
2025GC304 Method Blank-1	E2502896.d	C33495	Blank			1.134			5.3%	Pass	ND
M325B CCV 5	E2502908.d	C69795	Check	1.090		1.134	-8.8%		9.4%	Pass	
M325B CCV 5 REC	E2502915.d	C70783	Check	1.180		1.134	-1.2%		7.9%	Pass	
M325B CCV 5 REC	E2502921.d	B27830	Check	1.190		1.134	-0.40%		5.4%	Pass	

m-/p-Xylenes Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5 REC	E2502872.d	C68611	Cal	0.812		0.812	-22%	-1.5%		Pass	
M325B CCV 5	E2502895.d	B27830	Check	0.791		0.812	-24%		3.9%	Pass	
2025GC304 Method Blank-1	E2502896.d	C33495	Blank			0.812			5.3%	Pass	ND
M325B CCV 5	E2502908.d	C69795	Check	0.749		0.812	-28%		9.4%	Pass	
M325B CCV 5 REC	E2502915.d	C70783	Check	0.812		0.812	-22%		7.9%	Pass	
M325B CCV 5 REC	E2502921.d	B27830	Check	0.772		0.812	-26%		5.4%	Pass	

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey
Job No.: 2025GC304-1 EPA Method 325B Analysis
Client No.: PROJ-027966 Site: Sprague - Searsport


o-Xylene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5 REC	E2502872.d	C68611	Cal	0.845		0.845	-11%	-1.5%		Pass	
M325B CCV 5	E2502895.d	B27830	Check	0.760		0.845	-20%		3.9%	Pass	
2025GC304 Method Blank-1	E2502896.d	C33495	Blank			0.845			5.3%	Pass	ND
M325B CCV 5	E2502908.d	C69795	Check	0.770		0.845	-19%		9.4%	Pass	
M325B CCV 5 REC	E2502915.d	C70783	Check	0.820		0.845	-14%		7.9%	Pass	
M325B CCV 5 REC	E2502921.d	B27830	Check	0.753		0.845	-21%		5.4%	Pass	

Chromatograms



Name 2025GC304 Method Blank-1
Comment C33495
Data File E2502896.d
Acq. Date-Time 8/26/2025 12:56:26 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	513,970	
Benzene	benzene-d6 (IS)	8.103	8.110	6,081	
Toluene-d8 (IS)		10.746	10.753	564,477	
Toluene	Toluene-d8 (IS)	10.839	10.846	8,063	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	1,909	
m-/p-Xylenes	Toluene-d8 (IS)	13.203	13.217	1,989	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	986	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502896.d




+ Scan (10.696-10.817 min, 18 scans) E2502896.d

**Toluene**


+ EIC (91.1) Scan E2502896.d




+ Scan (10.810-10.882 min, 11 scans) E2502896.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502896.d




+ Scan (13.004-13.052 min, 6 scans) E2502896.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502896.d




+ Scan (13.178-13.252 min, 10 scans) E2502896.d




o-Xylene

+ EIC (91.1) Scan E2502896.d




+ Scan (13.692-13.732 min, 5 scans) E2502896.d





Name SPRSEA-6-B-20250807
Comment B15753
Data File E2502897.d
Acq. Date-Time 8/26/2025 1:22:00 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	511,459	
Benzene	benzene-d6 (IS)	8.102	8.110	13,116	
Toluene-d8 (IS)		10.745	10.753	561,514	
Toluene	Toluene-d8 (IS)	10.838	10.846	13,542	
Ethylbenzene	Toluene-d8 (IS)	13.030	13.038	1,918	
m-/p-Xylenes	Toluene-d8 (IS)	13.202	13.217	1,491	m
o-Xylene	Toluene-d8 (IS)	13.704	13.718	662	m

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502897.d




+ Scan (10.695-10.824 min, 19 scans) E2502897.d

**Toluene**


+ EIC (91.1) Scan E2502897.d




+ Scan (10.803-10.881 min, 11 scans) E2502897.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502897.d




+ Scan (13.004-13.076 min, 10 scans) E2502897.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502897.d




+ Scan (13.183-13.262 min, 11 scans) E2502897.d




o-Xylene

+ EIC (91.1) Scan E2502897.d




+ Scan (13.692-13.736 min, 6 scans) E2502897.d





Name SPRSEA-1-S-20250807
Comment C43715
Data File E2502898.d
Acq. Date-Time 8/26/2025 1:47:43 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	514,952	
Benzene	benzene-d6 (IS)	8.102	8.110	65,850	
Toluene-d8 (IS)		10.746	10.753	562,393	
Toluene	Toluene-d8 (IS)	10.839	10.846	142,953	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	26,849	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	47,363	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	19,111	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502898.d




+ Scan (10.696-10.824 min, 19 scans) E2502898.d

**Toluene**


+ EIC (91.1) Scan E2502898.d




+ Scan (10.796-10.917 min, 17 scans) E2502898.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502898.d




+ Scan (12.989-13.071 min, 11 scans) E2502898.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502898.d




+ Scan (13.167-13.281 min, 17 scans) E2502898.d




o-Xylene

+ EIC (91.1) Scan E2502898.d




+ Scan (13.683-13.765 min, 11 scans) E2502898.d





Name SPRSEA-2-S-20250807
Comment C00562
Data File E2502899.d
Acq. Date-Time 8/26/2025 2:13:17 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	509,514	
Benzene	benzene-d6 (IS)	8.102	8.110	92,722	
Toluene-d8 (IS)		10.746	10.753	568,451	
Toluene	Toluene-d8 (IS)	10.839	10.846	222,381	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	45,178	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	102,086	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	39,696	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502899.d




+ Scan (10.696-10.824 min, 19 scans) E2502899.d

**Toluene**


+ EIC (91.1) Scan E2502899.d




+ Scan (10.796-10.931 min, 19 scans) E2502899.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502899.d




+ Scan (12.995-13.077 min, 12 scans) E2502899.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502899.d




+ Scan (13.174-13.286 min, 15 scans) E2502899.d




o-Xylene

+ EIC (91.1) Scan E2502899.d




+ Scan (13.672-13.747 min, 10 scans) E2502899.d





Name SPRSEA-3-S-20250807
Comment C20428
Data File E2502900.d
Acq. Date-Time 8/26/2025 2:38:53 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	516,046	
Benzene	benzene-d6 (IS)	8.102	8.110	94,856	
Toluene-d8 (IS)		10.746	10.753	568,359	
Toluene	Toluene-d8 (IS)	10.839	10.846	220,381	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	47,018	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	79,101	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	30,015	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502900.d




+ Scan (10.703-10.824 min, 18 scans) E2502900.d

**Toluene**


+ EIC (91.1) Scan E2502900.d




+ Scan (10.796-10.925 min, 19 scans) E2502900.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502900.d




+ Scan (12.995-13.074 min, 11 scans) E2502900.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502900.d




+ Scan (13.160-13.272 min, 16 scans) E2502900.d




o-Xylene

+ EIC (91.1) Scan E2502900.d




+ Scan (13.671-13.754 min, 12 scans) E2502900.d





Name SPRSEA-4-S-20250807
Comment C01992
Data File E2502901.d
Acq. Date-Time 8/26/2025 3:05:22 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	514,023	
Benzene	benzene-d6 (IS)	8.103	8.110	46,138	
Toluene-d8 (IS)		10.746	10.753	571,519	
Toluene	Toluene-d8 (IS)	10.839	10.846	78,079	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	23,362	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	39,509	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	16,923	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502901.d




+ Scan (10.696-10.832 min, 20 scans) E2502901.d

**Toluene**


+ EIC (91.1) Scan E2502901.d




+ Scan (10.798-10.903 min, 14 scans) E2502901.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502901.d




+ Scan (12.996-13.074 min, 11 scans) E2502901.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502901.d




+ Scan (13.175-13.272 min, 13 scans) E2502901.d




o-Xylene

+ EIC (91.1) Scan E2502901.d




+ Scan (13.680-13.761 min, 11 scans) E2502901.d





Name SPRSEA-5-S-20250807
Comment C61499
Data File E2502902.d
Acq. Date-Time 8/26/2025 3:31:00 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	520,171	
Benzene	benzene-d6 (IS)	8.103	8.110	44,064	
Toluene-d8 (IS)		10.746	10.753	575,281	
Toluene	Toluene-d8 (IS)	10.839	10.846	73,305	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	15,565	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	46,511	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	18,625	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502902.d




+ Scan (10.696-10.832 min, 20 scans) E2502902.d

**Toluene**


+ EIC (91.1) Scan E2502902.d




+ Scan (10.797-10.890 min, 13 scans) E2502902.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502902.d




+ Scan (13.002-13.071 min, 9 scans) E2502902.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502902.d




+ Scan (13.177-13.276 min, 14 scans) E2502902.d




o-Xylene

+ EIC (91.1) Scan E2502902.d




+ Scan (13.668-13.753 min, 12 scans) E2502902.d





Name SPRSEA-6-S-20250807
Comment C01402
Data File E2502903.d
Acq. Date-Time 8/26/2025 3:56:37 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	521,834	
Benzene	benzene-d6 (IS)	8.102	8.110	48,341	
Toluene-d8 (IS)		10.746	10.753	570,190	
Toluene	Toluene-d8 (IS)	10.839	10.846	69,902	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	15,311	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	23,115	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	9,642	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502903.d




+ Scan (10.695-10.832 min, 20 scans) E2502903.d

**Toluene**


+ EIC (91.1) Scan E2502903.d




+ Scan (10.796-10.887 min, 13 scans) E2502903.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502903.d




+ Scan (12.998-13.074 min, 11 scans) E2502903.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502903.d




+ Scan (13.169-13.283 min, 16 scans) E2502903.d




o-Xylene

+ EIC (91.1) Scan E2502903.d




+ Scan (13.675-13.745 min, 10 scans) E2502903.d





Name SPRSEA-6-D-20250807
Comment B20636
Data File E2502904.d
Acq. Date-Time 8/26/2025 4:22:05 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	525,789	
Benzene	benzene-d6 (IS)	8.103	8.110	49,038	
Toluene-d8 (IS)		10.746	10.753	573,177	
Toluene	Toluene-d8 (IS)	10.839	10.846	82,232	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	45,838	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	27,955	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	14,018	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502904.d




+ Scan (10.696-10.832 min, 20 scans) E2502904.d

**Toluene**


+ EIC (91.1) Scan E2502904.d




+ Scan (10.800-10.889 min, 13 scans) E2502904.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502904.d




+ Scan (12.995-13.074 min, 11 scans) E2502904.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502904.d




+ Scan (13.177-13.259 min, 11 scans) E2502904.d




o-Xylene

+ EIC (91.1) Scan E2502904.d




+ Scan (13.677-13.746 min, 9 scans) E2502904.d





Name SPRSEA-7-S-20250807
Comment C39271
Data File E2502905.d
Acq. Date-Time 8/26/2025 4:47:43 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	524,471	
Benzene	benzene-d6 (IS)	8.103	8.110	32,771	
Toluene-d8 (IS)		10.746	10.753	571,071	
Toluene	Toluene-d8 (IS)	10.839	10.846	41,287	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	9,670	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	15,001	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	6,478	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502905.d




+ Scan (10.696-10.839 min, 21 scans) E2502905.d

**Toluene**


+ EIC (91.1) Scan E2502905.d




+ Scan (10.803-10.874 min, 10 scans) E2502905.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502905.d




+ Scan (13.002-13.078 min, 11 scans) E2502905.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502905.d




+ Scan (13.182-13.259 min, 10 scans) E2502905.d




o-Xylene

+ EIC (91.1) Scan E2502905.d




+ Scan (13.683-13.745 min, 9 scans) E2502905.d





Name SPRSEA-8-S-20250807
Comment B14681
Data File E2502906.d
Acq. Date-Time 8/26/2025 5:13:22 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	526,214	
Benzene	benzene-d6 (IS)	8.103	8.110	69,513	
Toluene-d8 (IS)		10.746	10.753	567,879	
Toluene	Toluene-d8 (IS)	10.839	10.846	58,952	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	15,370	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	17,118	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	8,401	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502906.d




+ Scan (10.696-10.839 min, 21 scans) E2502906.d

**Toluene**


+ EIC (91.1) Scan E2502906.d




+ Scan (10.796-10.896 min, 15 scans) E2502906.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502906.d




+ Scan (13.001-13.081 min, 12 scans) E2502906.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502906.d




+ Scan (13.182-13.260 min, 11 scans) E2502906.d




o-Xylene

+ EIC (91.1) Scan E2502906.d




+ Scan (13.668-13.743 min, 11 scans) E2502906.d





Name SPRSEA-9-S-20250807
Comment B46315
Data File E2502907.d
Acq. Date-Time 8/26/2025 5:38:56 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	525,513	
Benzene	benzene-d6 (IS)	8.102	8.110	74,571	
Toluene-d8 (IS)		10.745	10.753	582,874	
Toluene	Toluene-d8 (IS)	10.839	10.846	179,652	
Ethylbenzene	Toluene-d8 (IS)	13.030	13.038	51,396	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	86,142	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	39,164	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502907.d




+ Scan (10.695-10.831 min, 20 scans) E2502907.d

**Toluene**


+ EIC (91.1) Scan E2502907.d




+ Scan (10.796-10.895 min, 13 scans) E2502907.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502907.d




+ Scan (12.992-13.070 min, 11 scans) E2502907.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502907.d




+ Scan (13.174-13.273 min, 14 scans) E2502907.d




o-Xylene

+ EIC (91.1) Scan E2502907.d




+ Scan (13.668-13.763 min, 13 scans) E2502907.d





Name SPRSEA-10-S-20250807
Comment C37447; Recollect
Data File E2502916.d
Acq. Date-Time 8/26/2025 10:38:42 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	518,598	
Benzene	benzene-d6 (IS)	8.103	8.110	248,622	
Toluene-d8 (IS)		10.746	10.753	569,372	
Toluene	Toluene-d8 (IS)	10.839	10.846	792,536	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	156,544	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	360,398	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	147,876	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502916.d




+ Scan (10.696-10.839 min, 21 scans) E2502916.d

**Toluene**


+ EIC (91.1) Scan E2502916.d




+ Scan (10.796-10.961 min, 24 scans) E2502916.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502916.d




+ Scan (12.995-13.079 min, 12 scans) E2502916.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502916.d




+ Scan (13.166-13.295 min, 18 scans) E2502916.d




o-Xylene

+ EIC (91.1) Scan E2502916.d




+ Scan (13.675-13.783 min, 16 scans) E2502916.d





Name SPRSEA-11-S-20250807
Comment C37014; Recollect
Data File E2502917.d
Acq. Date-Time 8/26/2025 11:03:16 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	519,043	
Benzene	benzene-d6 (IS)	8.110	8.110	314,323	
Toluene-d8 (IS)		10.746	10.753	568,737	
Toluene	Toluene-d8 (IS)	10.839	10.846	1,036,316	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	179,281	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	410,052	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	161,045	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502917.d




+ Scan (10.695-10.832 min, 20 scans) E2502917.d

**Toluene**


+ EIC (91.1) Scan E2502917.d




+ Scan (10.796-10.961 min, 24 scans) E2502917.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502917.d




+ Scan (12.995-13.093 min, 14 scans) E2502917.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502917.d




+ Scan (13.170-13.310 min, 19 scans) E2502917.d




o-Xylene

+ EIC (91.1) Scan E2502917.d




+ Scan (13.668-13.804 min, 20 scans) E2502917.d





Name SPRSEA-12-S-20250807
Comment C57499; Recollect
Data File E2502918.d
Acq. Date-Time 8/26/2025 11:28:01 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	518,898	
Benzene	benzene-d6 (IS)	8.102	8.110	173,197	
Toluene-d8 (IS)		10.746	10.753	566,495	
Toluene	Toluene-d8 (IS)	10.846	10.846	445,835	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	71,629	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	137,864	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	54,994	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502918.d




+ Scan (10.703-10.832 min, 19 scans) E2502918.d

**Toluene**


+ EIC (91.1) Scan E2502918.d




+ Scan (10.789-10.939 min, 22 scans) E2502918.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502918.d




+ Scan (12.996-13.080 min, 11 scans) E2502918.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502918.d




+ Scan (13.167-13.292 min, 17 scans) E2502918.d




o-Xylene

+ EIC (91.1) Scan E2502918.d




+ Scan (13.675-13.767 min, 13 scans) E2502918.d





Name SPRSEA-12-D-20250807
Comment C36957; Recollect
Data File E2502919.d
Acq. Date-Time 8/26/2025 11:52:48 AM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	515,759	
Benzene	benzene-d6 (IS)	8.102	8.110	159,734	
Toluene-d8 (IS)		10.746	10.753	577,249	
Toluene	Toluene-d8 (IS)	10.846	10.846	438,996	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	104,237	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	176,245	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	65,921	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502919.d




+ Scan (10.703-10.824 min, 18 scans) E2502919.d

**Toluene**


+ EIC (91.1) Scan E2502919.d




+ Scan (10.796-10.932 min, 20 scans) E2502919.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502919.d




+ Scan (12.996-13.081 min, 12 scans) E2502919.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502919.d




+ Scan (13.174-13.274 min, 15 scans) E2502919.d




o-Xylene

+ EIC (91.1) Scan E2502919.d




+ Scan (13.677-13.753 min, 10 scans) E2502919.d





Name SPRSEA-12-B-20250807
Comment C40632; Recollect
Data File E2502920.d
Acq. Date-Time 8/26/2025 12:17:34 PM
Acq. Method File M325B-MTD
Tube Sorbent CarboxpackX
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.045	8.052	522,322	
Benzene	benzene-d6 (IS)	8.102	8.110	21,343	
Toluene-d8 (IS)		10.745	10.753	577,148	
Toluene	Toluene-d8 (IS)	10.839	10.846	33,067	
Ethylbenzene	Toluene-d8 (IS)	13.031	13.038	7,878	
m-/p-Xylenes	Toluene-d8 (IS)	13.210	13.217	3,689	
o-Xylene	Toluene-d8 (IS)	13.711	13.718	1,503	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan E2502920.d




+ Scan (10.703-10.839 min, 20 scans) E2502920.d

**Toluene**


+ EIC (91.1) Scan E2502920.d




+ Scan (10.803-10.887 min, 12 scans) E2502920.d

**Ethylbenzene**


+ EIC (91.1) Scan E2502920.d




+ Scan (13.009-13.081 min, 11 scans) E2502920.d

**m-/p-Xylenes**

+ EIC (91.1) Scan E2502920.d




+ Scan (13.189-13.259 min, 9 scans) E2502920.d




o-Xylene

+ EIC (91.1) Scan E2502920.d



+ Scan (13.679-13.737 min, 8 scans) E2502920.d



Initial Calibration



Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC304-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Calibration Curves

Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
E050525A_CC252679_QT_CC185154	Benzene	1	E2500043.d	4.94	48593	54.2	531168	1.004	0.14
E050525A_CC252679_QT_CC185154	Benzene	2	E2500044.d	10.06	88289	54.2	525272	0.905	0.032
E050525A_CC252679_QT_CC185154	Benzene	3	E2500045.d	20.12	167722	54.2	519365	0.869	-0.0086
E050525A_CC252679_QT_CC185154	Benzene	4	E2500046.d	40.25	314317	54.2	495891	0.853	-0.027
E050525A_CC252679_QT_CC185154	Benzene	5	E2500047.d	100.62	771973	54.2	516636	0.804	-0.083
E050525A_CC252679_QT_CC185154	Benzene	6	E2500048.d	201.24	1566630	54.2	517234	0.815	-0.07
E050525A_CC252679_QT_CC185154	Benzene	7	E2500049.d	603.73	5084416	54.2	514456	0.887	0.011
E050525A_CC252679_QT_CC185154	Benzene	QT		120.51		54.2			
						Avg:	517146	0.877	
						%RSD:	2.1%	7.6%	
E050525A_CC252679_QT_CC185154	Toluene	1	E2500043.d	5.34	57611	63.2	569179	1.198	0.21
E050525A_CC252679_QT_CC185154	Toluene	2	E2500044.d	10.88	96506	63.2	555358	1.009	0.02
E050525A_CC252679_QT_CC185154	Toluene	3	E2500045.d	21.76	183454	63.2	550972	0.967	-0.023
E050525A_CC252679_QT_CC185154	Toluene	4	E2500046.d	43.52	328667	63.2	514484	0.927	-0.063
E050525A_CC252679_QT_CC185154	Toluene	5	E2500047.d	108.80	817722	63.2	541924	0.876	-0.11
E050525A_CC252679_QT_CC185154	Toluene	6	E2500048.d	217.60	1753197	63.2	536168	0.949	-0.04
E050525A_CC252679_QT_CC185154	Toluene	7	E2500049.d	652.80	5564266	63.2	538797	0.999	0.01
E050525A_CC252679_QT_CC185154	Toluene	QT		105.84		63.2			
						Avg:	543840	0.989	
						%RSD:	3.2%	10.3%	
E050525A_CC252679_QT_CC185154	Ethylbenzene	1	E2500043.d	5.13	60169	63.2	569179	1.301	0.089
E050525A_CC252679_QT_CC185154	Ethylbenzene	2	E2500044.d	10.46	105835	63.2	555358	1.151	-0.037
E050525A_CC252679_QT_CC185154	Ethylbenzene	3	E2500045.d	20.92	208215	63.2	550972	1.141	-0.045
E050525A_CC252679_QT_CC185154	Ethylbenzene	4	E2500046.d	41.83	392499	63.2	514484	1.152	-0.036
E050525A_CC252679_QT_CC185154	Ethylbenzene	5	E2500047.d	104.58	995735	63.2	541924	1.110	-0.071
E050525A_CC252679_QT_CC185154	Ethylbenzene	6	E2500048.d	209.16	2327735	63.2	536168	1.311	0.097
E050525A_CC252679_QT_CC185154	Ethylbenzene	7	E2500049.d	627.49	6406466	63.2	538797	1.197	0.0018
E050525A_CC252679_QT_CC185154	Ethylbenzene	QT		110.01		63.2			
						Avg:	543840	1.195	
						%RSD:	3.2%	6.7%	

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC304-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Calibration Curves

Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	1	E2500043.d	4.85	58394	63.2	569179	1.337	0.29
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	2	E2500044.d	9.88	88399	63.2	555358	1.018	-0.019
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	3	E2500045.d	19.75	172744	63.2	550972	1.003	-0.034
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	4	E2500046.d	39.51	305031	63.2	514484	0.948	-0.086
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	5	E2500047.d	98.77	730099	63.2	541924	0.862	-0.17
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	6	E2500048.d	197.54	1739458	63.2	536168	1.037	-4.1E-05
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	7	E2500049.d	592.63	5345328	63.2	538797	1.057	0.019
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	QT		123.29		63.2			
						Avg:	543840	1.037	
						%RSD:	3.2%	14.2%	
E050525A_CC252679_QT_CC185154	o-Xylene	1	E2500043.d	5.05	50277	63.2	569179	1.104	0.16
E050525A_CC252679_QT_CC185154	o-Xylene	2	E2500044.d	10.30	81187	63.2	555358	0.896	-0.056
E050525A_CC252679_QT_CC185154	o-Xylene	3	E2500045.d	20.60	161444	63.2	550972	0.898	-0.054
E050525A_CC252679_QT_CC185154	o-Xylene	4	E2500046.d	41.21	299918	63.2	514484	0.894	-0.059
E050525A_CC252679_QT_CC185154	o-Xylene	5	E2500047.d	103.02	797704	63.2	541924	0.903	-0.05
E050525A_CC252679_QT_CC185154	o-Xylene	6	E2500048.d	206.03	1692581	63.2	536168	0.968	0.019
E050525A_CC252679_QT_CC185154	o-Xylene	7	E2500049.d	618.10	5206672	63.2	538797	0.988	0.04
E050525A_CC252679_QT_CC185154	o-Xylene	QT		114.65		63.2			
						Avg:	543840	0.950	
						%RSD:	3.2%	8.2%	

Calibration Curves

Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
E050525A_CC252679_QT_CC185154	Benzene	ICV	E2500050.d	64.90	531194	54.2	505187	0.878	0.1%
E050525A_CC252679_QT_CC185154	Toluene	ICV	E2500050.d	77.40	626119	63.2	533535	0.958	-3.2%
E050525A_CC252679_QT_CC185154	Ethylbenzene	ICV	E2500050.d	87.14	736386	63.2	533535	1.001	-16.0%
E050525A_CC252679_QT_CC185154	m-/p-Xylenes	ICV	E2500050.d	90.70	588321	63.2	533535	0.768	-26.0%
E050525A_CC252679_QT_CC185154	o-Xylene	ICV	E2500050.d	89.27	625693	63.2	533535	0.830	-13.0%

M325B PDF Report ver.20250822

Sample Custody

**This Is The Last Page
Of This Report.**

Sprague - Searsport

70 Trundy Road
Searsport, ME 04974

Sampling Event 29 Sprague - Searsport

Client Project# PROJ-027966

Samples Received: 9/10/2025

Analytical Report 2025GC305

EPA Method 325B Analysis

Report Issue Date: 9/19/2025

I certify that to the best of my knowledge all analytical data presented in this report have been checked for completeness, accuracy, errors and legibility in addition to having been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s). This report shall not be reproduced except in full without approval of the laboratory. This will provide assurance that parts of the report are not taken out of context.

Amendment(s):

Signature:



QA Review by Isabel Obando Marrero, Data Reviewer



Matt Cavanaugh
Matthew.Cavanaugh@enthalpy.com / www.enthalpy.com
O: (919) 850-4392
Enthalpy Analytical
800 Capitola Drive Suite 1 Durham, NC 27713

Table of Contents

Case Narrative	3
Results	6
Summary of Results	7
Detailed Results	8
QC Data	11
Chromatograms	14
Initial Calibration	66
Sample Custody	69
Chain of Custody	70

Narrative Summary



Enthalpy Analytical Narrative Summary

Company	Montrose Air Quality Services, LLC - New Jersey
Job No.	2025GC305-1
Client ID.	PROJ-027966 Site: Sprague - Searsport

1. Custody

The samples were received at Enthalpy Analytical on September 10, 2025 at 23.5 °C, which is above the method recommended 23.0 °C. The samples were received in good condition. Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC

Table 1 - Sample Inventory

Sample ID	Tube ID	Sample Type
SPRSEA-1-S-20250821	B27217	Sample
SPRSEA-2-S-20250821	B46982	Sample
SPRSEA-3-S-20250821	B50949	Sample
SPRSEA-4-S-20250821	C40149	Sample
SPRSEA-5-S-20250821	B49635	Sample
SPRSEA-6-S-20250821	B46127	Sample
SPRSEA-6-D-20250821	B49528	Duplicate
SPRSEA-6-B-20250821	C00595	Blank
SPRSEA-7-S-20250821	B17520	Sample
SPRSEA-8-S-20250821	C55400	Sample
SPRSEA-9-S-20250821	C56980	Sample
SPRSEA-10-S-20250821	C01665	Sample
SPRSEA-11-S-20250821	C01392	Sample
SPRSEA-12-S-20250821	C70525	Sample
SPRSEA-12-D-20250821	C61742	Duplicate
SPRSEA-12-B-20250821	B44953	Blank

2. Analysis

The samples were analyzed for Benzene, Toluene, Ethylbenzene, m-/p-Xylenes, and o-Xylene using EPA Method 325B – Volatile Organic Compounds from Fugitive and Area Sources by Thermal Desorption and GC/MS. A copy of the acquisition method M325B-TD35 is not included in this report but may be available upon request.

The sample tube media used for this sampling period was CarbopackX. All calibration standards and laboratory QC were prepared using the same media.

3. Calibration

All BFB tune criteria have been met for this analysis.

The initial calibration (P041125A_CC252679_R1_QT_CC185154) met all 30% RSD criteria. The initial calibration verification met 30% difference criteria. The continuing calibration verifications met 30% difference criteria. The initial and continuing calibration raw data are not included in this report but are available upon request.

Enthalpy Analytical Narrative Summary

Company	Montrose Air Quality Services, LLC - New Jersey
Job No.	2025GC305-1
Client ID.	PROJ-027966 Site: Sprague - Searsport

5. QC Notes

All quality control criteria required by the method and/or the laboratory SOP have been met unless noted otherwise below.

6. Reporting Notes

All tubes used for this sampling period met the method criteria for number of uses; no tube exceeded 50 field uses.

As specified in EPA Method 325B, the response factor of the daily continuing calibration standard was used to quantitate all field samples and blanks.

All samples were reported as amount in ng catch, and concentration in ug/m³ and ppbv.

The results presented in this report are representative of the samples as provided to the laboratory. These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, located at 800 Capitola Drive, Suite 1, Durham NC, 27713 is accredited by the Louisiana Department of Environmental Quality (LDEQ) for EPA Method 325B for all analytes included in this report under **Certificate Number 04010**.

Results

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC305-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Summary

Sample Code	Tube ID	Benzene		Toluene		Ethylbenzene		m-/p-Xylenes		o-Xylene	
		(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag
SPRSEA-1-S-20250821	B27217	0.993		4.57		1.15		3.54		1.34	
SPRSEA-2-S-20250821	B46982	0.977		3.75		0.842		2.82		1.10	
SPRSEA-3-S-20250821	B50949	0.687		2.90		0.646		2.17		0.930	
SPRSEA-4-S-20250821	C40149	0.805		3.72		1.01		3.05		1.19	
SPRSEA-5-S-20250821	B49635	0.977		4.95		1.06		4.12		1.60	
SPRSEA-6-S-20250821	B46127	0.870		4.21		0.809		3.56		1.43	
SPRSEA-6-D-20250821	B49528	1.00		4.47		1.04		3.59		1.50	
SPRSEA-6-B-20250821	C00595	0.188	ND	0.242	ND	0.273	ND	0.273	ND	0.273	ND
SPRSEA-7-S-20250821	B17520	0.978		4.17		1.01		3.73		1.47	
SPRSEA-8-S-20250821	C55400	0.944		4.19		0.879		3.23		1.37	
SPRSEA-9-S-20250821	C56980	1.22		5.84		1.25		5.39		2.15	
SPRSEA-10-S-20250821	C01665	4.68		17.8		3.18		12.2		4.91	
SPRSEA-11-S-20250821	C01392	1.55		7.07		1.54		5.62		2.33	
SPRSEA-12-S-20250821	C70525	1.82		8.04		1.58		6.43		2.60	
SPRSEA-12-D-20250821	C61742	1.79		7.59		1.54		6.47		2.51	
SPRSEA-12-B-20250821	B44953	0.188	ND	0.242	ND	0.273	ND	0.273	ND	0.273	ND

ND: The analyte was not present above the Method Detection Limit

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC305-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Benzene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp (°F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250821	B27217	0.993	0.311	13.2	64.3	0.662	20140	0.188	0.372	0.0587	0.117		P2505745.D	2025-09-10 18:54	1.000	8.402	119783	816486	90.3	8.349	-3.7%
SPRSEA-2-S-20250821	B46982	0.977	0.306	13.0	64.3	0.662	20140	0.188	0.372	0.0587	0.117		P2505746.D	2025-09-10 19:32	1.000	8.402	114774	795444	90.3	8.349	-6.2%
SPRSEA-3-S-20250821	B50949	0.687	0.215	9.16	64.3	0.662	20140	0.188	0.372	0.0587	0.117		P2505747.D	2025-09-10 20:10	1.000	8.408	84437	831935	90.3	8.349	-1.9%
SPRSEA-4-S-20250821	C40149	0.805	0.252	10.7	64.3	0.662	20140	0.188	0.372	0.0587	0.117		P2505748.D	2025-09-10 20:47	1.000	8.397	97234	817739	90.3	8.343	-3.6%
SPRSEA-5-S-20250821	B49635	0.977	0.306	13.0	64.3	0.662	20140	0.188	0.372	0.0587	0.117		P2505749.D	2025-09-10 21:25	1.000	8.403	121174	839556	90.3	8.343	-1.0%
SPRSEA-6-S-20250821	B46127	0.870	0.272	11.6	64.3	0.662	20140	0.188	0.372	0.0587	0.117		P2505750.D	2025-09-10 22:03	1.000	8.402	105782	823521	90.3	8.343	-2.9%
SPRSEA-6-D-20250821	B49528	1.00	0.315	13.4	64.3	0.662	20140	0.188	0.372	0.0587	0.117		P2505751.D	2025-09-10 22:41	1.000	8.402	122266	823778	90.3	8.337	-2.9%
SPRSEA-6-B-20250821	C00595	0.188	0.0587		64.3	0.662	20140	0.188	0.372	0.0587	0.117	ND	P2505743.D	2025-09-10 17:38	1.000	8.391	12448	810912	90.3	8.343	-4.4%
SPRSEA-7-S-20250821	B17520	0.978	0.306	13.0	64.3	0.662	20140	0.188	0.372	0.0587	0.117		P2505752.D	2025-09-10 23:19	1.000	8.397	118356	819639	90.3	8.343	-3.4%
SPRSEA-8-S-20250821	C55400	0.944	0.296	12.6	64.3	0.662	20140	0.188	0.372	0.0587	0.117		P2505754.D	2025-09-11 00:59	1.000	8.391	115064	825538	90.3	8.343	-2.7%
SPRSEA-9-S-20250821	C56980	1.22	0.383	16.3	64.3	0.662	20140	0.188	0.372	0.0587	0.117		P2505755.D	2025-09-11 01:37	1.000	8.403	148673	824020	90.3	8.343	-2.8%
SPRSEA-10-S-20250821	C01665	4.68	1.47	62.4	64.3	0.662	20140	0.188	0.372	0.0587	0.117		P2505756.D	2025-09-11 02:15	1.000	8.402	578608	837447	90.3	8.343	-1.3%
SPRSEA-11-S-20250821	C01392	1.55	0.485	20.6	64.3	0.662	20140	0.188	0.372	0.0587	0.117		P2505757.D	2025-09-11 02:53	1.000	8.408	194881	852233	90.3	8.343	0.5%
SPRSEA-12-S-20250821	C70525	1.82	0.569	24.2	64.3	0.662	20140	0.188	0.372	0.0587	0.117		P2505758.D	2025-09-11 03:31	1.000	8.403	226494	844195	90.3	8.343	-0.5%
SPRSEA-12-D-20250821	C61742	1.79	0.561	23.9	64.3	0.662	20140	0.188	0.372	0.0587	0.117		P2505759.D	2025-09-11 04:08	1.000	8.397	213580	807606	90.3	8.343	-4.8%
SPRSEA-12-B-20250821	B44953	0.188	0.0587		64.3	0.662	20140	0.188	0.372	0.0587	0.117	ND	P2505744.D	2025-09-10 18:16	1.000	8.402	13128	810026	90.3	8.349	-4.5%

Toluene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp (°F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250821	B27217	4.57	1.21	47.3	64.3	0.514	20140	0.242	0.519	0.0641	0.138		P2505745.D	2025-09-10 18:54	1.057	11.002	465353	979282	105.3	10.907	-1.2%
SPRSEA-2-S-20250821	B46982	3.75	0.996	38.8	64.3	0.514	20140	0.242	0.519	0.0641	0.138		P2505746.D	2025-09-10 19:32	1.057	11.002	370773	951628	105.3	10.907	-4.0%
SPRSEA-3-S-20250821	B50949	2.90	0.771	30.0	64.3	0.514	20140	0.242	0.519	0.0642	0.138		P2505747.D	2025-09-10 20:10	1.057	11.002	290620	964015	105.3	10.907	-2.7%
SPRSEA-4-S-20250821	C40149	3.72	0.988	38.5	64.3	0.514	20140	0.242	0.519	0.0642	0.138		P2505748.D	2025-09-10 20:47	1.057	11.002	374790	969250	105.3	10.907	-2.2%
SPRSEA-5-S-20250821	B49635	4.95	1.32	51.3	64.3	0.514	20140	0.242	0.519	0.0641	0.138		P2505749.D	2025-09-10 21:25	1.057	10.996	499735	970954	105.3	10.907	-2.0%
SPRSEA-6-S-20250821	B46127	4.21	1.12	43.6	64.3	0.514	20140	0.242	0.519	0.0641	0.138		P2505750.D	2025-09-10 22:03	1.057	10.996	429121	980366	105.3	10.907	-1.1%
SPRSEA-6-D-20250821	B49528	4.47	1.19	46.2	64.3	0.514	20140	0.242	0.519	0.0641	0.138		P2505751.D	2025-09-10 22:41	1.057	10.996	460817	993234	105.3	10.901	0.2%
SPRSEA-6-B-20250821	C00595	0.242	0.0641		64.3	0.514	20140	0.242	0.519	0.0641	0.138	ND	P2505743.D	2025-09-10 17:38	1.057	11.002	8247	940547	105.3	10.907	-5.1%
SPRSEA-7-S-20250821	B17520	4.17	1.11	43.2	64.3	0.514	20140	0.242	0.519	0.0641	0.138		P2505752.D	2025-09-10 23:19	1.057	11.002	426546	983850	105.3	10.907	-0.7%
SPRSEA-8-S-20250821	C55400	4.19	1.11	43.4	64.3	0.514	20140	0.242	0.519	0.0641	0.138		P2505754.D	2025-09-11 00:59	1.057	10.996	433230	995352	105.3	10.907	0.4%
SPRSEA-9-S-20250821	C56980	5.84	1.55	60.4	64.3	0.514	20140	0.242	0.519	0.0641	0.138		P2505755.D	2025-09-11 01:37	1.057	10.996	613489	1011499	105.3	10.907	2.1%
SPRSEA-10-S-20250821	C01665	17.8	4.72	184	64.3	0.514	20140	0.242	0.519	0.0641	0.138		P2505756.D	2025-09-11 02:15	1.057	11.002	1866476	1010908	105.3	10.907	2.0%
SPRSEA-11-S-20250821	C01392	7.07	1.88	73.2	64.3	0.514	20140	0.242	0.519	0.0641	0.138		P2505757.D	2025-09-11 02:53	1.057	11.002	743161	1011606	105.3	10.907	2.1%
SPRSEA-12-S-20250821	C70525	8.04	2.13	83.2	64.3	0.514	20140	0.242	0.519	0.0641	0.138		P2505758.D	2025-09-11 03:31	1.057	10.996	846434	1014076	105.3	10.907	2.3%
SPRSEA-12-D-20250821	C61742	7.59	2.02	78.6	64.3	0.514	20140	0.242	0.519	0.0641	0.138		P2505759.D	2025-09-11 04:08	1.057	11.002	781781	991179	105.3	10.907	0.0%
SPRSEA-12-B-20250821	B44953	0.242	0.0641		64.3	0.514	20140	0.242	0.519	0.0641	0.138	ND	P2505744.D	2025-09-10 18:16	1.057	11.008	7906	973333	105.3	10.907	-1.8%

Ethylbenzene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp (°F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250821	B27217	1.15	0.264	10.5	64.3	0.455	20140	0.273	0.563	0.0629	0.130		P2505745.D	2025-09-10 18:54	1.092	13.127	106559	979282	105.3	10.907	-1.2%

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey
Job No.: 2025GC305-1 EPA Method 325B Analysis
Client No.: PROJ-027966 Site: Sprague - Searsport

Ethylbenzene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-2-S-20250821	B46982	0.842	0.194	7.71	64.3	0.455	20140	0.273	0.563	0.0629	0.130		P2505746.D	2025-09-10 19:32	1.092	13.127	76062	951628	105.3	10.907	-4.0%
SPRSEA-3-S-20250821	B50949	0.646	0.149	5.91	64.3	0.455	20140	0.273	0.563	0.0629	0.130		P2505747.D	2025-09-10 20:10	1.092	13.127	59115	964015	105.3	10.907	-2.7%
SPRSEA-4-S-20250821	C40149	1.01	0.233	9.26	64.3	0.455	20140	0.273	0.563	0.0629	0.130		P2505748.D	2025-09-10 20:47	1.092	13.127	93088	969250	105.3	10.907	-2.2%
SPRSEA-5-S-20250821	B49635	1.06	0.243	9.67	64.3	0.455	20140	0.273	0.563	0.0629	0.130		P2505749.D	2025-09-10 21:25	1.092	13.127	97347	970954	105.3	10.907	-2.0%
SPRSEA-6-S-20250821	B46127	0.809	0.186	7.40	64.3	0.455	20140	0.273	0.563	0.0629	0.130		P2505750.D	2025-09-10 22:03	1.092	13.127	75285	980366	105.3	10.907	-1.1%
SPRSEA-6-D-20250821	B49528	1.04	0.240	9.54	64.3	0.455	20140	0.273	0.563	0.0629	0.130		P2505751.D	2025-09-10 22:41	1.092	13.127	98231	993234	105.3	10.901	0.2%
SPRSEA-6-B-20250821	C00595	0.273	0.0629		64.3	0.455	20140	0.273	0.563	0.0629	0.130	ND	P2505743.D	2025-09-10 17:38	1.092	13.157	2510	940547	105.3	10.907	-5.1%
SPRSEA-7-S-20250821	B17520	1.01	0.233	9.26	64.3	0.455	20140	0.273	0.563	0.0629	0.130		P2505752.D	2025-09-10 23:19	1.092	13.127	94478	983850	105.3	10.907	-0.7%
SPRSEA-8-S-20250821	C55400	0.879	0.203	8.05	64.3	0.455	20140	0.273	0.563	0.0629	0.130		P2505754.D	2025-09-11 00:59	1.092	13.127	83080	995352	105.3	10.907	0.4%
SPRSEA-9-S-20250821	C56980	1.25	0.289	11.5	64.3	0.455	20140	0.273	0.563	0.0629	0.130		P2505755.D	2025-09-11 01:37	1.092	13.127	120346	1011499	105.3	10.907	2.1%
SPRSEA-10-S-20250821	C01665	3.18	0.734	29.1	64.3	0.455	20140	0.273	0.563	0.0629	0.130		P2505756.D	2025-09-11 02:15	1.092	13.127	305551	1010908	105.3	10.907	2.0%
SPRSEA-11-S-20250821	C01392	1.54	0.356	14.1	64.3	0.455	20140	0.273	0.563	0.0629	0.130		P2505757.D	2025-09-11 02:53	1.092	13.127	148190	1011606	105.3	10.907	2.1%
SPRSEA-12-S-20250821	C70525	1.58	0.365	14.5	64.3	0.455	20140	0.273	0.563	0.0629	0.130		P2505758.D	2025-09-11 03:31	1.092	13.127	152403	1014076	105.3	10.907	2.3%
SPRSEA-12-D-20250821	C61742	1.54	0.354	14.1	64.3	0.455	20140	0.273	0.563	0.0629	0.130		P2505759.D	2025-09-11 04:08	1.092	13.127	144499	991179	105.3	10.907	0.0%
SPRSEA-12-B-20250821	B44953	0.273	0.0629		64.3	0.455	20140	0.273	0.563	0.0629	0.130	ND	P2505744.D	2025-09-10 18:16	1.092	13.305	0	973333	105.3	10.907	-1.8%

m-/p-Xylenes

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250821	B27217	3.54	0.815	32.4	64.3	0.455	20140	0.273	0.532	0.0629	0.123		P2505745.D	2025-09-10 18:54	0.737	13.299	221994	979282	105.3	10.907	-1.2%
SPRSEA-2-S-20250821	B46982	2.82	0.651	25.9	64.3	0.455	20140	0.273	0.532	0.0629	0.123		P2505746.D	2025-09-10 19:32	0.737	13.299	172282	951628	105.3	10.907	-4.0%
SPRSEA-3-S-20250821	B50949	2.17	0.500	19.9	64.3	0.455	20140	0.273	0.532	0.0629	0.123		P2505747.D	2025-09-10 20:10	0.737	13.299	133989	964015	105.3	10.907	-2.7%
SPRSEA-4-S-20250821	C40149	3.05	0.703	27.9	64.3	0.455	20140	0.273	0.532	0.0629	0.123		P2505748.D	2025-09-10 20:47	0.737	13.299	189392	969250	105.3	10.907	-2.2%
SPRSEA-5-S-20250821	B49635	4.12	0.949	37.7	64.3	0.455	20140	0.273	0.532	0.0629	0.123		P2505749.D	2025-09-10 21:25	0.737	13.299	256181	970954	105.3	10.907	-2.0%
SPRSEA-6-S-20250821	B46127	3.56	0.820	32.6	64.3	0.455	20140	0.273	0.532	0.0629	0.123		P2505750.D	2025-09-10 22:03	0.737	13.299	223556	980366	105.3	10.907	-1.1%
SPRSEA-6-D-20250821	B49528	3.59	0.828	32.9	64.3	0.455	20140	0.273	0.532	0.0629	0.123		P2505751.D	2025-09-10 22:41	0.737	13.299	228796	993234	105.3	10.901	0.2%
SPRSEA-6-B-20250821	C00595	0.273	0.0629		64.3	0.455	20140	0.273	0.532	0.0629	0.123	ND	P2505743.D	2025-09-10 17:38	0.737	13.329	625	940547	105.3	10.907	-5.1%
SPRSEA-7-S-20250821	B17520	3.73	0.859	34.1	64.3	0.455	20140	0.273	0.532	0.0629	0.123		P2505752.D	2025-09-10 23:19	0.737	13.299	234986	983850	105.3	10.907	-0.7%
SPRSEA-8-S-20250821	C55400	3.23	0.745	29.6	64.3	0.455	20140	0.273	0.532	0.0629	0.123		P2505754.D	2025-09-11 00:59	0.737	13.299	206195	995352	105.3	10.907	0.4%
SPRSEA-9-S-20250821	C56980	5.39	1.24	49.4	64.3	0.455	20140	0.273	0.532	0.0629	0.123		P2505755.D	2025-09-11 01:37	0.737	13.299	349709	1011499	105.3	10.907	2.1%
SPRSEA-10-S-20250821	C01665	12.2	2.82	112	64.3	0.455	20140	0.273	0.532	0.0629	0.123		P2505756.D	2025-09-11 02:15	0.737	13.299	792177	1010908	105.3	10.907	2.0%
SPRSEA-11-S-20250821	C01392	5.62	1.30	51.5	64.3	0.455	20140	0.273	0.532	0.0629	0.123		P2505757.D	2025-09-11 02:53	0.737	13.299	364391	1011606	105.3	10.907	2.1%
SPRSEA-12-S-20250821	C70525	6.43	1.48	58.8	64.3	0.455	20140	0.273	0.532	0.0629	0.123		P2505758.D	2025-09-11 03:31	0.737	13.299	417635	1014076	105.3	10.907	2.3%
SPRSEA-12-D-20250821	C61742	6.47	1.49	59.2	64.3	0.455	20140	0.273	0.532	0.0629	0.123		P2505759.D	2025-09-11 04:08	0.737	13.299	410978	991179	105.3	10.907	0.0%
SPRSEA-12-B-20250821	B44953	0.273	0.0629		64.3	0.455	20140	0.273	0.532	0.0629	0.123	ND	P2505744.D	2025-09-10 18:16	0.737	13.305	1475	973333	105.3	10.907	-1.8%

o-Xylene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250821	B27217	1.34	0.309	12.3	64.3	0.455	20140	0.273	0.555	0.0629	0.128		P2505745.D	2025-09-10 18:54	0.693	13.792	79077	979282	105.3	10.907	-1.2%
SPRSEA-2-S-20250821	B46982	1.10	0.253	10.1	64.3	0.455	20140	0.273	0.555	0.0629	0.128		P2505746.D	2025-09-10 19:32	0.693	13.791	63046	951628	105.3	10.907	-4.0%

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC305-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

o-Xylene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-3-S-20250821	B50949	0.930	0.214	8.51	64.3	0.455	20140	0.273	0.555	0.0629	0.128		P2505747.D	2025-09-10 20:10	0.693	13.792	53983	964015	105.3	10.907	-2.7%
SPRSEA-4-S-20250821	C40149	1.19	0.275	10.9	64.3	0.455	20140	0.273	0.555	0.0629	0.128		P2505748.D	2025-09-10 20:47	0.693	13.786	69685	969250	105.3	10.907	-2.2%
SPRSEA-5-S-20250821	B49635	1.60	0.368	14.6	64.3	0.455	20140	0.273	0.555	0.0629	0.128		P2505749.D	2025-09-10 21:25	0.693	13.792	93430	970954	105.3	10.907	-2.0%
SPRSEA-6-S-20250821	B46127	1.43	0.329	13.1	64.3	0.455	20140	0.273	0.555	0.0629	0.128		P2505750.D	2025-09-10 22:03	0.693	13.786	84232	980366	105.3	10.907	-1.1%
SPRSEA-6-D-20250821	B49528	1.50	0.346	13.7	64.3	0.455	20140	0.273	0.555	0.0629	0.128		P2505751.D	2025-09-10 22:41	0.693	13.786	89807	993234	105.3	10.901	0.2%
SPRSEA-6-B-20250821	C00595	0.273	0.0629		64.3	0.455	20140	0.273	0.555	0.0629	0.128	ND	P2505743.D	2025-09-10 17:38	0.693	13.982	0	940547	105.3	10.907	-5.1%
SPRSEA-7-S-20250821	B17520	1.47	0.339	13.5	64.3	0.455	20140	0.273	0.555	0.0629	0.128		P2505752.D	2025-09-10 23:19	0.693	13.792	87273	983850	105.3	10.907	-0.7%
SPRSEA-8-S-20250821	C55400	1.37	0.315	12.5	64.3	0.455	20140	0.273	0.555	0.0629	0.128		P2505754.D	2025-09-11 00:59	0.693	13.792	81972	995352	105.3	10.907	0.4%
SPRSEA-9-S-20250821	C56980	2.15	0.496	19.7	64.3	0.455	20140	0.273	0.555	0.0629	0.128		P2505755.D	2025-09-11 01:37	0.693	13.792	131121	1011499	105.3	10.907	2.1%
SPRSEA-10-S-20250821	C01665	4.91	1.13	45.0	64.3	0.455	20140	0.273	0.555	0.0629	0.128		P2505756.D	2025-09-11 02:15	0.693	13.791	299166	1010908	105.3	10.907	2.0%
SPRSEA-11-S-20250821	C01392	2.33	0.538	21.4	64.3	0.455	20140	0.273	0.555	0.0629	0.128		P2505757.D	2025-09-11 02:53	0.693	13.792	142191	1011606	105.3	10.907	2.1%
SPRSEA-12-S-20250821	C70525	2.60	0.599	23.8	64.3	0.455	20140	0.273	0.555	0.0629	0.128		P2505758.D	2025-09-11 03:31	0.693	13.786	158768	1014076	105.3	10.907	2.3%
SPRSEA-12-D-20250821	C61742	2.51	0.577	22.9	64.3	0.455	20140	0.273	0.555	0.0629	0.128		P2505759.D	2025-09-11 04:08	0.693	13.792	149609	991179	105.3	10.907	0.0%
SPRSEA-12-B-20250821	B44953	0.273	0.0629		64.3	0.455	20140	0.273	0.555	0.0629	0.128	ND	P2505744.D	2025-09-10 18:16	0.693	13.910	0	973333	105.3	10.907	-1.8%

ND: The analyte was not present above the Method Detection Limit

QC Data



Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey
Job No.: 2025GC305-1 EPA Method 325B Analysis
Client No.: PROJ-027966 Site: Sprague - Searsport

QC Samples

Field Sample Type	Sample Code	Benzene		Toluene		Ethylbenzene		m-/p-Xylenes		o-Xylene	
Blanks (ug/m³)	SPRSEA-6-B-20250821	ND	Pass	ND	Pass	ND	Pass	ND	Pass	ND	Pass
	SPRSEA-12-B-20250821	ND	Pass	ND	Pass	ND	Pass	ND	Pass	ND	Pass
Duplicates (difference)	SPRSEA-6-D-20250821	14%	Pass	5.8%	Pass	25%	Pass	1.0%	Pass	5.1%	Pass
	SPRSEA-12-D-20250821	1.4%	Pass	5.7%	Pass	3.0%	Pass	0.68%	Pass	3.7%	Pass

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC305-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Benzene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	P2505741.D	C70811	Cal	1.000		1.000	12%	10%		Pass	
2025GC305 Method Blank-1	P2505742.D	B20616	Blank			1.000			-10%	Pass	ND
M325B CCV 5 REC	P2505753.D	B48053	Check	0.999		1.000	12%		6.2%	Pass	
M325B CCV 5 REC	P2505760.D	C00767	Check	1.007		1.000	12%		2.9%	Pass	

Toluene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	P2505741.D	C70811	Cal	1.057		1.057	19%	3.8%		Pass	
2025GC305 Method Blank-1	P2505742.D	B20616	Blank			1.057			-4.3%	Pass	ND
M325B CCV 5 REC	P2505753.D	B48053	Check	1.055		1.057	19%		9.0%	Pass	
M325B CCV 5 REC	P2505760.D	C00767	Check	1.054		1.057	19%		8.6%	Pass	

Ethylbenzene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	P2505741.D	C70811	Cal	1.092		1.092	5.7%	3.8%		Pass	
2025GC305 Method Blank-1	P2505742.D	B20616	Blank			1.092			-4.3%	Pass	ND
M325B CCV 5 REC	P2505753.D	B48053	Check	1.039		1.092	0.50%		9.0%	Pass	
M325B CCV 5 REC	P2505760.D	C00767	Check	1.052		1.092	1.8%		8.6%	Pass	

m-/p-Xylenes Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	P2505741.D	C70811	Cal	0.737		0.737	-6.9%	3.8%		Pass	
2025GC305 Method Blank-1	P2505742.D	B20616	Blank			0.737			-4.3%	Pass	ND
M325B CCV 5 REC	P2505753.D	B48053	Check	0.649		0.737	-18%		9.0%	Pass	
M325B CCV 5 REC	P2505760.D	C00767	Check	0.654		0.737	-17%		8.6%	Pass	


o-Xylene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	P2505741.D	C70811	Cal	0.693		0.693	-17%	3.8%		Pass	
2025GC305 Method Blank-1	P2505742.D	B20616	Blank			0.693			-4.3%	Pass	ND
M325B CCV 5 REC	P2505753.D	B48053	Check	0.683		0.693	-18%		9.0%	Pass	
M325B CCV 5 REC	P2505760.D	C00767	Check	0.681		0.693	-18%		8.6%	Pass	

Chromatograms



Name 2025GC305 Method Blank-1
Comment B20616
Data File P2505742.D
Acq. Date-Time 9/10/2025 5:00:23 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.343	8.367	761,375	
Benzene	benzene-d6 (IS)	8.408	8.426	12,508	
Toluene-d8 (IS)		10.907	10.931	948,119	
Toluene	Toluene-d8 (IS)	11.008	11.020	14,815	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	8,666	
m-/p-Xylenes	Toluene-d8 (IS)	13.317	13.341	8,827	
o-Xylene	Toluene-d8 (IS)	13.780	13.809	3,778	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2505742.D




+ Scan (10.864-11.032 min, 29 scans) P2505742.D

**Toluene**


+ EIC (91.1) Scan P2505742.D




+ Scan (10.964-11.074 min, 19 scans) P2505742.D

**Ethylbenzene**


+ EIC (91.1) Scan P2505742.D




+ Scan (13.103-13.163 min, 11 scans) P2505742.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2505742.D




+ Scan (13.275-13.372 min, 17 scans) P2505742.D




o-Xylene

+ EIC (91.1) Scan P2505742.D




+ Scan (13.774-13.820 min, 7 scans) P2505742.D





Name SPRSEA-6-B-20250821
Comment C00595
Data File P2505743.D
Acq. Date-Time 9/10/2025 5:38:18 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.343	8.367	810,912	
Benzene	benzene-d6 (IS)	8.391	8.426	12,448	
Toluene-d8 (IS)		10.907	10.931	940,547	
Toluene	Toluene-d8 (IS)	11.002	11.020	8,247	
Ethylbenzene	Toluene-d8 (IS)	13.157	13.151	2,510	
m-/p-Xylenes	Toluene-d8 (IS)	13.329	13.341	625	
o-Xylene	Toluene-d8 (IS)	13.982	13.809	ND	m

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2505743.D




+ Scan (10.865-10.990 min, 22 scans) P2505743.D

**Toluene**


+ EIC (91.1) Scan P2505743.D




+ Scan (10.973-11.060 min, 14 scans) P2505743.D

**Ethylbenzene**


+ EIC (91.1) Scan P2505743.D




+ Scan (13.100-13.167 min, 11 scans) P2505743.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2505743.D




+ Scan (13.307-13.339 min, 5 scans) P2505743.D




o-Xylene

+ EIC (91.1) Scan P2505743.D




+ Scan (13.982-13.982 min, 1 scans) P2505743.D





Name SPRSEA-12-B-20250821
Comment B44953
Data File P2505744.D
Acq. Date-Time 9/10/2025 6:16:16 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.349	8.367	810,026	
Benzene	benzene-d6 (IS)	8.402	8.426	13,128	
Toluene-d8 (IS)		10.907	10.931	973,333	
Toluene	Toluene-d8 (IS)	11.008	11.020	7,906	
Ethylbenzene	Toluene-d8 (IS)	13.305	13.151	ND	m
m-/p-Xylenes	Toluene-d8 (IS)	13.305	13.341	1,475	m
o-Xylene	Toluene-d8 (IS)	13.910	13.809	ND	m

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2505744.D




+ Scan (10.859-11.043 min, 32 scans) P2505744.D

**Toluene**


+ EIC (91.1) Scan P2505744.D




+ Scan (10.961-11.043 min, 14 scans) P2505744.D

**Ethylbenzene**


+ EIC (91.1) Scan P2505744.D




+ Scan (13.305-13.305 min, 1 scans) P2505744.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2505744.D




+ Scan (13.295-13.355 min, 10 scans) P2505744.D




o-Xylene

+ EIC (91.1) Scan P2505744.D




+ Scan (13.910-13.910 min, 1 scans) P2505744.D





Name SPRSEA-1-S-20250821
Comment B27217
Data File P2505745.D
Acq. Date-Time 9/10/2025 6:54:09 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.349	8.367	816,486	
Benzene	benzene-d6 (IS)	8.402	8.426	119,783	
Toluene-d8 (IS)		10.907	10.931	979,282	
Toluene	Toluene-d8 (IS)	11.002	11.020	465,353	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	106,559	
m-/p-Xylenes	Toluene-d8 (IS)	13.299	13.341	221,994	
o-Xylene	Toluene-d8 (IS)	13.792	13.809	79,077	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2505745.D




+ Scan (10.860-11.055 min, 33 scans) P2505745.D

**Toluene**


+ EIC (91.1) Scan P2505745.D




+ Scan (10.955-11.115 min, 27 scans) P2505745.D

**Ethylbenzene**


+ EIC (91.1) Scan P2505745.D




+ Scan (13.093-13.210 min, 20 scans) P2505745.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2505745.D




+ Scan (13.264-13.434 min, 28 scans) P2505745.D




o-Xylene

+ EIC (91.1) Scan P2505745.D




+ Scan (13.762-13.856 min, 15 scans) P2505745.D




Name SPRSEA-2-S-20250821
Comment B46982
Data File P2505746.D
Acq. Date-Time 9/10/2025 7:32:04 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxen X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram




Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.349	8.367	795,444	
Benzene	benzene-d6 (IS)	8.402	8.426	114,774	
Toluene-d8 (IS)		10.907	10.931	951,628	
Toluene	Toluene-d8 (IS)	11.002	11.020	370,773	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	76,062	
m-/p-Xylenes	Toluene-d8 (IS)	13.299	13.341	172,282	
o-Xylene	Toluene-d8 (IS)	13.791	13.809	63,046	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2505746.D




+ Scan (10.859-11.014 min, 27 scans) P2505746.D

**Toluene**


+ EIC (91.1) Scan P2505746.D




+ Scan (10.961-11.109 min, 25 scans) P2505746.D

**Ethylbenzene**


+ EIC (91.1) Scan P2505746.D




+ Scan (13.094-13.228 min, 23 scans) P2505746.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2505746.D




+ Scan (13.265-13.429 min, 27 scans) P2505746.D




o-Xylene

+ EIC (91.1) Scan P2505746.D




+ Scan (13.757-13.898 min, 23 scans) P2505746.D





Name SPRSEA-3-S-20250821
Comment B50949
Data File P2505747.D
Acq. Date-Time 9/10/2025 8:10:00 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.349	8.367	831,935	
Benzene	benzene-d6 (IS)	8.408	8.426	84,437	
Toluene-d8 (IS)		10.907	10.931	964,015	
Toluene	Toluene-d8 (IS)	11.002	11.020	290,620	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	59,115	
m-/p-Xylenes	Toluene-d8 (IS)	13.299	13.341	133,989	
o-Xylene	Toluene-d8 (IS)	13.792	13.809	53,983	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2505747.D




+ Scan (10.860-11.014 min, 27 scans) P2505747.D

**Toluene**


+ EIC (91.1) Scan P2505747.D




+ Scan (10.961-11.097 min, 23 scans) P2505747.D

**Ethylbenzene**


+ EIC (91.1) Scan P2505747.D




+ Scan (13.094-13.186 min, 16 scans) P2505747.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2505747.D




+ Scan (13.264-13.418 min, 26 scans) P2505747.D




o-Xylene

+ EIC (91.1) Scan P2505747.D




+ Scan (13.762-13.851 min, 15 scans) P2505747.D





Name SPRSEA-4-S-20250821
Comment C40149
Data File P2505748.D
Acq. Date-Time 9/10/2025 8:47:53 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxen X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.343	8.367	817,739	
Benzene	benzene-d6 (IS)	8.397	8.426	97,234	
Toluene-d8 (IS)		10.907	10.931	969,250	
Toluene	Toluene-d8 (IS)	11.002	11.020	374,790	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	93,088	
m-/p-Xylenes	Toluene-d8 (IS)	13.299	13.341	189,392	
o-Xylene	Toluene-d8 (IS)	13.786	13.809	69,685	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2505748.D




+ Scan (10.854-11.014 min, 27 scans) P2505748.D

**Toluene**


+ EIC (91.1) Scan P2505748.D




+ Scan (10.955-11.115 min, 28 scans) P2505748.D

**Ethylbenzene**


+ EIC (91.1) Scan P2505748.D




+ Scan (13.094-13.192 min, 17 scans) P2505748.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2505748.D




+ Scan (13.255-13.424 min, 29 scans) P2505748.D




o-Xylene

+ EIC (91.1) Scan P2505748.D




+ Scan (13.756-13.861 min, 17 scans) P2505748.D





Name SPRSEA-5-S-20250821
Comment B49635
Data File P2505749.D
Acq. Date-Time 9/10/2025 9:25:54 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.343	8.367	839,556	
Benzene	benzene-d6 (IS)	8.403	8.426	121,174	
Toluene-d8 (IS)		10.907	10.931	970,954	
Toluene	Toluene-d8 (IS)	10.996	11.020	499,735	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	97,347	
m-/p-Xylenes	Toluene-d8 (IS)	13.299	13.341	256,181	
o-Xylene	Toluene-d8 (IS)	13.792	13.809	93,430	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2505749.D




+ Scan (10.862-11.020 min, 27 scans) P2505749.D

**Toluene**


+ EIC (91.1) Scan P2505749.D




+ Scan (10.949-11.145 min, 34 scans) P2505749.D

**Ethylbenzene**


+ EIC (91.1) Scan P2505749.D




+ Scan (13.089-13.209 min, 20 scans) P2505749.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2505749.D




+ Scan (13.252-13.418 min, 28 scans) P2505749.D




o-Xylene

+ EIC (91.1) Scan P2505749.D




+ Scan (13.750-13.893 min, 25 scans) P2505749.D





Name SPRSEA-6-S-20250821
Comment B46127
Data File P2505750.D
Acq. Date-Time 9/10/2025 10:03:46 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.343	8.367	823,521	
Benzene	benzene-d6 (IS)	8.402	8.426	105,782	
Toluene-d8 (IS)		10.907	10.931	980,366	
Toluene	Toluene-d8 (IS)	10.996	11.020	429,121	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	75,285	
m-/p-Xylenes	Toluene-d8 (IS)	13.299	13.341	223,556	
o-Xylene	Toluene-d8 (IS)	13.786	13.809	84,232	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2505750.D




+ Scan (10.857-11.020 min, 28 scans) P2505750.D

**Toluene**


+ EIC (91.1) Scan P2505750.D




+ Scan (10.950-11.139 min, 32 scans) P2505750.D

**Ethylbenzene**


+ EIC (91.1) Scan P2505750.D




+ Scan (13.090-13.198 min, 19 scans) P2505750.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2505750.D




+ Scan (13.246-13.441 min, 34 scans) P2505750.D




o-Xylene

+ EIC (91.1) Scan P2505750.D




+ Scan (13.750-13.875 min, 22 scans) P2505750.D





Name SPRSEA-6-D-20250821
Comment B49528
Data File P2505751.D
Acq. Date-Time 9/10/2025 10:41:41 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.337	8.367	823,778	
Benzene	benzene-d6 (IS)	8.402	8.426	122,266	
Toluene-d8 (IS)		10.901	10.931	993,234	
Toluene	Toluene-d8 (IS)	10.996	11.020	460,817	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	98,231	
m-/p-Xylenes	Toluene-d8 (IS)	13.299	13.341	228,796	
o-Xylene	Toluene-d8 (IS)	13.786	13.809	89,807	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2505751.D




+ Scan (10.860-11.050 min, 32 scans) P2505751.D

**Toluene**


+ EIC (91.1) Scan P2505751.D




+ Scan (10.950-11.127 min, 30 scans) P2505751.D

**Ethylbenzene**


+ EIC (91.1) Scan P2505751.D




+ Scan (13.092-13.234 min, 24 scans) P2505751.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2505751.D




+ Scan (13.264-13.405 min, 23 scans) P2505751.D




o-Xylene

+ EIC (91.1) Scan P2505751.D




+ Scan (13.752-13.881 min, 22 scans) P2505751.D





Name SPRSEA-7-S-20250821
Comment B17520
Data File P2505752.D
Acq. Date-Time 9/10/2025 11:19:33 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.343	8.367	819,639	
Benzene	benzene-d6 (IS)	8.397	8.426	118,356	
Toluene-d8 (IS)		10.907	10.931	983,850	
Toluene	Toluene-d8 (IS)	11.002	11.020	426,546	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	94,478	
m-/p-Xylenes	Toluene-d8 (IS)	13.299	13.341	234,986	
o-Xylene	Toluene-d8 (IS)	13.792	13.809	87,273	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2505752.D




+ Scan (10.860-11.008 min, 26 scans) P2505752.D

**Toluene**


+ EIC (91.1) Scan P2505752.D




+ Scan (10.958-11.103 min, 25 scans) P2505752.D

**Ethylbenzene**


+ EIC (91.1) Scan P2505752.D




+ Scan (13.089-13.221 min, 22 scans) P2505752.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2505752.D




+ Scan (13.258-13.474 min, 36 scans) P2505752.D




o-Xylene

+ EIC (91.1) Scan P2505752.D




+ Scan (13.753-13.863 min, 19 scans) P2505752.D





Name SPRSEA-8-S-20250821
Comment C55400
Data File P2505754.D
Acq. Date-Time 9/11/2025 12:59:29 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.343	8.367	825,538	
Benzene	benzene-d6 (IS)	8.391	8.426	115,064	
Toluene-d8 (IS)		10.907	10.931	995,352	
Toluene	Toluene-d8 (IS)	10.996	11.020	433,230	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	83,080	
m-/p-Xylenes	Toluene-d8 (IS)	13.299	13.341	206,195	
o-Xylene	Toluene-d8 (IS)	13.792	13.809	81,972	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2505754.D




+ Scan (10.855-11.008 min, 26 scans) P2505754.D

**Toluene**


+ EIC (91.1) Scan P2505754.D




+ Scan (10.955-11.145 min, 32 scans) P2505754.D

**Ethylbenzene**


+ EIC (91.1) Scan P2505754.D




+ Scan (13.087-13.215 min, 21 scans) P2505754.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2505754.D




+ Scan (13.260-13.388 min, 22 scans) P2505754.D




o-Xylene

+ EIC (91.1) Scan P2505754.D




+ Scan (13.748-13.863 min, 20 scans) P2505754.D





Name SPRSEA-9-S-20250821
Comment C56980
Data File P2505755.D
Acq. Date-Time 9/11/2025 1:37:23 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.343	8.367	824,020	
Benzene	benzene-d6 (IS)	8.403	8.426	148,673	
Toluene-d8 (IS)		10.907	10.931	1,011,499	
Toluene	Toluene-d8 (IS)	10.996	11.020	613,489	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	120,346	
m-/p-Xylenes	Toluene-d8 (IS)	13.299	13.341	349,709	
o-Xylene	Toluene-d8 (IS)	13.792	13.809	131,121	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2505755.D




+ Scan (10.854-11.020 min, 29 scans) P2505755.D

**Toluene**


+ EIC (91.1) Scan P2505755.D




+ Scan (10.955-11.127 min, 29 scans) P2505755.D

**Ethylbenzene**


+ EIC (91.1) Scan P2505755.D




+ Scan (13.093-13.181 min, 15 scans) P2505755.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2505755.D




+ Scan (13.262-13.440 min, 30 scans) P2505755.D




o-Xylene

+ EIC (91.1) Scan P2505755.D




+ Scan (13.757-13.869 min, 19 scans) P2505755.D





Name SPRSEA-10-S-20250821
Comment C01665
Data File P2505756.D
Acq. Date-Time 9/11/2025 2:15:15 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.343	8.367	837,447	
Benzene	benzene-d6 (IS)	8.402	8.426	578,608	
Toluene-d8 (IS)		10.907	10.931	1,010,908	
Toluene	Toluene-d8 (IS)	11.002	11.020	1,866,476	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	305,551	
m-/p-Xylenes	Toluene-d8 (IS)	13.299	13.341	792,177	
o-Xylene	Toluene-d8 (IS)	13.791	13.809	299,166	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2505756.D




+ Scan (10.859-11.014 min, 27 scans) P2505756.D

**Toluene**


+ EIC (91.1) Scan P2505756.D




+ Scan (10.954-11.115 min, 28 scans) P2505756.D

**Ethylbenzene**


+ EIC (91.1) Scan P2505756.D




+ Scan (13.091-13.228 min, 24 scans) P2505756.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2505756.D




+ Scan (13.257-13.471 min, 37 scans) P2505756.D




o-Xylene

+ EIC (91.1) Scan P2505756.D




+ Scan (13.744-13.857 min, 20 scans) P2505756.D





Name SPRSEA-11-S-20250821
Comment C01392
Data File P2505757.D
Acq. Date-Time 9/11/2025 2:53:07 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.343	8.367	852,233	
Benzene	benzene-d6 (IS)	8.408	8.426	194,881	
Toluene-d8 (IS)		10.907	10.931	1,011,606	
Toluene	Toluene-d8 (IS)	11.002	11.020	743,161	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	148,190	
m-/p-Xylenes	Toluene-d8 (IS)	13.299	13.341	364,391	
o-Xylene	Toluene-d8 (IS)	13.792	13.809	142,191	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2505757.D




+ Scan (10.862-11.008 min, 25 scans) P2505757.D

**Toluene**


+ EIC (91.1) Scan P2505757.D




+ Scan (10.955-11.097 min, 24 scans) P2505757.D

**Ethylbenzene**


+ EIC (91.1) Scan P2505757.D




+ Scan (13.094-13.196 min, 17 scans) P2505757.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2505757.D




+ Scan (13.265-13.434 min, 28 scans) P2505757.D




o-Xylene

+ EIC (91.1) Scan P2505757.D




+ Scan (13.746-13.887 min, 24 scans) P2505757.D





Name SPRSEA-12-S-20250821
Comment C70525
Data File P2505758.D
Acq. Date-Time 9/11/2025 3:31:00 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.343	8.367	844,195	
Benzene	benzene-d6 (IS)	8.403	8.426	226,494	
Toluene-d8 (IS)		10.907	10.931	1,014,076	
Toluene	Toluene-d8 (IS)	10.996	11.020	846,434	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	152,403	
m-/p-Xylenes	Toluene-d8 (IS)	13.299	13.341	417,635	
o-Xylene	Toluene-d8 (IS)	13.786	13.809	158,768	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2505758.D




+ Scan (10.855-11.014 min, 27 scans) P2505758.D

**Toluene**


+ EIC (91.1) Scan P2505758.D




+ Scan (10.955-11.139 min, 31 scans) P2505758.D

**Ethylbenzene**


+ EIC (91.1) Scan P2505758.D




+ Scan (13.090-13.192 min, 17 scans) P2505758.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2505758.D




+ Scan (13.258-13.412 min, 27 scans) P2505758.D




o-Xylene

+ EIC (91.1) Scan P2505758.D




+ Scan (13.756-13.881 min, 22 scans) P2505758.D





Name SPRSEA-12-D-20250821
Comment C61742
Data File P2505759.D
Acq. Date-Time 9/11/2025 4:08:51 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
benzene-d6 (IS)		8.343	8.367	807,606	
Benzene	benzene-d6 (IS)	8.397	8.426	213,580	
Toluene-d8 (IS)		10.907	10.931	991,179	
Toluene	Toluene-d8 (IS)	11.002	11.020	781,781	
Ethylbenzene	Toluene-d8 (IS)	13.127	13.151	144,499	
m-/p-Xylenes	Toluene-d8 (IS)	13.299	13.341	410,978	
o-Xylene	Toluene-d8 (IS)	13.792	13.809	149,609	

benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan P2505759.D




+ Scan (10.860-11.050 min, 32 scans) P2505759.D

**Toluene**


+ EIC (91.1) Scan P2505759.D




+ Scan (10.956-11.115 min, 27 scans) P2505759.D

**Ethylbenzene**


+ EIC (91.1) Scan P2505759.D




+ Scan (13.092-13.203 min, 18 scans) P2505759.D

**m-/p-Xylenes**

+ EIC (91.1) Scan P2505759.D




+ Scan (13.264-13.471 min, 36 scans) P2505759.D




o-Xylene

+ EIC (91.1) Scan P2505759.D



+ Scan (13.757-13.857 min, 17 scans) P2505759.D



Initial Calibration



Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC305-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Calibration Curves

Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
P041125A_CC252679_R1_QT_CC185154	Benzene	1	P2502258.D	4.96	47440	90.3	798559	1.081	0.21
P041125A_CC252679_R1_QT_CC185154	Benzene	2	P2502259.D	9.92	81580	90.3	784474	0.946	0.056
P041125A_CC252679_R1_QT_CC185154	Benzene	3	P2502260.D	19.85	154990	90.3	766291	0.920	0.027
P041125A_CC252679_R1_QT_CC185154	Benzene	4	P2502261.D	39.70	307006	90.3	776135	0.900	0.0043
P041125A_CC252679_R1_QT_CC185154	Benzene	5	P2502262.D	99.25	682063	90.3	768529	0.807	-0.099
P041125A_CC252679_R1_QT_CC185154	Benzene	6	P2502263.D	198.49	1383026	90.3	766391	0.821	-0.084
P041125A_CC252679_R1_QT_CC185154	Benzene	7	P2502264.D	595.47	3847989	90.3	733331	0.796	-0.11
P041125A_CC252679_R1_QT_CC185154	Benzene	QT		120.51		90.3			
						Avg:	770530	0.896	
						%RSD:	2.6%	11.2%	
P041125A_CC252679_R1_QT_CC185154	Toluene	1	P2502258.D	5.37	53462	105.3	981002	1.069	0.21
P041125A_CC252679_R1_QT_CC185154	Toluene	2	P2502259.D	10.73	93718	105.3	977723	0.940	0.061
P041125A_CC252679_R1_QT_CC185154	Toluene	3	P2502260.D	21.46	177517	105.3	952641	0.914	0.031
P041125A_CC252679_R1_QT_CC185154	Toluene	4	P2502261.D	42.92	366700	105.3	958584	0.938	0.058
P041125A_CC252679_R1_QT_CC185154	Toluene	5	P2502262.D	107.31	751020	105.3	937774	0.786	-0.11
P041125A_CC252679_R1_QT_CC185154	Toluene	6	P2502263.D	214.62	1544431	105.3	950212	0.797	-0.1
P041125A_CC252679_R1_QT_CC185154	Toluene	7	P2502264.D	643.86	4297415	105.3	923104	0.761	-0.14
P041125A_CC252679_R1_QT_CC185154	Toluene	QT		105.84		105.3			
						Avg:	954434	0.887	
						%RSD:	2.2%	12.5%	
P041125A_CC252679_R1_QT_CC185154	Ethylbenzene	1	P2502258.D	5.16	56966	105.3	981002	1.185	0.15
P041125A_CC252679_R1_QT_CC185154	Ethylbenzene	2	P2502259.D	10.32	108870	105.3	977723	1.137	0.1
P041125A_CC252679_R1_QT_CC185154	Ethylbenzene	3	P2502260.D	20.63	213920	105.3	952641	1.146	0.11
P041125A_CC252679_R1_QT_CC185154	Ethylbenzene	4	P2502261.D	41.26	422263	105.3	958584	1.124	0.088
P041125A_CC252679_R1_QT_CC185154	Ethylbenzene	5	P2502262.D	103.15	801782	105.3	937774	0.873	-0.16
P041125A_CC252679_R1_QT_CC185154	Ethylbenzene	6	P2502263.D	206.30	1736097	105.3	950212	0.932	-0.098
P041125A_CC252679_R1_QT_CC185154	Ethylbenzene	7	P2502264.D	618.90	4543663	105.3	923104	0.837	-0.19
P041125A_CC252679_R1_QT_CC185154	Ethylbenzene	QT		110.01		105.3			
						Avg:	954434	1.034	
						%RSD:	2.2%	14.2%	

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC305-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Calibration Curves


Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
P041125A_CC252679_R1_QT_CC185154	m-/p-Xylenes	1	P2502258.D	4.87	38899	105.3	981002	0.857	0.083
P041125A_CC252679_R1_QT_CC185154	m-/p-Xylenes	2	P2502259.D	9.74	73807	105.3	977723	0.816	0.03
P041125A_CC252679_R1_QT_CC185154	m-/p-Xylenes	3	P2502260.D	19.48	153415	105.3	952641	0.870	0.099
P041125A_CC252679_R1_QT_CC185154	m-/p-Xylenes	4	P2502261.D	38.97	310930	105.3	958584	0.876	0.11
P041125A_CC252679_R1_QT_CC185154	m-/p-Xylenes	5	P2502262.D	97.42	591523	105.3	937774	0.682	-0.14
P041125A_CC252679_R1_QT_CC185154	m-/p-Xylenes	6	P2502263.D	194.84	1339681	105.3	950212	0.762	-0.038
P041125A_CC252679_R1_QT_CC185154	m-/p-Xylenes	7	P2502264.D	584.52	3480142	105.3	923104	0.679	-0.14
P041125A_CC252679_R1_QT_CC185154	m-/p-Xylenes	QT		123.29		105.3			
						Avg:	954434	0.792	
						%RSD:	2.2%	10.8%	
P041125A_CC252679_R1_QT_CC185154	o-Xylene	1	P2502258.D	5.08	44039	105.3	981002	0.930	0.12
P041125A_CC252679_R1_QT_CC185154	o-Xylene	2	P2502259.D	10.16	84765	105.3	977723	0.898	0.078
P041125A_CC252679_R1_QT_CC185154	o-Xylene	3	P2502260.D	20.32	174712	105.3	952641	0.950	0.14
P041125A_CC252679_R1_QT_CC185154	o-Xylene	4	P2502261.D	40.64	330869	105.3	958584	0.894	0.073
P041125A_CC252679_R1_QT_CC185154	o-Xylene	5	P2502262.D	101.61	648711	105.3	937774	0.717	-0.14
P041125A_CC252679_R1_QT_CC185154	o-Xylene	6	P2502263.D	203.21	1413541	105.3	950212	0.771	-0.075
P041125A_CC252679_R1_QT_CC185154	o-Xylene	7	P2502264.D	609.64	3599376	105.3	923104	0.673	-0.19
P041125A_CC252679_R1_QT_CC185154	o-Xylene	QT		114.65		105.3			
						Avg:	954434	0.833	
						%RSD:	2.2%	13.3%	

Calibration Curves

Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
P041125A_CC252679_R1_QT_CC185154	Benzene	ICV	P2502265.D	64.22	465786	90.3	753743	0.869	-3.0%
P041125A_CC252679_R1_QT_CC185154	Toluene	ICV	P2502265.D	76.59	610878	105.3	957560	0.877	-1.1%
P041125A_CC252679_R1_QT_CC185154	Ethylbenzene	ICV	P2502265.D	86.22	753297	105.3	957560	0.961	-7.1%
P041125A_CC252679_R1_QT_CC185154	m-/p-Xylenes	ICV	P2502265.D	89.75	613496	105.3	957560	0.752	-5.1%
P041125A_CC252679_R1_QT_CC185154	o-Xylene	ICV	P2502265.D	88.34	625434	105.3	957560	0.778	-6.6%

M325B PDF Report ver.20250822

Sample Custody



EA Job No. 2025GC305 Page 70 of 71

**This Is The Last Page
Of This Report.**



Sprague - Searsport

70 Trundy Road
Searsport, ME 04974

Sampling Event 30 Sprague - Searsport

Client Project# PROJ-027966

Samples Received: 10/1/2025

Analytical Report 2025GC306

EPA Method 325B Analysis

Report Issue Date: 10/10/2025

I certify that to the best of my knowledge all analytical data presented in this report have been checked for completeness, accuracy, errors and legibility in addition to having been conducted in accordance with approved protocol, and that all deviations and analytical problems are summarized in the appropriate narrative(s). This report shall not be reproduced except in full without approval of the laboratory. This will provide assurance that parts of the report are not taken out of context.

Amendment(s):

Signature:


QA Review by Isabel Obando Marrero, Data Reviewer



Matt Cavanaugh
Matthew.Cavanaugh@enthalpy.com / www.enthalpy.com
O: (919) 850-4392
Enthalpy Analytical
800 Capitola Drive Suite 1 Durham, NC 27713

Table of Contents

Case Narrative	3
Results	6
Summary of Results	7
Detailed Results	8
QC Data	11
Chromatograms	14
Initial Calibration	66
Sample Custody	69
Chain of Custody	70

Narrative Summary



Enthalpy Analytical Narrative Summary

Company	Montrose Air Quality Services, LLC - New Jersey
Job No.	2025GC306-1
Client ID.	PROJ-027966 Site: Sprague - Searsport

1. Custody

The samples were received at Enthalpy Analytical on October 1, 2025 at 21 °C. The samples were received in good condition. Prior to, during, and after analysis, the samples were kept under lock with access only to authorized personnel by Enthalpy Analytical, LLC

Table 1 - Sample Inventory

Sample ID	Tube ID	Sample Type
SPRSEA-1-S-20250904	C33505	Sample
SPRSEA-2-S-20250904	C70563	Sample
SPRSEA-3-S-20250904	C01621	Sample
SPRSEA-4-S-20250904	C38576	Sample
SPRSEA-5-S-20250904	C70808	Sample
SPRSEA-6-S-20250904	C65284	Sample
SPRSEA-6-D-20250904	C31427	Duplicate
SPRSEA-6-B-20250904	C55482	Blank
SPRSEA-7-S-20250904	B53254	Sample
SPRSEA-8-S-20250904	C17220	Sample
SPRSEA-9-S-20250904	C33447	Sample
SPRSEA-10-S-20250904	C33770	Sample
SPRSEA-11-S-20250904	C70575	Sample
SPRSEA-12-S-20250904	C69685	Sample
SPRSEA-12-D-20250904	C43700	Duplicate
SPRSEA-12-B-20250904	C40528	Blank

2. Analysis

The samples were analyzed for Benzene, Toluene, Ethylbenzene, m-/p-Xylenes, and o-Xylene using EPA Method 325B – Volatile Organic Compounds from Fugitive and Area Sources by Thermal Desorption and GC/MS. A copy of the acquisition method M325B-TD35 is not included in this report but may be available upon request.

The sample tube media used for this sampling period was CarbopackX. All calibration standards and laboratory QC were prepared using the same media.

3. Calibration

All BFB tune criteria have been met for this analysis.

The initial calibration (N061725A_CC252679_QT_CC185154) met all 30% RSD criteria. The initial calibration verification met 30% difference criteria. The continuing calibration verifications met 30% difference criteria. The initial and continuing calibration raw data are not included in this report but are available upon request.

Enthalpy Analytical Narrative Summary

Company	Montrose Air Quality Services, LLC - New Jersey
Job No.	2025GC306-1
Client ID.	PROJ-027966 Site: Sprague - Searsport

5. QC Notes

All quality control criteria required by the method and/or the laboratory SOP have been met unless noted otherwise below.

The primary sample SPRSEA-6-S-20250904 (tube ID C65284) and its corresponding duplicate SPRSEA-6-D-20250904 (tube ID C31427) failed to meet the 30% difference criterion for m-/p-Xylenes as specified by the method. However, the concentrations of the analyte in both the sample and the duplicate were less than two times the reporting limit of the instrument's calibration curve. Therefore, the percent difference observed may not suggest the data set has been negatively affected. All samples in the data set have been flagged "Pc" for m-/p-Xylenes to denote this failure.

6. Reporting Notes

All tubes used for this sampling period met the method criteria for number of uses; no tube exceeded 50 field uses.

As specified in EPA Method 325B, the response factor of the daily continuing calibration standard was used to quantitate all field samples and blanks.

All samples were reported as amount in ng catch, and concentration in ug/m³ and ppbv.

The results presented in this report are representative of the samples as provided to the laboratory. These analyses met the requirements of the TNI Standard. Any deviations from the requirements of the reference method or TNI Standard have been stated above.

Enthalpy Analytical, located at 800 Capitola Drive, Suite 1, Durham NC, 27713 is accredited by the Louisiana Department of Environmental Quality (LDEQ) for EPA Method 325B for all analytes included in this report under **Certificate Number 04010**.

Results

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC306-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Summary

Sample Code	Tube ID	Benzene		Toluene		Ethylbenzene		m-/p-Xylenes		o-Xylene	
		(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag	(ug/m³)	Flag
SPRSEA-1-S-20250904	C33505	1.22		2.68		0.518	J	1.59	Pc	0.659	
SPRSEA-2-S-20250904	C70563	1.09		2.40		0.463	J	1.77	Pc	0.712	
SPRSEA-3-S-20250904	C01621	1.91		2.09		0.384	J	1.21	Pc	0.483	J
SPRSEA-4-S-20250904	C38576	1.75		1.33		0.274	ND	0.666	Pc	0.274	ND
SPRSEA-5-S-20250904	C70808	0.754		1.33		0.274	ND	0.934	Pc	0.369	J
SPRSEA-6-S-20250904	C65284	0.675		1.12		0.274	ND	0.859	Pc	0.354	J
SPRSEA-6-D-20250904	C31427	0.834		0.995		0.274	ND	0.591	Pc	0.274	ND
SPRSEA-6-B-20250904	C55482	0.188	ND	0.242	ND	0.274	ND	0.274	ND,Pc	0.274	ND
SPRSEA-7-S-20250904	B53254	0.968		0.924		0.274	ND	0.464	J,Pc	0.274	ND
SPRSEA-8-S-20250904	C17220	0.914		1.18		0.274	ND	0.605	Pc	0.274	ND
SPRSEA-9-S-20250904	C33447	1.12		2.55		0.527	J	1.62	Pc	0.695	
SPRSEA-10-S-20250904	C33770	3.80		11.1		2.04		7.27	Pc	2.99	
SPRSEA-11-S-20250904	C70575	2.80		8.02		1.48		5.91	Pc	2.34	
SPRSEA-12-S-20250904	C69685	2.52		7.36		1.37		5.33	Pc	2.16	
SPRSEA-12-D-20250904	C43700	2.40		6.71		1.16		4.16	Pc	1.66	
SPRSEA-12-B-20250904	C40528	0.188	ND	0.242	ND	0.274	ND	0.274	ND,Pc	0.274	ND

J: Estimated Value - The analyte was detected between the Method Detection Limit and Reporting Limit

ND: The analyte was not present above the Method Detection Limit

Pc: Field duplicate(s) exceed 30%RPD. Concentrations of both samples in duplicate are near the reporting limit

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC306-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Benzene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp (°F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250904	C33505	1.22	0.382	16.2	60.3	0.659	20160	0.188	0.374	0.0589	0.117		N2505940.D	2025-10-01 21:26	1.050	8.088	136941	740539	92.1	8.039	3.3%
SPRSEA-2-S-20250904	C70563	1.09	0.341	14.5	60.3	0.659	20160	0.188	0.374	0.0589	0.117		N2505941.D	2025-10-01 22:06	1.050	8.089	122443	742355	92.1	8.034	3.6%
SPRSEA-3-S-20250904	C01621	1.91	0.599	25.4	60.3	0.659	20160	0.188	0.374	0.0589	0.117		N2505942.D	2025-10-01 22:46	1.050	8.089	205205	707053	92.1	8.034	-1.3%
SPRSEA-4-S-20250904	C38576	1.75	0.549	23.3	60.3	0.659	20160	0.188	0.374	0.0589	0.117		N2505943.D	2025-10-01 23:26	1.050	8.089	200972	755593	92.1	8.027	5.4%
SPRSEA-5-S-20250904	C70808	0.754	0.236	10.0	60.3	0.659	20160	0.188	0.374	0.0589	0.117		N2505944.D	2025-10-02 00:06	1.050	8.089	88057	769649	92.1	8.034	7.4%
SPRSEA-6-S-20250904	C65284	0.675	0.212	8.98	60.3	0.659	20160	0.188	0.374	0.0589	0.117		N2505945.D	2025-10-02 00:46	1.050	8.083	74321	725611	92.1	8.027	1.3%
SPRSEA-6-D-20250904	C31427	0.834	0.261	11.1	60.3	0.659	20160	0.188	0.374	0.0589	0.117		N2505946.D	2025-10-02 01:26	1.050	8.083	97723	772873	92.1	8.028	7.9%
SPRSEA-6-B-20250904	C55482	0.188	0.0589		60.3	0.659	20160	0.188	0.374	0.0589	0.117	ND	N2505939.D	2025-10-01 20:46	1.050	8.034	4460	753045	92.1	8.034	5.1%
SPRSEA-7-S-20250904	B53254	0.968	0.303	12.9	60.3	0.659	20160	0.188	0.374	0.0589	0.117		N2505947.D	2025-10-02 02:06	1.050	8.083	103874	707515	92.1	8.028	-1.3%
SPRSEA-8-S-20250904	C17220	0.914	0.286	12.2	60.3	0.659	20160	0.188	0.374	0.0589	0.117		N2505948.D	2025-10-02 02:46	1.050	8.083	106122	765118	92.1	8.028	6.8%
SPRSEA-9-S-20250904	C33447	1.12	0.352	14.9	60.3	0.659	20160	0.188	0.374	0.0589	0.117		N2505950.D	2025-10-02 04:25	1.050	8.083	134395	788471	92.1	8.027	10.0%
SPRSEA-10-S-20250904	C33770	3.80	1.19	50.5	60.3	0.659	20160	0.188	0.374	0.0589	0.117		N2505951.D	2025-10-02 05:05	1.050	8.083	458294	795899	92.1	8.028	11.1%
SPRSEA-11-S-20250904	C70575	2.80	0.876	37.2	60.3	0.659	20160	0.188	0.374	0.0589	0.117		N2505952.D	2025-10-02 05:45	1.050	8.083	337463	795621	92.1	8.028	11.0%
SPRSEA-12-S-20250904	C69685	2.52	0.789	33.5	60.3	0.659	20160	0.188	0.374	0.0589	0.117		N2505953.D	2025-10-02 06:25	1.050	8.083	311194	814899	92.1	8.028	13.7%
SPRSEA-12-D-20250904	C43700	2.40	0.752	31.9	60.3	0.659	20160	0.188	0.374	0.0589	0.117		N2505954.D	2025-10-02 07:05	1.050	8.083	254340	698802	92.1	8.027	-2.5%
SPRSEA-12-B-20250904	C40528	0.188	0.0589		60.3	0.659	20160	0.188	0.374	0.0589	0.117	ND	N2505955.D	2025-10-02 07:45	1.050	8.077	4332	692556	92.1	8.028	-3.3%

Toluene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp (°F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250904	C33505	2.68	0.711	27.6	60.3	0.512	20160	0.242	0.521	0.0643	0.138		N2505940.D	2025-10-01 21:26	0.952	10.694	253599	1047332	108.6	10.603	0.8%
SPRSEA-2-S-20250904	C70563	2.40	0.638	24.8	60.3	0.512	20160	0.242	0.521	0.0643	0.138		N2505941.D	2025-10-01 22:06	0.952	10.695	233864	1075486	108.6	10.597	3.5%
SPRSEA-3-S-20250904	C01621	2.09	0.555	21.6	60.3	0.512	20160	0.242	0.521	0.0643	0.138		N2505942.D	2025-10-01 22:46	0.952	10.695	191107	1010630	108.6	10.597	-2.7%
SPRSEA-4-S-20250904	C38576	1.33	0.354	13.8	60.3	0.512	20160	0.242	0.521	0.0643	0.138		N2505943.D	2025-10-01 23:26	0.952	10.695	128400	1065027	108.6	10.603	2.5%
SPRSEA-5-S-20250904	C70808	1.33	0.352	13.7	60.3	0.512	20160	0.242	0.521	0.0643	0.138		N2505944.D	2025-10-02 00:06	0.952	10.695	129820	1081595	108.6	10.603	4.1%
SPRSEA-6-S-20250904	C65284	1.12	0.298	11.6	60.3	0.512	20160	0.242	0.521	0.0643	0.138		N2505945.D	2025-10-02 00:46	0.952	10.689	104476	1027506	108.6	10.597	-1.1%
SPRSEA-6-D-20250904	C31427	0.995	0.264	10.3	60.3	0.512	20160	0.242	0.521	0.0643	0.138		N2505946.D	2025-10-02 01:26	0.952	10.689	97705	1085860	108.6	10.597	4.5%
SPRSEA-6-B-20250904	C55482	0.242	0.0643		60.3	0.512	20160	0.242	0.521	0.0643	0.138	ND	N2505939.D	2025-10-01 20:46	0.952	10.695	8524	1072817	108.6	10.603	3.3%
SPRSEA-7-S-20250904	B53254	0.924	0.245	9.54	60.3	0.512	20160	0.242	0.521	0.0643	0.138		N2505947.D	2025-10-02 02:06	0.952	10.689	85645	1024547	108.6	10.597	-1.4%
SPRSEA-8-S-20250904	C17220	1.18	0.313	12.2	60.3	0.512	20160	0.242	0.521	0.0643	0.138		N2505948.D	2025-10-02 02:46	0.952	10.689	115162	1079519	108.6	10.597	3.9%
SPRSEA-9-S-20250904	C33447	2.55	0.678	26.3	60.3	0.512	20160	0.242	0.521	0.0643	0.138		N2505950.D	2025-10-02 04:25	0.952	10.689	252734	1094994	108.6	10.597	5.4%
SPRSEA-10-S-20250904	C33770	11.1	2.94	114	60.3	0.512	20160	0.242	0.521	0.0643	0.138		N2505951.D	2025-10-02 05:05	0.952	10.689	1126282	1123364	108.6	10.597	8.2%
SPRSEA-11-S-20250904	C70575	8.02	2.13	82.7	60.3	0.512	20160	0.242	0.521	0.0643	0.138		N2505952.D	2025-10-02 05:45	0.952	10.689	799784	1102567	108.6	10.597	6.2%
SPRSEA-12-S-20250904	C69685	7.36	1.95	76.0	60.3	0.512	20160	0.242	0.521	0.0643	0.138		N2505953.D	2025-10-02 06:25	0.952	10.689	754233	1132793	108.6	10.597	9.1%
SPRSEA-12-D-20250904	C43700	6.71	1.78	69.2	60.3	0.512	20160	0.242	0.521	0.0643	0.138		N2505954.D	2025-10-02 07:05	0.952	10.689	610970	1006644	108.6	10.597	-3.1%
SPRSEA-12-B-20250904	C40528	0.242	0.0643		60.3	0.512	20160	0.242	0.521	0.0643	0.138	ND	N2505955.D	2025-10-02 07:45	0.952	10.683	10813	1026034	108.6	10.591	-1.2%

Ethylbenzene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp (°F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250904	C33505	0.518	0.119	4.73	60.3	0.453	20160	0.274	0.566	0.0631	0.131	J	N2505940.D	2025-10-01 21:26	1.011	12.817	46105	1047332	108.6	10.603	0.8%

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC306-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Ethylbenzene

Sample Code	Tube ID	Conc (ug/m ³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m ³)	LOQ (ug/m ³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-2-S-20250904	C70563	0.463	0.107	4.23	60.3	0.453	20160	0.274	0.566	0.0631	0.131	J	N2505941.D	2025-10-01 22:06	1.011	12.818	42337	1075486	108.6	10.597	3.5%
SPRSEA-3-S-20250904	C01621	0.384	0.0886	3.51	60.3	0.453	20160	0.274	0.566	0.0631	0.131	J	N2505942.D	2025-10-01 22:46	1.011	12.818	32996	1010630	108.6	10.597	-2.7%
SPRSEA-4-S-20250904	C38576	0.274	0.0631		60.3	0.453	20160	0.274	0.566	0.0631	0.131	ND	N2505943.D	2025-10-01 23:26	1.011	12.818	22015	1065027	108.6	10.603	2.5%
SPRSEA-5-S-20250904	C70808	0.274	0.0631		60.3	0.453	20160	0.274	0.566	0.0631	0.131	ND	N2505944.D	2025-10-02 00:06	1.011	12.818	23279	1081595	108.6	10.603	4.1%
SPRSEA-6-S-20250904	C65284	0.274	0.0631		60.3	0.453	20160	0.274	0.566	0.0631	0.131	ND	N2505945.D	2025-10-02 00:46	1.011	12.812	19494	1027506	108.6	10.597	-1.1%
SPRSEA-6-D-20250904	C31427	0.274	0.0631		60.3	0.453	20160	0.274	0.566	0.0631	0.131	ND	N2505946.D	2025-10-02 01:26	1.011	12.812	19801	1085860	108.6	10.597	4.5%
SPRSEA-6-B-20250904	C55482	0.274	0.0631		60.3	0.453	20160	0.274	0.566	0.0631	0.131	ND	N2505939.D	2025-10-01 20:46	1.011	12.818	1545	1072817	108.6	10.603	3.3%
SPRSEA-7-S-20250904	B53254	0.274	0.0631		60.3	0.453	20160	0.274	0.566	0.0631	0.131	ND	N2505947.D	2025-10-02 02:06	1.011	12.812	15260	1024547	108.6	10.597	-1.4%
SPRSEA-8-S-20250904	C17220	0.274	0.0631		60.3	0.453	20160	0.274	0.566	0.0631	0.131	ND	N2505948.D	2025-10-02 02:46	1.011	12.818	18811	1079519	108.6	10.597	3.9%
SPRSEA-9-S-20250904	C33447	0.527	0.121	4.81	60.3	0.453	20160	0.274	0.566	0.0631	0.131	J	N2505950.D	2025-10-02 04:25	1.011	12.812	48991	1094994	108.6	10.597	5.4%
SPRSEA-10-S-20250904	C33770	2.04	0.470	18.6	60.3	0.453	20160	0.274	0.566	0.0631	0.131		N2505951.D	2025-10-02 05:05	1.011	12.812	194603	1123364	108.6	10.597	8.2%
SPRSEA-11-S-20250904	C70575	1.48	0.342	13.6	60.3	0.453	20160	0.274	0.566	0.0631	0.131		N2505952.D	2025-10-02 05:45	1.011	12.812	139016	1102567	108.6	10.597	6.2%
SPRSEA-12-S-20250904	C69685	1.37	0.315	12.5	60.3	0.453	20160	0.274	0.566	0.0631	0.131		N2505953.D	2025-10-02 06:25	1.011	12.812	131621	1132793	108.6	10.597	9.1%
SPRSEA-12-D-20250904	C43700	1.16	0.268	10.6	60.3	0.453	20160	0.274	0.566	0.0631	0.131		N2505954.D	2025-10-02 07:05	1.011	12.812	99598	1006644	108.6	10.597	-3.1%
SPRSEA-12-B-20250904	C40528	0.274	0.0631		60.3	0.453	20160	0.274	0.566	0.0631	0.131	ND	N2505955.D	2025-10-02 07:45	1.011	12.812	2245	1026034	108.6	10.591	-1.2%

m-/p-Xylenes

Sample Code	Tube ID	Conc (ug/m ³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m ³)	LOQ (ug/m ³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250904	C33505	1.59	0.367	14.5	60.3	0.453	20160	0.274	0.535	0.0631	0.123	Pc	N2505940.D	2025-10-01 21:26	0.679	12.995	95076	1047332	108.6	10.603	0.8%
SPRSEA-2-S-20250904	C70563	1.77	0.408	16.2	60.3	0.453	20160	0.274	0.535	0.0631	0.123	Pc	N2505941.D	2025-10-01 22:06	0.679	12.989	108664	1075486	108.6	10.597	3.5%
SPRSEA-3-S-20250904	C01621	1.21	0.278	11.0	60.3	0.453	20160	0.274	0.535	0.0631	0.123	Pc	N2505942.D	2025-10-01 22:46	0.679	12.989	69575	1010630	108.6	10.597	-2.7%
SPRSEA-4-S-20250904	C38576	0.666	0.154	6.08	60.3	0.453	20160	0.274	0.535	0.0631	0.123	Pc	N2505943.D	2025-10-01 23:26	0.679	12.989	40485	1065027	108.6	10.603	2.5%
SPRSEA-5-S-20250904	C70808	0.934	0.215	8.53	60.3	0.453	20160	0.274	0.535	0.0631	0.123	Pc	N2505944.D	2025-10-02 00:06	0.679	12.989	57642	1081595	108.6	10.603	4.1%
SPRSEA-6-S-20250904	C65284	0.859	0.198	7.84	60.3	0.453	20160	0.274	0.535	0.0631	0.123	Pc	N2505945.D	2025-10-02 00:46	0.679	12.989	50360	1027506	108.6	10.597	-1.1%
SPRSEA-6-D-20250904	C31427	0.591	0.136	5.39	60.3	0.453	20160	0.274	0.535	0.0631	0.123	Pc	N2505946.D	2025-10-02 01:26	0.679	12.989	36599	1085860	108.6	10.597	4.5%
SPRSEA-6-B-20250904	C55482	0.274	0.0631		60.3	0.453	20160	0.274	0.535	0.0631	0.123	ND,Pc	N2505939.D	2025-10-01 20:46	0.679	13.001	1454	1072817	108.6	10.603	3.3%
SPRSEA-7-S-20250904	B53254	0.464	0.107	4.24	60.3	0.453	20160	0.274	0.535	0.0631	0.123	J,Pc	N2505947.D	2025-10-02 02:06	0.679	12.989	27135	1024547	108.6	10.597	-1.4%
SPRSEA-8-S-20250904	C17220	0.605	0.139	5.52	60.3	0.453	20160	0.274	0.535	0.0631	0.123	Pc	N2505948.D	2025-10-02 02:46	0.679	12.989	37241	1079519	108.6	10.597	3.9%
SPRSEA-9-S-20250904	C33447	1.62	0.374	14.8	60.3	0.453	20160	0.274	0.535	0.0631	0.123	Pc	N2505950.D	2025-10-02 04:25	0.679	12.989	101296	1094994	108.6	10.597	5.4%
SPRSEA-10-S-20250904	C33770	7.27	1.68	66.4	60.3	0.453	20160	0.274	0.535	0.0631	0.123	Pc	N2505951.D	2025-10-02 05:05	0.679	12.983	465858	1123364	108.6	10.597	8.2%
SPRSEA-11-S-20250904	C70575	5.91	1.36	53.9	60.3	0.453	20160	0.274	0.535	0.0631	0.123	Pc	N2505952.D	2025-10-02 05:45	0.679	12.983	371636	1102567	108.6	10.597	6.2%
SPRSEA-12-S-20250904	C69685	5.33	1.23	48.7	60.3	0.453	20160	0.274	0.535	0.0631	0.123	Pc	N2505953.D	2025-10-02 06:25	0.679	12.983	344710	1132793	108.6	10.597	9.1%
SPRSEA-12-D-20250904	C43700	4.16	0.958	37.9	60.3	0.453	20160	0.274	0.535	0.0631	0.123	Pc	N2505954.D	2025-10-02 07:05	0.679	12.983	238690	1006644	108.6	10.597	-3.1%
SPRSEA-12-B-20250904	C40528	0.274	0.0631		60.3	0.453	20160	0.274	0.535	0.0631	0.123	ND,Pc	N2505955.D	2025-10-02 07:45	0.679	12.983	3510	1026034	108.6	10.591	-1.2%

o-Xylene

Sample Code	Tube ID	Conc (ug/m ³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m ³)	LOQ (ug/m ³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-1-S-20250904	C33505	0.659	0.152	6.01	60.3	0.453	20160	0.274	0.558	0.0631	0.129		N2505940.D	2025-10-01 21:26	0.657	13.484	38122	1047332	108.6	10.603	0.8%
SPRSEA-2-S-20250904	C70563	0.712	0.164	6.50	60.3	0.453	20160	0.274	0.558	0.0631	0.129		N2505941.D	2025-10-01 22:06	0.657	13.485	42286	1075486	108.6	10.597	3.5%

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC306-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

o-Xylene

Sample Code	Tube ID	Conc (ug/m³)	Conc (ppbv)	Calc Amt (ng)	Temp ("F)	Uptake Rate (mL/min)	Sample Time (min)	LOD (ug/m³)	LOQ (ug/m³)	LOD (ppbv)	LOQ (ppbv)	Flags	Data File	Inj DateTime	CCV RRF	Ret Time (min)	Target Area	ISTD Area	ISTD Amt	ISTD RT	ISTD Change
SPRSEA-3-S-20250904	C01621	0.483	0.111	4.41	60.3	0.453	20160	0.274	0.558	0.0631	0.129	J	N2505942.D	2025-10-01 22:46	0.657	13.485	26981	1010630	108.6	10.597	-2.7%
SPRSEA-4-S-20250904	C38576	0.274	0.0631		60.3	0.453	20160	0.274	0.558	0.0631	0.129	ND	N2505943.D	2025-10-01 23:26	0.657	13.484	15994	1065027	108.6	10.603	2.5%
SPRSEA-5-S-20250904	C70808	0.369	0.0851	3.37	60.3	0.453	20160	0.274	0.558	0.0631	0.129	J	N2505944.D	2025-10-02 00:06	0.657	13.485	22068	1081595	108.6	10.603	4.1%
SPRSEA-6-S-20250904	C65284	0.354	0.0815	3.23	60.3	0.453	20160	0.274	0.558	0.0631	0.129	J	N2505945.D	2025-10-02 00:46	0.657	13.478	20070	1027506	108.6	10.597	-1.1%
SPRSEA-6-D-20250904	C31427	0.274	0.0631		60.3	0.453	20160	0.274	0.558	0.0631	0.129	ND	N2505946.D	2025-10-02 01:26	0.657	13.485	13844	1085860	108.6	10.597	4.5%
SPRSEA-6-B-20250904	C55482	0.274	0.0631		60.3	0.453	20160	0.274	0.558	0.0631	0.129	ND	N2505939.D	2025-10-01 20:46	0.657	13.485	680	1072817	108.6	10.603	3.3%
SPRSEA-7-S-20250904	B53254	0.274	0.0631		60.3	0.453	20160	0.274	0.558	0.0631	0.129	ND	N2505947.D	2025-10-02 02:06	0.657	13.485	10626	1024547	108.6	10.597	-1.4%
SPRSEA-8-S-20250904	C17220	0.274	0.0631		60.3	0.453	20160	0.274	0.558	0.0631	0.129	ND	N2505948.D	2025-10-02 02:46	0.657	13.485	14836	1079519	108.6	10.597	3.9%
SPRSEA-9-S-20250904	C33447	0.695	0.160	6.34	60.3	0.453	20160	0.274	0.558	0.0631	0.129		N2505950.D	2025-10-02 04:25	0.657	13.478	42024	1094994	108.6	10.597	5.4%
SPRSEA-10-S-20250904	C33770	2.99	0.689	27.3	60.3	0.453	20160	0.274	0.558	0.0631	0.129		N2505951.D	2025-10-02 05:05	0.657	13.479	185655	1123364	108.6	10.597	8.2%
SPRSEA-11-S-20250904	C70575	2.34	0.540	21.4	60.3	0.453	20160	0.274	0.558	0.0631	0.129		N2505952.D	2025-10-02 05:45	0.657	13.479	142742	1102567	108.6	10.597	6.2%
SPRSEA-12-S-20250904	C69685	2.16	0.499	19.7	60.3	0.453	20160	0.274	0.558	0.0631	0.129		N2505953.D	2025-10-02 06:25	0.657	13.478	135371	1132793	108.6	10.597	9.1%
SPRSEA-12-D-20250904	C43700	1.66	0.383	15.2	60.3	0.453	20160	0.274	0.558	0.0631	0.129		N2505954.D	2025-10-02 07:05	0.657	13.478	92510	1006644	108.6	10.597	-3.1%
SPRSEA-12-B-20250904	C40528	0.274	0.0631		60.3	0.453	20160	0.274	0.558	0.0631	0.129	ND	N2505955.D	2025-10-02 07:45	0.657	13.479	2035	1026034	108.6	10.591	-1.2%

J: Estimated Value - The analyte was detected between the Method Detection Limit and Reporting Limit

ND: The analyte was not present above the Method Detection Limit

Pc: Field duplicate(s) exceed 30%RPD. Concentrations of both samples in duplicate are near the reporting limit

QC Data



Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey
Job No.: 2025GC306-1 EPA Method 325B Analysis
Client No.: PROJ-027966 Site: Sprague - Searsport

QC Samples

Field Sample Type	Sample Code	Benzene		Toluene		Ethylbenzene		m-/p-Xylenes		o-Xylene	
Blanks (ug/m³)	SPRSEA-6-B-20250904	ND	Pass	ND	Pass	ND	Pass	ND	Pass	ND	Pass
	SPRSEA-12-B-20250904	ND	Pass	ND	Pass	ND	Pass	ND	Pass	ND	Pass
Duplicates (difference)	SPRSEA-6-D-20250904	21%	Pass	12%	Pass	ND	Pass	37%	Fail	ND	Pass
	SPRSEA-12-D-20250904	4.8%	Pass	9.3%	Pass	16%	Pass	25%	Pass	26%	Pass

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC306-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Benzene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	N2505937.D	C67323	Cal	1.050		1.050	3.5%	-1.8%		Pass	
2025GC306 Method Blank-1	N2505938.D	C35790	Blank			1.050			0.86%	Pass	ND
M325B CCV 5 REC	N2505949.D	B50962	Check	1.054		1.050	3.9%		16%	Pass	
M325B CCV 5	N2505956.D	C67416	Check	1.004		1.050	-0.98%		-4.7%	Pass	

Toluene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	N2505937.D	C67323	Cal	0.952		0.952	21%	-20%		Pass	
2025GC306 Method Blank-1	N2505938.D	C35790	Blank			0.952			1.5%	Pass	ND
M325B CCV 5 REC	N2505949.D	B50962	Check	0.934		0.952	18%		19%	Pass	
M325B CCV 5	N2505956.D	C67416	Check	0.857		0.952	8.6%		-1.6%	Pass	

Ethylbenzene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	N2505937.D	C67323	Cal	1.011		1.011	1.7%	-20%		Pass	
2025GC306 Method Blank-1	N2505938.D	C35790	Blank			1.011			1.5%	Pass	ND
M325B CCV 5 REC	N2505949.D	B50962	Check	1.033		1.011	4.0%		19%	Pass	
M325B CCV 5	N2505956.D	C67416	Check	0.971		1.011	-2.3%		-1.6%	Pass	

m-/p-Xylenes Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	N2505937.D	C67323	Cal	0.679		0.679	-7.8%	-20%		Pass	
2025GC306 Method Blank-1	N2505938.D	C35790	Blank			0.679			1.5%	Pass	ND
M325B CCV 5 REC	N2505949.D	B50962	Check	0.740		0.679	0.56%		19%	Pass	
M325B CCV 5	N2505956.D	C67416	Check	0.693		0.679	-5.8%		-1.6%	Pass	


o-Xylene Calibration and Blanks

Sample Code	Data File	Tube ID	Type	RRF	ICAL RRF	Last CCV RRF	RRF Change	ISTD Change vs ICal	ISTD Change vs Concal	Pass/Fail	Flags
M325B CCV 5	N2505937.D	C67323	Cal	0.657		0.657	-18%	-20%		Pass	
2025GC306 Method Blank-1	N2505938.D	C35790	Blank			0.657			1.5%	Pass	ND
M325B CCV 5 REC	N2505949.D	B50962	Check	0.802		0.657	0.64%		19%	Pass	
M325B CCV 5	N2505956.D	C67416	Check	0.714		0.657	-10%		-1.6%	Pass	

Chromatograms



Name 2025GC306 Method Blank-1
Comment C35790
Data File N2505938.D
Acq. Date-Time 10/1/2025 8:07:02 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
Benzene-d6 (IS)		8.040	8.064	722,741	
Benzene	Benzene-d6 (IS)	8.089	8.119	5,949	
Toluene-d8 (IS)		10.603	10.627	1,054,490	
Toluene	Toluene-d8 (IS)	10.695	10.719	12,668	
Ethylbenzene	Toluene-d8 (IS)	12.818	12.842	6,029	
m-/p-Xylenes	Toluene-d8 (IS)	12.995	13.038	5,361	
o-Xylene	Toluene-d8 (IS)	13.485	13.509	3,304	

Benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan N2505938.D




+ Scan (10.551-10.719 min, 28 scans) N2505938.D

**Toluene**


+ EIC (91.1) Scan N2505938.D




+ Scan (10.655-10.757 min, 17 scans) N2505938.D

**Ethylbenzene**


+ EIC (91.1) Scan N2505938.D




+ Scan (12.787-12.871 min, 14 scans) N2505938.D

**m-/p-Xylenes**

+ EIC (91.1) Scan N2505938.D




+ Scan (12.953-13.075 min, 20 scans) N2505938.D




o-Xylene

+ EIC (91.1) Scan N2505938.D




+ Scan (13.461-13.531 min, 11 scans) N2505938.D





Name SPRSEA-6-B-20250904
Comment C55482
Data File N2505939.D
Acq. Date-Time 10/1/2025 8:46:57 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
Benzene-d6 (IS)		8.034	8.064	753,045	
Benzene	Benzene-d6 (IS)	8.034	8.119	4,460	
Toluene-d8 (IS)		10.603	10.627	1,072,817	
Toluene	Toluene-d8 (IS)	10.695	10.719	8,524	
Ethylbenzene	Toluene-d8 (IS)	12.818	12.842	1,545	
m-/p-Xylenes	Toluene-d8 (IS)	13.001	13.038	1,454	
o-Xylene	Toluene-d8 (IS)	13.485	13.509	680	

Benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan N2505939.D




+ Scan (10.549-10.719 min, 28 scans) N2505939.D

**Toluene**


+ EIC (91.1) Scan N2505939.D




+ Scan (10.660-10.758 min, 16 scans) N2505939.D

**Ethylbenzene**


+ EIC (91.1) Scan N2505939.D




+ Scan (12.793-12.854 min, 10 scans) N2505939.D

**m-/p-Xylenes**

+ EIC (91.1) Scan N2505939.D




+ Scan (12.959-13.049 min, 14 scans) N2505939.D




o-Xylene

+ EIC (91.1) Scan N2505939.D




+ Scan (13.461-13.514 min, 8 scans) N2505939.D





Name SPRSEA-1-S-20250904
Comment C33505
Data File N2505940.D
Acq. Date-Time 10/1/2025 9:26:53 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
Benzene-d6 (IS)		8.039	8.064	740,539	
Benzene	Benzene-d6 (IS)	8.088	8.119	136,941	
Toluene-d8 (IS)		10.603	10.627	1,047,332	
Toluene	Toluene-d8 (IS)	10.694	10.719	253,599	
Ethylbenzene	Toluene-d8 (IS)	12.817	12.842	46,105	
m-/p-Xylenes	Toluene-d8 (IS)	12.995	13.038	95,076	
o-Xylene	Toluene-d8 (IS)	13.484	13.509	38,122	

Benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan N2505940.D




+ Scan (10.548-10.719 min, 29 scans) N2505940.D

**Toluene**


+ EIC (91.1) Scan N2505940.D




+ Scan (10.646-10.798 min, 26 scans) N2505940.D

**Ethylbenzene**


+ EIC (91.1) Scan N2505940.D




+ Scan (12.781-12.890 min, 18 scans) N2505940.D

**m-/p-Xylenes**

+ EIC (91.1) Scan N2505940.D




+ Scan (12.949-13.135 min, 31 scans) N2505940.D




o-Xylene

+ EIC (91.1) Scan N2505940.D




+ Scan (13.449-13.576 min, 21 scans) N2505940.D





Name SPRSEA-2-S-20250904
Comment C70563
Data File N2505941.D
Acq. Date-Time 10/1/2025 10:06:48 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
Benzene-d6 (IS)		8.034	8.064	742,355	
Benzene	Benzene-d6 (IS)	8.089	8.119	122,443	
Toluene-d8 (IS)		10.597	10.627	1,075,486	
Toluene	Toluene-d8 (IS)	10.695	10.719	233,864	
Ethylbenzene	Toluene-d8 (IS)	12.818	12.842	42,337	
m-/p-Xylenes	Toluene-d8 (IS)	12.989	13.038	108,664	
o-Xylene	Toluene-d8 (IS)	13.485	13.509	42,286	

Benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan N2505941.D




+ Scan (10.548-10.738 min, 32 scans) N2505941.D

**Toluene**


+ EIC (91.1) Scan N2505941.D




+ Scan (10.643-10.793 min, 25 scans) N2505941.D

**Ethylbenzene**


+ EIC (91.1) Scan N2505941.D




+ Scan (12.779-12.889 min, 18 scans) N2505941.D

**m-/p-Xylenes**

+ EIC (91.1) Scan N2505941.D




+ Scan (12.951-13.105 min, 26 scans) N2505941.D




o-Xylene

+ EIC (91.1) Scan N2505941.D




+ Scan (13.444-13.540 min, 16 scans) N2505941.D





Name SPRSEA-3-S-20250904
Comment C01621
Data File N2505942.D
Acq. Date-Time 10/1/2025 10:46:43 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
Benzene-d6 (IS)		8.034	8.064	707,053	
Benzene	Benzene-d6 (IS)	8.089	8.119	205,205	
Toluene-d8 (IS)		10.597	10.627	1,010,630	
Toluene	Toluene-d8 (IS)	10.695	10.719	191,107	
Ethylbenzene	Toluene-d8 (IS)	12.818	12.842	32,996	
m-/p-Xylenes	Toluene-d8 (IS)	12.989	13.038	69,575	
o-Xylene	Toluene-d8 (IS)	13.485	13.509	26,981	

Benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan N2505942.D




+ Scan (10.549-10.744 min, 32 scans) N2505942.D

**Toluene**


+ EIC (91.1) Scan N2505942.D




+ Scan (10.646-10.811 min, 28 scans) N2505942.D

**Ethylbenzene**


+ EIC (91.1) Scan N2505942.D




+ Scan (12.776-12.897 min, 20 scans) N2505942.D

**m-/p-Xylenes**

+ EIC (91.1) Scan N2505942.D




+ Scan (12.946-13.087 min, 23 scans) N2505942.D




o-Xylene

+ EIC (91.1) Scan N2505942.D




+ Scan (13.436-13.561 min, 21 scans) N2505942.D





Name SPRSEA-4-S-20250904
Comment C38576
Data File N2505943.D
Acq. Date-Time 10/1/2025 11:26:39 PM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
Benzene-d6 (IS)		8.027	8.064	755,593	
Benzene	Benzene-d6 (IS)	8.089	8.119	200,972	
Toluene-d8 (IS)		10.603	10.627	1,065,027	
Toluene	Toluene-d8 (IS)	10.695	10.719	128,400	
Ethylbenzene	Toluene-d8 (IS)	12.818	12.842	22,015	
m-/p-Xylenes	Toluene-d8 (IS)	12.989	13.038	40,485	
o-Xylene	Toluene-d8 (IS)	13.484	13.509	15,994	

Benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan N2505943.D




+ Scan (10.542-10.707 min, 28 scans) N2505943.D

**Toluene**


+ EIC (91.1) Scan N2505943.D




+ Scan (10.646-10.805 min, 27 scans) N2505943.D

**Ethylbenzene**


+ EIC (91.1) Scan N2505943.D




+ Scan (12.781-12.878 min, 16 scans) N2505943.D

**m-/p-Xylenes**

+ EIC (91.1) Scan N2505943.D




+ Scan (12.952-13.117 min, 28 scans) N2505943.D




o-Xylene

+ EIC (91.1) Scan N2505943.D




+ Scan (13.448-13.550 min, 17 scans) N2505943.D





Name SPRSEA-5-S-20250904
Comment C70808
Data File N2505944.D
Acq. Date-Time 10/2/2025 12:06:37 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
Benzene-d6 (IS)		8.034	8.064	769,649	
Benzene	Benzene-d6 (IS)	8.089	8.119	88,057	
Toluene-d8 (IS)		10.603	10.627	1,081,595	
Toluene	Toluene-d8 (IS)	10.695	10.719	129,820	
Ethylbenzene	Toluene-d8 (IS)	12.818	12.842	23,279	
m-/p-Xylenes	Toluene-d8 (IS)	12.989	13.038	57,642	
o-Xylene	Toluene-d8 (IS)	13.485	13.509	22,068	

Benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan N2505944.D




+ Scan (10.543-10.707 min, 27 scans) N2505944.D

**Toluene**


+ EIC (91.1) Scan N2505944.D




+ Scan (10.646-10.793 min, 25 scans) N2505944.D

**Ethylbenzene**


+ EIC (91.1) Scan N2505944.D




+ Scan (12.778-12.876 min, 16 scans) N2505944.D

**m-/p-Xylenes**

+ EIC (91.1) Scan N2505944.D




+ Scan (12.946-13.093 min, 24 scans) N2505944.D




o-Xylene

+ EIC (91.1) Scan N2505944.D




+ Scan (13.448-13.595 min, 24 scans) N2505944.D





Name SPRSEA-6-S-20250904
Comment C65284
Data File N2505945.D
Acq. Date-Time 10/2/2025 12:46:33 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
Benzene-d6 (IS)		8.027	8.064	725,611	
Benzene	Benzene-d6 (IS)	8.083	8.119	74,321	
Toluene-d8 (IS)		10.597	10.627	1,027,506	
Toluene	Toluene-d8 (IS)	10.689	10.719	104,476	
Ethylbenzene	Toluene-d8 (IS)	12.812	12.842	19,494	
m-/p-Xylenes	Toluene-d8 (IS)	12.989	13.038	50,360	
o-Xylene	Toluene-d8 (IS)	13.478	13.509	20,070	

Benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan N2505945.D




+ Scan (10.542-10.713 min, 29 scans) N2505945.D

**Toluene**


+ EIC (91.1) Scan N2505945.D




+ Scan (10.641-10.787 min, 24 scans) N2505945.D

**Ethylbenzene**


+ EIC (91.1) Scan N2505945.D




+ Scan (12.775-12.882 min, 17 scans) N2505945.D

**m-/p-Xylenes**

+ EIC (91.1) Scan N2505945.D




+ Scan (12.942-13.102 min, 26 scans) N2505945.D




o-Xylene

+ EIC (91.1) Scan N2505945.D




+ Scan (13.439-13.582 min, 24 scans) N2505945.D





Name SPRSEA-6-D-20250904
Comment C31427
Data File N2505946.D
Acq. Date-Time 10/2/2025 1:26:28 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
Benzene-d6 (IS)		8.028	8.064	772,873	
Benzene	Benzene-d6 (IS)	8.083	8.119	97,723	
Toluene-d8 (IS)		10.597	10.627	1,085,860	
Toluene	Toluene-d8 (IS)	10.689	10.719	97,705	
Ethylbenzene	Toluene-d8 (IS)	12.812	12.842	19,801	
m-/p-Xylenes	Toluene-d8 (IS)	12.989	13.038	36,599	
o-Xylene	Toluene-d8 (IS)	13.485	13.509	13,844	

Benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan N2505946.D




+ Scan (10.542-10.713 min, 29 scans) N2505946.D

**Toluene**


+ EIC (91.1) Scan N2505946.D




+ Scan (10.640-10.805 min, 28 scans) N2505946.D

**Ethylbenzene**


+ EIC (91.1) Scan N2505946.D




+ Scan (12.776-12.879 min, 16 scans) N2505946.D

**m-/p-Xylenes**

+ EIC (91.1) Scan N2505946.D




+ Scan (12.949-13.116 min, 27 scans) N2505946.D




o-Xylene

+ EIC (91.1) Scan N2505946.D




+ Scan (13.448-13.544 min, 15 scans) N2505946.D





Name SPRSEA-7-S-20250904
Comment B53254
Data File N2505947.D
Acq. Date-Time 10/2/2025 2:06:23 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxen X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
Benzene-d6 (IS)		8.028	8.064	707,515	
Benzene	Benzene-d6 (IS)	8.083	8.119	103,874	
Toluene-d8 (IS)		10.597	10.627	1,024,547	
Toluene	Toluene-d8 (IS)	10.689	10.719	85,645	
Ethylbenzene	Toluene-d8 (IS)	12.812	12.842	15,260	
m-/p-Xylenes	Toluene-d8 (IS)	12.989	13.038	27,135	
o-Xylene	Toluene-d8 (IS)	13.485	13.509	10,626	

Benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan N2505947.D




+ Scan (10.548-10.707 min, 27 scans) N2505947.D

**Toluene**


+ EIC (91.1) Scan N2505947.D




+ Scan (10.643-10.793 min, 25 scans) N2505947.D

**Ethylbenzene**


+ EIC (91.1) Scan N2505947.D




+ Scan (12.777-12.887 min, 18 scans) N2505947.D

**m-/p-Xylenes**

+ EIC (91.1) Scan N2505947.D




+ Scan (12.953-13.075 min, 20 scans) N2505947.D




o-Xylene

+ EIC (91.1) Scan N2505947.D




+ Scan (13.449-13.546 min, 15 scans) N2505947.D





Name SPRSEA-8-S-20250904
Comment C17220
Data File N2505948.D
Acq. Date-Time 10/2/2025 2:46:17 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
Benzene-d6 (IS)		8.028	8.064	765,118	
Benzene	Benzene-d6 (IS)	8.083	8.119	106,122	
Toluene-d8 (IS)		10.597	10.627	1,079,519	
Toluene	Toluene-d8 (IS)	10.689	10.719	115,162	
Ethylbenzene	Toluene-d8 (IS)	12.818	12.842	18,811	
m-/p-Xylenes	Toluene-d8 (IS)	12.989	13.038	37,241	
o-Xylene	Toluene-d8 (IS)	13.485	13.509	14,836	

Benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan N2505948.D




+ Scan (10.543-10.750 min, 34 scans) N2505948.D

**Toluene**


+ EIC (91.1) Scan N2505948.D




+ Scan (10.643-10.842 min, 33 scans) N2505948.D

**Ethylbenzene**


+ EIC (91.1) Scan N2505948.D




+ Scan (12.782-12.866 min, 13 scans) N2505948.D

**m-/p-Xylenes**

+ EIC (91.1) Scan N2505948.D




+ Scan (12.950-13.081 min, 22 scans) N2505948.D




o-Xylene

+ EIC (91.1) Scan N2505948.D




+ Scan (13.444-13.540 min, 16 scans) N2505948.D





Name SPRSEA-9-S-20250904
Comment C33447
Data File N2505950.D
Acq. Date-Time 10/2/2025 4:25:39 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
Benzene-d6 (IS)		8.027	8.064	788,471	
Benzene	Benzene-d6 (IS)	8.083	8.119	134,395	
Toluene-d8 (IS)		10.597	10.627	1,094,994	
Toluene	Toluene-d8 (IS)	10.689	10.719	252,734	
Ethylbenzene	Toluene-d8 (IS)	12.812	12.842	48,991	
m-/p-Xylenes	Toluene-d8 (IS)	12.989	13.038	101,296	
o-Xylene	Toluene-d8 (IS)	13.478	13.509	42,024	

Benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan N2505950.D




+ Scan (10.542-10.713 min, 29 scans) N2505950.D

**Toluene**


+ EIC (91.1) Scan N2505950.D




+ Scan (10.641-10.799 min, 26 scans) N2505950.D

**Ethylbenzene**


+ EIC (91.1) Scan N2505950.D




+ Scan (12.769-12.891 min, 21 scans) N2505950.D

**m-/p-Xylenes**

+ EIC (91.1) Scan N2505950.D




+ Scan (12.946-13.087 min, 22 scans) N2505950.D




o-Xylene

+ EIC (91.1) Scan N2505950.D




+ Scan (13.442-13.558 min, 20 scans) N2505950.D





Name SPRSEA-10-S-20250904
Comment C33770
Data File N2505951.D
Acq. Date-Time 10/2/2025 5:05:34 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxen X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
Benzene-d6 (IS)		8.028	8.064	795,899	
Benzene	Benzene-d6 (IS)	8.083	8.119	458,294	
Toluene-d8 (IS)		10.597	10.627	1,123,364	
Toluene	Toluene-d8 (IS)	10.689	10.719	1,126,282	
Ethylbenzene	Toluene-d8 (IS)	12.812	12.842	194,603	
m-/p-Xylenes	Toluene-d8 (IS)	12.983	13.038	465,858	
o-Xylene	Toluene-d8 (IS)	13.479	13.509	185,655	

Benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan N2505951.D




+ Scan (10.542-10.738 min, 32 scans) N2505951.D

**Toluene**


+ EIC (91.1) Scan N2505951.D




+ Scan (10.640-10.836 min, 33 scans) N2505951.D

**Ethylbenzene**


+ EIC (91.1) Scan N2505951.D




+ Scan (12.770-12.922 min, 25 scans) N2505951.D

**m-/p-Xylenes**

+ EIC (91.1) Scan N2505951.D




+ Scan (12.940-13.173 min, 39 scans) N2505951.D




o-Xylene

+ EIC (91.1) Scan N2505951.D




+ Scan (13.436-13.594 min, 26 scans) N2505951.D





Name SPRSEA-11-S-20250904
Comment C70575
Data File N2505952.D
Acq. Date-Time 10/2/2025 5:45:29 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
Benzene-d6 (IS)		8.028	8.064	795,621	
Benzene	Benzene-d6 (IS)	8.083	8.119	337,463	
Toluene-d8 (IS)		10.597	10.627	1,102,567	
Toluene	Toluene-d8 (IS)	10.689	10.719	799,784	
Ethylbenzene	Toluene-d8 (IS)	12.812	12.842	139,016	
m-/p-Xylenes	Toluene-d8 (IS)	12.983	13.038	371,636	
o-Xylene	Toluene-d8 (IS)	13.479	13.509	142,742	

Benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan N2505952.D




+ Scan (10.542-10.713 min, 29 scans) N2505952.D

**Toluene**


+ EIC (91.1) Scan N2505952.D




+ Scan (10.640-10.799 min, 27 scans) N2505952.D

**Ethylbenzene**


+ EIC (91.1) Scan N2505952.D




+ Scan (12.770-12.897 min, 21 scans) N2505952.D

**m-/p-Xylenes**

+ EIC (91.1) Scan N2505952.D




+ Scan (12.943-13.111 min, 28 scans) N2505952.D




o-Xylene

+ EIC (91.1) Scan N2505952.D




+ Scan (13.436-13.583 min, 25 scans) N2505952.D





Name SPRSEA-12-S-20250904
Comment C69685
Data File N2505953.D
Acq. Date-Time 10/2/2025 6:25:22 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxen X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
Benzene-d6 (IS)		8.028	8.064	814,899	
Benzene	Benzene-d6 (IS)	8.083	8.119	311,194	
Toluene-d8 (IS)		10.597	10.627	1,132,793	
Toluene	Toluene-d8 (IS)	10.689	10.719	754,233	
Ethylbenzene	Toluene-d8 (IS)	12.812	12.842	131,621	
m-/p-Xylenes	Toluene-d8 (IS)	12.983	13.038	344,710	
o-Xylene	Toluene-d8 (IS)	13.478	13.509	135,371	

Benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan N2505953.D




+ Scan (10.542-10.713 min, 28 scans) N2505953.D

**Toluene**


+ EIC (91.1) Scan N2505953.D




+ Scan (10.640-10.805 min, 28 scans) N2505953.D

**Ethylbenzene**


+ EIC (91.1) Scan N2505953.D




+ Scan (12.771-12.873 min, 17 scans) N2505953.D

**m-/p-Xylenes**

+ EIC (91.1) Scan N2505953.D




+ Scan (12.941-13.111 min, 28 scans) N2505953.D




o-Xylene

+ EIC (91.1) Scan N2505953.D




+ Scan (13.433-13.595 min, 27 scans) N2505953.D





Name SPRSEA-12-D-20250904
Comment C43700
Data File N2505954.D
Acq. Date-Time 10/2/2025 7:05:17 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxpack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
Benzene-d6 (IS)		8.027	8.064	698,802	
Benzene	Benzene-d6 (IS)	8.083	8.119	254,340	
Toluene-d8 (IS)		10.597	10.627	1,006,644	
Toluene	Toluene-d8 (IS)	10.689	10.719	610,970	
Ethylbenzene	Toluene-d8 (IS)	12.812	12.842	99,598	
m-/p-Xylenes	Toluene-d8 (IS)	12.983	13.038	238,690	
o-Xylene	Toluene-d8 (IS)	13.478	13.509	92,510	

Benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan N2505954.D




+ Scan (10.542-10.713 min, 29 scans) N2505954.D

**Toluene**


+ EIC (91.1) Scan N2505954.D




+ Scan (10.640-10.805 min, 28 scans) N2505954.D

**Ethylbenzene**


+ EIC (91.1) Scan N2505954.D




+ Scan (12.770-12.919 min, 24 scans) N2505954.D

**m-/p-Xylenes**

+ EIC (91.1) Scan N2505954.D




+ Scan (12.940-13.160 min, 36 scans) N2505954.D




o-Xylene

+ EIC (91.1) Scan N2505954.D




+ Scan (13.436-13.587 min, 24 scans) N2505954.D





Name SPRSEA-12-B-20250904
Comment C40528
Data File N2505955.D
Acq. Date-Time 10/2/2025 7:45:12 AM
Acq. Method File M325B-TD35
Tube Sorbent Carboxipack X
Analyze Quant Version 12.1
Report Quant Version 12.1

Sample Chromatogram





Name	ISTD	RT	ICAL RT	Resp.	Int. Flag
Benzene-d6 (IS)		8.028	8.064	692,556	
Benzene	Benzene-d6 (IS)	8.077	8.119	4,332	
Toluene-d8 (IS)		10.591	10.627	1,026,034	
Toluene	Toluene-d8 (IS)	10.683	10.719	10,813	
Ethylbenzene	Toluene-d8 (IS)	12.812	12.842	2,245	
m-/p-Xylenes	Toluene-d8 (IS)	12.983	13.038	3,510	
o-Xylene	Toluene-d8 (IS)	13.479	13.509	2,035	

Benzene-d6 (IS)




Benzene




Toluene-d8 (IS)


+ EIC (98.1) Scan N2505955.D




+ Scan (10.536-10.732 min, 33 scans) N2505955.D

**Toluene**


+ EIC (91.1) Scan N2505955.D




+ Scan (10.644-10.756 min, 18 scans) N2505955.D

**Ethylbenzene**


+ EIC (91.1) Scan N2505955.D




+ Scan (12.771-12.849 min, 13 scans) N2505955.D

**m-/p-Xylenes**

+ EIC (91.1) Scan N2505955.D




+ Scan (12.953-13.038 min, 14 scans) N2505955.D




o-Xylene

+ EIC (91.1) Scan N2505955.D



+ Scan (13.448-13.577 min, 22 scans) N2505955.D



Initial Calibration



Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC306-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Calibration Curves

Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
N061725A_CC252679_QT_CC185154	Benzene	1	N2503571.D	4.97	48303	92.1	707017	1.264	0.25
N061725A_CC252679_QT_CC185154	Benzene	2	N2503572.D	9.95	79140	92.1	714169	1.025	0.011
N061725A_CC252679_QT_CC185154	Benzene	3	N2503573.D	19.90	157152	92.1	739545	0.983	-0.031
N061725A_CC252679_QT_CC185154	Benzene	4	N2503574.D	39.80	299422	92.1	724978	0.955	-0.058
N061725A_CC252679_QT_CC185154	Benzene	5	N2503575.D	99.50	702411	92.1	737415	0.881	-0.13
N061725A_CC252679_QT_CC185154	Benzene	6	N2503576.D	198.99	1439381	92.1	730431	0.912	-0.1
N061725A_CC252679_QT_CC185154	Benzene	7	N2503577.D	596.98	5287554	92.1	755598	1.079	0.064
N061725A_CC252679_QT_CC185154	Benzene	QT		120.21		92.1			
						Avg:	729879	1.014	
						%RSD:	2.2%	12.7%	
N061725A_CC252679_QT_CC185154	Toluene	1	N2503571.D	5.38	59233	108.6	1246056	0.960	0.22
N061725A_CC252679_QT_CC185154	Toluene	2	N2503572.D	10.76	106197	108.6	1281424	0.837	0.06
N061725A_CC252679_QT_CC185154	Toluene	3	N2503573.D	21.52	189922	108.6	1314802	0.729	-0.076
N061725A_CC252679_QT_CC185154	Toluene	4	N2503574.D	43.03	379471	108.6	1294356	0.740	-0.062
N061725A_CC252679_QT_CC185154	Toluene	5	N2503575.D	107.58	847879	108.6	1312916	0.652	-0.17
N061725A_CC252679_QT_CC185154	Toluene	6	N2503576.D	215.16	1723866	108.6	1308216	0.665	-0.16
N061725A_CC252679_QT_CC185154	Toluene	7	N2503577.D	645.49	7209247	108.6	1288332	0.942	0.19
N061725A_CC252679_QT_CC185154	Toluene	QT		105.58		108.6			
						Avg:	1292300	0.789	
						%RSD:	1.9%	15.9%	
N061725A_CC252679_QT_CC185154	Ethylbenzene	1	N2503571.D	5.17	70546	108.6	1246056	1.189	0.2
N061725A_CC252679_QT_CC185154	Ethylbenzene	2	N2503572.D	10.34	128094	108.6	1281424	1.050	0.057
N061725A_CC252679_QT_CC185154	Ethylbenzene	3	N2503573.D	20.68	240657	108.6	1314802	0.961	-0.032
N061725A_CC252679_QT_CC185154	Ethylbenzene	4	N2503574.D	41.36	507093	108.6	1294356	1.029	0.036
N061725A_CC252679_QT_CC185154	Ethylbenzene	5	N2503575.D	103.41	961768	108.6	1312916	0.769	-0.23
N061725A_CC252679_QT_CC185154	Ethylbenzene	6	N2503576.D	206.82	2121196	108.6	1308216	0.852	-0.14
N061725A_CC252679_QT_CC185154	Ethylbenzene	7	N2503577.D	620.46	8117786	108.6	1288332	1.103	0.11
N061725A_CC252679_QT_CC185154	Ethylbenzene	QT		109.73		108.6			
						Avg:	1292300	0.993	
						%RSD:	1.9%	14.6%	
N061725A_CC252679_QT_CC185154	m-/p-Xylenes	1	N2503571.D	4.88	52754	108.6	1246056	0.942	0.28
N061725A_CC252679_QT_CC185154	m-/p-Xylenes	2	N2503572.D	9.77	92124	108.6	1281424	0.800	0.086
N061725A_CC252679_QT_CC185154	m-/p-Xylenes	3	N2503573.D	19.53	172515	108.6	1314802	0.730	-0.0091
N061725A_CC252679_QT_CC185154	m-/p-Xylenes	4	N2503574.D	39.07	370413	108.6	1294356	0.796	0.081
N061725A_CC252679_QT_CC185154	m-/p-Xylenes	5	N2503575.D	97.67	661384	108.6	1312916	0.560	-0.24
N061725A_CC252679_QT_CC185154	m-/p-Xylenes	6	N2503576.D	195.33	1391174	108.6	1308216	0.591	-0.2
N061725A_CC252679_QT_CC185154	m-/p-Xylenes	QT		122.98		108.6			
						Avg:	1292962	0.736	
						%RSD:	2.0%	19.4%	

Enthalpy Analytical

Company: Montrose Air Quality Services, LLC - New Jersey

Job No.: 2025GC306-1 EPA Method 325B Analysis

Client No.: PROJ-027966 Site: Sprague - Searsport

Calibration Curves

Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
N061725A_CC252679_QT_CC185154	o-Xylene	1	N2503571.D	5.09	58550	108.6	1246056	1.002	0.26
N061725A_CC252679_QT_CC185154	o-Xylene	2	N2503572.D	10.19	103745	108.6	1281424	0.863	0.083
N061725A_CC252679_QT_CC185154	o-Xylene	3	N2503573.D	20.37	195592	108.6	1314802	0.793	-0.0052
N061725A_CC252679_QT_CC185154	o-Xylene	4	N2503574.D	40.75	419953	108.6	1294356	0.865	0.085
N061725A_CC252679_QT_CC185154	o-Xylene	5	N2503575.D	101.86	720803	108.6	1312916	0.585	-0.27
N061725A_CC252679_QT_CC185154	o-Xylene	6	N2503576.D	203.73	1504617	108.6	1308216	0.613	-0.23
N061725A_CC252679_QT_CC185154	o-Xylene	7	N2503577.D	611.18	6228161	108.6	1288332	0.859	0.078
N061725A_CC252679_QT_CC185154	o-Xylene	QT		114.37		108.6			
						Avg:	1292300	0.797	
						%RSD:	1.9%	18.7%	

Calibration Curves

Method	Compound	Level	Cal File	Amount (ng)	Area	ISTD Amt (ng)	ISTD Area	RRF	Dev
N061725A_CC252679_QT_CC185154	Benzene	ICV	N2503582.D	63.84	448559	92.1	660511	0.979	-3.5%
N061725A_CC252679_QT_CC185154	Toluene	ICV	N2503582.D	76.15	626235	108.6	1150784	0.776	-1.6%
N061725A_CC252679_QT_CC185154	Ethylbenzene	ICV	N2503582.D	85.72	790798	108.6	1150784	0.871	-12.0%
N061725A_CC252679_QT_CC185154	m-/p-Xylenes	ICV	N2503582.D	89.23	607838	108.6	1150784	0.643	-13.0%
N061725A_CC252679_QT_CC185154	o-Xylene	ICV	N2503582.D	87.83	624336	108.6	1150784	0.671	-16.0%

M325B PDF Report ver.20250917

Sample Custody

[illegible]

**This Is The Last Page
Of This Report.**