



ANALYTICAL REPORT

Lab Number:	L2449150
Client:	Maine DEP-Div. of Technical Services 17 State House Station Augusta, ME 04333
ATTN:	Molly King
Phone:	(207) 287-8169
Project Name:	NASB-HANGAR 4 AFFF
Project Number:	83336
Report Date:	09/12/24

The original project report/data package is held by Alpha Analytical. This report/data package is paginated and should be reproduced only in its entirety. Alpha Analytical holds no responsibility for results and/or data that are not consistent with the original.

Certifications & Approvals: MA (M-MA030), NH NELAP (2062), CT (PH-0825), DoD (L2474), FL (E87814), IL (200081), IN (C-MA-04), KY (KY98046), LA (85084), ME (MA00030), MD (350), MI (9110), MN (025-999-495), NJ (MA015), NY (11627), NC (685), OR (MA-0262), PA (68-02089), RI (LAO00299), TX (T104704419), VT (VT-0015), VA (460194), WA (C954), US Army Corps of Engineers, USDA (Permit #525-23-107-88708A1), USFWS (Permit #A24920).

320 Forbes Boulevard, Mansfield, MA 02048-1806
508-822-9300 (Fax) 508-822-3288 800-624-9220 - www.alphalab.com



Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

Alpha Sample ID	Client ID	Matrix	Sample Location	Collection Date/Time	Receive Date
L2449150-01	SW68-20240827	WATER	BRUNSWICK	08/27/24 11:00	08/28/24
L2449150-02	SW67-20240827	WATER	BRUNSWICK	08/27/24 12:00	08/28/24
L2449150-03	ARB2-SW-2024-08-27	WATER	BRUNSWICK	08/27/24 14:30	08/28/24
L2449150-04	ARB1-SW-2024-08-27	WATER	BRUNSWICK	08/27/24 14:10	08/28/24
L2449150-05	ART-SW-2024-08-27	WATER	BRUNSWICK	08/27/24 13:15	08/28/24

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

Case Narrative

The samples were received in accordance with the Chain of Custody and no significant deviations were encountered during the preparation or analysis unless otherwise noted. Sample Receipt, Container Information, and the Chain of Custody are located at the back of the report.

Results contained within this report relate only to the samples submitted under this Alpha Lab Number and meet NELAP requirements for all NELAP accredited parameters unless otherwise noted in the following narrative. The data presented in this report is organized by parameter (i.e. VOC, SVOC, etc.). Sample specific Quality Control data (i.e. Surrogate Spike Recovery) is reported at the end of the target analyte list for each individual sample, followed by the Laboratory Batch Quality Control at the end of each parameter. Tentatively Identified Compounds (TICs), if requested, are reported for compounds identified to be present and are not part of the method/program Target Compound List, even if only a subset of the TCL are being reported. If a sample was re-analyzed or re-extracted due to a required quality control corrective action and if both sets of data are reported, the Laboratory ID of the re-analysis or re-extraction is designated with an "R" or "RE", respectively.

When multiple Batch Quality Control elements are reported (e.g. more than one LCS), the associated samples for each element are noted in the grey shaded header line of each data table. Any Laboratory Batch, Sample Specific % recovery or RPD value that is outside the listed Acceptance Criteria is bolded in the report. In reference to questions H (CAM) or 4 (RCP) when "NO" is checked, the performance criteria for CAM and RCP methods allow for some quality control failures to occur and still be within method compliance. In these instances, the specific failure is not narrated but noted in the associated QC Outlier Summary Report, located directly after the Case Narrative. QC information is also incorporated in the Data Usability Assessment table (Format 11) of our Data Merger tool, where it can be reviewed in conjunction with the sample result, associated regulatory criteria and any associated data usability implications.

Soil/sediments and solids are reported on a dry weight basis unless otherwise noted. Tissues are reported "as received" or on a wet weight basis, unless otherwise noted. Definitions of all data qualifiers and acronyms used in this report are provided in the Glossary located at the back of the report.

HOLD POLICY - For samples submitted on hold, Alpha's policy is to hold samples (with the exception of Air canisters) free of charge for 21 calendar days from the date the project is completed. After 21 calendar days, we will dispose of all samples submitted including those put on hold unless you have contacted your Alpha Project Manager and made arrangements for Alpha to continue to hold the samples. Air canisters will be disposed after 3 business days from the date the project is completed.

Please contact Project Management at 800-624-9220 with any questions.

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

Case Narrative (continued)

Report Submission

All non-detect (ND) or estimated concentrations (J-qualified) have been quantitated to the limit noted in the MDL column.

Perfluorinated Alkyl Acids by 1633

L2449150-01R, -02R, WG1969890-1R, WG1969890-2R, and WG1969890-3R: The sample was re-analyzed due to QC failures in the original analysis. The results of the re-analysis are reported.

L2449150-02R: The sample has elevated detection limits due to the dilution required by the elevated concentrations of target compounds in the sample.

L2449150-03RE: The sample was re-extracted within the method required holding time due to overdilution of the original extraction. The re-extraction is reported.

L2449150-04RE and -05RE: The sample was re-extracted within holding time due to overdilution of the original extraction. The results of the re-extraction are reported.

I, the undersigned, attest under the pains and penalties of perjury that, to the best of my knowledge and belief and based upon my personal inquiry of those responsible for providing the information contained in this analytical report, such information is accurate and complete. This certificate of analysis is not complete unless this page accompanies any and all pages of this report.

Authorized Signature:

 Ashley Leitao

Title: Technical Director/Representative

Date: 09/12/24

ORGANICS

SEMIVOLATILES

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

SAMPLE RESULTS

Lab ID: L2449150-01 R
 Client ID: SW68-20240827
 Sample Location: BRUNSWICK

Date Collected: 08/27/24 11:00
 Date Received: 08/28/24
 Field Prep: Not Specified

Sample Depth:

Matrix: Water
 Analytical Method: 144,1633
 Analytical Date: 09/11/24 15:15
 Analyst: ANH

Extraction Method: EPA 1633
 Extraction Date: 09/10/24 18:45

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab						
Perfluorobutanoic Acid (PFBA)	5.23	J	ng/l	6.25	1.00	1
Perfluoropentanoic Acid (PFPeA)	2.96	J	ng/l	3.12	0.836	1
Perfluorobutanesulfonic Acid (PFBS)	4.86		ng/l	1.56	0.523	1
1H,1H,2H,2H-Perfluorohexanesulfonic Acid (4:2FTS)	ND		ng/l	6.25	1.63	1
Perfluorohexanoic Acid (PFHxA)	5.24		ng/l	1.56	0.461	1
Perfluoropentanesulfonic Acid (PFPeS)	5.25		ng/l	1.56	0.273	1
Perfluoroheptanoic Acid (PFHpA)	1.44	J	ng/l	1.56	0.312	1
Perfluorohexanesulfonic Acid (PFHxS)	16.0		ng/l	1.56	0.375	1
Perfluorooctanoic Acid (PFOA)	3.47		ng/l	1.56	0.680	1
1H,1H,2H,2H-Perfluorooctanesulfonic Acid (6:2FTS)	10.2		ng/l	6.25	2.11	1
Perfluoroheptanesulfonic Acid (PFHpS)	1.54	J	ng/l	1.56	0.422	1
Perfluorononanoic Acid (PFNA)	ND		ng/l	1.56	0.492	1
Perfluorooctanesulfonic Acid (PFOS)	49.3		ng/l	1.56	0.711	1
Perfluorodecanoic Acid (PFDA)	ND		ng/l	1.56	0.633	1
1H,1H,2H,2H-Perfluorodecanesulfonic Acid (8:2FTS)	ND		ng/l	6.25	2.43	1
Perfluoronanesulfonic Acid (PFNS)	ND		ng/l	1.56	0.484	1
N-Methyl Perfluorooctanesulfonamidoacetic Acid (NMeFOSAA)	ND		ng/l	1.56	0.852	1
Perfluoroundecanoic Acid (PFUnA)	ND		ng/l	1.56	0.680	1
Perfluorodecanesulfonic Acid (PFDS)	ND		ng/l	1.56	0.359	1
Perfluorooctanesulfonamide (PFOSA)	ND		ng/l	1.56	0.422	1
N-Ethyl Perfluorooctanesulfonamidoacetic Acid (NEtFOSAA)	ND		ng/l	1.56	0.844	1
Perfluorododecanoic Acid (PFDoA)	ND		ng/l	1.56	0.719	1
Perfluorotridecanoic Acid (PFTrDA)	ND		ng/l	1.56	0.586	1
Perfluorotetradecanoic Acid (PFTeDA)	ND		ng/l	1.56	0.414	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	ND		ng/l	6.25	0.875	1
4,8-Dioxa-3h-Perfluorononanoic Acid (ADONA)	ND		ng/l	6.25	0.984	1
Perfluorododecanesulfonic Acid (PFDoS)	ND		ng/l	1.56	0.594	1

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

SAMPLE RESULTS

Lab ID: L2449150-01 R
Client ID: SW68-20240827
Sample Location: BRUNSWICK

Date Collected: 08/27/24 11:00
Date Received: 08/28/24
Field Prep: Not Specified

Sample Depth:

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab						
9-Chlorohexadecafluoro-3-Oxanone-1-Sulfonic Acid (9Cl-PF3ONS)	ND		ng/l	6.25	1.29	1
11-Chloroeicosafuoro-3-Oxaundecane-1-Sulfonic Acid (11Cl-PF3OUdS)	ND		ng/l	6.25	1.29	1
N-Methyl Perfluorooctane Sulfonamide (NMeFOSA)	ND		ng/l	1.56	0.680	1
N-Ethyl Perfluorooctane Sulfonamide (NEtFOSA)	ND		ng/l	1.56	0.719	1
N-Methyl Perfluorooctanesulfonamido Ethanol (NMeFOSE)	ND		ng/l	15.6	3.67	1
N-Ethyl Perfluorooctanesulfonamido Ethanol (NEtFOSE)	ND		ng/l	15.6	1.91	1
Perfluoro-3-Methoxypropanoic Acid (PFMPA)	ND		ng/l	3.12	0.445	1
Perfluoro-4-Methoxybutanoic Acid (PFMBA)	ND		ng/l	3.12	0.414	1
Perfluoro(2-Ethoxyethane)Sulfonic Acid (PFEEESA)	ND		ng/l	3.12	0.344	1
Nonafluoro-3,6-Dioxaheptanoic Acid (NFDHA)	ND		ng/l	3.12	1.84	1
3-Perfluoropropyl Propanoic Acid (3:3FTCA)	ND		ng/l	7.81	2.58	1
2H,2H,3H,3H-Perfluorooctanoic Acid (5:3FTCA)	ND		ng/l	39.1	9.14	1
3-Perfluoroheptyl Propanoic Acid (7:3FTCA)	ND		ng/l	39.1	6.16	1

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

SAMPLE RESULTS

Lab ID: L2449150-01 R
 Client ID: SW68-20240827
 Sample Location: BRUNSWICK

Date Collected: 08/27/24 11:00
 Date Received: 08/28/24
 Field Prep: Not Specified

Sample Depth:

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab						

Surrogate	% Recovery	Qualifier	Acceptance Criteria
Perfluoro-n-[13C4]Butanoic Acid (13C4-PFBA)	76		41-123
Perfluoro-n-[13C5]Pentanoic Acid (13C5-PFPeA)	84		29-123
Perfluoro-1-[2,3,4-13C3]Butanesulfonic Acid (13C3-PFBS)	88		41-125
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Hexanesulfonic Acid (13C2-4:2FTS)	89		10-290
Perfluoro-n-[1,2,3,4,6-13C5]Hexanoic Acid (13C5-PFHxA)	83		40-121
Perfluoro-n-[1,2,3,4-13C4]Heptanoic Acid (13C4-PFHpA)	83		27-156
Perfluoro-1-[1,2,3-13C3]Hexanesulfonic Acid (13C3-PFHxS)	80		46-115
Perfluoro-n-[13C8]Octanoic Acid (13C8-PFOA)	79		39-121
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Octanesulfonic Acid (13C2-6:2FTS)	73		10-261
Perfluoro-n-[13C9]Nonanoic Acid (13C9-PFNA)	74		38-114
Perfluoro-1-[13C8]Octanesulfonic Acid (13C8-PFOS)	71		32-114
Perfluoro-n-[1,2,3,4,5,6-13C6]Decanoic Acid (13C6-PFDA)	69		28-115
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Decanesulfonic Acid (13C2-8:2FTS)	54		10-213
N-Methyl-d3-perfluoro-1-octanesulfonamidoacetic Acid (D3-NMeFOSAA)	63		10-172
Perfluoro-n-[1,2,3,4,5,6,7-13C7]Undecanoic Acid (13C7-PFUnA)	89		16-123
Perfluoro-1-[13C8]Octanesulfonamide (13C8-PFOSA)	61		14-108
N-Ethyl-d5-perfluoro-1-octanesulfonamidoacetic Acid (D5-NEtFOSAA)	75		10-150
Perfluoro-n-[1,2-13C2]Dodecanoic Acid (13C2-PFDoA)	81		10-126
Perfluoro-n-[1,2-13C2]Tetradecanoic Acid (13C2-PFTeDA)	92		10-145
Tetrafluoro-2-heptafluoropropoxy-[13C3]-propanoic acid (13C3-HFPO-DA)	88		35-142
N-Methyl-d3-Perfluoro-1-Octanesulfonamide (D3-NMeFOSA)	78		11-94
N-Ethyl-d5-Perfluoro-1-Octanesulfonamide (D5-NEtFOSA)	91		11-97
N-Methyl-d7-Perfluorooctanesulfonamidoethanol (D7-NMeFOSE)	88		10-137
N-Ethyl-d9-Perfluorooctanesulfonamidoethanol (D9-NEtFOSE)	92		10-130

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

SAMPLE RESULTS

Lab ID: L2449150-02 R
 Client ID: SW67-20240827
 Sample Location: BRUNSWICK

Date Collected: 08/27/24 12:00
 Date Received: 08/28/24
 Field Prep: Not Specified

Sample Depth:

Matrix: Water
 Analytical Method: 144,1633
 Analytical Date: 09/11/24 15:28
 Analyst: ANH

Extraction Method: EPA 1633
 Extraction Date: 09/10/24 18:45

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab						
Perfluorobutanoic Acid (PFBA)	257	J	ng/l	320	51.2	1
Perfluoropentanoic Acid (PFPeA)	180		ng/l	160	42.8	1
Perfluorobutanesulfonic Acid (PFBS)	527		ng/l	80.0	26.8	1
1H,1H,2H,2H-Perfluorohexanesulfonic Acid (4:2FTS)	ND		ng/l	320	83.6	1
Perfluorohexanoic Acid (PFHxA)	396		ng/l	80.0	23.6	1
Perfluoropentanesulfonic Acid (PFPeS)	538		ng/l	80.0	14.0	1
Perfluoroheptanoic Acid (PFHpA)	101		ng/l	80.0	16.0	1
Perfluorohexanesulfonic Acid (PFHxS)	1690		ng/l	80.0	19.2	1
Perfluorooctanoic Acid (PFOA)	358		ng/l	80.0	34.8	1
1H,1H,2H,2H-Perfluorooctanesulfonic Acid (6:2FTS)	930		ng/l	320	108.	1
Perfluoroheptanesulfonic Acid (PFHpS)	188		ng/l	80.0	21.6	1
Perfluorononanoic Acid (PFNA)	ND		ng/l	80.0	25.2	1
Perfluorooctanesulfonic Acid (PFOS)	8160		ng/l	80.0	36.4	1
Perfluorodecanoic Acid (PFDA)	ND		ng/l	80.0	32.4	1
1H,1H,2H,2H-Perfluorodecanesulfonic Acid (8:2FTS)	ND		ng/l	320	124.	1
Perfluoronanesulfonic Acid (PFNS)	ND		ng/l	80.0	24.8	1
N-Methyl Perfluorooctanesulfonamidoacetic Acid (NMeFOSAA)	ND		ng/l	80.0	43.6	1
Perfluoroundecanoic Acid (PFUnA)	ND		ng/l	80.0	34.8	1
Perfluorodecanesulfonic Acid (PFDS)	ND		ng/l	80.0	18.4	1
Perfluorooctanesulfonamide (PFOSA)	ND		ng/l	80.0	21.6	1
N-Ethyl Perfluorooctanesulfonamidoacetic Acid (NEtFOSAA)	ND		ng/l	80.0	43.2	1
Perfluorododecanoic Acid (PFDoA)	ND		ng/l	80.0	36.8	1
Perfluorotridecanoic Acid (PFTrDA)	ND		ng/l	80.0	30.0	1
Perfluorotetradecanoic Acid (PFTeDA)	ND		ng/l	80.0	21.2	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	ND		ng/l	320	44.8	1
4,8-Dioxa-3h-Perfluorononanoic Acid (ADONA)	ND		ng/l	320	50.4	1
Perfluorododecanesulfonic Acid (PFDoS)	ND		ng/l	80.0	30.4	1

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

SAMPLE RESULTS

Lab ID: L2449150-02 R
Client ID: SW67-20240827
Sample Location: BRUNSWICK

Date Collected: 08/27/24 12:00
Date Received: 08/28/24
Field Prep: Not Specified

Sample Depth:

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab						
9-Chlorohexadecafluoro-3-Oxanone-1-Sulfonic Acid (9Cl-PF3ONS)	ND		ng/l	320	66.0	1
11-Chloroeicosafuoro-3-Oxaundecane-1-Sulfonic Acid (11Cl-PF3OUdS)	ND		ng/l	320	66.0	1
N-Methyl Perfluorooctane Sulfonamide (NMeFOSA)	ND		ng/l	80.0	34.8	1
N-Ethyl Perfluorooctane Sulfonamide (NEtFOSA)	ND		ng/l	80.0	36.8	1
N-Methyl Perfluorooctanesulfonamido Ethanol (NMeFOSE)	ND		ng/l	800	188.	1
N-Ethyl Perfluorooctanesulfonamido Ethanol (NEtFOSE)	ND		ng/l	800	98.0	1
Perfluoro-3-Methoxypropanoic Acid (PFMPA)	ND		ng/l	160	22.8	1
Perfluoro-4-Methoxybutanoic Acid (PFMBA)	ND		ng/l	160	21.2	1
Perfluoro(2-Ethoxyethane)Sulfonic Acid (PFEEESA)	ND		ng/l	160	17.6	1
Nonafluoro-3,6-Dioxaheptanoic Acid (NFDHA)	ND		ng/l	160	94.4	1
3-Perfluoropropyl Propanoic Acid (3:3FTCA)	ND		ng/l	400	132.	1
2H,2H,3H,3H-Perfluorooctanoic Acid (5:3FTCA)	ND		ng/l	2000	468.	1
3-Perfluoroheptyl Propanoic Acid (7:3FTCA)	ND		ng/l	2000	316.	1

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

SAMPLE RESULTS

Lab ID: L2449150-02 R
Client ID: SW67-20240827
Sample Location: BRUNSWICK

Date Collected: 08/27/24 12:00
Date Received: 08/28/24
Field Prep: Not Specified

Sample Depth:

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab						

Surrogate	% Recovery	Qualifier	Acceptance Criteria
Perfluoro-n-[13C4]Butanoic Acid (13C4-PFBA)	83		41-123
Perfluoro-n-[13C5]Pentanoic Acid (13C5-PFPeA)	93		29-123
Perfluoro-1-[2,3,4-13C3]Butanesulfonic Acid (13C3-PFBS)	92		41-125
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Hexanesulfonic Acid (13C2-4:2FTS)	87		10-290
Perfluoro-n-[1,2,3,4,6-13C5]Hexanoic Acid (13C5-PFHxA)	90		40-121
Perfluoro-n-[1,2,3,4-13C4]Heptanoic Acid (13C4-PFHpA)	89		27-156
Perfluoro-1-[1,2,3-13C3]Hexanesulfonic Acid (13C3-PFHxS)	86		46-115
Perfluoro-n-[13C8]Octanoic Acid (13C8-PFOA)	83		39-121
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Octanesulfonic Acid (13C2-6:2FTS)	76		10-261
Perfluoro-n-[13C9]Nonanoic Acid (13C9-PFNA)	89		38-114
Perfluoro-1-[13C8]Octanesulfonic Acid (13C8-PFOS)	81		32-114
Perfluoro-n-[1,2,3,4,5,6-13C6]Decanoic Acid (13C6-PFDA)	85		28-115
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Decanesulfonic Acid (13C2-8:2FTS)	53		10-213
N-Methyl-d3-perfluoro-1-octanesulfonamidoacetic Acid (D3-NMeFOSAA)	78		10-172
Perfluoro-n-[1,2,3,4,5,6,7-13C7]Undecanoic Acid (13C7-PFUnA)	91		16-123
Perfluoro-1-[13C8]Octanesulfonamide (13C8-PFOSA)	74		14-108
N-Ethyl-d5-perfluoro-1-octanesulfonamidoacetic Acid (D5-NEtFOSAA)	94		10-150
Perfluoro-n-[1,2-13C2]Dodecanoic Acid (13C2-PFDoA)	97		10-126
Perfluoro-n-[1,2-13C2]Tetradecanoic Acid (13C2-PFTeDA)	115		10-145
Tetrafluoro-2-heptafluoropropoxy-[13C3]-propanoic acid (13C3-HFPO-DA)	95		35-142
N-Methyl-d3-Perfluoro-1-Octanesulfonamide (D3-NMeFOSA)	77		11-94
N-Ethyl-d5-Perfluoro-1-Octanesulfonamide (D5-NEtFOSA)	90		11-97
N-Methyl-d7-Perfluorooctanesulfonamidoethanol (D7-NMeFOSE)	96		10-137
N-Ethyl-d9-Perfluorooctanesulfonamidoethanol (D9-NEtFOSE)	102		10-130

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

SAMPLE RESULTS

Lab ID: L2449150-03 RE
 Client ID: ARB2-SW-2024-08-27
 Sample Location: BRUNSWICK

Date Collected: 08/27/24 14:30
 Date Received: 08/28/24
 Field Prep: Not Specified

Sample Depth:

Matrix: Water
 Analytical Method: 144,1633
 Analytical Date: 09/12/24 10:09
 Analyst: AC

Extraction Method: EPA 1633
 Extraction Date: 09/11/24 18:15

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab						
Perfluorobutanoic Acid (PFBA)	4.26	J	ng/l	5.84	0.935	1
Perfluoropentanoic Acid (PFPeA)	1.12	J	ng/l	2.92	0.781	1
Perfluorobutanesulfonic Acid (PFBS)	0.533	J	ng/l	1.46	0.489	1
1H,1H,2H,2H-Perfluorohexanesulfonic Acid (4:2FTS)	ND		ng/l	5.84	1.53	1
Perfluorohexanoic Acid (PFHxA)	1.08	J	ng/l	1.46	0.431	1
Perfluoropentanesulfonic Acid (PFPeS)	ND		ng/l	1.46	0.256	1
Perfluoroheptanoic Acid (PFHpA)	0.774	J	ng/l	1.46	0.292	1
Perfluorohexanesulfonic Acid (PFHxS)	0.584	J	ng/l	1.46	0.350	1
Perfluorooctanoic Acid (PFOA)	1.44	J	ng/l	1.46	0.635	1
1H,1H,2H,2H-Perfluorooctanesulfonic Acid (6:2FTS)	ND		ng/l	5.84	1.97	1
Perfluoroheptanesulfonic Acid (PFHpS)	ND		ng/l	1.46	0.394	1
Perfluorononanoic Acid (PFNA)	ND		ng/l	1.46	0.460	1
Perfluorooctanesulfonic Acid (PFOS)	3.68		ng/l	1.46	0.664	1
Perfluorodecanoic Acid (PFDA)	ND		ng/l	1.46	0.591	1
1H,1H,2H,2H-Perfluorodecanesulfonic Acid (8:2FTS)	ND		ng/l	5.84	2.27	1
Perfluoronanesulfonic Acid (PFNS)	ND		ng/l	1.46	0.453	1
N-Methyl Perfluorooctanesulfonamidoacetic Acid (NMeFOSAA)	ND		ng/l	1.46	0.796	1
Perfluoroundecanoic Acid (PFUnA)	ND		ng/l	1.46	0.635	1
Perfluorodecanesulfonic Acid (PFDS)	ND		ng/l	1.46	0.336	1
Perfluorooctanesulfonamide (PFOSA)	ND		ng/l	1.46	0.394	1
N-Ethyl Perfluorooctanesulfonamidoacetic Acid (NEtFOSAA)	1.28	J	ng/l	1.46	0.789	1
Perfluorododecanoic Acid (PFDoA)	ND		ng/l	1.46	0.672	1
Perfluorotridecanoic Acid (PFTrDA)	ND		ng/l	1.46	0.548	1
Perfluorotetradecanoic Acid (PFTeDA)	ND		ng/l	1.46	0.387	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	ND		ng/l	5.84	0.818	1
4,8-Dioxa-3h-Perfluorononanoic Acid (ADONA)	ND		ng/l	5.84	0.920	1
Perfluorododecanesulfonic Acid (PFDoS)	ND		ng/l	1.46	0.555	1

Project Name: NASB-HANGAR 4 AFFF

Lab Number: L2449150

Project Number: 83336

Report Date: 09/12/24

SAMPLE RESULTS

Lab ID: L2449150-03 RE

Date Collected: 08/27/24 14:30

Client ID: ARB2-SW-2024-08-27

Date Received: 08/28/24

Sample Location: BRUNSWICK

Field Prep: Not Specified

Sample Depth:

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab						
9-Chlorohexadecafluoro-3-Oxanone-1-Sulfonic Acid (9Cl-PF3ONS)	ND		ng/l	5.84	1.20	1
11-Chloroeicosafuoro-3-Oxaundecane-1-Sulfonic Acid (11Cl-PF3OUdS)	ND		ng/l	5.84	1.20	1
N-Methyl Perfluorooctane Sulfonamide (NMeFOSA)	ND		ng/l	1.46	0.635	1
N-Ethyl Perfluorooctane Sulfonamide (NEtFOSA)	ND		ng/l	1.46	0.672	1
N-Methyl Perfluorooctanesulfonamido Ethanol (NMeFOSE)	ND		ng/l	14.6	3.43	1
N-Ethyl Perfluorooctanesulfonamido Ethanol (NEtFOSE)	ND		ng/l	14.6	1.79	1
Perfluoro-3-Methoxypropanoic Acid (PFMPA)	ND		ng/l	2.92	0.416	1
Perfluoro-4-Methoxybutanoic Acid (PFMBA)	ND		ng/l	2.92	0.387	1
Perfluoro(2-Ethoxyethane)Sulfonic Acid (PFEEESA)	ND		ng/l	2.92	0.321	1
Nonafluoro-3,6-Dioxaheptanoic Acid (NFDHA)	ND		ng/l	2.92	1.72	1
3-Perfluoropropyl Propanoic Acid (3:3FTCA)	ND		ng/l	7.30	2.41	1
2H,2H,3H,3H-Perfluorooctanoic Acid (5:3FTCA)	ND		ng/l	36.5	8.54	1
3-Perfluoroheptyl Propanoic Acid (7:3FTCA)	ND		ng/l	36.5	5.76	1

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

SAMPLE RESULTS

Lab ID: L2449150-03 RE
 Client ID: ARB2-SW-2024-08-27
 Sample Location: BRUNSWICK

Date Collected: 08/27/24 14:30
 Date Received: 08/28/24
 Field Prep: Not Specified

Sample Depth:

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab						

Surrogate	% Recovery	Qualifier	Acceptance Criteria
Perfluoro-n-[13C4]Butanoic Acid (13C4-PFBA)	72		41-123
Perfluoro-n-[13C5]Pentanoic Acid (13C5-PFPeA)	73		29-123
Perfluoro-1-[2,3,4-13C3]Butanesulfonic Acid (13C3-PFBS)	64		41-125
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Hexanesulfonic Acid (13C2-4:2FTS)	106		10-290
Perfluoro-n-[1,2,3,4,6-13C5]Hexanoic Acid (13C5-PFHxA)	75		40-121
Perfluoro-n-[1,2,3,4-13C4]Heptanoic Acid (13C4-PFHpA)	81		27-156
Perfluoro-1-[1,2,3-13C3]Hexanesulfonic Acid (13C3-PFHxS)	70		46-115
Perfluoro-n-[13C8]Octanoic Acid (13C8-PFOA)	75		39-121
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Octanesulfonic Acid (13C2-6:2FTS)	73		10-261
Perfluoro-n-[13C9]Nonanoic Acid (13C9-PFNA)	80		38-114
Perfluoro-1-[13C8]Octanesulfonic Acid (13C8-PFOS)	74		32-114
Perfluoro-n-[1,2,3,4,5,6-13C6]Decanoic Acid (13C6-PFDA)	66		28-115
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Decanesulfonic Acid (13C2-8:2FTS)	51		10-213
N-Methyl-d3-perfluoro-1-octanesulfonamidoacetic Acid (D3-NMeFOSAA)	71		10-172
Perfluoro-n-[1,2,3,4,5,6,7-13C7]Undecanoic Acid (13C7-PFUnA)	76		16-123
Perfluoro-1-[13C8]Octanesulfonamide (13C8-PFOSA)	64		14-108
N-Ethyl-d5-perfluoro-1-octanesulfonamidoacetic Acid (D5-NEtFOSAA)	61		10-150
Perfluoro-n-[1,2-13C2]Dodecanoic Acid (13C2-PFDoA)	69		10-126
Perfluoro-n-[1,2-13C2]Tetradecanoic Acid (13C2-PFTeDA)	46		10-145
Tetrafluoro-2-heptafluoropropoxy-[13C3]-propanoic acid (13C3-HFPO-DA)	73		35-142
N-Methyl-d3-Perfluoro-1-Octanesulfonamide (D3-NMeFOSA)	53		11-94
N-Ethyl-d5-Perfluoro-1-Octanesulfonamide (D5-NEtFOSA)	53		11-97
N-Methyl-d7-Perfluorooctanesulfonamidoethanol (D7-NMeFOSE)	62		10-137
N-Ethyl-d9-Perfluorooctanesulfonamidoethanol (D9-NEtFOSE)	63		10-130

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

SAMPLE RESULTS

Lab ID: L2449150-04 RE
Client ID: ARB1-SW-2024-08-27
Sample Location: BRUNSWICK

Date Collected: 08/27/24 14:10
Date Received: 08/28/24
Field Prep: Not Specified

Sample Depth:

Matrix: Water
Analytical Method: 144,1633
Analytical Date: 09/12/24 10:22
Analyst: AC

Extraction Method: EPA 1633
Extraction Date: 09/11/24 18:15

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab						
Perfluorobutanoic Acid (PFBA)	3.76	J	ng/l	6.00	0.960	1
Perfluoropentanoic Acid (PFPeA)	1.07	J	ng/l	3.00	0.802	1
Perfluorobutanesulfonic Acid (PFBS)	ND		ng/l	1.50	0.502	1
1H,1H,2H,2H-Perfluorohexanesulfonic Acid (4:2FTS)	ND		ng/l	6.00	1.57	1
Perfluorohexanoic Acid (PFHxA)	0.982	J	ng/l	1.50	0.442	1
Perfluoropentanesulfonic Acid (PFPeS)	ND		ng/l	1.50	0.262	1
Perfluoroheptanoic Acid (PFHpA)	0.675	J	ng/l	1.50	0.300	1
Perfluorohexanesulfonic Acid (PFHxS)	0.390	J	ng/l	1.50	0.360	1
Perfluorooctanoic Acid (PFOA)	1.20	J	ng/l	1.50	0.652	1
1H,1H,2H,2H-Perfluorooctanesulfonic Acid (6:2FTS)	ND		ng/l	6.00	2.02	1
Perfluoroheptanesulfonic Acid (PFHpS)	ND		ng/l	1.50	0.405	1
Perfluorononanoic Acid (PFNA)	0.855	J	ng/l	1.50	0.472	1
Perfluorooctanesulfonic Acid (PFOS)	2.72		ng/l	1.50	0.682	1
Perfluorodecanoic Acid (PFDA)	ND		ng/l	1.50	0.607	1
1H,1H,2H,2H-Perfluorodecanesulfonic Acid (8:2FTS)	ND		ng/l	6.00	2.33	1
Perfluoronanesulfonic Acid (PFNS)	ND		ng/l	1.50	0.465	1
N-Methyl Perfluorooctanesulfonamidoacetic Acid (NMeFOSAA)	ND		ng/l	1.50	0.817	1
Perfluoroundecanoic Acid (PFUnA)	ND		ng/l	1.50	0.652	1
Perfluorodecanesulfonic Acid (PFDS)	ND		ng/l	1.50	0.345	1
Perfluorooctanesulfonamide (PFOSA)	ND		ng/l	1.50	0.405	1
N-Ethyl Perfluorooctanesulfonamidoacetic Acid (NEtFOSAA)	1.15	JF	ng/l	1.50	0.810	1
Perfluorododecanoic Acid (PFDoA)	ND		ng/l	1.50	0.690	1
Perfluorotridecanoic Acid (PFTrDA)	ND		ng/l	1.50	0.562	1
Perfluorotetradecanoic Acid (PFTeDA)	ND		ng/l	1.50	0.397	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	ND		ng/l	6.00	0.840	1
4,8-Dioxa-3h-Perfluorononanoic Acid (ADONA)	ND		ng/l	6.00	0.945	1
Perfluorododecanesulfonic Acid (PFDoS)	ND		ng/l	1.50	0.570	1

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

SAMPLE RESULTS

Lab ID: L2449150-04 RE
Client ID: ARB1-SW-2024-08-27
Sample Location: BRUNSWICK

Date Collected: 08/27/24 14:10
Date Received: 08/28/24
Field Prep: Not Specified

Sample Depth:

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab						
9-Chlorohexadecafluoro-3-Oxanone-1-Sulfonic Acid (9Cl-PF3ONS)	ND		ng/l	6.00	1.24	1
11-Chloroeicosafuoro-3-Oxaundecane-1-Sulfonic Acid (11Cl-PF3OUdS)	ND		ng/l	6.00	1.24	1
N-Methyl Perfluorooctane Sulfonamide (NMeFOSA)	ND		ng/l	1.50	0.652	1
N-Ethyl Perfluorooctane Sulfonamide (NEtFOSA)	ND		ng/l	1.50	0.690	1
N-Methyl Perfluorooctanesulfonamido Ethanol (NMeFOSE)	ND		ng/l	15.0	3.52	1
N-Ethyl Perfluorooctanesulfonamido Ethanol (NEtFOSE)	ND		ng/l	15.0	1.84	1
Perfluoro-3-Methoxypropanoic Acid (PFMPA)	ND		ng/l	3.00	0.427	1
Perfluoro-4-Methoxybutanoic Acid (PFMBA)	ND		ng/l	3.00	0.397	1
Perfluoro(2-Ethoxyethane)Sulfonic Acid (PFEEESA)	ND		ng/l	3.00	0.330	1
Nonafluoro-3,6-Dioxaheptanoic Acid (NFDHA)	ND		ng/l	3.00	1.77	1
3-Perfluoropropyl Propanoic Acid (3:3FTCA)	ND		ng/l	7.50	2.47	1
2H,2H,3H,3H-Perfluorooctanoic Acid (5:3FTCA)	ND		ng/l	37.5	8.77	1
3-Perfluoroheptyl Propanoic Acid (7:3FTCA)	ND		ng/l	37.5	5.92	1

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

SAMPLE RESULTS

Lab ID: L2449150-04 RE
 Client ID: ARB1-SW-2024-08-27
 Sample Location: BRUNSWICK

Date Collected: 08/27/24 14:10
 Date Received: 08/28/24
 Field Prep: Not Specified

Sample Depth:

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab						

Surrogate	% Recovery	Qualifier	Acceptance Criteria
Perfluoro-n-[13C4]Butanoic Acid (13C4-PFBA)	64		41-123
Perfluoro-n-[13C5]Pentanoic Acid (13C5-PFPeA)	66		29-123
Perfluoro-1-[2,3,4-13C3]Butanesulfonic Acid (13C3-PFBS)	62		41-125
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Hexanesulfonic Acid (13C2-4:2FTS)	97		10-290
Perfluoro-n-[1,2,3,4,6-13C5]Hexanoic Acid (13C5-PFHxA)	67		40-121
Perfluoro-n-[1,2,3,4-13C4]Heptanoic Acid (13C4-PFHpA)	77		27-156
Perfluoro-1-[1,2,3-13C3]Hexanesulfonic Acid (13C3-PFHxS)	63		46-115
Perfluoro-n-[13C8]Octanoic Acid (13C8-PFOA)	67		39-121
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Octanesulfonic Acid (13C2-6:2FTS)	63		10-261
Perfluoro-n-[13C9]Nonanoic Acid (13C9-PFNA)	69		38-114
Perfluoro-1-[13C8]Octanesulfonic Acid (13C8-PFOS)	63		32-114
Perfluoro-n-[1,2,3,4,5,6-13C6]Decanoic Acid (13C6-PFDA)	56		28-115
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Decanesulfonic Acid (13C2-8:2FTS)	52		10-213
N-Methyl-d3-perfluoro-1-octanesulfonamidoacetic Acid (D3-NMeFOSAA)	64		10-172
Perfluoro-n-[1,2,3,4,5,6,7-13C7]Undecanoic Acid (13C7-PFUnA)	58		16-123
Perfluoro-1-[13C8]Octanesulfonamide (13C8-PFOSA)	56		14-108
N-Ethyl-d5-perfluoro-1-octanesulfonamidoacetic Acid (D5-NEtFOSAA)	53		10-150
Perfluoro-n-[1,2-13C2]Dodecanoic Acid (13C2-PFDoA)	49		10-126
Perfluoro-n-[1,2-13C2]Tetradecanoic Acid (13C2-PFTeDA)	35		10-145
Tetrafluoro-2-heptafluoropropoxy-[13C3]-propanoic acid (13C3-HFPO-DA)	67		35-142
N-Methyl-d3-Perfluoro-1-Octanesulfonamide (D3-NMeFOSA)	46		11-94
N-Ethyl-d5-Perfluoro-1-Octanesulfonamide (D5-NEtFOSA)	44		11-97
N-Methyl-d7-Perfluorooctanesulfonamidoethanol (D7-NMeFOSE)	53		10-137
N-Ethyl-d9-Perfluorooctanesulfonamidoethanol (D9-NEtFOSE)	54		10-130

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

SAMPLE RESULTS

Lab ID: L2449150-05 RE
 Client ID: ART-SW-2024-08-27
 Sample Location: BRUNSWICK

Date Collected: 08/27/24 13:15
 Date Received: 08/28/24
 Field Prep: Not Specified

Sample Depth:

Matrix: Water
 Analytical Method: 144,1633
 Analytical Date: 09/12/24 10:35
 Analyst: AC

Extraction Method: EPA 1633
 Extraction Date: 09/11/24 18:15

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab						
Perfluorobutanoic Acid (PFBA)	1.64	J	ng/l	5.90	0.943	1
Perfluoropentanoic Acid (PFPeA)	0.818	J	ng/l	2.95	0.788	1
Perfluorobutanesulfonic Acid (PFBS)	ND		ng/l	1.47	0.494	1
1H,1H,2H,2H-Perfluorohexanesulfonic Acid (4:2FTS)	ND		ng/l	5.90	1.54	1
Perfluorohexanoic Acid (PFHxA)	0.825	J	ng/l	1.47	0.435	1
Perfluoropentanesulfonic Acid (PFPeS)	ND		ng/l	1.47	0.258	1
Perfluoroheptanoic Acid (PFHpA)	0.597	J	ng/l	1.47	0.295	1
Perfluorohexanesulfonic Acid (PFHxS)	0.405	J	ng/l	1.47	0.354	1
Perfluorooctanoic Acid (PFOA)	1.28	J	ng/l	1.47	0.641	1
1H,1H,2H,2H-Perfluorooctanesulfonic Acid (6:2FTS)	ND		ng/l	5.90	1.99	1
Perfluoroheptanesulfonic Acid (PFHpS)	ND		ng/l	1.47	0.398	1
Perfluorononanoic Acid (PFNA)	0.604	J	ng/l	1.47	0.464	1
Perfluorooctanesulfonic Acid (PFOS)	2.25		ng/l	1.47	0.670	1
Perfluorodecanoic Acid (PFDA)	ND		ng/l	1.47	0.597	1
1H,1H,2H,2H-Perfluorodecanesulfonic Acid (8:2FTS)	ND		ng/l	5.90	2.29	1
Perfluoronanesulfonic Acid (PFNS)	ND		ng/l	1.47	0.457	1
N-Methyl Perfluorooctanesulfonamidoacetic Acid (NMeFOSAA)	ND		ng/l	1.47	0.803	1
Perfluoroundecanoic Acid (PFUnA)	ND		ng/l	1.47	0.641	1
Perfluorodecanesulfonic Acid (PFDS)	ND		ng/l	1.47	0.339	1
Perfluorooctanesulfonamide (PFOSA)	ND		ng/l	1.47	0.398	1
N-Ethyl Perfluorooctanesulfonamidoacetic Acid (NEtFOSAA)	1.20	JF	ng/l	1.47	0.796	1
Perfluorododecanoic Acid (PFDoA)	ND		ng/l	1.47	0.678	1
Perfluorotridecanoic Acid (PFTrDA)	ND		ng/l	1.47	0.553	1
Perfluorotetradecanoic Acid (PFTeDA)	ND		ng/l	1.47	0.390	1
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	ND		ng/l	5.90	0.825	1
4,8-Dioxa-3h-Perfluorononanoic Acid (ADONA)	ND		ng/l	5.90	0.928	1
Perfluorododecanesulfonic Acid (PFDoS)	ND		ng/l	1.47	0.560	1

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

SAMPLE RESULTS

Lab ID: L2449150-05 RE
Client ID: ART-SW-2024-08-27
Sample Location: BRUNSWICK

Date Collected: 08/27/24 13:15
Date Received: 08/28/24
Field Prep: Not Specified

Sample Depth:

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab						
9-Chlorohexadecafluoro-3-Oxanone-1-Sulfonic Acid (9Cl-PF3ONS)	ND		ng/l	5.90	1.22	1
11-Chloroeicosafuoro-3-Oxaundecane-1-Sulfonic Acid (11Cl-PF3OUdS)	ND		ng/l	5.90	1.22	1
N-Methyl Perfluorooctane Sulfonamide (NMeFOSA)	ND		ng/l	1.47	0.641	1
N-Ethyl Perfluorooctane Sulfonamide (NEtFOSA)	ND		ng/l	1.47	0.678	1
N-Methyl Perfluorooctanesulfonamido Ethanol (NMeFOSE)	ND		ng/l	14.7	3.46	1
N-Ethyl Perfluorooctanesulfonamido Ethanol (NEtFOSE)	ND		ng/l	14.7	1.80	1
Perfluoro-3-Methoxypropanoic Acid (PFMPA)	ND		ng/l	2.95	0.420	1
Perfluoro-4-Methoxybutanoic Acid (PFMBA)	ND		ng/l	2.95	0.390	1
Perfluoro(2-Ethoxyethane)Sulfonic Acid (PFEEESA)	ND		ng/l	2.95	0.324	1
Nonafluoro-3,6-Dioxaheptanoic Acid (NFDHA)	ND		ng/l	2.95	1.74	1
3-Perfluoropropyl Propanoic Acid (3:3FTCA)	ND		ng/l	7.37	2.43	1
2H,2H,3H,3H-Perfluorooctanoic Acid (5:3FTCA)	ND		ng/l	36.8	8.62	1
3-Perfluoroheptyl Propanoic Acid (7:3FTCA)	ND		ng/l	36.8	5.81	1

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

SAMPLE RESULTS

Lab ID: L2449150-05 RE
 Client ID: ART-SW-2024-08-27
 Sample Location: BRUNSWICK

Date Collected: 08/27/24 13:15
 Date Received: 08/28/24
 Field Prep: Not Specified

Sample Depth:

Parameter	Result	Qualifier	Units	RL	MDL	Dilution Factor
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab						

Surrogate	% Recovery	Qualifier	Acceptance Criteria
Perfluoro-n-[13C4]Butanoic Acid (13C4-PFBA)	63		41-123
Perfluoro-n-[13C5]Pentanoic Acid (13C5-PFPeA)	64		29-123
Perfluoro-1-[2,3,4-13C3]Butanesulfonic Acid (13C3-PFBS)	59		41-125
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Hexanesulfonic Acid (13C2-4:2FTS)	147		10-290
Perfluoro-n-[1,2,3,4,6-13C5]Hexanoic Acid (13C5-PFHxA)	72		40-121
Perfluoro-n-[1,2,3,4-13C4]Heptanoic Acid (13C4-PFHpA)	73		27-156
Perfluoro-1-[1,2,3-13C3]Hexanesulfonic Acid (13C3-PFHxS)	68		46-115
Perfluoro-n-[13C8]Octanoic Acid (13C8-PFOA)	70		39-121
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Octanesulfonic Acid (13C2-6:2FTS)	73		10-261
Perfluoro-n-[13C9]Nonanoic Acid (13C9-PFNA)	67		38-114
Perfluoro-1-[13C8]Octanesulfonic Acid (13C8-PFOS)	63		32-114
Perfluoro-n-[1,2,3,4,5,6-13C6]Decanoic Acid (13C6-PFDA)	59		28-115
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Decanesulfonic Acid (13C2-8:2FTS)	64		10-213
N-Methyl-d3-perfluoro-1-octanesulfonamidoacetic Acid (D3-NMeFOSAA)	69		10-172
Perfluoro-n-[1,2,3,4,5,6,7-13C7]Undecanoic Acid (13C7-PFUnA)	56		16-123
Perfluoro-1-[13C8]Octanesulfonamide (13C8-PFOSA)	55		14-108
N-Ethyl-d5-perfluoro-1-octanesulfonamidoacetic Acid (D5-NEtFOSAA)	55		10-150
Perfluoro-n-[1,2-13C2]Dodecanoic Acid (13C2-PFDoA)	47		10-126
Perfluoro-n-[1,2-13C2]Tetradecanoic Acid (13C2-PFTeDA)	37		10-145
Tetrafluoro-2-heptafluoropropoxy-[13C3]-propanoic acid (13C3-HFPO-DA)	67		35-142
N-Methyl-d3-Perfluoro-1-Octanesulfonamide (D3-NMeFOSA)	46		11-94
N-Ethyl-d5-Perfluoro-1-Octanesulfonamide (D5-NEtFOSA)	45		11-97
N-Methyl-d7-Perfluorooctanesulfonamidoethanol (D7-NMeFOSE)	54		10-137
N-Ethyl-d9-Perfluorooctanesulfonamidoethanol (D9-NEtFOSE)	52		10-130

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

Method Blank Analysis
Batch Quality Control

Analytical Method: 144,1633
Analytical Date: 09/11/24 13:18
Analyst: ANH

Extraction Method: EPA 1633
Extraction Date: 09/10/24 18:45

Parameter	Result	Qualifier	Units	RL	MDL
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab for sample(s): 01-02 Batch: WG1969890-1 R					
Perfluorobutanoic Acid (PFBA)	ND		ng/l	6.40	1.02
Perfluoropentanoic Acid (PFPeA)	ND		ng/l	3.20	0.856
Perfluorobutanesulfonic Acid (PFBS)	ND		ng/l	1.60	0.536
1H,1H,2H,2H-Perfluorohexanesulfonic Acid (4:2FTS)	ND		ng/l	6.40	1.67
Perfluorohexanoic Acid (PFHxA)	ND		ng/l	1.60	0.472
Perfluoropentanesulfonic Acid (PFPeS)	ND		ng/l	1.60	0.280
Perfluoroheptanoic Acid (PFHpA)	ND		ng/l	1.60	0.320
Perfluorohexanesulfonic Acid (PFHxS)	ND		ng/l	1.60	0.384
Perfluorooctanoic Acid (PFOA)	ND		ng/l	1.60	0.696
1H,1H,2H,2H-Perfluorooctanesulfonic Acid (6:2FTS)	ND		ng/l	6.40	2.16
Perfluoroheptanesulfonic Acid (PFHpS)	ND		ng/l	1.60	0.432
Perfluorononanoic Acid (PFNA)	ND		ng/l	1.60	0.504
Perfluorooctanesulfonic Acid (PFOS)	ND		ng/l	1.60	0.728
Perfluorodecanoic Acid (PFDA)	ND		ng/l	1.60	0.648
1H,1H,2H,2H-Perfluorodecanesulfonic Acid (8:2FTS)	ND		ng/l	6.40	2.49
Perfluorononanesulfonic Acid (PFNS)	ND		ng/l	1.60	0.496
N-Methyl Perfluorooctanesulfonamidoacetic Acid (NMeFOSAA)	ND		ng/l	1.60	0.872
Perfluoroundecanoic Acid (PFUnA)	ND		ng/l	1.60	0.696
Perfluorodecanesulfonic Acid (PFDS)	ND		ng/l	1.60	0.368
Perfluorooctanesulfonamide (PFOSA)	ND		ng/l	1.60	0.432
N-Ethyl Perfluorooctanesulfonamidoacetic Acid (NEtFOSAA)	ND		ng/l	1.60	0.864
Perfluorododecanoic Acid (PFDoA)	ND		ng/l	1.60	0.736
Perfluorotridecanoic Acid (PFTrDA)	ND		ng/l	1.60	0.600
Perfluorotetradecanoic Acid (PFTeDA)	ND		ng/l	1.60	0.424
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	ND		ng/l	6.40	0.896
4,8-Dioxa-3h-Perfluorononanoic Acid (ADONA)	ND		ng/l	6.40	1.01
Perfluorododecanesulfonic Acid (PFDoS)	ND		ng/l	1.60	0.608

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

Method Blank Analysis Batch Quality Control

Analytical Method: 144,1633
Analytical Date: 09/11/24 13:18
Analyst: ANH

Extraction Method: EPA 1633
Extraction Date: 09/10/24 18:45

Parameter	Result	Qualifier	Units	RL	MDL
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab for sample(s): 01-02 Batch: WG1969890-1 R					
9-Chlorohexadecafluoro-3-Oxanone-1-Sulfonic Acid (9Cl-PF3ONS)	ND		ng/l	6.40	1.32
11-Chloroeicosafluoro-3-Oxaundecane-1-Sulfonic Acid (11Cl-PF3OUdS)	ND		ng/l	6.40	1.32
N-Methyl Perfluorooctane Sulfonamide (NMeFOSA)	ND		ng/l	1.60	0.696
N-Ethyl Perfluorooctane Sulfonamide (NEtFOSA)	ND		ng/l	1.60	0.736
N-Methyl Perfluorooctanesulfonamido Ethanol (NMeFOSE)	ND		ng/l	16.0	3.76
N-Ethyl Perfluorooctanesulfonamido Ethanol (NEtFOSE)	ND		ng/l	16.0	1.96
Perfluoro-3-Methoxypropanoic Acid (PFMPA)	ND		ng/l	3.20	0.456
Perfluoro-4-Methoxybutanoic Acid (PFMBA)	ND		ng/l	3.20	0.424
Perfluoro(2-Ethoxyethane)Sulfonic Acid (PFEESA)	ND		ng/l	3.20	0.352
Nonafluoro-3,6-Dioxaheptanoic Acid (NFDHA)	ND		ng/l	3.20	1.89
3-Perfluoropropyl Propanoic Acid (3:3FTCA)	ND		ng/l	8.00	2.64
2H,2H,3H,3H-Perfluorooctanoic Acid (5:3FTCA)	ND		ng/l	40.0	9.36
3-Perfluoroheptyl Propanoic Acid (7:3FTCA)	ND		ng/l	40.0	6.31

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

Method Blank Analysis
Batch Quality Control

Analytical Method: 144,1633
Analytical Date: 09/11/24 13:18
Analyst: ANH

Extraction Method: EPA 1633
Extraction Date: 09/10/24 18:45

Parameter	Result	Qualifier	Units	RL	MDL
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab for sample(s): 01-02 Batch: WG1969890-1 R					

Surrogate	%Recovery	Qualifier	Acceptance Criteria
Perfluoro-n-[13C4]Butanoic Acid (13C4-PFBA)	86		41-123
Perfluoro-n-[13C5]Pentanoic Acid (13C5-PFPeA)	81		29-123
Perfluoro-1-[2,3,4-13C3]Butanesulfonic Acid (13C3-PFBS)	84		41-125
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Hexanesulfonic Acid (13C2-4:2FTS)	145		10-290
Perfluoro-n-[1,2,3,4,6-13C5]Hexanoic Acid (13C5-PFHxA)	82		40-121
Perfluoro-n-[1,2,3,4-13C4]Heptanoic Acid (13C4-PFHpA)	80		27-156
Perfluoro-1-[1,2,3-13C3]Hexanesulfonic Acid (13C3-PFHxS)	87		46-115
Perfluoro-n-[13C8]Octanoic Acid (13C8-PFOA)	86		39-121
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Octanesulfonic Acid (13C2-6:2FTS)	84		10-261
Perfluoro-n-[13C9]Nonanoic Acid (13C9-PFNA)	84		38-114
Perfluoro-1-[13C8]Octanesulfonic Acid (13C8-PFOS)	83		32-114
Perfluoro-n-[1,2,3,4,5,6-13C6]Decanoic Acid (13C6-PFDA)	82		28-115
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Decanesulfonic Acid (13C2-8:2FTS)	121		10-213
N-Methyl-d3-perfluoro-1-octanesulfonamidoacetic Acid (D3-NMeFOSAA)	72		10-172
Perfluoro-n-[1,2,3,4,5,6,7-13C7]Undecanoic Acid (13C7-PFUnA)	87		16-123
Perfluoro-1-[13C8]Octanesulfonamide (13C8-PFOSA)	61		14-108
N-Ethyl-d5-perfluoro-1-octanesulfonamidoacetic Acid (D5-NEtFOSAA)	76		10-150
Perfluoro-n-[1,2-13C2]Dodecanoic Acid (13C2-PFDoA)	91		10-126
Perfluoro-n-[1,2-13C2]Tetradecanoic Acid (13C2-PFTeDA)	99		10-145
Tetrafluoro-2-heptafluoropropoxy-[13C3]-propanoic acid (13C3-HFPO-DA)	81		35-142
N-Methyl-d3-Perfluoro-1-Octanesulfonamide (D3-NMeFOSA)	54		11-94
N-Ethyl-d5-Perfluoro-1-Octanesulfonamide (D5-NEtFOSA)	67		11-97
N-Methyl-d7-Perfluorooctanesulfonamidoethanol (D7-NMeFOSE)	73		10-137
N-Ethyl-d9-Perfluorooctanesulfonamidoethanol (D9-NEtFOSE)	72		10-130

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

Method Blank Analysis Batch Quality Control

Analytical Method: 144,1633
Analytical Date: 09/12/24 09:05
Analyst: AC

Extraction Method: EPA 1633
Extraction Date: 09/11/24 18:15

Parameter	Result	Qualifier	Units	RL	MDL
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab for sample(s): 03-05 Batch: WG1970425-1					
Perfluorobutanoic Acid (PFBA)	ND		ng/l	6.40	1.02
Perfluoropentanoic Acid (PFPeA)	ND		ng/l	3.20	0.856
Perfluorobutanesulfonic Acid (PFBS)	ND		ng/l	1.60	0.536
1H,1H,2H,2H-Perfluorohexanesulfonic Acid (4:2FTS)	ND		ng/l	6.40	1.67
Perfluorohexanoic Acid (PFHxA)	ND		ng/l	1.60	0.472
Perfluoropentanesulfonic Acid (PFPeS)	ND		ng/l	1.60	0.280
Perfluoroheptanoic Acid (PFHpA)	ND		ng/l	1.60	0.320
Perfluorohexanesulfonic Acid (PFHxS)	ND		ng/l	1.60	0.384
Perfluorooctanoic Acid (PFOA)	ND		ng/l	1.60	0.696
1H,1H,2H,2H-Perfluorooctanesulfonic Acid (6:2FTS)	ND		ng/l	6.40	2.16
Perfluoroheptanesulfonic Acid (PFHpS)	ND		ng/l	1.60	0.432
Perfluorononanoic Acid (PFNA)	ND		ng/l	1.60	0.504
Perfluorooctanesulfonic Acid (PFOS)	ND		ng/l	1.60	0.728
Perfluorodecanoic Acid (PFDA)	ND		ng/l	1.60	0.648
1H,1H,2H,2H-Perfluorodecanesulfonic Acid (8:2FTS)	ND		ng/l	6.40	2.49
Perfluorononanesulfonic Acid (PFNS)	ND		ng/l	1.60	0.496
N-Methyl Perfluorooctanesulfonamidoacetic Acid (NMeFOSAA)	ND		ng/l	1.60	0.872
Perfluoroundecanoic Acid (PFUnA)	ND		ng/l	1.60	0.696
Perfluorodecanesulfonic Acid (PFDS)	ND		ng/l	1.60	0.368
Perfluorooctanesulfonamide (PFOSA)	ND		ng/l	1.60	0.432
N-Ethyl Perfluorooctanesulfonamidoacetic Acid (NEtFOSAA)	ND		ng/l	1.60	0.864
Perfluorododecanoic Acid (PFDoA)	ND		ng/l	1.60	0.736
Perfluorotridecanoic Acid (PFTrDA)	ND		ng/l	1.60	0.600
Perfluorotetradecanoic Acid (PFTeDA)	ND		ng/l	1.60	0.424
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	ND		ng/l	6.40	0.896
4,8-Dioxa-3h-Perfluorononanoic Acid (ADONA)	ND		ng/l	6.40	1.01
Perfluorododecanesulfonic Acid (PFDoS)	ND		ng/l	1.60	0.608

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

Method Blank Analysis Batch Quality Control

Analytical Method: 144,1633
Analytical Date: 09/12/24 09:05
Analyst: AC

Extraction Method: EPA 1633
Extraction Date: 09/11/24 18:15

Parameter	Result	Qualifier	Units	RL	MDL
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab for sample(s): 03-05 Batch: WG1970425-1					
9-Chlorohexadecafluoro-3-Oxanone-1-Sulfonic Acid (9Cl-PF3ONS)	ND		ng/l	6.40	1.32
11-Chloroeicosafluoro-3-Oxaundecane-1-Sulfonic Acid (11Cl-PF3OUdS)	ND		ng/l	6.40	1.32
N-Methyl Perfluorooctane Sulfonamide (NMeFOSA)	ND		ng/l	1.60	0.696
N-Ethyl Perfluorooctane Sulfonamide (NEtFOSA)	ND		ng/l	1.60	0.736
N-Methyl Perfluorooctanesulfonamido Ethanol (NMeFOSE)	ND		ng/l	16.0	3.76
N-Ethyl Perfluorooctanesulfonamido Ethanol (NEtFOSE)	ND		ng/l	16.0	1.96
Perfluoro-3-Methoxypropanoic Acid (PFMPA)	ND		ng/l	3.20	0.456
Perfluoro-4-Methoxybutanoic Acid (PFMBA)	ND		ng/l	3.20	0.424
Perfluoro(2-Ethoxyethane)Sulfonic Acid (PFEESA)	ND		ng/l	3.20	0.352
Nonafluoro-3,6-Dioxaheptanoic Acid (NFDHA)	ND		ng/l	3.20	1.89
3-Perfluoropropyl Propanoic Acid (3:3FTCA)	ND		ng/l	8.00	2.64
2H,2H,3H,3H-Perfluorooctanoic Acid (5:3FTCA)	ND		ng/l	40.0	9.36
3-Perfluoroheptyl Propanoic Acid (7:3FTCA)	ND		ng/l	40.0	6.31

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

Method Blank Analysis
Batch Quality Control

Analytical Method: 144,1633
Analytical Date: 09/12/24 09:05
Analyst: AC

Extraction Method: EPA 1633
Extraction Date: 09/11/24 18:15

Parameter	Result	Qualifier	Units	RL	MDL
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab for sample(s): 03-05 Batch: WG1970425-1					

Surrogate	%Recovery	Qualifier	Acceptance Criteria
Perfluoro-n-[13C4]Butanoic Acid (13C4-PFBA)	60		41-123
Perfluoro-n-[13C5]Pentanoic Acid (13C5-PFPeA)	67		29-123
Perfluoro-1-[2,3,4-13C3]Butanesulfonic Acid (13C3-PFBS)	60		41-125
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Hexanesulfonic Acid (13C2-4:2FTS)	63		10-290
Perfluoro-n-[1,2,3,4,6-13C5]Hexanoic Acid (13C5-PFHxA)	63		40-121
Perfluoro-n-[1,2,3,4-13C4]Heptanoic Acid (13C4-PFHpA)	63		27-156
Perfluoro-1-[1,2,3-13C3]Hexanesulfonic Acid (13C3-PFHxS)	56		46-115
Perfluoro-n-[13C8]Octanoic Acid (13C8-PFOA)	63		39-121
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Octanesulfonic Acid (13C2-6:2FTS)	51		10-261
Perfluoro-n-[13C9]Nonanoic Acid (13C9-PFNA)	59		38-114
Perfluoro-1-[13C8]Octanesulfonic Acid (13C8-PFOS)	54		32-114
Perfluoro-n-[1,2,3,4,5,6-13C6]Decanoic Acid (13C6-PFDA)	57		28-115
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Decanesulfonic Acid (13C2-8:2FTS)	47		10-213
N-Methyl-d3-perfluoro-1-octanesulfonamidoacetic Acid (D3-NMeFOSAA)	70		10-172
Perfluoro-n-[1,2,3,4,5,6,7-13C7]Undecanoic Acid (13C7-PFUnA)	53		16-123
Perfluoro-1-[13C8]Octanesulfonamide (13C8-PFOSA)	51		14-108
N-Ethyl-d5-perfluoro-1-octanesulfonamidoacetic Acid (D5-NEtFOSAA)	52		10-150
Perfluoro-n-[1,2-13C2]Dodecanoic Acid (13C2-PFDoA)	51		10-126
Perfluoro-n-[1,2-13C2]Tetradecanoic Acid (13C2-PFTeDA)	46		10-145
Tetrafluoro-2-heptafluoropropoxy-[13C3]-propanoic acid (13C3-HFPO-DA)	59		35-142
N-Methyl-d3-Perfluoro-1-Octanesulfonamide (D3-NMeFOSA)	33		11-94
N-Ethyl-d5-Perfluoro-1-Octanesulfonamide (D5-NEtFOSA)	37		11-97
N-Methyl-d7-Perfluorooctanesulfonamidoethanol (D7-NMeFOSE)	50		10-137
N-Ethyl-d9-Perfluorooctanesulfonamidoethanol (D9-NEtFOSE)	52		10-130

Lab Control Sample Analysis

Batch Quality Control

Project Name: NASB-HANGAR 4 AFFF

Lab Number: L2449150

Project Number: 83336

Report Date: 09/12/24

Parameter	Low Level	Qual	Low Level	Qual	%Recovery	RPD	Qual	RPD
	LCS		LCS		Limits			Limits
	%Recovery		%Recovery					
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab Associated sample(s): 01-02 Batch: WG1969890-2 LOW LEVEL								
Perfluorobutanoic Acid (PFBA)	119		-		40-150	-		30
Perfluoropentanoic Acid (PFPeA)	105		-		40-150	-		30
Perfluorobutanesulfonic Acid (PFBS)	104		-		40-150	-		30
1H,1H,2H,2H-Perfluorohexanesulfonic Acid (4:2FTS)	108		-		40-150	-		30
Perfluorohexanoic Acid (PFHxA)	96		-		40-150	-		30
Perfluoropentanesulfonic Acid (PFPeS)	105		-		40-150	-		30
Perfluoroheptanoic Acid (PFHpA)	99		-		40-150	-		30
Perfluorohexanesulfonic Acid (PFHxS)	98		-		40-150	-		30
Perfluorooctanoic Acid (PFOA)	102		-		40-150	-		30
1H,1H,2H,2H-Perfluorooctanesulfonic Acid (6:2FTS)	105		-		40-150	-		30
Perfluoroheptanesulfonic Acid (PFHpS)	111		-		40-150	-		30
Perfluorononanoic Acid (PFNA)	112		-		40-150	-		30
Perfluorooctanesulfonic Acid (PFOS)	99		-		40-150	-		30
Perfluorodecanoic Acid (PFDA)	102		-		40-150	-		30
1H,1H,2H,2H-Perfluorodecanesulfonic Acid (8:2FTS)	102		-		40-150	-		30
Perfluorononanesulfonic Acid (PFNS)	82		-		40-150	-		30
N-Methyl Perfluorooctanesulfonamidoacetic Acid (NMeFOSAA)	126		-		40-150	-		30
Perfluoroundecanoic Acid (PFUnA)	103		-		40-150	-		30
Perfluorodecanesulfonic Acid (PFDS)	91		-		40-150	-		30
Perfluorooctanesulfonamide (PFOSA)	104		-		40-150	-		30
N-Ethyl Perfluorooctanesulfonamidoacetic Acid (NEtFOSAA)	105		-		40-150	-		30
Perfluorododecanoic Acid (PFDoA)	95		-		40-150	-		30

Lab Control Sample Analysis

Batch Quality Control

Project Name: NASB-HANGAR 4 AFFF

Lab Number: L2449150

Project Number: 83336

Report Date: 09/12/24

Parameter	Low Level LCS %Recovery	Qual	Low Level LCSD %Recovery	Qual	%Recovery Limits	RPD	Qual	RPD Limits
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab Associated sample(s): 01-02 Batch: WG1969890-2 LOW LEVEL								
Perfluorotridecanoic Acid (PFTTrDA)	116		-		40-150	-		30
Perfluorotetradecanoic Acid (PFTeDA)	104		-		40-150	-		30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	103		-		40-150	-		30
4,8-Dioxa-3h-Perfluorononanoic Acid (ADONA)	92		-		40-150	-		30
Perfluorododecanesulfonic Acid (PFDoS)	106		-		40-150	-		30
9-Chlorohexadecafluoro-3-Oxanone-1-Sulfonic Acid (9Cl-PF3ONS)	68		-		40-150	-		30
11-Chloroeicosafluoro-3-Oxaundecane-1-Sulfonic Acid (11Cl-PF3OUS)	71		-		40-150	-		30
N-Methyl Perfluorooctane Sulfonamide (NMeFOSA)	96		-		40-150	-		30
N-Ethyl Perfluorooctane Sulfonamide (NEtFOSA)	94		-		40-150	-		30
N-Methyl Perfluorooctanesulfonamido Ethanol (NMeFOSE)	96		-		40-150	-		30
N-Ethyl Perfluorooctanesulfonamido Ethanol (NEtFOSE)	94		-		40-150	-		30
Perfluoro-3-Methoxypropanoic Acid (PFMPA)	103		-		40-150	-		30
Perfluoro-4-Methoxybutanoic Acid (PFMBA)	97		-		40-150	-		30
Perfluoro(2-Ethoxyethane)Sulfonic Acid (PFEESA)	93		-		40-150	-		30
Nonafluoro-3,6-Dioxaheptanoic Acid (NFDHA)	105		-		40-150	-		30
3-Perfluoropropyl Propanoic Acid (3:3FTCA)	110		-		40-150	-		30
2H,2H,3H,3H-Perfluorooctanoic Acid (5:3FTCA)	79		-		40-150	-		30
3-Perfluoroheptyl Propanoic Acid (7:3FTCA)	53		-		40-150	-		30

Lab Control Sample Analysis

Batch Quality Control

Project Name: NASB-HANGAR 4 AFFF

Lab Number: L2449150

Project Number: 83336

Report Date: 09/12/24

Parameter	Low Level LCS		Low Level LCSD		%Recovery Limits		RPD	
	%Recovery	Qual	%Recovery	Qual			Qual	Limits
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab Associated sample(s): 01-02 Batch: WG1969890-2 LOW LEVEL								

Surrogate	LCS		LCSD		Acceptance Criteria
	%Recovery	Qual	%Recovery	Qual	
Perfluoro-n-[13C4]Butanoic Acid (13C4-PFBA)	68				41-123
Perfluoro-n-[13C5]Pentanoic Acid (13C5-PFPeA)	74				29-123
Perfluoro-1-[2,3,4-13C3]Butanesulfonic Acid (13C3-PFBS)	66				41-125
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Hexanesulfonic Acid (13C2-4:2FTS)	66				10-290
Perfluoro-n-[1,2,3,4,6-13C5]Hexanoic Acid (13C5-PFHxA)	73				40-121
Perfluoro-n-[1,2,3,4-13C4]Heptanoic Acid (13C4-PFHpA)	70				27-156
Perfluoro-1-[1,2,3-13C3]Hexanesulfonic Acid (13C3-PFHxS)	66				46-115
Perfluoro-n-[13C8]Octanoic Acid (13C8-PFOA)	68				39-121
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Octanesulfonic Acid (13C2-6:2FTS)	62				10-261
Perfluoro-n-[13C9]Nonanoic Acid (13C9-PFNA)	59				38-114
Perfluoro-1-[13C8]Octanesulfonic Acid (13C8-PFOS)	71				32-114
Perfluoro-n-[1,2,3,4,5,6-13C6]Decanoic Acid (13C6-PFDA)	68				28-115
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Decanesulfonic Acid (13C2-8:2FTS)	58				10-213
N-Methyl-d3-perfluoro-1-octanesulfonamidoacetic Acid (D3-NMeFOSAA)	59				10-172
Perfluoro-n-[1,2,3,4,5,6,7-13C7]Undecanoic Acid (13C7-PFUnA)	69				16-123
Perfluoro-1-[13C8]Octanesulfonamide (13C8-PFOSA)	57				14-108
N-Ethyl-d5-perfluoro-1-octanesulfonamidoacetic Acid (D5-NEtFOSAA)	61				10-150
Perfluoro-n-[1,2-13C2]Dodecanoic Acid (13C2-PFDoA)	81				10-126
Perfluoro-n-[1,2-13C2]Tetradecanoic Acid (13C2-PFTeDA)	88				10-145
Tetrafluoro-2-heptafluoropropoxy-[13C3]-propanoic acid (13C3-HFPO-DA)	75				35-142
N-Methyl-d3-Perfluoro-1-Octanesulfonamide (D3-NMeFOSA)	58				11-94
N-Ethyl-d5-Perfluoro-1-Octanesulfonamide (D5-NEtFOSA)	69				11-97
N-Methyl-d7-Perfluorooctanesulfonamidoethanol (D7-NMeFOSE)	73				10-137
N-Ethyl-d9-Perfluorooctanesulfonamidoethanol (D9-NEtFOSE)	78				10-130

Lab Control Sample Analysis

Batch Quality Control

Project Name: NASB-HANGAR 4 AFFF

Lab Number: L2449150

Project Number: 83336

Report Date: 09/12/24

Parameter	LCS %Recovery	Qual	LCSD %Recovery	Qual	%Recovery Limits	RPD	Qual	RPD Limits
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab Associated sample(s): 01-02 Batch: WG1969890-3								
Perfluorobutanoic Acid (PFBA)	107		-		40-150	-		30
Perfluoropentanoic Acid (PFPeA)	108		-		40-150	-		30
Perfluorobutanesulfonic Acid (PFBS)	106		-		40-150	-		30
1H,1H,2H,2H-Perfluorohexanesulfonic Acid (4:2FTS)	120		-		40-150	-		30
Perfluorohexanoic Acid (PFHxA)	106		-		40-150	-		30
Perfluoropentanesulfonic Acid (PFPeS)	117		-		40-150	-		30
Perfluoroheptanoic Acid (PFHpA)	108		-		40-150	-		30
Perfluorohexanesulfonic Acid (PFHxS)	110		-		40-150	-		30
Perfluorooctanoic Acid (PFOA)	100		-		40-150	-		30
1H,1H,2H,2H-Perfluorooctanesulfonic Acid (6:2FTS)	108		-		40-150	-		30
Perfluoroheptanesulfonic Acid (PFHpS)	120		-		40-150	-		30
Perfluorononanoic Acid (PFNA)	112		-		40-150	-		30
Perfluorooctanesulfonic Acid (PFOS)	102		-		40-150	-		30
Perfluorodecanoic Acid (PFDA)	110		-		40-150	-		30
1H,1H,2H,2H-Perfluorodecanesulfonic Acid (8:2FTS)	112		-		40-150	-		30
Perfluorononanesulfonic Acid (PFNS)	95		-		40-150	-		30
N-Methyl Perfluorooctanesulfonamidoacetic Acid (NMeFOSAA)	108		-		40-150	-		30
Perfluoroundecanoic Acid (PFUnA)	105		-		40-150	-		30
Perfluorodecanesulfonic Acid (PFDS)	107		-		40-150	-		30
Perfluorooctanesulfonamide (PFOSA)	111		-		40-150	-		30
N-Ethyl Perfluorooctanesulfonamidoacetic Acid (NEtFOSAA)	119		-		40-150	-		30
Perfluorododecanoic Acid (PFDoA)	110		-		40-150	-		30

Lab Control Sample Analysis

Batch Quality Control

Project Name: NASB-HANGAR 4 AFFF

Lab Number: L2449150

Project Number: 83336

Report Date: 09/12/24

Parameter	LCS %Recovery	Qual	LCSD %Recovery	Qual	%Recovery Limits	RPD	Qual	RPD Limits
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab Associated sample(s): 01-02 Batch: WG1969890-3								
Perfluorotridecanoic Acid (PFTTrDA)	129		-		40-150	-		30
Perfluorotetradecanoic Acid (PFTTeDA)	118		-		40-150	-		30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	112		-		40-150	-		30
4,8-Dioxa-3h-Perfluorononanoic Acid (ADONA)	104		-		40-150	-		30
Perfluorododecanesulfonic Acid (PFDoS)	133		-		40-150	-		30
9-Chlorohexadecafluoro-3-Oxanone-1-Sulfonic Acid (9Cl-PF3ONS)	76		-		40-150	-		30
11-Chloroeicosafluoro-3-Oxaundecane-1-Sulfonic Acid (11Cl-PF3OUdS)	82		-		40-150	-		30
N-Methyl Perfluorooctane Sulfonamide (NMeFOSA)	109		-		40-150	-		30
N-Ethyl Perfluorooctane Sulfonamide (NEtFOSA)	101		-		40-150	-		30
N-Methyl Perfluorooctanesulfonamido Ethanol (NMeFOSE)	107		-		40-150	-		30
N-Ethyl Perfluorooctanesulfonamido Ethanol (NEtFOSE)	112		-		40-150	-		30
Perfluoro-3-Methoxypropanoic Acid (PFMPA)	106		-		40-150	-		30
Perfluoro-4-Methoxybutanoic Acid (PFMBA)	108		-		40-150	-		30
Perfluoro(2-Ethoxyethane)Sulfonic Acid (PFEESA)	108		-		40-150	-		30
Nonafluoro-3,6-Dioxaheptanoic Acid (NFDHA)	114		-		40-150	-		30
3-Perfluoropropyl Propanoic Acid (3:3FTCA)	106		-		40-150	-		30
2H,2H,3H,3H-Perfluorooctanoic Acid (5:3FTCA)	96		-		40-150	-		30
3-Perfluoroheptyl Propanoic Acid (7:3FTCA)	62		-		40-150	-		30

Lab Control Sample Analysis

Batch Quality Control

Project Name: NASB-HANGAR 4 AFFF

Lab Number: L2449150

Project Number: 83336

Report Date: 09/12/24

Parameter	LCS		LCSD		%Recovery		RPD	RPD	
	%Recovery	Qual	%Recovery	Qual	Limits	Qual		Limits	
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab Associated sample(s): 01-02 Batch: WG1969890-3									

Surrogate	LCS		LCSD		Acceptance Criteria
	%Recovery	Qual	%Recovery	Qual	
Perfluoro-n-[13C4]Butanoic Acid (13C4-PFBA)	80				41-123
Perfluoro-n-[13C5]Pentanoic Acid (13C5-PFPeA)	89				29-123
Perfluoro-1-[2,3,4-13C3]Butanesulfonic Acid (13C3-PFBS)	88				41-125
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Hexanesulfonic Acid (13C2-4:2FTS)	81				10-290
Perfluoro-n-[1,2,3,4,6-13C5]Hexanoic Acid (13C5-PFHxA)	85				40-121
Perfluoro-n-[1,2,3,4-13C4]Heptanoic Acid (13C4-PFHpA)	82				27-156
Perfluoro-1-[1,2,3-13C3]Hexanesulfonic Acid (13C3-PFHxS)	79				46-115
Perfluoro-n-[13C8]Octanoic Acid (13C8-PFOA)	85				39-121
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Octanesulfonic Acid (13C2-6:2FTS)	77				10-261
Perfluoro-n-[13C9]Nonanoic Acid (13C9-PFNA)	82				38-114
Perfluoro-1-[13C8]Octanesulfonic Acid (13C8-PFOS)	81				32-114
Perfluoro-n-[1,2,3,4,5,6-13C6]Decanoic Acid (13C6-PFDA)	87				28-115
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Decanesulfonic Acid (13C2-8:2FTS)	69				10-213
N-Methyl-d3-perfluoro-1-octanesulfonamidoacetic Acid (D3-NMeFOSAA)	63				10-172
Perfluoro-n-[1,2,3,4,5,6,7-13C7]Undecanoic Acid (13C7-PFUnA)	87				16-123
Perfluoro-1-[13C8]Octanesulfonamide (13C8-PFOSA)	67				14-108
N-Ethyl-d5-perfluoro-1-octanesulfonamidoacetic Acid (D5-NEtFOSAA)	65				10-150
Perfluoro-n-[1,2-13C2]Dodecanoic Acid (13C2-PFDoA)	96				10-126
Perfluoro-n-[1,2-13C2]Tetradecanoic Acid (13C2-PFTeDA)	104				10-145
Tetrafluoro-2-heptafluoropropoxy-[13C3]-propanoic acid (13C3-HFPO-DA)	91				35-142
N-Methyl-d3-Perfluoro-1-Octanesulfonamide (D3-NMeFOSA)	73				11-94
N-Ethyl-d5-Perfluoro-1-Octanesulfonamide (D5-NEtFOSA)	82				11-97
N-Methyl-d7-Perfluorooctanesulfonamidoethanol (D7-NMeFOSE)	83				10-137
N-Ethyl-d9-Perfluorooctanesulfonamidoethanol (D9-NEtFOSE)	86				10-130

Lab Control Sample Analysis

Batch Quality Control

Project Name: NASB-HANGAR 4 AFFF

Lab Number: L2449150

Project Number: 83336

Report Date: 09/12/24

Parameter	Low Level	Qual	Low Level	Qual	%Recovery Limits	RPD	Qual	RPD Limits
	LCS %Recovery		LCSD %Recovery					
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab Associated sample(s): 03-05 Batch: WG1970425-2 LOW LEVEL								
Perfluorobutanoic Acid (PFBA)	119		-		40-150	-		30
Perfluoropentanoic Acid (PFPeA)	123		-		40-150	-		30
Perfluorobutanesulfonic Acid (PFBS)	105		-		40-150	-		30
1H,1H,2H,2H-Perfluorohexanesulfonic Acid (4:2FTS)	115		-		40-150	-		30
Perfluorohexanoic Acid (PFHxA)	118		-		40-150	-		30
Perfluoropentanesulfonic Acid (PFPeS)	119		-		40-150	-		30
Perfluoroheptanoic Acid (PFHpA)	114		-		40-150	-		30
Perfluorohexanesulfonic Acid (PFHxS)	109		-		40-150	-		30
Perfluorooctanoic Acid (PFOA)	122		-		40-150	-		30
1H,1H,2H,2H-Perfluorooctanesulfonic Acid (6:2FTS)	117		-		40-150	-		30
Perfluoroheptanesulfonic Acid (PFHpS)	125		-		40-150	-		30
Perfluorononanoic Acid (PFNA)	123		-		40-150	-		30
Perfluorooctanesulfonic Acid (PFOS)	147		-		40-150	-		30
Perfluorodecanoic Acid (PFDA)	146		-		40-150	-		30
1H,1H,2H,2H-Perfluorodecanesulfonic Acid (8:2FTS)	119		-		40-150	-		30
Perfluorononanesulfonic Acid (PFNS)	134		-		40-150	-		30
N-Methyl Perfluorooctanesulfonamidoacetic Acid (NMeFOSAA)	94		-		40-150	-		30
Perfluoroundecanoic Acid (PFUnA)	136		-		40-150	-		30
Perfluorodecanesulfonic Acid (PFDS)	112		-		40-150	-		30
Perfluorooctanesulfonamide (PFOSA)	116		-		40-150	-		30
N-Ethyl Perfluorooctanesulfonamidoacetic Acid (NEtFOSAA)	120		-		40-150	-		30
Perfluorododecanoic Acid (PFDoA)	132		-		40-150	-		30

Lab Control Sample Analysis

Batch Quality Control

Project Name: NASB-HANGAR 4 AFFF

Lab Number: L2449150

Project Number: 83336

Report Date: 09/12/24

Parameter	Low Level LCS %Recovery	Qual	Low Level LCSD %Recovery	Qual	%Recovery Limits	RPD	Qual	RPD Limits
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab Associated sample(s): 03-05 Batch: WG1970425-2 LOW LEVEL								
Perfluorotridecanoic Acid (PFTTrDA)	115		-		40-150	-		30
Perfluorotetradecanoic Acid (PFTeDA)	122		-		40-150	-		30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	119		-		40-150	-		30
4,8-Dioxa-3h-Perfluorononanoic Acid (ADONA)	128		-		40-150	-		30
Perfluorododecanesulfonic Acid (PFDoS)	118		-		40-150	-		30
9-Chlorohexadecafluoro-3-Oxanone-1-Sulfonic Acid (9Cl-PF3ONS)	119		-		40-150	-		30
11-Chloroeicosafluoro-3-Oxaundecane-1-Sulfonic Acid (11Cl-PF3OUdS)	101		-		40-150	-		30
N-Methyl Perfluorooctane Sulfonamide (NMeFOSA)	98		-		40-150	-		30
N-Ethyl Perfluorooctane Sulfonamide (NEtFOSA)	123		-		40-150	-		30
N-Methyl Perfluorooctanesulfonamido Ethanol (NMeFOSE)	100		-		40-150	-		30
N-Ethyl Perfluorooctanesulfonamido Ethanol (NEtFOSE)	106		-		40-150	-		30
Perfluoro-3-Methoxypropanoic Acid (PFMPA)	78		-		40-150	-		30
Perfluoro-4-Methoxybutanoic Acid (PFMBA)	111		-		40-150	-		30
Perfluoro(2-Ethoxyethane)Sulfonic Acid (PFEESA)	130		-		40-150	-		30
Nonafluoro-3,6-Dioxaheptanoic Acid (NFDHA)	122		-		40-150	-		30
3-Perfluoropropyl Propanoic Acid (3:3FTCA)	121		-		40-150	-		30
2H,2H,3H,3H-Perfluorooctanoic Acid (5:3FTCA)	102		-		40-150	-		30
3-Perfluoroheptyl Propanoic Acid (7:3FTCA)	90		-		40-150	-		30

Lab Control Sample Analysis

Batch Quality Control

Project Name: NASB-HANGAR 4 AFFF

Lab Number: L2449150

Project Number: 83336

Report Date: 09/12/24

Parameter	Low Level LCS		Low Level LCSD		%Recovery Limits	RPD	Qual	RPD Limits
	%Recovery	Qual	%Recovery	Qual				
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab Associated sample(s): 03-05 Batch: WG1970425-2 LOW LEVEL								

Surrogate	LCS		LCSD		Acceptance Criteria
	%Recovery	Qual	%Recovery	Qual	
Perfluoro-n-[13C4]Butanoic Acid (13C4-PFBA)	69				41-123
Perfluoro-n-[13C5]Pentanoic Acid (13C5-PFPeA)	80				29-123
Perfluoro-1-[2,3,4-13C3]Butanesulfonic Acid (13C3-PFBS)	67				41-125
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Hexanesulfonic Acid (13C2-4:2FTS)	72				10-290
Perfluoro-n-[1,2,3,4,6-13C5]Hexanoic Acid (13C5-PFHxA)	75				40-121
Perfluoro-n-[1,2,3,4-13C4]Heptanoic Acid (13C4-PFHpA)	75				27-156
Perfluoro-1-[1,2,3-13C3]Hexanesulfonic Acid (13C3-PFHxS)	66				46-115
Perfluoro-n-[13C8]Octanoic Acid (13C8-PFOA)	71				39-121
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Octanesulfonic Acid (13C2-6:2FTS)	64				10-261
Perfluoro-n-[13C9]Nonanoic Acid (13C9-PFNA)	68				38-114
Perfluoro-1-[13C8]Octanesulfonic Acid (13C8-PFOS)	63				32-114
Perfluoro-n-[1,2,3,4,5,6-13C6]Decanoic Acid (13C6-PFDA)	64				28-115
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Decanesulfonic Acid (13C2-8:2FTS)	56				10-213
N-Methyl-d3-perfluoro-1-octanesulfonamidoacetic Acid (D3-NMeFOSAA)	78				10-172
Perfluoro-n-[1,2,3,4,5,6,7-13C7]Undecanoic Acid (13C7-PFUnA)	59				16-123
Perfluoro-1-[13C8]Octanesulfonamide (13C8-PFOSA)	59				14-108
N-Ethyl-d5-perfluoro-1-octanesulfonamidoacetic Acid (D5-NEtFOSAA)	59				10-150
Perfluoro-n-[1,2-13C2]Dodecanoic Acid (13C2-PFDoA)	53				10-126
Perfluoro-n-[1,2-13C2]Tetradecanoic Acid (13C2-PFTeDA)	44				10-145
Tetrafluoro-2-heptafluoropropoxy-[13C3]-propanoic acid (13C3-HFPO-DA)	71				35-142
N-Methyl-d3-Perfluoro-1-Octanesulfonamide (D3-NMeFOSA)	44				11-94
N-Ethyl-d5-Perfluoro-1-Octanesulfonamide (D5-NEtFOSA)	40				11-97
N-Methyl-d7-Perfluorooctanesulfonamidoethanol (D7-NMeFOSE)	57				10-137
N-Ethyl-d9-Perfluorooctanesulfonamidoethanol (D9-NEtFOSE)	56				10-130

Lab Control Sample Analysis

Batch Quality Control

Project Name: NASB-HANGAR 4 AFFF

Lab Number: L2449150

Project Number: 83336

Report Date: 09/12/24

Parameter	LCS %Recovery	Qual	LCSD %Recovery	Qual	%Recovery Limits	RPD	Qual	RPD Limits
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab Associated sample(s): 03-05 Batch: WG1970425-3								
Perfluorobutanoic Acid (PFBA)	134		-		40-150	-		30
Perfluoropentanoic Acid (PFPeA)	128		-		40-150	-		30
Perfluorobutanesulfonic Acid (PFBS)	123		-		40-150	-		30
1H,1H,2H,2H-Perfluorohexanesulfonic Acid (4:2FTS)	119		-		40-150	-		30
Perfluorohexanoic Acid (PFHxA)	123		-		40-150	-		30
Perfluoropentanesulfonic Acid (PFPeS)	126		-		40-150	-		30
Perfluoroheptanoic Acid (PFHpA)	119		-		40-150	-		30
Perfluorohexanesulfonic Acid (PFHxS)	116		-		40-150	-		30
Perfluorooctanoic Acid (PFOA)	123		-		40-150	-		30
1H,1H,2H,2H-Perfluorooctanesulfonic Acid (6:2FTS)	134		-		40-150	-		30
Perfluoroheptanesulfonic Acid (PFHpS)	119		-		40-150	-		30
Perfluorononanoic Acid (PFNA)	131		-		40-150	-		30
Perfluorooctanesulfonic Acid (PFOS)	126		-		40-150	-		30
Perfluorodecanoic Acid (PFDA)	135		-		40-150	-		30
1H,1H,2H,2H-Perfluorodecanesulfonic Acid (8:2FTS)	112		-		40-150	-		30
Perfluorononanesulfonic Acid (PFNS)	139		-		40-150	-		30
N-Methyl Perfluorooctanesulfonamidoacetic Acid (NMeFOSAA)	116		-		40-150	-		30
Perfluoroundecanoic Acid (PFUnA)	133		-		40-150	-		30
Perfluorodecanesulfonic Acid (PFDS)	131		-		40-150	-		30
Perfluorooctanesulfonamide (PFOSA)	122		-		40-150	-		30
N-Ethyl Perfluorooctanesulfonamidoacetic Acid (NEtFOSAA)	108		-		40-150	-		30
Perfluorododecanoic Acid (PFDoA)	130		-		40-150	-		30

Lab Control Sample Analysis

Batch Quality Control

Project Name: NASB-HANGAR 4 AFFF

Lab Number: L2449150

Project Number: 83336

Report Date: 09/12/24

Parameter	LCS %Recovery	Qual	LCSD %Recovery	Qual	%Recovery Limits	RPD	Qual	RPD Limits
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab Associated sample(s): 03-05 Batch: WG1970425-3								
Perfluorotridecanoic Acid (PFTrDA)	124		-		40-150	-		30
Perfluorotetradecanoic Acid (PFTeDA)	133		-		40-150	-		30
Hexafluoropropylene Oxide Dimer Acid (HFPO-DA)	122		-		40-150	-		30
4,8-Dioxa-3h-Perfluorononanoic Acid (ADONA)	137		-		40-150	-		30
Perfluorododecanesulfonic Acid (PFDoS)	143		-		40-150	-		30
9-Chlorohexadecafluoro-3-Oxanone-1-Sulfonic Acid (9Cl-PF3ONS)	107		-		40-150	-		30
11-Chloroeicosafluoro-3-Oxaundecane-1-Sulfonic Acid (11Cl-PF3OUS)	109		-		40-150	-		30
N-Methyl Perfluorooctane Sulfonamide (NMeFOSA)	134		-		40-150	-		30
N-Ethyl Perfluorooctane Sulfonamide (NEtFOSA)	133		-		40-150	-		30
N-Methyl Perfluorooctanesulfonamido Ethanol (NMeFOSE)	117		-		40-150	-		30
N-Ethyl Perfluorooctanesulfonamido Ethanol (NEtFOSE)	110		-		40-150	-		30
Perfluoro-3-Methoxypropanoic Acid (PFMPA)	101		-		40-150	-		30
Perfluoro-4-Methoxybutanoic Acid (PFMBA)	110		-		40-150	-		30
Perfluoro(2-Ethoxyethane)Sulfonic Acid (PFEESA)	128		-		40-150	-		30
Nonafluoro-3,6-Dioxaheptanoic Acid (NFDHA)	111		-		40-150	-		30
3-Perfluoropropyl Propanoic Acid (3:3FTCA)	110		-		40-150	-		30
2H,2H,3H,3H-Perfluorooctanoic Acid (5:3FTCA)	114		-		40-150	-		30
3-Perfluoroheptyl Propanoic Acid (7:3FTCA)	105		-		40-150	-		30

Lab Control Sample Analysis

Batch Quality Control

Project Name: NASB-HANGAR 4 AFFF

Lab Number: L2449150

Project Number: 83336

Report Date: 09/12/24

Parameter	LCS		LCSD		%Recovery		RPD	RPD	
	%Recovery	Qual	%Recovery	Qual	Limits	Qual		Limits	
Perfluorinated Alkyl Acids by EPA 1633 - Mansfield Lab Associated sample(s): 03-05 Batch: WG1970425-3									

Surrogate	LCS		LCSD		Acceptance Criteria
	%Recovery	Qual	%Recovery	Qual	
Perfluoro-n-[13C4]Butanoic Acid (13C4-PFBA)	81				41-123
Perfluoro-n-[13C5]Pentanoic Acid (13C5-PFPeA)	98				29-123
Perfluoro-1-[2,3,4-13C3]Butanesulfonic Acid (13C3-PFBS)	78				41-125
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Hexanesulfonic Acid (13C2-4:2FTS)	80				10-290
Perfluoro-n-[1,2,3,4,6-13C5]Hexanoic Acid (13C5-PFHxA)	86				40-121
Perfluoro-n-[1,2,3,4-13C4]Heptanoic Acid (13C4-PFHpA)	82				27-156
Perfluoro-1-[1,2,3-13C3]Hexanesulfonic Acid (13C3-PFHxS)	74				46-115
Perfluoro-n-[13C8]Octanoic Acid (13C8-PFOA)	79				39-121
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Octanesulfonic Acid (13C2-6:2FTS)	69				10-261
Perfluoro-n-[13C9]Nonanoic Acid (13C9-PFNA)	80				38-114
Perfluoro-1-[13C8]Octanesulfonic Acid (13C8-PFOS)	67				32-114
Perfluoro-n-[1,2,3,4,5,6-13C6]Decanoic Acid (13C6-PFDA)	76				28-115
1H,1H,2H,2H-Perfluoro-1-[1,2-13C2]Decanesulfonic Acid (13C2-8:2FTS)	64				10-213
N-Methyl-d3-perfluoro-1-octanesulfonamidoacetic Acid (D3-NMeFOSAA)	80				10-172
Perfluoro-n-[1,2,3,4,5,6,7-13C7]Undecanoic Acid (13C7-PFUnA)	69				16-123
Perfluoro-1-[13C8]Octanesulfonamide (13C8-PFOSA)	68				14-108
N-Ethyl-d5-perfluoro-1-octanesulfonamidoacetic Acid (D5-NEtFOSAA)	66				10-150
Perfluoro-n-[1,2-13C2]Dodecanoic Acid (13C2-PFDoA)	70				10-126
Perfluoro-n-[1,2-13C2]Tetradecanoic Acid (13C2-PFTeDA)	57				10-145
Tetrafluoro-2-heptafluoropropoxy-[13C3]-propanoic acid (13C3-HFPO-DA)	86				35-142
N-Methyl-d3-Perfluoro-1-Octanesulfonamide (D3-NMeFOSA)	51				11-94
N-Ethyl-d5-Perfluoro-1-Octanesulfonamide (D5-NEtFOSA)	55				11-97
N-Methyl-d7-Perfluorooctanesulfonamidoethanol (D7-NMeFOSE)	65				10-137
N-Ethyl-d9-Perfluorooctanesulfonamidoethanol (D9-NEtFOSE)	67				10-130

Project Name: NASB-HANGAR 4 AFFF**Lab Number:** L2449150**Project Number:** 83336**Report Date:** 09/12/24**Sample Receipt and Container Information**

Were project specific reporting limits specified?

YES

Cooler Information

Cooler	Custody Seal
A	Absent

Container Information

Container ID	Container Type	Cooler	Initial pH	Final pH	Temp deg C	Pres	Seal	Frozen Date/Time	Analysis(*)
L2449150-01A	Plastic 500ml unpreserved	A	NA		4.8	Y	Absent		A2-1633-DRAFT(28)
L2449150-01B	Plastic 500ml unpreserved	A	NA		4.8	Y	Absent		A2-1633-DRAFT(28)
L2449150-01C	Plastic 500ml unpreserved	A	NA		4.8	Y	Absent		A2-1633-DRAFT(28)
L2449150-02A	Plastic 500ml unpreserved	A	NA		4.8	Y	Absent		A2-1633-DRAFT(28)
L2449150-02B	Plastic 500ml unpreserved	A	NA		4.8	Y	Absent		A2-1633-DRAFT(28)
L2449150-02C	Plastic 500ml unpreserved	A	NA		4.8	Y	Absent		A2-1633-DRAFT(28)
L2449150-03A	Plastic 500ml unpreserved	A	NA		4.8	Y	Absent		A2-1633-DRAFT(28)
L2449150-03B	Plastic 500ml unpreserved	A	NA		4.8	Y	Absent		A2-1633-DRAFT(28)
L2449150-03C	Plastic 500ml unpreserved	A	NA		4.8	Y	Absent		A2-1633-DRAFT(28)
L2449150-04A	Plastic 500ml unpreserved	A	NA		4.8	Y	Absent		A2-1633-DRAFT(28)
L2449150-04B	Plastic 500ml unpreserved	A	NA		4.8	Y	Absent		A2-1633-DRAFT(28)
L2449150-04C	Plastic 500ml unpreserved	A	NA		4.8	Y	Absent		A2-1633-DRAFT(28)
L2449150-05A	Plastic 500ml unpreserved	A	NA		4.8	Y	Absent		A2-1633-DRAFT(28)
L2449150-05B	Plastic 500ml unpreserved	A	NA		4.8	Y	Absent		A2-1633-DRAFT(28)
L2449150-05C	Plastic 500ml unpreserved	A	NA		4.8	Y	Absent		A2-1633-DRAFT(28)

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Serial_No:09122416:23
Lab Number: L2449150
Report Date: 09/12/24

PFAS PARAMETER SUMMARY

Parameter	Acronym	CAS Number
PERFLUOROALKYL CARBOXYLIC ACIDS (PFCAs)		
Perfluorooctadecanoic Acid	PFODA	16517-11-6
Perfluorohexadecanoic Acid	PFHxDA	67905-19-5
Perfluorotetradecanoic Acid	PFTA/PFTeDA	376-06-7
Perfluorotridecanoic Acid	PFTrDA	72629-94-8
Perfluorododecanoic Acid	PFDoA	307-55-1
Perfluoroundecanoic Acid	PFUnA	2058-94-8
Perfluorodecanoic Acid	PFDA	335-76-2
Perfluorononanoic Acid	PFNA	375-95-1
Perfluorooctanoic Acid	PFOA	335-67-1
Perfluoroheptanoic Acid	PFHpA	375-85-9
Perfluorohexanoic Acid	PFHxA	307-24-4
Perfluoropentanoic Acid	PFPeA	2706-90-3
Perfluorobutanoic Acid	PFBA	375-22-4
PERFLUOROALKYL SULFONIC ACIDS (PFSAs)		
Perfluorododecanesulfonic Acid	PFDoDS/PFDoS	79780-39-5
Perfluorodecanesulfonic Acid	PFDS	335-77-3
Perfluorononanesulfonic Acid	PFNS	68259-12-1
Perfluorooctanesulfonic Acid	PFOS	1763-23-1
Perfluoroheptanesulfonic Acid	PFHpS	375-92-8
Perfluorohexanesulfonic Acid	PFHxS	355-46-4
Perfluoropentanesulfonic Acid	PFPeS	2706-91-4
Perfluorobutanesulfonic Acid	PFBS	375-73-5
Perfluoropropanesulfonic Acid	PFPrS	423-41-6
FLUOROTELOMERS		
1H,1H,2H,2H-Perfluorododecanesulfonic Acid	10:2FTS	120226-60-0
1H,1H,2H,2H-Perfluorodecanesulfonic Acid	8:2FTS	39108-34-4
1H,1H,2H,2H-Perfluorooctanesulfonic Acid	6:2FTS	27619-97-2
1H,1H,2H,2H-Perfluorohexanesulfonic Acid	4:2FTS	757124-72-4
PERFLUOROALKANE SULFONAMIDES (FASAs)		
Perfluorooctanesulfonamide	FOSA/PFOSA	754-91-6
N-Ethyl Perfluorooctane Sulfonamide	NEtFOSA	4151-50-2
N-Methyl Perfluorooctane Sulfonamide	NMeFOSA	31506-32-8
PERFLUOROALKANE SULFONYL SUBSTANCES		
N-Ethyl Perfluorooctanesulfonamido Ethanol	NEtFOSE	1691-99-2
N-Methyl Perfluorooctanesulfonamido Ethanol	NMeFOSE	24448-09-7
N-Ethyl Perfluorooctanesulfonamidoacetic Acid	NEtFOSAA	2991-50-6
N-Methyl Perfluorooctanesulfonamidoacetic Acid	NMeFOSAA	2355-31-9
PER- and POLYFLUOROALKYL ETHER CARBOXYLIC ACIDS		
2,3,3,3-Tetrafluoro-2-[1,1,2,2,3,3,3-Heptafluoropropoxy]-Propanoic Acid	HFPO-DA	13252-13-6
4,8-Dioxa-3h-Perfluorononanoic Acid	ADONA	919005-14-4
CHLORO-PERFLUOROALKYL SULFONIC ACIDS		
11-Chloroeicosafuoro-3-Oxaundecane-1-Sulfonic Acid	11Cl-PF3OUdS	763051-92-9
9-Chlorohexadecafluoro-3-Oxanone-1-Sulfonic Acid	9Cl-PF3ONS	756426-58-1
PERFLUOROETHER SULFONIC ACIDS (PFESAs)		
Perfluoro(2-Ethoxyethane)Sulfonic Acid	PFEEA	113507-82-7
PERFLUOROETHER/POLYETHER CARBOXYLIC ACIDS (PFPCAs)		
Perfluoro-3-Methoxypropanoic Acid	PFMPA	377-73-1
Perfluoro-4-Methoxybutanoic Acid	PFMBA	863090-89-5
Nonafluoro-3,6-Dioxaheptanoic Acid	NFDHA	151772-58-6

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Serial_No:09122416:23
Lab Number: L2449150
Report Date: 09/12/24

PFAS PARAMETER SUMMARY

Parameter	Acronym	CAS Number
FLUOROTELOMER CARBOXYLIC ACIDS (FTCAs)		
3-Perfluoroheptyl Propanoic Acid	7:3FTCA	812-70-4
2H,2H,3H,3H-Perfluorooctanoic Acid	5:3FTCA	914637-49-3
3-Perfluoropropyl Propanoic Acid	3:3FTCA	356-02-5

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

GLOSSARY

Acronyms

DL	- Detection Limit: This value represents the level to which target analyte concentrations are reported as estimated values, when those target analyte concentrations are quantified below the limit of quantitation (LOQ). The DL includes any adjustments from dilutions, concentrations or moisture content, where applicable. (DoD report formats only.)
EDL	- Estimated Detection Limit: This value represents the level to which target analyte concentrations are reported as estimated values, when those target analyte concentrations are quantified below the reporting limit (RL). The EDL includes any adjustments from dilutions, concentrations or moisture content, where applicable. The use of EDLs is specific to the analysis of PAHs using Solid-Phase Microextraction (SPME).
EMPC	- Estimated Maximum Possible Concentration: The concentration that results from the signal present at the retention time of an analyte when the ions meet all of the identification criteria except the ion abundance ratio criteria. An EMPC is a worst-case estimate of the concentration.
EPA	- Environmental Protection Agency.
LCS	- Laboratory Control Sample: A sample matrix, free from the analytes of interest, spiked with verified known amounts of analytes or a material containing known and verified amounts of analytes.
LCSD	- Laboratory Control Sample Duplicate: Refer to LCS.
LFB	- Laboratory Fortified Blank: A sample matrix, free from the analytes of interest, spiked with verified known amounts of analytes or a material containing known and verified amounts of analytes.
LOD	- Limit of Detection: This value represents the level to which a target analyte can reliably be detected for a specific analyte in a specific matrix by a specific method. The LOD includes any adjustments from dilutions, concentrations or moisture content, where applicable. (DoD report formats only.)
LOQ	- Limit of Quantitation: The value at which an instrument can accurately measure an analyte at a specific concentration. The LOQ includes any adjustments from dilutions, concentrations or moisture content, where applicable. (DoD report formats only.) Limit of Quantitation: The value at which an instrument can accurately measure an analyte at a specific concentration. The LOQ includes any adjustments from dilutions, concentrations or moisture content, where applicable. (DoD report formats only.)
MDL	- Method Detection Limit: This value represents the level to which target analyte concentrations are reported as estimated values, when those target analyte concentrations are quantified below the reporting limit (RL). The MDL includes any adjustments from dilutions, concentrations or moisture content, where applicable.
MS	- Matrix Spike Sample: A sample prepared by adding a known mass of target analyte to a specified amount of matrix sample for which an independent estimate of target analyte concentration is available. For Method 332.0, the spike recovery is calculated using the native concentration, including estimated values.
MSD	- Matrix Spike Sample Duplicate: Refer to MS.
NA	- Not Applicable.
NC	- Not Calculated: Term is utilized when one or more of the results utilized in the calculation are non-detect at the parameter's reporting unit.
NDPA/DPA	- N-Nitrosodiphenylamine/Diphenylamine.
NI	- Not Ignitable.
NP	- Non-Plastic: Term is utilized for the analysis of Atterberg Limits in soil.
NR	- No Results: Term is utilized when 'No Target Compounds Requested' is reported for the analysis of Volatile or Semivolatile Organic TIC only requests.
RL	- Reporting Limit: The value at which an instrument can accurately measure an analyte at a specific concentration. The RL includes any adjustments from dilutions, concentrations or moisture content, where applicable.
RPD	- Relative Percent Difference: The results from matrix and/or matrix spike duplicates are primarily designed to assess the precision of analytical results in a given matrix and are expressed as relative percent difference (RPD). Values which are less than five times the reporting limit for any individual parameter are evaluated by utilizing the absolute difference between the values; although the RPD value will be provided in the report.
SRM	- Standard Reference Material: A reference sample of a known or certified value that is of the same or similar matrix as the associated field samples.
STLP	- Semi-dynamic Tank Leaching Procedure per EPA Method 1315.
TEF	- Toxic Equivalency Factors: The values assigned to each dioxin and furan to evaluate their toxicity relative to 2,3,7,8-TCDD.
TEQ	- Toxic Equivalent: The measure of a sample's toxicity derived by multiplying each dioxin and furan by its corresponding TEF and then summing the resulting values.
TIC	- Tentatively Identified Compound: A compound that has been identified to be present and is not part of the target compound list (TCL) for the method and/or program. All TICs are qualitatively identified and reported as estimated concentrations.

Report Format: DU Report with 'J' Qualifiers



Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

Footnotes

- 1 - The reference for this analyte should be considered modified since this analyte is absent from the target analyte list of the original method.

Terms

Analytical Method: Both the document from which the method originates and the analytical reference method. (Example: EPA 8260B is shown as 1,8260B.) The codes for the reference method documents are provided in the References section of the Addendum.

Chlordane: The target compound Chlordane (CAS No. 57-74-9) is reported for GC ECD analyses. Per EPA, this compound "refers to a mixture of chlordane isomers, other chlorinated hydrocarbons and numerous other components." (Reference: USEPA Toxicological Review of Chlordane, In Support of Summary Information on the Integrated Risk Information System (IRIS), December 1997.)

Difference: With respect to Total Oxidizable Precursor (TOP) Assay analysis, the difference is defined as the Post-Treatment value minus the Pre-Treatment value.

Final pH: As it pertains to Sample Receipt & Container Information section of the report, Final pH reflects pH of container determined after adjustment at the laboratory, if applicable. If no adjustment required, value reflects Initial pH.

Frozen Date/Time: With respect to Volatile Organics in soil, Frozen Date/Time reflects the date/time at which associated Reagent Water-preserved vials were initially frozen. Note: If frozen date/time is beyond 48 hours from sample collection, value will be reflected in 'bold'.

Gasoline Range Organics (GRO): Gasoline Range Organics (GRO) results include all chromatographic peaks eluting from Methyl tert butyl ether through Naphthalene, with the exception of GRO analysis in support of State of Ohio programs, which includes all chromatographic peaks eluting from Hexane through Dodecane.

Initial pH: As it pertains to Sample Receipt & Container Information section of the report, Initial pH reflects pH of container determined upon receipt, if applicable.

PAH Total: With respect to Alkylated PAH analyses, the 'PAHs, Total' result is defined as the summation of results for all or a subset of the following compounds: Naphthalene, C1-C4 Naphthalenes, 2-Methylnaphthalene, 1-Methylnaphthalene, Biphenyl, Acenaphthylene, Acenaphthene, Fluorene, C1-C3 Fluorenes, Phenanthrene, C1-C4 Phenanthrenes/Anthracenes, Anthracene, Fluoranthene, Pyrene, C1-C4 Fluoranthenes/Pyrenes, Benz(a)anthracene, Chrysene, C1-C4 Chrysenes, Benzo(b)fluoranthene, Benzo(j)+(k)fluoranthene, Benzo(e)pyrene, Benzo(a)pyrene, Perylene, Indeno(1,2,3-cd)pyrene, Dibenz(ah)+(ac)anthracene, Benzo(g,h,i)perylene. If a 'Total' result is requested, the results of its individual components will also be reported.

PFAS Total: With respect to PFAS analyses, the 'PFAS, Total (5)' result is defined as the summation of results for: PFHpA, PFHxS, PFOA, PFNA and PFOS. In addition, the 'PFAS, Total (6)' result is defined as the summation of results for: PFHpA, PFHxS, PFOA, PFNA, PFDA and PFOS. For MassDEP DW compliance analysis only, the 'PFAS, Total (6)' result is defined as the summation of results at or above the RL. Note: If a 'Total' result is requested, the results of its individual components will also be reported.

Total: With respect to Organic analyses, a 'Total' result is defined as the summation of results for individual isomers or Aroclors. If a 'Total' result is requested, the results of its individual components will also be reported. This is applicable to 'Total' results for methods 8260, 8081 and 8082.

Data Qualifiers

- A** - Spectra identified as "Aldol Condensates" are byproducts of the extraction/concentration procedures when acetone is introduced in the process.
- B** - The analyte was detected above the reporting limit in the associated method blank. Flag only applies to associated field samples that have detectable concentrations of the analyte at less than ten times (10x) the concentration found in the blank. For MCP-related projects, flag only applies to associated field samples that have detectable concentrations of the analyte at less than ten times (10x) the concentration found in the blank. For DOD-related projects, flag only applies to associated field samples that have detectable concentrations of the analyte at less than ten times (10x) the concentration found in the blank AND the analyte was detected above one-half the reporting limit (or above the reporting limit for common lab contaminants) in the associated method blank. For NJ-Air-related projects, flag only applies to associated field samples that have detectable concentrations of the analyte above the reporting limit. For NJ-related projects (excluding Air), flag only applies to associated field samples that have detectable concentrations of the analyte, which was detected above the reporting limit in the associated method blank or above five times the reporting limit for common lab contaminants (Phthalates, Acetone, Methylene Chloride, 2-Butanone).
- C** - Co-elution: The target analyte co-elutes with a known lab standard (i.e. surrogate, internal standards, etc.) for co-extracted analyses.
- D** - Concentration of analyte was quantified from diluted analysis. Flag only applies to field samples that have detectable concentrations of the analyte.
- E** - Concentration of analyte exceeds the range of the calibration curve and/or linear range of the instrument.
- F** - The ratio of quantifier ion response to qualifier ion response falls outside of the laboratory criteria. Results are considered to be an estimated maximum concentration.
- G** - The concentration may be biased high due to matrix interferences (i.e. co-elution) with non-target compound(s). The result should be considered estimated.
- H** - The analysis of pH was performed beyond the regulatory-required holding time of 15 minutes from the time of sample collection.
- I** - The lower value for the two columns has been reported due to obvious interference.
- J** - Estimated value. The Target analyte concentration is below the quantitation limit (RL), but above the Method Detection Limit (MDL) or Estimated Detection Limit (EDL) for SPME-related analyses. This represents an estimated concentration for Tentatively

Report Format: DU Report with 'J' Qualifiers



Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

Data Qualifiers

Identified Compounds (TICs). For calculated parameters, this represents that one or more values used in the calculation were estimated.

- M** - Reporting Limit (RL) exceeds the MCP CAM Reporting Limit for this analyte.
- ND** - Not detected at the method detection limit (MDL) for the sample, or estimated detection limit (EDL) for SPME-related analyses.
- NJ** - Presumptive evidence of compound. This represents an estimated concentration for Tentatively Identified Compounds (TICs), where the identification is based on a mass spectral library search.
- P** - The RPD between the results for the two columns exceeds the method-specified criteria.
- Q** - The quality control sample exceeds the associated acceptance criteria. For DOD-related projects, LCS and/or Continuing Calibration Standard exceedences are also qualified on all associated sample results. Note: This flag is not applicable for matrix spike recoveries when the sample concentration is greater than 4x the spike added or for batch duplicate RPD when the sample concentrations are less than 5x the RL. (Metals only.)
- R** - Analytical results are from sample re-analysis.
- RE** - Analytical results are from sample re-extraction.
- S** - Analytical results are from modified screening analysis.
- V** - The surrogate associated with this target analyte has a recovery outside the QC acceptance limits. (Applicable to MassDEP DW Compliance samples only.)
- Z** - The batch matrix spike and/or duplicate associated with this target analyte has a recovery/RPD outside the QC acceptance limits. (Applicable to MassDEP DW Compliance samples only.)

Project Name: NASB-HANGAR 4 AFFF
Project Number: 83336

Lab Number: L2449150
Report Date: 09/12/24

REFERENCES

- 144 Analysis of Per- and Polyfluoroalkyl Substances (PFAS) in Aqueous, Solid, Biosolids, and Tissue Samples by LC-MS/MS. Draft EPA Method 1633, EPA Document 821-D-22-001, June 2022.

LIMITATION OF LIABILITIES

Alpha Analytical performs services with reasonable care and diligence normal to the analytical testing laboratory industry. In the event of an error, the sole and exclusive responsibility of Alpha Analytical shall be to re-perform the work at it's own expense. In no event shall Alpha Analytical be held liable for any incidental, consequential or special damages, including but not limited to, damages in any way connected with the use of, interpretation of, information or analysis provided by Alpha Analytical.

We strongly urge our clients to comply with EPA protocol regarding sample volume, preservation, cooling, containers, sampling procedures, holding time and splitting of samples in the field.



Certification Information

The following analytes are not included in our Primary NELAP Scope of Accreditation:

Westborough Facility

EPA 624.1: m/p-xylene, o-xylene, Naphthalene

EPA 625.1: alpha-Terpineol

EPA 8260D: NPW: 1,2,4,5-Tetramethylbenzene; 4-Ethyltoluene; SCM: Iodomethane (methyl iodide), 1,2,4,5-Tetramethylbenzene; 4-Ethyltoluene.

EPA 8270E: NPW: Dimethylnaphthalene,1,4-Diphenylhydrazine, alpha-Terpineol, Azobenzene; SCM: Dimethylnaphthalene,1,4-Diphenylhydrazine.

SM4500: NPW: Amenable Cyanide; SCM: Total Phosphorus, TKN, NO₂, NO₃.

Mansfield Facility

SM 2540D: TSS.

EPA TO-15: Halothane, 2,4,4-Trimethyl-2-pentene, 2,4,4-Trimethyl-1-pentene, Thiophene, 2-Methylthiophene,

3-Methylthiophene, 2-Ethylthiophene, 1,2,3-Trimethylbenzene, Indan, Indene, 1,2,4,5-Tetramethylbenzene, Benzothiophene, 1-Methylnaphthalene.

Nonpotable Water: EPA RSK-175 Dissolved Gases

Biological Tissue Matrix: EPA 3050B

The following analytes are included in our Massachusetts DEP Scope of Accreditation

Westborough Facility:

Drinking Water

EPA 300.0: Chloride, Nitrate-N, Fluoride, Sulfate; **EPA 353.2:** Nitrate-N, Nitrite-N; **SM4500NO3-F:** Nitrate-N, Nitrite-N; **SM4500F-C, SM4500CN-CE,**

EPA 180.1, SM2130B, SM4500Cl-D, SM2320B, SM2540C, SM4500H-B, SM4500NO2-B

EPA 524.2: THMs and VOCs; **EPA 504.1:** EDB, DBCP.

Microbiology: SM9215B; SM9223-P/A, SM9223B-Colilert-QT, SM9222D.

Non-Potable Water

SM4500H,B, EPA 120.1, SM2510B, SM2540C, SM2320B, SM4500CL-E, SM4500F-BC, SM4500NH3-BH: Ammonia-N and Kjeldahl-N, **EPA 350.1:**

Ammonia-N, **LACHAT 10-107-06-1-B:** Ammonia-N, **EPA 351.1, SM4500NO3-F, EPA 353.2:** Nitrate-N, **SM4500P-E, SM4500P-B, E, SM4500SO4-E,**

SM5220D, EPA 410.4, SM5210B, SM5310C, SM4500CL-D, EPA 1664, EPA 420.1, SM4500-CN-CE, SM2540D, EPA 300: Chloride, Sulfate, Nitrate.

EPA 624.1: Volatile Halocarbons & Aromatics,

EPA 608.3: Chlordane, Toxaphene, Aldrin, alpha-BHC, beta-BHC, gamma-BHC, delta-BHC, Dieldrin, DDD, DDE, DDT, Endosulfan I, Endosulfan II, Endosulfan sulfate, Endrin, Endrin Aldehyde, Heptachlor, Heptachlor Epoxide, PCBs

EPA 625.1: SVOC (Acid/Base/Neutral Extractables).

Microbiology: SM9223B-Colilert-QT; Enterolert-QT, EPA 1600, EPA 1603, SM9222D.

Mansfield Facility:

Drinking Water

EPA 200.7: Al, Ba, Cd, Cr, Cu, Fe, Mn, Ni, Na, Ag, Ca, Zn. **EPA 200.8:** Al, Sb, As, Ba, Be, Cd, Cr, Cu, Pb, Mn, Ni, Se, Ag, TL, Zn. **EPA 245.1** Hg.

EPA 522, EPA 537.1.

Non-Potable Water

EPA 200.7: Al, Sb, As, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Mo, Ni, K, Se, Ag, Na, Sr, TL, Ti, V, Zn.

EPA 200.8: Al, Sb, As, Be, Cd, Cr, Cu, Fe, Pb, Mn, Ni, K, Se, Ag, Na, TL, Zn.

EPA 245.1 Hg.

SM2340B

For a complete listing of analytes and methods, please contact your Alpha Project Manager.



CHAIN OF CUSTODY

PAGE 1 OF 1

8 Walkup Drive
Westboro, MA 01581
Tel: 508-896-9220

320 Forbes Blvd
Mansfield, MA 02048
Tel: 508-822-9300

Project Information

Project Name: **NASB-Hanger 4 AFFF**

Project Location: **Brunswick**

Project #: **EGAD: 83336**

Project Manager: **Molly King**

ALPHA Quote #:

Date Rec'd in Lab: **8/28/24**

ALPHA Job #: **L2449150**

Report Information - Data Deliverables

ADEx EMAIL

Billing Information

Same as Client info PO #:

Client Information

Client: **Maine Department of Environmental Protection**

Address: **17 State House STA**

Augusta, ME 04333

Phone: **207-458-8839**

Email: **Molly.King@maine.gov**

Regulatory Requirements & Project Information Requirements

Yes No MA MCP Analytical Methods Yes No CT RCP Analytical Methods

Yes No Matrix Spike Required on this SDG? (Required for MCP Inorganics)

Yes No GW1 Standards (Info Required for Metals & EPH with Targets)

Yes No NPDES RGP

Other State /Fed Program Criteria

Turn-Around Time

Standard RUSH (only confirmed if pre-approved)

Date Due:

ANALYSIS	VOC: <input type="checkbox"/> 8260 <input type="checkbox"/> 624 <input type="checkbox"/> 524.2	SAMPLE INFO	
	SVOC: <input type="checkbox"/> ABN <input type="checkbox"/> PAH		
	METALS: <input type="checkbox"/> MCP 13 <input type="checkbox"/> MCP 14 <input type="checkbox"/> RCP 15		
	METALS: <input type="checkbox"/> RCRA5 <input type="checkbox"/> RCRA8		
	EPH: <input type="checkbox"/> Ranges & Targets <input type="checkbox"/> Ranges Only		
	VPH: <input type="checkbox"/> Ranges & Targets <input type="checkbox"/> Ranges Only		
	PCB <input type="checkbox"/> PEST		
	TPH: <input type="checkbox"/> Quant Only <input type="checkbox"/> Fingerprint		
	FFAS - 1633		
	Filtration <input type="checkbox"/> Field <input type="checkbox"/> Lab to do		
Preservation <input type="checkbox"/> Lab to do			
Sample Comments			

Additional Project Information:

contact Tom Danielson with question about sites or samples

Copies to: Finn.Whiting@maine.gov and iver.j.mcleod@maine.gov

ALPHA Lab ID (Lab Use Only)	Sample ID	Collection		Sample Matrix	Sampler Initials
		Date	Time		
49150-01	SW68-20240827	8/27/24	11:00	SW	JG
02	SW67-20240827	8/27/24	12:00	SW	JG
03	ARB2-SW-2024-08-27	8/27/24	17:30	SW	TD
04	ARB1-SW-2024-08-27	8/27/24	14:40	SW	TD
05	ART-SW-2024-08-27	8/27/24	13:15	SW	TD

Container Type

P= Plastic
A= Amber glass
V= Vial
G= Glass
B= Bacteria cup
C= Cube
O= Other
E= Encore
D= BOD Bottle

Preservative

A= None
B= HCl
C= HNO₃
D= H₂SO₄
E= NaOH
F= MeOH
G= NaHSO₄
H= Na₂S₂O₃
I= Ascorbic Acid
J= NH₄Cl
K= Zn Acetate
O= Other

Container Type	P
Preservative	A

Relinquished By:	Date/Time	Received By:	Date/Time
<i>[Signature]</i>	8/28/24 10:30	<i>[Signature]</i>	8/28/24 13:30
<i>[Signature]</i>	8/28/24 14:00	<i>[Signature]</i>	8/28/24 14:49
<i>[Signature]</i>	8/28/24 18:03	<i>[Signature]</i>	

All samples submitted are subject to Alpha's Terms and Conditions. See reverse side.

FORM NO: 01-01 (rev. 12-Mar-2012)

TOTAL # BOTTLES 8/28/24 20:30

8/28/24 19:30